

issued by

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testing performed at the above address (Unit 14) only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS	<u>Molecular Biology examination activities for the purpose of clinical diagnosis:</u>	Documented In-house methods using analysers specified
Urines, genital, throat and rectal swabs	Detection of all variants of <i>Chlamydia trachomatis</i> cryptic plasmid and genomic DNA	Roche Cobas 6800 multiplex screening system utilising amplification of target DNA by real time PCR and nucleic acid hybridisation PCR03
Anogenital swabs	Detection of DR-9 specific <i>Neisseria gonorrhoea</i> DNA <i>Herpes simplex</i> 1 and 2	Roche Cobas 4800 multiplex screening system utilising amplification of target DNA by real time PCR and nucleic acid hybridization PCR02



4680
Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Acculabs Diagnostics UK Ltd
Issue No: 013 Issue date: 09 July 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<u>Serology examination activities for the purpose of clinical diagnosis:</u>	Documented In-house methods using analyser and methods specified
Serum/Plasma	Hepatitis B surface antigen (HBsAg)	Roche Cobas e601 using electrochemiluminescence immunoassay SER01
	Hepatitis B surface antibody (Anti HBs)	
	Hepatitis B core antibody (Anti HBc)	
	Hepatitis C antibody (Anti HCV)	
	Hepatitis A total antibody	
Serum/Plasma (including capillary)	HIV Ag/Ab Combo (Anti HIV1/2 and HIV1p24 Ag)	CAPTIATM Syphilis Total Antibody EIA SER16 Biochrom EZ read Microplate 400 reader SER06
Serum/Plasma (including capillary)	Syphilis total antibody	
	Confirmation testing of Syphilis serology:	
Serum/Plasma	Syphilis Total antibody	New-Bio TPHA haem-agglutination SER16
Serum/Plasma	<i>Treponema pallidum</i> antibody	
Serum/Plasma	<i>Treponema pallidum</i> antigen	
Serum/Plasma	Syphilis IgM	Mikrogen Syphilis IgM ELISA SER16 Biochrom EZ read Microplate 400 reader SER06



4680

Accredited to
ISO 15189:2022

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Acculabs Diagnostics UK Ltd

Issue No: 013 Issue date: 09 July 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<u>Clinical Biochemistry examination activities for the purpose of clinical diagnosis</u>	In house documented methods incorporating manufacturers' instructions where relevant
Serum/Plasma	Albumin (ALB) ALP ALT AST Calcium CK Creatinine GGT Glucose HDL Cholesterol Iron LDL Cholesterol Magnesium (Mg) Phosphate Total-Bilirubin Total Cholesterol Total Protein Triglycerides UBIC Urea Uric Acid HbA1c	Roche Cobas C311 Using suppliers methods unless otherwise stated and SOP CHEM02 Spectrophotometric
Whole Blood		
Serum/Plasma	Apolipoprotein (APO) A Apolipoprotein (APO) B CRP (including hs-CRP)	Immunoturbidimetric
	Adjusted Calcium e-GFR Globulins	Roche Cobas C311 Calculations from the results of the methods above using SOP GUI 115 v1.0 :



4680

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Acculabs Diagnostics UK Ltd
Issue No: 013 Issue date: 09 July 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<u>Clinical Biochemistry examination activities for the purpose of clinical diagnosis (cont'd)</u>	In house documented methods incorporating manufacturers' instructions where relevant
Serum/Plasma	Sodium Potassium	Roche Cobas C311 ISE Module Using suppliers methods unless otherwise stated and SOP CHEM02 Indirect ISE
	Active B12 Ferritin Folate FSH FT3 FT4 HCG LH Oestradiol Progesterone Prolactin PSA SHGB Testosterone TSH Vitamin D	Roche Cobas e801 Using suppliers methods unless otherwise stated and SOP SER13 ECL Immunoassay
Serum	Cortisol DHEAS	Roche Cobas e411 Using suppliers methods unless otherwise stated and SOP CHEM 03 ECLIA
END		