

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



4688

Accredited to
ISO/IEC 17025:2017

Instrument Repairs and Calibration A trading company of IRC Limited

Issue No: 019 Issue date: 16 May 2024

Unit 5

Howard Court Industrial Estate
East Kilbride
Glasgow
G74 4QZ

Contact: Mr Craig Moore

Tel: +44 (0)1355 264120

Fax: +44 (0)1355 264150

E-Mail: service@instrument-repairs.com

Website: www.instrument-repairs.com

Calibration performed by the Organisation at the locations specified

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 5 Howard Court Industrial Estate East Kilbride Glasgow G74 4QZ Local contact Contact: Mr Craig Moore Tel: +44 (0)1355 264120 E-Mail: service@instrument-repairs.com Website: www.instrument-repairs.com	Head Office Permanent Laboratory Electrical Dimensional	East Kilbride
Address IRC Ltd 7A Ferguson Centre 53-57 Manse Road Glengormley Newtonabbey BT36 6RW United Kingdom Local contact Contact: Mr Frank Silo Tel: +44 (0)2890 837300 E-Mail: Belfast@instrument-repairs.com Website: www.instrument-repairs.com	Permanent Laboratory Electrical	Belfast and site calibration

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. Local contact Contact: Mr Craig Moore Tel: +44 (0)1355 264120 Contact: Mr Frank Silo Tel: +44 (0)2890 837300	Site Dimensional Site Electrical	East Kilbride and site calibration Belfast and site calibration



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL			Method used: All electrical calibrations are performed as a direct comparison against a reference standard unless otherwise stated.	East Kilbride
DC Voltage Measurement	0 V to 100 mV	12 $\mu\text{V/V} + 0.50 \mu\text{V}$	For measurement of instrument outputs	
	100 mV to 1 V	10 $\mu\text{V/V} + 1.1 \mu\text{V}$		
	1 V to 10 V	11 $\mu\text{V/V} + 10 \mu\text{V}$		
	10 V to 100 V	12 $\mu\text{V/V} + 110 \mu\text{V}$		
	100 V to 1000 V	12 $\mu\text{V/V} + 1.1 \text{ mV}$		
	1 kV to 40 kV	0.24 %		
	40 kV to 60 kV	0.66 kV		
Generation	0 mV to 30 mV	70 $\mu\text{V/V} + 3.6 \mu\text{V}$	Values can be generated for the calibration of measuring instruments	
	30 to 300 mV	70 $\mu\text{V/V} + 4.6 \mu\text{V}$		
	300 mV to 3 V	58 $\mu\text{V/V} + 16 \mu\text{V}$		
	3 V to 30 V	58 $\mu\text{V/V} + 190 \mu\text{V}$		
	30 V to 300 V	64 $\mu\text{V/V} + 2.5 \text{ mV}$		
	300 V to 1000 V	64 $\mu\text{V/V} + 8.3 \text{ V}$		
	1 kV to 40 kV	0.24 %		
	40 kV to 60 kV	0.66 kV		
DC Current Measurement	0 μA to 1 μA	63 $\mu\text{A/A} + 54 \text{ pA}$	For measurement of instrument outputs	East Kilbride
	1 μA to 10 μA	25 $\mu\text{A/A} + 71 \text{ pA}$		
	10 μA to 100 μA	25 $\mu\text{A/A} + 0.42 \text{ nA}$		
	100 μA to 1 mA	25 $\mu\text{A/A} + 7.0 \text{ nA}$		
	1 mA to 10 mA	25 $\mu\text{A/A} + 70 \text{ nA}$		
	10 mA to 100 mA	42 $\mu\text{A/A} + 0.70 \mu\text{A}$		
	100 mA to 1 A	130 $\mu\text{A/A} + 19 \mu\text{A}$		
	1 A to 10 A	0.10 %		
	10 A to 60 A	0.20 %		
	60 A to 300A	0.61%		



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued) DC Current (continued) Generation	0 μ A to 3 mA 3 mA to 30 mA 30 mA to 300 mA 300 mA to 2 A 2 A to 10 A 10 A to 60 A 60 A to 300A 60 A to 550 A 550 A to 1025 A	150 μ A/A + 95 nA 120 μ A/A + 0.82 μ A 120 μ A/A + 8.5 μ A 350 μ A/A + 110 μ A 700 μ A/A + 0.90 mA 0.20 % 0.86 % 0.50 % 1.0 %	Method used: All electrical calibrations are performed as a direct comparison against a reference standard unless otherwise stated. Values can be generated for the calibration of measuring instruments Appropriate for the calibration of clamp-on ammeters	East Kilbride



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
ELECTRICAL (continued)			Method used: All electrical calibrations are performed as a direct comparison against a reference standard unless otherwise stated.	East Kilbride
DC Resistance				
Measurement	0 Ω to 10 Ω	19 $\mu\Omega/\Omega$ + 58 $\mu\Omega$	For measurement of instrument outputs	
	10 Ω to 100 Ω	15 $\mu\Omega/\Omega$ + 0.60 m Ω		
	100 Ω to 1 k Ω	13 $\mu\Omega/\Omega$ + 0.60 m Ω		
	1 k Ω to 10 k Ω	16 $\mu\Omega/\Omega$ + 6.0 m Ω		
	10 k Ω to 100 k Ω	13 $\mu\Omega/\Omega$ + 58 m Ω		
	100 k Ω to 1 M Ω	20 $\mu\Omega/\Omega$ + 3.0 Ω		
	1 M Ω to 10 M Ω	61 $\mu\Omega/\Omega$ + 120 Ω		
	10 M Ω to 100 M Ω	700 $\mu\Omega/\Omega$ + 1.2 k Ω		
	100 M Ω to 1 G Ω	0.65 % + 12 k Ω		
Generation	100 $\mu\Omega$	170 n Ω	Values can be sourced / generated for the calibration of measuring instruments	
Specific Values	1 m Ω	0.21 $\mu\Omega$		
	10 m Ω	4.0 $\mu\Omega$		
	100 m Ω	92 $\mu\Omega$		
	0.25 Ω	7.3 m Ω		
	0.3 Ω	12 m Ω		
	0.4 Ω	26 m Ω		
	0.5 Ω	5.0 m Ω		
	0.6 Ω	15 m Ω		
	0.7 Ω	9.0 m Ω		
	0.8 Ω	8.0 m Ω		
	0.9 Ω	16 m Ω		
	1 Ω	72 $\mu\Omega$		
	2 Ω	3.0 m Ω		
	4 Ω	3.0 m Ω		
	6 Ω	3.0 m Ω		
	8 Ω	3.0 m Ω		
	10 Ω	660 $\mu\Omega$		
	100 Ω	7.0 m Ω		
	1 k Ω	120 m Ω		
	10 k Ω	1.2 Ω		
	100 k Ω	12 Ω		
	1 M Ω	1.0 k Ω		
	10 M Ω	0.034 %		
	100 M Ω	0.16 %		
	1 G Ω	0.13 %		
	10 G Ω	0.31 %		
	100 G Ω	1.7 %		
	1 T Ω	1.3 %		



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
AC Voltage (continued) Generation	<i>10 Hz to 45 Hz</i> 1 mV to 30 mV 30 mV to 300 mV 300 mV to 3 V 3 V to 30 V <i>45 Hz to 1 kHz</i> 300 V to 1000 V <i>1 kHz to 5 kHz</i> 300 V to 1000 V <i>5 kHz to 10 kHz</i> 300 V to 1000 V <i>45 Hz to 10 kHz</i> 1 mV to 30 mV 30 mV to 300 mV 300 mV to 3 V 3 V to 30 V 30 V to 300 V <i>10 kHz to 20 kHz</i> 30 V to 300 V <i>10 kHz to 100 kHz</i> 1 mV to 30 mV <i>10 kHz to 50 kHz</i> 30 mV to 300 mV 300 mV to 3 V 3 V to 30 V <i>50 kHz to 100 kHz</i> 30 mV to 300 mV 300 mV to 3 V 3 V to 30 V <i>100 kHz to 500 kHz</i> 1 mV to 30 mV 30 mV to 300 mV 300 mV to 3 V <i>50 Hz</i> 1 kV to 28 kV 28 kV to 50 kV	 0.40 % + 23 μ V 0.29 % + 59 μ V 0.17 % + 420 μ V 0.18 % + 3.2 mV 0.060 % + 98 mV 0.23 % + 32 mV 0.23 % + 32 mV 0.18 % + 23 μ V 0.060 % + 25 μ V 0.040 % + 110 μ V 0.050 % + 1.1 mV 0.10 % + 3.2 mV 0.10 % + 3.2 mV 0.40 % + 39 μ V 0.19 % + 47 μ V 0.16 % + 360 μ V 0.22 % + 5.9 mV 0.28 % + 200 μ V 0.28 % + 2.0 mV 0.28 % + 20 mV 1.2 % + 70 μ V 0.81 % + 380 μ V 0.63 % + 3.9 mV 0.50 % 0.69 kV	Values can be generated for the calibration of measuring instruments	East Kilbride



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
AC Power	45 Hz to 400 Hz 1 W to 11 kW	0.45 %	At unity power factor only	East Kilbride
AC phase angle	50 Hz 0° to 360°	0.28°	Voltage : Current Phase	
Tachometer calibration	600 rpm to 50000 rpm	0.60 rpm		
Inductance				
Generation	At 1 kHz:		Values can be generated for the calibration of measuring instruments	
	100 µH	2.6 µH		
	1 mH	26 µH		
	10 mH	380 µH		
	100 mH	2.2 mH		
	1 H	22 mH		
	10 H	98 mH		
Capacitance				
Generation	At 5 kHz:		Values can be generated for the calibration of measuring instruments	
	400 pF	13 pF		
	At 1 kHz:			
	1 nF	17 pF		
	2 nF	23 pF		
	3 nF	29 pF		
	10 nF	70 pF		
	20 nF	180 pF		
	30 nF	210 pF		
	100 nF	420 pF		
	200 nF	960 pF		
	300 nF	1.3 nF		
	1 µF	4.2 nF		
	At 100 Hz:			
	2 µF	11 nF		
	3 µF	15 nF		
	10 µF	53 nF		
	20 µF	130 nF		
	30 µF	180 nF		
	At 50 Hz:			
	100 µF	0.72 µF		
	200 µF	2.0 µF		
	300 µF	3.8 µF		



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
EQUIPMENT FOR IEE 17 TH EDITION WIRING TESTING			Other, similar, equipment that includes the same quantities and ranges can also be calibrated.	East Kilbride
<u>RCD Testers</u>				
Trip Time	20 ms to 390 ms 390 ms to 900 ms	0.90 ms 8.2 ms		
Trip Current	At 50 Hz: 10 mA 30 mA 100 mA 300 mA 1 A 2 A	34 μ A 0.26 mA 0.17 mA 2.0 mA 3.3 mA 10 mA	For trip times < 200 ms	
	At 50 Hz: 10 mA 30 mA 100 mA 300 mA 1 A 2 A	34 μ A 0.28 mA 0.17 mA 2.0 mA 3.3 mA 10 mA	For trip times > 200 ms	
<u>Loop Testers</u>				
AC Resistance	At 50 Hz: 0.05 Ω 0.1 Ω 0.22 Ω 0.33 Ω 0.5 Ω 1 Ω 5 Ω 10 Ω 100 Ω 1 k Ω	2.9 m Ω 1.5 m Ω 1.3 m Ω 3.5 m Ω 3.1 m Ω 4.0 m Ω 9.2 m Ω 25 m Ω 100 m Ω 140 m Ω		
<u>Continuity Testers</u>				
DC Resistance	See Specific Values on Page 3.			
DC Continuity Current	10 mA to 100 mA 100 mA to 200 mA 200 mA to 300 mA	0.98 mA 0.65 mA 1.5 mA		



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code	
<u>Insulation Testers</u>					
DC Resistance	See <i>Specific Values</i> on Page 4.			East Kilbride	
DC Voltage	50 V 100 V 250 V 500 V 1000 V 1000 V to 10 kV	1.0 V 1.0 V 1.0 V 1.4 V 1.6 V See <i>DC Voltage</i> measurement capability on Page 2.			
<u>Portable Appliance Testers</u>					
Earth bond resistance	<i>At 50 Hz:</i> 40 mΩ 100 mΩ 200 mΩ 500 mΩ 1 Ω 5 Ω 10 Ω	2.6 mΩ 1.3 mΩ 2.0 mΩ 2.2 mΩ 1.8 m Ω 7.5 mΩ 35 mΩ			
Earth bond current	<i>At 50 Hz:</i> 0 mA to 100 mA 100 mA to 300 mA 1 A to 30 A	4.7 mA 9.0 mA 420 mA			
Leakage current	<i>At 50 Hz:</i> 2 mA 4.7 mA 7.7 mA	15 μA 25 μA 59 μA			
Flash test	<i>At 50 Hz:</i> 1 kV to 7 kV	See <i>AC Voltage</i> measurement capability on Page 4.			
Load test	<i>At 50 Hz:</i> 3 kVA	2.0 %			
OSCILLOSCOPE CALIBRATION					
Vertical coefficients	5 mV to 50 V p-p	0.47 %			
Horizontal coefficients	2 ns to 1 μs 1 μs to 5 s	29 μs/s 32 μs/s			



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
ELECTRICAL SIMULATION OF TEMPERATURE				
<u>Calibration of thermocouple indicators</u>			Including cold junction compensation	
Thermocouple type				
Type B	600 °C to 800 °C 800 °C to 1820 °C	1.1 °C 1.1 °C		
Type C	0 °C to 150 °C 150 °C to 650 °C 650 °C to 1000 °C 1000 °C to 1800 °C 1800 °C to 2316 °C	0.55 °C 0.58 °C 0.67 °C 0.76 °C 1.2 °C		
Type E	-250 °C to -100 °C -100 °C to +650 °C 650 °C to 1000 °C	0.59 °C 0.24 °C 0.29 °C		
Type J	-210 °C to -100 °C -100 °C to -30 °C -30 °C to +150 °C 150 °C to 760 °C 760 °C to 1200 °C	0.36 °C 0.24 °C 0.22 °C 0.25 °C 0.32 °C		
Type K	-200 °C to -100 °C -100 °C to -25 °C -25 °C to +120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.41 °C 0.26 °C 0.24 °C 0.34 °C 0.49 °C		
Type N	-200 °C to -100 °C -100 °C to -25 °C -25 °C to +410 °C 410 °C to 1300 °C	0.53 °C 0.35 °C 0.32 °C 0.39 °C		
Type R	0 °C to 250 °C 250 °C to 1767 °C	1.2 °C 1.1 °C		
Type S	0 °C to 250 °C 250 °C to 1767 °C	1.2 °C 1.1 °C		
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.73 °C 0.31 °C 0.24 °C 0.22 °C		
Calibration of RTD indicators	-200 °C to +800 °C	0.0031 °C		
Temperature of reference junction / Cold junction compensation	At ambient temperature of 20 °C ± 3 °C, or nominal 0°C	0.030 °C		

East Kilbride



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
DIMENSIONAL CALIBRATION LENGTH Plain plug gauges (parallel) Length gauges, flat and spherical ended Feeler gauges MEASURING INSTRUMENTS AND MACHINES External micrometer Internal micrometer	1 to 50 diameter 50 to 100 diameter 100 to 200 diameter 200 to 300 diameter 25 to 1000 BS 957:2008 0.025 to 1 BS 870:2008 and above 0 to 1000 Heads: (Zero) Setting, 0 to 25: (Zero) Setting, 25 to 1000: Flatness of anvils: Parallelism of anvils: Spindle alignment: BS 959:2008 5 to 900	1.0 1.1 on diameter 1.6 1.8 1 + (7.0 x length in m) 1.0 2.0 between any two points 1.0 1.0 + (8.0 x length in m) 0.58 1.0 10 Heads 2.0 between any two points. Setting and extension rods 1.0 + (8.0 x length in m)	Using a length measuring machine. Using a length measuring machine or by comparison with end standards. Instrument entries in this section of the schedule also cover digital and dial type gauges which are calibrated based on the quoted standards.	East Kilbride



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
Depth micrometer	Based on BS 6468:2008 0 to 300 Heads: Setting and Extension rods: Base Flatness: Rod Flatness: Parallelism: Rod axis of rotation: Squareness of Face to spindle / rod axis: Rod axis to datum face:	2.0 1.0 + (8.0 x length in m) 0.40 0.58 3.0 10 5.0 10		East Kilbride
Vernier type gauges including dial and digital				
Caliper	BS 887:2008 0 to 1000 Overall performance Flatness Parallelism Squareness Width of internal jaws	10 + (30 x length in m) 3.5 3.5 5.5 3.5		
Height	BS 1643:2008 0 to 1000 Overall performance	10 + (30 x length in m)		
Depth	BS 6365:2008 0 to 600 Overall performance	10 + (30 x length in m)		
Dial gauges and dial test indicators	BS 2795:1981, BS 907:2008 and above. 0 to 25 0 to 50 0 to 75 0 to 100	1.1 1.3 1.6 1.8		
FORM Surface plates & tables Granite and Cast iron	BS 817:2008 (and above) 160 x 100 to 4000 x 4000 Flatness of working surface (Note 1): Local variation of working surface:	1.5 + (0.80 x diagonal in m) 2.2	Note 1) The uncertainty quoted is for the departure from flatness, straightness, or squareness, i.e. the distance separating the two parallel planes, which just enclose the surface under consideration.	East Kilbride and site calibration



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
ELECTRICAL CALIBRATION				Belfast and site calibration
DC Voltage				
Measurement	0 V to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV 40 kV to 60 kV	12 μ V/V + 0.50 μ V 10 μ V/V + 1.1 μ V 11 μ V/V + 10 μ V 12 μ V/V + 110 μ V 12 μ V/V + 1.1 mV 0.24 % 0.66 kV		
Generation	0 mV to 30 mV 30 to 300 mV 300 mV to 3 V 3 V to 30 V 30 V to 300 V 300 V to 1000 V 1 kV to 40 kV 40 kV to 60 kV	70 μ V/V + 3.6 μ V 70 μ V/V + 4.6 μ V 58 μ V/V + 16 μ V 58 μ V/V + 190 μ V 64 μ V/V + 2.5 mV 64 μ V/V + 8.3 V 0.24 % 0.66 kV		
DC Current				
Measurement	0 μ A to 1 μ A 1 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 60 A 60 A to 300A	63 μ A/A + 54 pA 25 μ A/A + 71 pA 25 μ A/A + 0.42 nA 25 μ A/A + 7.0 nA 25 μ A/A + 70 nA 42 μ A/A + 0.70 μ A 130 μ A/A + 19 μ A 0.20 % 0.61%		
Generation	0 μ A to 3 mA 3 mA to 30 mA 30 mA to 300 mA 300 mA to 2 A 2 A to 10 A 10 A to 60 A 60 A to 550 A 550 A to 1025 A	150 μ A/A + 95 nA 120 μ A/A + 0.82 μ A 120 μ A/A + 8.5 μ A 350 μ A/V + 110 μ A 700 μ A/A + 0.90 mA 0.20 % 0.50 % 1.0 %	Appropriate for the calibration of clamp-on ammeters	



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
DC Resistance Measurement	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	19 $\mu\Omega/\Omega$ + 58 $\mu\Omega$ 15 $\mu\Omega/\Omega$ + 0.60 m Ω 13 $\mu\Omega/\Omega$ + 0.60 m Ω 16 $\mu\Omega/\Omega$ + 6.0 m Ω 13 $\mu\Omega/\Omega$ + 58 m Ω 20 $\mu\Omega/\Omega$ + 3.0 Ω 61 $\mu\Omega/\Omega$ + 120 Ω 700 $\mu\Omega/\Omega$ + 1.2 k Ω 0.65 % + 12 k Ω		Belfast and site calibration
Generation				
Specific Values	100 $\mu\Omega$ 1 m Ω 10 m Ω 100 m Ω 0.25 Ω 0.3 Ω 0.4 Ω 0.5 Ω 0.6 Ω 0.7 Ω 0.8 Ω 0.9 Ω 1 Ω 2 Ω 4 Ω 6 Ω 8 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1 G Ω 10 G Ω 100 G Ω 1 T Ω	170 n Ω 0.21 $\mu\Omega$ 4.0 $\mu\Omega$ 92 $\mu\Omega$ 64 m Ω 33 m Ω 34 m Ω 35 m Ω 36 m Ω 37 m Ω 39 m Ω 40 m Ω 75 $\mu\Omega$ 52 m Ω 75 m Ω 99 m Ω 120 m Ω 670 $\mu\Omega$ 7.0 m Ω 120 m Ω 1.2 Ω 12 Ω 1.6 % 0.060 % 0.15 % 0.13 % 0.40 % 2.7 % 1.7 %		



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
DC Resistance (continued) Generation (continued) Other values	0 Ω to 10 Ω 10 Ω to 30 Ω 30 Ω to 100 Ω 100 Ω to 300 Ω 300 Ω to 1 k Ω 1 k Ω to 3 k Ω 3 k Ω to 10 k Ω 10 k Ω to 30 k Ω 30 k Ω to 100 k Ω 100 k Ω to 300 k Ω 300 k Ω to 1 M Ω 1 M Ω to 3 M Ω 3 M Ω to 10 M Ω 10 M Ω to 30 M Ω 30 M Ω to 100 M Ω 100 M Ω to 300 M Ω	140 $\mu\Omega/\Omega$ + 9.5 m Ω 140 $\mu\Omega/\Omega$ + 18 m Ω 100 $\mu\Omega/\Omega$ + 18 m Ω 100 $\mu\Omega/\Omega$ + 18 m Ω 100 $\mu\Omega/\Omega$ + 71 m Ω 100 $\mu\Omega/\Omega$ + 75 m Ω 100 $\mu\Omega/\Omega$ + 0.70 Ω 100 $\mu\Omega/\Omega$ + 0.75 Ω 130 $\mu\Omega/\Omega$ + 7.0 Ω 140 $\mu\Omega/\Omega$ + 8.5 Ω 150 $\mu\Omega/\Omega$ + 67 Ω 150 $\mu\Omega/\Omega$ + 99 Ω 690 $\mu\Omega/\Omega$ + 700 Ω 0.12 % + 1.7 k Ω 0.58 % + 16 k Ω 0.58 % + 120 k Ω	<div>The CMCs shown are for 4-terminal measurements. The quoted uncertainty may be increased for 2-terminal measurements</div> <div>2-terminal measurements</div>	Belfast and site calibration
AC Voltage Measurement	40 Hz to 1 kHz 1 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 700 V 700 V to 1000 V 1 kHz to 100 kHz 100 mV to 1 V 1 V to 10 V 10 V to 100 V 1 kHz to 20 kHz 100 V to 700 V 50 Hz 1 kV to 28 kV 28 kV to 50 kV	240 $\mu\text{V/V}$ + 1.5 μV 82 $\mu\text{V/V}$ + 5.2 μV 80 $\mu\text{V/V}$ + 46 μV 81 $\mu\text{V/V}$ + 2.4 mV 230 $\mu\text{V/V}$ + 5.3 mV 500 $\mu\text{V/V}$ + 30 mV 0.13 % + 120 mV 930 $\mu\text{V/V}$ + 43 μV 930 $\mu\text{V/V}$ + 0.40 mV 0.14 % + 5.0 mV 700 $\mu\text{V/V}$ + 30 mV 0.50 % 0.69 kV		



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
AC Voltage (continued) Generation	<i>10 Hz to 45 Hz</i> 1 mV to 30 mV 30 mV to 300 mV 300 mV to 3 V 3 V to 30 V <i>45 Hz to 1 kHz</i> 300 V to 1000 V <i>1 kHz to 5 kHz</i> 300 V to 1000 V <i>5 kHz to 10 kHz</i> 300 V to 1000 V <i>45 Hz to 10 kHz</i> 1 mV to 30 mV 30 mV to 300 mV 300 mV to 3 V 3 V to 30 V 30 V to 300 V <i>10 kHz to 20 kHz</i> 30 V to 300 V <i>10 kHz to 100 kHz</i> 1 mV to 30 mV <i>10 kHz to 50 kHz</i> 30 mV to 300 mV 300 mV to 3 V 3 V to 30 V <i>50 kHz to 100 kHz</i> 30 mV to 300 mV 300 mV to 3 V 3 V to 30 V <i>100 kHz to 500 kHz</i> 1 mV to 30 mV 30 mV to 300 mV 300 mV to 3 V <i>50 Hz</i> 1 kV to 28 kV 28 kV to 50 kV	0.40 % + 23 μ V 0.29 % + 59 μ V 0.17 % + 420 μ V 0.18 % + 3.2 mV 0.060 % + 98 mV 0.23 % + 32 mV 0.23 % + 32 mV 0.18 % + 23 μ V 0.060 % + 25 μ V 0.040 % + 110 μ V 0.050 % + 1.1 mV 0.10 % + 3.2 mV 0.10 % + 3.2 mV 0.40 % + 39 μ V 0.19 % + 47 μ V 0.16 % + 360 μ V 0.22 % + 5.9 mV 0.28 % + 200 μ V 0.28 % + 2.0 mV 0.28 % + 20 mV 1.2 % + 70 μ V 0.81 % + 380 μ V 0.63 % + 3.9 mV 0.50 % 0.69 kV		Belfast and site calibration



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
AC Current Measurement	<i>45 Hz to 1 kHz</i> 5 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	700 μ A/A + 48 nA 700 μ A/A + 400 nA 700 μ A/A + 2.6 μ A 400 μ A/A + 26 μ A 0.12 % + 280 μ A		Belfast and site calibration
	<i>1 kHz to 5 kHz</i> 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	700 μ A/A + 2.6 μ A 400 μ A/A + 26 μ A 0.12 % + 24 mA		
	<i>50 Hz</i> 10 A to 100 A	1.1 %		
Generation	<i>10 Hz to 5 kHz</i> 29 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA	0.47 % + 200 nA 0.23 % + 380 nA 0.23 % + 3.8 μ A 0.24 % + 38 μ A		
	<i>5 kHz to 10 kHz</i> 29 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10mA to 100 mA	1.5 % + 210 nA 0.70 % + 450 nA 0.70 % + 5.8 μ A 0.70 % + 58 μ A		
	<i>10 Hz to 1 kHz</i> 100 mA to 1 A	0.12 % + 390 μ A		
	<i>1 k Hz to 5 kHz</i> 100 mA to 1 A	0.87 % + 420 μ A		
	<i>45 Hz to 1 kHz</i> 1 A to 11 A	0.39 % + 4.7 mA		
	<i>50 Hz</i> 10 A to 100 A	1.1 %		
	<i>45 Hz to 1 kHz</i> 11 A to 16.5 A 16.5 A to 550 A 550 A to 1025 A	1.5 % 0.50 % 1.1 %	<div></div> For the calibration of clamp-on ammeters	



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
AC Power	45 Hz to 400 Hz 1 W to 11 kW	0.41 %	At unity power factor only	Belfast and site calibration
AC phase angle	50 Hz 0° to 360°	0.28°	Voltage : Current Phase	
Tachometer calibration	600 rpm to 50000 rpm	0.60 rpm		
Inductance				
Generation	At 1 kHz: 100 µH 1 mH 10 mH 100 mH 1 H 10 H	2.6 µH 26 µH 380 µH 2.2 mH 22 mH 98 mH		
Capacitance				
Generation	At 5 kHz: 400 pF	13 pF		
	At 1 kHz: 1 nF 2 nF 3 nF 10 nF 20 nF 30 nF 100 nF 200 nF 300 nF 1 µF	17 pF 23 pF 29 pF 70 pF 180 pF 210 pF 420 pF 960 pF 1.3 nF 4.2 nF		
	At 100 Hz: 2 µF 3 µF 10 µF 20 µF 30 µF	11 nF 15 nF 53 nF 130 nF 180 nF		
	At 50 Hz: 100 µF 200 µF 300 µF	0.72 µF 2.0 µF 3.8 µF		



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code	
EQUIPMENT FOR IEE 17 TH EDITION WIRING TESTING					
<u>RCD Testers</u>					
Trip Time	20 ms to 390 ms 390 ms to 900 ms	0.90 ms 8.2 ms	Other, similar, equipment that includes the same quantities and ranges can also be calibrated.	Belfast and site calibration	
Trip Current	At 50 Hz: 10 mA 30 mA 100 mA 300 mA 1 A 2 A	0.80 mA 2.0 mA 6.0 mA 20 mA 60 mA 120 mA	For trip times < 200 ms		
	At 50 Hz: 10 mA 30 mA 100 mA 300 mA 1 A 2 A	0.80 mA 2.0 mA 6.0 mA 20 mA 60 mA 120 mA 0.80 mA	For trip times > 200 ms		
<u>Loop Testers</u>					
AC Resistance	At 50 Hz: 0.05 Ω 0.1 Ω 0.22 Ω 0.33 Ω 0.5 Ω 1 Ω 5 Ω 10 Ω 100 Ω 1 kΩ	5.0 mΩ 6.0 mΩ 6.0 mΩ 6.7 mΩ 7.7 mΩ 11 mΩ 34 mΩ 63 mΩ 600 mΩ 5.8 Ω			
<u>Continuity Testers</u>					
DC Resistance	See <i>Specific Values</i> on Page 13.				
DC Continuity Current	10 mA to 100 mA 100 mA to 200 mA 200 mA to 300 mA	2.2 mA 4.0 mA 5.3 mA			



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
<u>Insulation Testers</u>				Belfast and site calibration
DC Resistance	See <i>Specific Values</i> on Page 14.			
DC Voltage	50 V 100 V 250 V 500 V 1000 V 1000 V to 10 kV	1.8 V 2.3 V 3.1 V 6.8 V 13 V See <i>DC Voltage</i> measurement capability on Page 12.		
<u>Portable Appliance Testers</u>				
Earth bond resistance	<i>At 50 Hz:</i> 40 mΩ 100 mΩ 200 mΩ 500 mΩ 1 Ω 5 Ω 10 Ω 100 Ω 1 kΩ	5.0 mΩ 5.3 mΩ 6.0 mΩ 7.7 mΩ 11 mΩ 34 mΩ 63 mΩ 0.60 Ω 5.8 Ω		
Earth bond current	<i>At 50 Hz:</i> 0 mA to 100 mA 100 mA to 300 mA 1 A to 30 A	4.7 mA 9.0 mA 420 mA		
Leakage current	<i>At 50 Hz:</i> 2 mA 4.7 mA 7.7 mA	15 μA 25 μA 59 μA		
Flash test	<i>At 50 Hz:</i> 1 kV to 7 kV	See <i>AC Voltage</i> measurement capability on Page 4.		
Load test	<i>At 50 Hz:</i> 3 kVA	2.0 %		
OSCILLOSCOPE CALIBRATION				
Vertical coefficients	5 mV to 50 V p-p	0.47 %		
Horizontal coefficients	2 ns to 1 μs 1 μs to 5 s	29 us/s 32 us/s		



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
ELECTRICAL SIMULATION OF TEMPERATURE				
<u>Calibration of thermocouple indicators</u>			Including cold junction compensation	
Thermocouple type				
Type B	600 °C to 800 °C 800 °C to 1820 °C	1.1 °C 1.1 °C		
Type C	0 °C to 150 °C 150 °C to 650 °C 650 °C to 1000 °C 1000 °C to 1800 °C 1800 °C to 2316 °C	0.55 °C 0.58 °C 0.67 °C 0.76 °C 1.2 °C		
Type E	-250 °C to -100 °C -100 °C to +650 °C 650 °C to 1000 °C	0.59 °C 0.24 °C 0.29 °C		
Type J	-210 °C to -100 °C -100 °C to -30 °C -30 °C to +150 °C 150 °C to 760 °C 760 °C to 1200 °C	0.36 °C 0.24 °C 0.22 °C 0.25 °C 0.32 °C		
Type K	-200 °C to -100 °C -100 °C to -25 °C -25 °C to +120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.41 °C 0.26 °C 0.24 °C 0.34 °C 0.49 °C		
Type N	-200 °C to -100 °C -100 °C to -25 °C -25 °C to +410 °C 410 °C to 1300 °C	0.53 °C 0.35 °C 0.32 °C 0.39 °C		
Type R	0 °C to 250 °C 250 °C to 1767 °C	1.2 °C 1.1 °C		
Type S	0 °C to 250 °C 250 °C to 1767 °C	1.2 °C 1.1 °C		
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.73 °C 0.31 °C 0.24 °C 0.22 °C		

Belfast and site calibration



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
Calibration of RTD indicators	-200 °C to +800 °C	0.0031 °C		Belfast and site calibration
Temperature of reference junction / Cold junction compensation	At ambient temperature of 20 °C ± 3 °C, or nominal 0°C	0.030 °C		
END				



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Instrument Repairs and Calibration
A trading company of IRC Limited
Issue No: 019 Issue date: 16 May 2024

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$