


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING</p> <p><b>4696</b></p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<p><b>Athena Environmental Solutions Limited</b></p> <p><b>Issue No: 008      Issue date: 30 April 2026</b></p>	
	<p><b>1 Walls Yard</b> <b>Riverside Industrial Estate</b> <b>Rochford</b> <b>SS4 1GR</b></p>	<p><b>Contact: Mr Brett Hopson</b> <b>Tel: +44 (0)1268 761171</b> <b>E-Mail: brett@athena-env.co.uk</b> <b>Website: www.athenaasbestosanalysis.com</b></p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos</p>	<p><u>Health and Hygiene</u></p> <p>Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite</p>	<p>Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021 Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)</p> <p>Documented In-House Method MAN 002 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248</p>
<p>END</p>		