

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 4707 Accredited to ISO/IEC 17025:2017	Leftfield Environmental Limited	
	Issue No: 011 Issue date: 19 February 2025	
	Laboratory 4 The Innovation Centre Baglan Energy Park Central Avenue Port Talbot SA12 7AX	Contact: Mr Karl Chambers Tel: +44 (0)1639 897700 E-Mail: karl@leftfield-environmental.com Web: www.leftfield-environmental.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Laboratory 4 The Innovation Centre Baglan Energy Park Central Avenue Port Talbot SA12 7AX	Local contact Mr Karl Chambers Tel: +44 (0)1639 897700 E-Mail: karl@leftfield-environmental.com Web: www.leftfield-environmental.com	Health and Hygiene Asbestos – All Support Functions A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



4707
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Leftfield Environmental Limited
Issue No: 011 **Issue date:** 19 February 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method LSOP6 based on HSG 248	A, B
	Fibre counting	Documented In-House Method LSOP6, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method LSOP6, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method LSOP1 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A

END