

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 7691 Accredited to ISO/IEC 17025:2017	North Star Environmental Ltd	
	Issue No: 013 Issue date: 29 May 2024	
	Navigation House Rowleys Park Evans Way Shotton CH5 1QJ	Contact: Mr Karak Burns Tel: +44 (0)151 538 3141 E-Mail: info@northstarencvironmental.co.uk Website: www.northstarencvironmental.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Navigation House Rowleys Park Evans Way Shotton CH5 1QJ	Health and Hygiene Asbestos – All Support Functions	A
Site Laboratory The Medical Centre Springfields Fuels Ltd Salwick Preston PR4 0XJ	Health and Hygiene	D

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



7691
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

North Star Environmental Ltd
Issue No: 013 **Issue date:** 29 May 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method NSDOC 47 based on HSG 248	A, B
	Fibre counting	Documented In-House Method NSDOC 46, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method NSDOC 48, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method NSDOC 62 Asbestos in Bulk Materials using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, C, D



7691
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

North Star Environmental Ltd
Issue No: 013 Issue date: 29 May 2024

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN SOILS – The Identification of Asbestos fibres in bulk samples of Soil, <i>specifically:</i> <i>Soil</i> <i>Ballast</i> <i>Aggregate</i> <i>Rail Ballast</i> <i>Slurry</i> <i>Sediment</i> <i>Made Ground</i>	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No. NSDOC169 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, C
ASBESTOS IN SOILS – The Identification and Quantification of Asbestos fibres in bulk samples of Soil, <i>specifically:</i> <i>Soil</i> <i>Ballast</i> <i>Aggregate</i> <i>Rail Ballast</i> <i>Slurry</i> <i>Sediment</i> <i>Made Ground</i>	Identification and Quantification of Asbestos content of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No. NSDOC169 for identification using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248. Documented In-House Method No. NSDOC169 for quantification of asbestos.	A

END