


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 7783 Accredited to ISO/IEC 17025:2017	Davidson and Hardy (Laboratory Supplies) Limited	
	Issue No: 013 Issue date: 10 July 2025	
	453 – 455 Antrim Road Belfast BT15 3BL	Contact: Mr J Featherstone Tel: +44 (0)7501 255405 Fax: +44 (0)28 9037 0115 E-Mail: joe@dhlab.com Website: www.dhlab.com
Calibration performed by the Organisation at the locations specified below		

Location details	Activity	Location code
Customer Premises – Any The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Local contact Mr J Featherstone (as above)	Electrical Temperature Rotational speed Non Automatic Weighing Instruments
		S



7783
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Davidson and Hardy (Laboratory Supplies), Limited
Issue No: 013 Issue date: 10 July 2025

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL Time Interval	10 s to 10 hours	1.0 s		S
TEMPERATURE Temperature controlled autoclaves, sterilizers, media preparators, incubators, environmental chambers, fridges, freezers and rooms (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	- 90 °C to - 30 °C -30°C to 100 °C 100 °C to 140 °C	0.30 °C 0.11 °C 0.14 °C	Single and multipoint monitoring probes. Time dependent temperature profiling.	S
Liquid baths	20 °C to 100 °C	0.42 °C	With reference thermometer	S
Digital Indicators with Probes	-90 °C to -30 °C -30 °C to 100 °C 100 °C to 140 C	0.15 °C 0.07 °C 0.08 °C	Calibration within a dry metal block	S
ROTATIONAL SPEED Centrifuge	0 rpm to 6000 rpm 6000 rpm to 30000 rpm 30000 rpm to 60000 rpm	2.0 rpm 4.0 rpm 5.0 rpm	Calibration with a digital tachometer	S



7783
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Davidson and Hardy (Laboratory Supplies), Limited
Issue No: 013 Issue date: 10 July 2025

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
WEIGHING INSTRUMENTS				
Non-automatic weighing instruments (Electronic self indicating)	50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg	0.009 mg 0.010 mg 0.012 mg 0.014 mg 0.016 mg 0.021 mg 0.027 mg 0.037 mg 0.046 mg 0.055 mg 0.084 mg 0.31 mg 0.76 mg 1.5 mg 3.1 mg 7.6 mg 8.9 mg	1. Weights are available in OIML classes: E2 from 1 mg to 100 g, Max. grouped load 200 g F1 from 1 mg to 5 kg, Max. grouped load 10 kg 2. Other loads within the overall listed range may also be used. Method based on the requirements of Euramet CG-18.	S
END				



7783
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Davidson and Hardy (Laboratory Supplies), Limited
Issue No: 013 Issue date: 10 July 2025

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$