


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 8146 Accredited to ISO 15189:2022	Northern Care Alliance NHS Foundation Trust	
	Issue No: 012 Issue date: 17 April 2025	
	Haematology Laboratory Royal Oldham Hospital Rochdale Road OL1 2JD	Contact: Heather Lees Tel: +44 (0)161 656 1678 E-Mail: Heather.Lees@nca.nhs.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Haematology/BT Laboratory Royal Oldham Hospital Rochdale Road OL1 2JH Local contact Michael Heaton	Diagnostic testing	A
Address Essential Service Lab Fairfield General Hospital Rochdale Old Road Bury BL9 7TD Local Contact Michael Heaton	Diagnostic testing	C



8146

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Northern Care Alliance NHS Foundation Trust

Issue No: 012 Issue date: 17 April 2025

Testing performed by the Organisation at the locations specified

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Blood issue fridge Blood Transfusion Department, Pathology Phase 1 Theatre Central Delivery Unit	Storage and issue of blood and blood products	A
Blood Issue Fridge Essential Service Laboratory	Storage and issue of blood and blood products	C
Blood Issue Fridge Level B Blood Fridge Room	Storage and issue of blood and blood products	Rochdale Infirmary Whitehall Street Rochdale OL12 0NB

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Northern Care Alliance NHS Foundation Trust

Issue No: 012 Issue date: 17 April 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

[illegible]



8146

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Northern Care Alliance NHS Foundation Trust

Issue No: 012 Issue date: 17 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Haematology examination for the purposes of clinical diagnosis (cont'd)</u>	In-house documented procedures based on equipment manuals as relevant	
Blood	Blood film assessment for detection of abnormal cell types, morphology and numbers	Blood film assessment using SP10 automated slide maker stainer and May Grunwald and Giemsa stain examined by light microscopy. HAE-SOP-SYS-2 (SP-10 TROH)	A
Bone marrow	Bone marrow assessment for detection of abnormal cell types, morphology and numbers	HAE-SOP-GEN-6 -microscopes HAE-SOP-GEN-7-film examination	A,
Bone Marrow	Iron stain	Bone marrow assessment using RA Lamb Stainmate and May Grunwald and Giemsa stains and examined by light microscopy SOP HAE-SOP-GEN-14, HAE-SOP-LabCentre-6	A
Blood Plasma	Factor V Leiden Prothrombin Gene Variant	Manual Perl's Prussian Blue stain examined by light microscopy SOP HAE-SOP-GEN-13	A
Blood	Antenatal Haemoglobinopathy screening HbA2/F/S	Cepheid GeneXpert PCR analyser SOP HAE-SOP-COAG-35	A
Blood	Haemoglobinopathy detection	Documented in-house methods to meet the requirements of the Sickle Cell and Thalassemia screening programme(s) as defined in the July 2018 'NHS Sickle Cell and thalassemia screening: Laboratory QA evidence requirements.	A
Blood	Sickle cell screen	BioRad Variant II using HPLC SOP HAE-SOP-HPLC-1, 2, 3	A
		BioRad Variant II using HPLC SOP HAE-SOP-HPLC-1, 2, 3	A
		Streck Sickledex qualitative solubility test SOP HAE-SOP-HPLC-4	A, C



8146

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Northern Care Alliance NHS Foundation Trust

Issue No: 012 Issue date: 17 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Haematology examination for the purposes of clinical diagnosis (cont'd)</u>	In-house documented procedures based on equipment manuals as relevant	
Blood	G6PD screen	G6PD Trinity Biotech using red cell enzyme rate reaction SOP HAE-SOP-HPLC-7	A
Blood	Malaria screen	MGG stained thin films prepared using SP10 automated slide maker stainer (A) and RAL staining machine (C) for preliminary malaria screen: HAE-SOP-GEN-6 -microscopes HAE-SOP-GEN-7-film examination	A, C
Blood	Malarial confirmation and species identification <ul style="list-style-type: none"> • P.Falciparum • Non-falciparum IM screen	BinaxNOW malaria kit using immunochromatographic membrane assay SOP HAE-SOP-GEN-3 Clearview IM kit using immunoassay SOP HAE-SOP-GEN-1	A, C
Blood	<u>Blood Transfusion</u> Cross match Direct antiglobulin test	Ortho Vision analyser using column agglutination technology SOP BT-SOP-TRANS-20 Ortho Vision analyser using column agglutination technology SOP BT-SOP-TRANS-21	A, C A, C



8146

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Northern Care Alliance NHS Foundation Trust

Issue No: 012 Issue date: 17 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	Haematology examination for the purposes of clinical diagnosis (cont'd)	In-house documented procedures based on equipment manuals as relevant	
Blood	Antibody panels of the following systems: Rh-hr, KELL, DUFFY, KIDD, LEWIS, MNS, P and LUTHERAN	Ortho Vision analyser using column agglutination technology Resolve Panel C and Resolve Panel B SOP BT-SOP-TRANS-5 Biovue Manual antibody screen BT SOP TRANS 24	A, C
Blood	Blood Group, A, B, O and AB groups Rh groups, antigen D	Ortho Vision analyser using column agglutination technology SOP BT-SOP-TRANS-19	A, C
Blood	Kleihauer	CLIN-TECH LTD Shepard's stain kit using acid elution. SOP BT-SOP-TRANS-27	A
END			