

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



8321

Accredited to
ISO 15189:2012

Magna Laboratories Ltd

Issue No: 007 Issue date: 31 January 2024

Unit 5

Chase Industrial Estate

Alton Road

Ross-on-Wye

HR9 5WA

United Kingdom

Contact: Patricia Collins

Tel: +44 (0) 1989763333

Fax: +44 (0) 1989763533

E-Mail: info@magnalabs.co.uk

Website: www.magnalabs.co.uk

Testing performed at the above address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Human whole blood (for the purpose of monitoring the impact of clozapine on patients)	Full blood counts, which include the following parameters: White Cell Count (WBC) Haemoglobin (HGB) Red Cell Count (RBC) Haematocrit (Hct) Mean Cell Volume (MCV) Mean Cell Haemoglobin (MCH) Mean cell Haemoglobin Concentration (MCHC) Platelets (Plts) Neutrophils (Neut) Lymphocytes (Lymph) Monocytes (Mono) Eosinophils (Eos) Basophils (Baso)	Sysmex XN-550 T.S108 T.S109 T.S110 T.S111 T.S112 T.S113 T.S114
Human Plasma	Determination of the level of Clozapine and Nor-Clozapine in Plasma	Protein crash by Protein Precipitation (T.S203) or Filter Plates (T.S208) followed by Liquid Chromatography Mass Spectrometry using the Agilent 1290 LC with a Sciex 6500 Qtrap mass spectrometer
Blood / plasma	Therapeutic Drug Monitoring: Olanzapine, Risperidone, 9-hydroxyrisperidone	Sciex 6500 LC-MS/MS and phospholipid extraction plate. SOP: T.S211 Determination of olanzapine, risperidone and 9-hydroxyrisperidone in plasma by filter plate extraction and LC-MS



8321
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Magna Laboratories Ltd
Issue No: 007 Issue date: 31 January 2024

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Blood / plasma	Therapeutic Drug Monitoring: Aripiprazole, Dehydroaripiprazole	Sciex 6500 LC-MS/MS and phospholipid extraction plate. SOP T.S213 Determination of aripiprazole and dehydroaripiprazole by filter plate extraction and LC-MS.
Blood / plasma	Therapeutic Drug Monitoring: Haloperidol	Sciex 6500 LC-MS/MS and phospholipid extraction plate. T.S210 Determination of haloperidol by filter plate extraction and LC-MS.
Blood / plasma	Therapeutic Drug Monitoring: Quetiapine, Nor-quetiapine	Sciex 6500 LC-MS/MS and phospholipid extraction plate. T.S212 Determination of quetiapine and norquetiapine by filter plate extraction and LC-MS.
END		