


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> MEDICAL</p> <p>8331</p> <p>Accredited to ISO 15189:2022</p>	<h3>Northern Care Alliance NHS Foundation Trust</h3> <p>Issue No: 012    Issue date: 19 March 2026</p>	
	<p><b>Salford Royal Hospital</b> Stott Lane Salford M6 8HD United Kingdom</p>	<p><b>Contact: Louise Bell</b> Tel: +44 (0)161 206 5016 E-Mail: <a href="mailto:louise.bell@nca.nhs.uk">louise.bell@nca.nhs.uk</a> Website: <a href="http://www.northernalliance.nhs.uk">http://www.northernalliance.nhs.uk</a></p>
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Salford Royal Hospital Stott Lane Salford M6 8HD United Kingdom</p> <p><b>Local contact:</b> Louise Bell</p>	<p>Routine automated Chemistry General / Special Chemistry Toxicology</p>	A
<p>Department of Biochemistry Essential Services Laboratory Wigan Royal Albert Infirmary Wigan Lane Wigan WN1 2NN United Kingdom</p> <p><b>Local contact:</b> Louise Bell</p>	<p>Routine automated Chemistry</p>	B
<p>Thomas Linacre Outpatient Centre Parson's Walk Wigan WN1 1RU United Kingdom</p> <p><b>Local contact:</b> Louise Bell</p>	<p>Specimen reception activities only</p>	C
<p>Leigh Infirmary The Avenue Leigh WN7 1HS United Kingdom</p> <p><b>Local contact:</b> Louise Bell</p>	<p>Specimen reception activities only</p>	D



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u>	In house documented procedures based on equipment manuals and standard methods as specified below:  Siemens Atellica IM by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	
Serum	AFP	Using chemiluminescent 2-site sandwich IA	A
Serum	Anti-TPO antibodies	Chemiluminescent competitive IA	A
	Beta-HCG	Using chemiluminescent 2-site sandwich IA	A & B
Serum	CA199	Using chemiluminescent 2-site sandwich IA	A
Serum	CA125	Using chemiluminescent 2-site sandwich IA	A
Serum	CA153	Using chemiluminescent 2-site sandwich IA	A
Serum	CEA	Using chemiluminescent 2-site sandwich IA	A
Serum	Cortisol	Chemiluminescent competitive IA	A & B
Serum	DHEA-sulphate	Chemiluminescent competitive IA	A
Serum	Digoxin	Latex enhanced turbidimetry	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:  Siemens Atellica IM by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	
Serum	Ferritin	Using chemiluminescent 2-site sandwich IA	A & B
Serum	Folate	Chemiluminescent competitive IA	A & B
Serum	Follicle Stimulating Hormone	Chemiluminescent 2-site sandwich IA	A
Serum	Free T3	Chemiluminescent competitive IA	A
Serum	Free T4	Chemiluminescent competitive IA	A & B
Serum	Luteinising Hormone	Chemiluminescent 2-site sandwich IA	A
Serum	Oestradiol	Chemiluminescent competitive IA	A
Serum	nt-proBNP	Using chemiluminescent 2-site sandwich IA	A & B
Serum	Progesterone	Chemiluminescent competitive IA	A
Serum	Procalcitonin	Using chemiluminescent 2-site sandwich IA	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:  Siemens Atellica IM by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	
Serum	Prolactin	Chemiluminescent 2-site sandwich IA	A
Serum	PSA	Using chemiluminescent 2-site sandwich IA	A
Serum	PTH	Using chemiluminescent 2-site sandwich IA	A
Serum	SHBG	Using sandwich IA	A
Serum	Testosterone (Male)	Chemiluminescent competitive IA	A
Serum	Thyroid Stimulating Hormone	Chemiluminescent IA	A & B
Serum	High Sensitivity Troponin I	using chemiluminescence 3-site sandwich IA	A & B
Serum	Vitamin B12	Chemiluminescent competitive IA	A & B
Serum	Vitamin D	Competitive IA	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:	
Serum	ACTH	Diasorin Liaison XL by the following measurement principles using SOP's: BIO-S-S-378  Chemiluminescent IA	A
Serum	Growth Hormone	Chemiluminescent IA	A
Serum	IGF-1	One site sandwich IA  Siemens Atellica CH by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	A
Serum & Urine	Chloride	using Potentiometry (Indirect IMT)	A & B
Serum & Urine	Potassium	using Potentiometry (Indirect IMT)	A & B
Serum & Urine	Sodium	using Potentiometry (Indirect IMT)	A & B
Serum	Alpha-1-antitrypsin	Siemens Atellica CH by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47  Immunoturbidimetry	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis: (cont'd)</u>	In house documented procedures based on equipment manuals and standard methods as specified below:  Siemens Atellica CH by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	
Serum	Alanine aminotransferase	using NADH with P-5-P	A & B
Serum	Aspartate aminotransferase	using NADH with P-5-P	A
Serum	Albumin	using Bromocresol Green	A & B
Urine	Albumin	using PEG enhanced immunoturbidimetry	A
Serum	Alkaline phosphatase	using enzymatic IFCC Paranitrophenyl phosphate	A & B
Serum	Ammonia	using Glutamate dehydrogenase	A & B
Serum	Amylase	using enzymatic	A & B
Urine	Amylase	using enzymatic	A
Serum	Angiotensin Converting Enzyme (ACE)	Sentinel diagnostics using FAPGG enzymatic method	A
Serum	Bicarbonate/TOC2	Using spectrophotometry	A & B
Serum	Beta-2-Microglobulin	Using latex turbidimetry	A
Serum	Calcium	using colourimetric arsenazo III	A & B



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:  Siemens Atellica CH by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	
Urine	Calcium	using colourimetric arsenazo III	A
Serum	Carbamazepine	Using PETINIA immunoassay	A
Serum	Ceruloplasmin	Sentinel diagnostics using immunoturbidimetry	A
Serum	Cholesterol	using end point enzymatic/colorimetric	A & B
Serum	Conjugated bilirubin	using colourimetry	A & B
Serum	Creatine kinase	using enzymatic	A & B
Serum	Cystatin C		A
Serum	Enzymatic Creatinine	SOP BIO-S-A-265	A & B
Urine	Enzymatic Creatinine	SOP BIO-S-A-265	A
Serum	C-reactive protein	using latex agglutination	A & B
Serum	Ethanol	using enzymatic	A & B
Urine also	Ethanol	using enzymatic	A
Serum	Fructosamine	Using enzymatic Trinder end point	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis: (cont'd)</u>	In house documented procedures based on equipment manuals and standard methods as specified below:  Siemens Atellica CH by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	
Serum	Gamma GT	using zero-order kinetic with L-γGlutamyl-3-Carboxy-4-Nitroanilide	A
Serum	Gentamicin	using PETINIA	A & B
Serum	Glucose	using hexokinase	A & B
CSF	Glucose	using hexokinase	A
Serum	HDL-cholesterol	Using enzymatic	A & B
Serum	Iron	using colourimetric	A & B
Serum	Immunoglobulin A	using PEG enhanced immunoturbidimetry	A
Serum	Immunoglobulin G	using PEG enhanced immunoturbidimetry	A
Serum	Immunoglobulin M	using PEG enhanced immunoturbidimetry	A
Serum	Lactate	using enzymatic/colorimetric	A & B
CSF	Lactate	using enzymatic/colorimetric	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis: (cont'd)</u>	In house documented procedures based on equipment manuals and standard methods as specified below:  Siemens Atellica CH by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	
Serum	Lactate dehydrogenase	using Lactate to pyruvate in the presence of NAD	A & B
Serum	Lithium	using colourimetric endpoint	A
Serum	Magnesium	using colourimetric	A & B
Urine	Magnesium	using colourimetric	A
Serum	Paracetamol	using colourimetric	A & B
Serum	Phenytoin	using colourimetric	A
Serum	Phosphate	using phosphomolybdate	A & B
Urine	Phosphate	using phosphomolybdate	A
Serum	Salicylate	using enzymatic	A & B
Serum	Theophylline	using PETINIA	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No:** 012 **Issue date:** 19 March 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:  Siemens Atellica CH by the following measurement principles using SOP's: BIO-S-A-375, BIO-S-386, BIO-S-B-47	
Serum	Total bilirubin	using end point Vanadate	A & B
Urine & CSF	Total protein	using colourimetric Biuret	A
Serum	Total protein	using colourimetric Biuret	A & B
Serum	Transferrin	using PEG enhanced immunoturbidimetry	A & B
Serum	Triglycerides	using end point Enzymatic	A & B
Serum	Urate	using uricase	A & B
Urine	Urate	using uricase	A
Serum	Urea	using enzymatic	A & B
Urine	Urea	using enzymatic	A
Serum	Valproate	using PETINIA	A
Serum	Vancomycin	using PETINIA	A & B



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:	
Urine	Porphyryns	HPLC with FLR detection using Waters UPLC SOP BIO-S-G-56	A
Whole Blood	Red Cell Porphyryns	Perkin Elmer FL6500 Fluouresence Spectrophotometer SOP BIO-S-E-74	A
Plasma	Plasma Porphyryns	Perkin Elmer FL6500 Fluouresence Spectrophotometer SOP BIO-S-E-34	A
Faeces	Faecal porphyryns	Northstar BIO-UV spectrophotometer BIO-S-E-75	A
Urine	Porphobilinogen (Qualitative) screen	Manual Ion Exchange Chromatography SOP BIO-S-E-26	A
Urine	Porphobilinogen (PBG) (Quantitative) + $\alpha$ -aminolaevulinic acid (ALA)	Manual Ion Exchange Chromatography SOP BIO-S-E-26	A
Plasma, serum & Urine	Osmolality	Freezing point depression using the Advanced Instruments Osmo 1 SOP BIO-S-A-367	B
Urine and serum	Osmolality	Freezing point depression using the Advanced Instruments Osmo PRO SOP BIO-S-A-367	A
Blood Only	Immunosuppressants: Cyclosporin, Sirolimus, Tacrolimus, Everolimus	Shimadzu 8060NX Mass Spectrometer with Nexera Liquid Chromatography BIO-S-MS-002	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:  Shimadzu 8060NX Mass Spectrometer with Nexera Liquid Chromatography	
Urine only	Androgens: Testosterone (female), Androstenedione	BIO-S-MS-006	A
Urine only	5Hydroxyindoleacetic acid (5HIAA)	BIO-S-MS-003	A
Urine only	Cortisol	BIO-S-MS-004	A
Urine only	Metadrenalines: Metadrenaline, Normetadrenaline	BIO-S-MS-005	A
Urine only	Toxicology Screen; Morphine, Codeine, Dihydrocodeine, 6- Monoacetylmorphine, Acetylcodeine, Buprenorphine, Nor-Buprenorphine, Methadone, EDDP, Nor- Fentanyl, Oxycodone, Tramadol, Amphetamine, MDMA, Metamphetamine, Benzoyllecognine, Pregabalin, Gabapentin, Ketamine, Zopiclone, Cannabis metabolite (THC- COOH), Temazepam, Nordiazepam, Oxazepam, Lorazepam, Nitrazepam, Clonazepam, Alprazolam	BIO-S-MS-009	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:	
Whole Blood only	Blood gases / acid base: PH, pCO <sub>2</sub> , pO <sub>2</sub> , Sodium, Oxygen saturation, Potassium, ionised calcium, bicarbonate, glucose , lactate	Werfen GEM 5000 Blood Gas Analyser Co-oximetry with reference to manufacturer's instructions POCT-SOP-35	A & B
Whole Blood only	Total Hb, Carboxy-Hb, Met-Hb, Oxy-Hb	Werfen GEM 5000 Blood Gas Analyser Co-oximetry with reference to manufacturer's instructions and POCT-SOP-35	A & B
CSFs	Xanthochromia	Northstar BIO-UV spectrophotometer SOP BIO-S-G-25	A
Plasma	Metadrenaline Normetadrenaline	Manual ELISA using documented in-house procedure SOP BIO-S-E-64	A
Faeces	Calprotectin	Immunodiagnostik ELISA using the DS2 SOP-BIO-S-G-42	A
Faeces	Faecal Immunochemical Test	HMJAKarc SOP BIO-S-G-54	A
Serum	Serum Free Light chains	Turbidimetry using the Binding site Optilite SOP BIO-S-I-126	A
Serum	Immunodisplacement	Capillary Electrophoresis using the Helena V8 SOP BIO-S-I-01	A
Urine also	Serum protein and Urine Bence Jones protein (immunoglobulin light chains)	Capillary Electrophoresis using the Helena V8 and SAS3 and SAS4* system SOP BIO-S-I-01 SOP BIO-S-I-02	A



8331  
Accredited to  
ISO 15189:2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northern Care Alliance NHS Foundation Trust**

**Issue No: 012 Issue date: 19 March 2026**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd)	In house documented procedures based on equipment manuals and standard methods as specified below:	
Serum and CSF	Blood Brain Barrier Ratio calculation from: Serum albumin Serum IgG CSF albumin CSF IgG	Turbidimetry using Binding site Optilite SOP BIO-S-I-27	A
Serum and CSF	Oligoclonal Bands -IgG	Helena SAS-3 Semi automated gel electrophoresis system SOP BIO-S-I-23	A
Fluids	Tau Protein (B2 Transferrin)	Helena SAS-3 Semi automated gel electrophoresis system SOP BIO-S-I-22	A
Serum	Trace Metals: Copper, Zinc and Selenium	Agilent 8700 ICPMS equipped with Agilent SPS4 autosampler SOP BIO-S-G-47	A
END			