


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-Upon-Thames TW18 3HR

 Accredited to ISO 15189:2022	NHS Tayside Issue No: 016 Issue date: 02 April 2025	
	Medical Microbiology Ninewells Hospital Ninewells Avenue Dundee DD1 9SY	Contact: Alisson McPherson Tel: +44 (0)1382 660111 ext. 36718 E-Mail: alisson.mcperson@nhs.scot Website: http://www.nhstayside.scot.nhs.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Microbiology Department Ninewells Hospital Ninewells Avenue Dundee DD1 9SY	Local contact As above	Bacteriology Mycology Enteric Parasitology Virology Serology Molecular Microbiology	A
Address Sexual and Reproductive Health Clinic Ninewells Hospital Ninewells Avenue Dundee DD1 9SY	Local contact As above	Plate inoculation and microscopy	B



8610
Accredited to
ISO 15189:20122

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS AND TISSUES	<u>Microbiological examination activities for the purposes of clinical diagnosis</u>	In-house documented procedures based on UKHSA UK SMIs	
	<u>General Bacteriology</u>		
Human samples including swabs, fluids, tissues, urine.	Detection and characterisation of micro-organisms of clinical significance	Use of Gram stain and SOP: NB029 Microscopy and Identification	A
Human samples including swabs, urine	Detection and characterisation of micro-organisms of clinical significance	Use of Gram stain and SOP: NB029 Microscopy and Identification	B
Swabs from eyes, wounds, mouth, nose, throat, ears, genital sites and intrauterine contraceptive devices	General isolation and characterisation of micro-organisms of clinical significance	Manual culture using SOPs: NB002 Exudates Manual NB036 Genital Swabs NB039 Acute Bench	A
CSF	Detection of cryptococcus	Use of India Ink stain and SOP: NB039 Acute Bench	A
Inoculated plates from Site B	General isolation and characterisation of micro-organisms of clinical significance	Processing of inoculated plates using SOPs: NB010 Sexual and Reproductive Health Clinic NB036 Genital Swabs	A
Respiratory fluids (sputum, bronchial washings)	General isolation and characterisation of micro-organisms of clinical significance	Manual culture using SOP: NB003 Respiratory Specimens	A
Swabs from nose, perineum and wound sites	Isolation of MRSA	Manual culture on chromogenic agar using SOP: NB040 MRSA Screens	A
Swabs from rectum and faeces	Isolation of Vancomycin Resistant Enterococcus	Manual culture on chromogenic agar using SOP: NB008 Faeces Manual	A



8610
Accredited to
ISO 15189:20122

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS AND TISSUES (cont'd)	<u>Microbiological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In-house documented procedures based on UKHSA UK SMIs	
	<u>General Bacteriology</u> (cont'd)		
Swabs from various body sites	Isolation of carbapenem resistant Gram negative rods	Screening for and confirmation of carbapenem resistant Gram negative rods using SOPs: NB030 Carbapenem Resistance Confirmation in GNR NB008 Faeces Manual NB072 Use and Maintenance of the Cepheid GeneXpert	A
Intravenous catheters	General isolation and characterisation of micro-organisms of clinical significance	Manual culture using SOPs: NB039 Acute Bench	A
Fluids and tissues from normally sterile sites	General isolation and characterisation of micro-organisms of clinical significance	Manual culture using SOPs: NB038 Orthopaedic Specimens NB039 Acute Bench	A
Fluids from CSF and CAPD fluids	Quantification of white and red blood cells and WBC differential	Manual microscopy using SOPs: NB029 Microscopy and Identification NB039 Acute Bench	A
Urine	Quantification of white and red blood cells, bacteria and epithelial cells	Manual microscopy using SOP: NB007 Urine Manual	A
Urine	General isolation and characterisation of micro-organisms of clinical significance	Manual culture using SOP: NB007 Urine Manual	A



8610
Accredited to
ISO 15189:20122

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS AND TISSUES (cont'd)	<u>Microbiological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In-house documented procedures based on UKHSA UK SMIs	
	<u>General Bacteriology</u> (cont'd)		
Faeces	General isolation and characterisation of <i>Salmonella</i> , <i>Shigella</i> , <i>Campylobacter</i> & <i>E. coli</i> O157 and other enteric pathogens	Manual culture methods using selective agar and SOP: NB008 Faeces Manual	A
Faeces	Detection of <i>Clostridioides difficile</i> GDH antigen and toxins A & B	Screening for C diff GDH and toxins using SOP: NB008 Faeces Manual in conjunction with manufacturer's instructions using DS2 analyser (GDH) and VIDAS (Toxin)	A
Faeces	Detection of <i>Helicobacter pylori</i>	Screening for <i>H pylori</i> faecal antigen using SOP: NB073 <i>Helicobacter pylori</i> Antigen in conjunction with the manufacturer's instructions for Amplified IDEIA HpStAR on DS2 analyser.	A
Blood cultures	Detection of microbial growth and general isolation and characterisation of micro-organisms of clinical significance	Detection of growth in cultures of blood using SOP: NB001 Blood Cultures in conjunction with manufacturer's instructions for automated system BacT/Alert.	A
Bacterial and Yeast cultures (isolated in-house)	Identification and antimicrobial susceptibility testing and determination of minimum inhibitory concentration (MIC) of bacteria	Identification and susceptibility testing using VITEK 2 automated antimicrobial testing system and MALDI-tof and SIRScan SOPs: NB027 Sensitivity Testing NB012 Etest MIC NB013 ESBL Testing NB029 Microscopy and Identification NB030 Carbapenem Resistance NB065 MALDI-tof Procedure NB078 SIRScan	A



8610
Accredited to
ISO 15189:20122

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS AND TISSUES (cont'd)	<u>Microbiological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In-house documented procedures based on UKHSA UK SMIs	
	<u>General Bacteriology</u> (cont'd)		
Yeast cultures (isolated in-house)	Identification of yeasts	Identification of yeasts using VITEK 2 automated testing system and MALDI-tof	A
Bronchoalveolar lavage/washings, sputum	Detection of Mycobacterium Tuberculosis	TBRif Ultra kit via Cepheid PCR, using SOP NB072, NB004	A
	<u>Parasitology</u>		
Human samples including faeces and urine as specified	Examination activities for the purposes of clinical diagnosis	In-house documented procedures based on UKHSA SMIs	
Faeces	Examination for the presence of parasites (ova, cysts, trophozoites and worms) of clinical significance	Screening for and identification of enteric parasites using SOP: NB008 Faeces Manual using concentration where required and manual microscopy.	A
Urine	Detection of schistosome ova	Screening for and identification of urinary tract parasites using SOP: NB007 Urine Manual using concentration where required and manual microscopy.	A
	<u>Mycology</u>		
Human samples including skin scrapings, hair and nail clippings	Examination activities for the purposes of clinical diagnosis	In-house documented procedures based on UKHSA SMIs	
Skin scrapings, hair and nail clippings	General isolation and characterisation of fungi of clinical significance	Detection of pathogenic fungi using SOP: NB006 Mycology manual and direct microscopy of KOH digested samples and culture on non-chromogenic selective media.	A



8610
Accredited to
ISO 15189:20122

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS AND TISSUES (cont'd)	<u>Microbiological examination activities for the purposes of clinical diagnosis (cont'd)</u>	In house documented procedures based on manufacturer's instructions	A
Serum/plasma	<u>Serology</u> Detection of: Hepatitis A IgM Hepatitis A total antibody Hepatitis B anti-HBs Hepatitis B core antibody Hepatitis B surface antigen Hepatitis C antibody HIV antibody/antigen combination Rubella IgG Syphilis antibody	Serology using the Atellica IM (Siemens)) in conjunction with the manufacturer's instructions & SOPs: NV032 Reporting of HBsAg Positives samples from pregnant women NV033 Hepatitis B Testing and Reporting NV037 Hepatitis A Testing and Reporting NV040 Rubella IgG Testing NV115 HIV Minimum Testing Algorithm NV121 Serological Diagnosis of Syphilis NV136 Handling of HIV Requests in Children <18 months of age NV197 Preparation of Dried Blood Spots NV295 – Operation of the Siemens Atellica Solution	



8610
Accredited to
ISO 15189:20122

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS AND TISSUES (cont'd)	<u>Microbiological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures based on manufacturer's instructions	
Serum/plasma	<u>Serology</u> (cont'd) Detection or confirmation of: CMV IgM EBV EBNA IgG EBV VCA IgG EBV VCA IgM Measles IgG Parvovirus B19 IgG Parvovirus B19 IgM VZV IgG	DiaSorin Liaison XL analyser in conjunction with the manufacturer's instructions & SOPs: NV111 Biotrin Parvovirus ELISA NV215 EBV Testing on the Liaison NV217 CMV IgM testing on the Liaison NV227 VZV IgG testing on the Liaison NV252 Measles IgG on the Liaison NV262 Use and Maintenance of the Liaison XL	A
Serum/plasma	Detection or confirmation of: CMV IgG Hepatitis B core antibody Hepatitis B core IgM Hepatitis B e antibody Hepatitis B e antigen Hepatitis B surface antigen Hepatitis C antibody HIV antibody/antigen combination Mumps IgG Rubella IgM Toxoplasma IgM Toxoplasma total antibody	Serology using the BioMerieux VIDAS analyser in conjunction with the manufacturer's instructions and SOPs: NV017 Toxoplasma Total Antibody and IgM NV032 Reporting of HBsAg Positives from ANC Patients in Tayside NV033 Hepatitis B Testing and Reporting NV041 VIDAS HBsAg Confirmation Protocol NV102 Mini-Vidas NV103 Hep C Confirmation Reporting Protocol NV042 Mumps and CMV Vidas Protocol NV115 HIV Minimum Testing Algorithm NV122 Rubella IgM Serology NV136 Handling of HIV Requests in Children <18 months of age NV258 Use and Maintenance of the VIDAS 3 analyser	A



8610
Accredited to
ISO 15189:20122

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS AND TISSUES (cont'd)	<u>Microbiological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures based on manufacturer's instructions	
Dried Blood Spots	<u>Serology</u> (cont'd) Detection of: Hepatitis B Hepatitis C HIV	Serology using SOPs: NV197 Preparation of Dried Blood Spots NV218 Dried Blood Spot HIV, HCV, and HBsAg Testing	A
Serum	Confirmation of HIV	Serology using SOP NV250 Geenius HIV 1 and 2 Confirmatory Assay	A
Serum/plasma	Syphilis confirmation using specific Reagin antibody tests, ASI RPR Reditest	Serology using SOPs: NV014 ASI RPR Reditest NV121 Serological Diagnosis of Syphilis	A
Nasopharyngeal swab/aspirate Bronchoalveolar lavage, sputum	Detection of: Panel 1: <ul style="list-style-type: none"> Influenza virus, Influenza B, Flu A-H3, Flu A-H3, Flu A-H1pdm09 RSV A, RSV B. Panel 2: <ul style="list-style-type: none"> Adenovirus, Enterovirus, MPV, Parainfluenza 1, 2, 3 & 4. Panel 3: <ul style="list-style-type: none"> Bocavirus, Rhinovirus, CoV OC43, CoV NL63, CoV 229E Panel 4: <ul style="list-style-type: none"> M.pneumoniae, S.pneumoniae, H.influenzae 	Allplex respiratory assays via the Seegene Starlet analyser using SOP NV290 Seegene Allplex Respiratory panel assay using Seegene Starlet NV 303 work instruction for triaging samples for the extended respiratory screen	A



8610
Accredited to
ISO 15189:20122

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>HUMAN BODY FLUIDS AND TISSUES (cont'd)</p> <p>Bloods, CSF, Faeces, Fluids, Respiratory aspirates, Swabs (various sites), Urines</p>	<p><u>Microbiological examination activities for the purposes of clinical diagnosis (cont'd)</u></p> <p><u>Molecular Biology</u></p> <p>Detection of: Adenovirus Respiratory BK Virus CMV Coronavirus (x4) EBV Enterovirus Hepatitis B Hepatitis C HIV HSV Influenza A and B Influenza Subtyping Measles Metapneumovirus Mumps <i>Mycoplasma pneumoniae</i> Norovirus Parainfluenza (x4) Parechovirus Rhinovirus RSV VZV</p>	<p>In house documented procedures based on manufacturer's instructions</p> <p>Real-time PCR using BioMerieux Easymag and ABI 7500 thermocyclers and SOPs: NV115 HIV Minimum Testing Algorithm NV186 Preparation of Norovirus RNA Control for PCR NV203 Preparation of BK Standard and DNA control for PCR NV204 Preparation of CMV Standard and DNA Control for PCR NV205 Preparation of DNA Control for HSV, VZV and Adenovirus PCR NV225 Non-respiratory IH PCR NV234 Preparation of EBV Standard and EBV Control NV237 Preparation of Nucleic Acid for Parecho-Entero virus PCR Controls NV255 Preparation of MMR Positive control NV140 Probe titration for Real time PCR NV165 Virus Detection examination Procedure NV163 Preparation of primer/probe mix for inhouse PCR NV266 Preparation of PCR mix for IH PCR NV 146 Use of the ABI 7500 NV 144 Maintenance of the ABI 7500 NV 243 Preparation of internal control for IH PCR NV 257 Verification of Quantitative Standards for PCR NV 254 Preparation of samples for virus Detection</p>	<p>A</p>



8610
Accredited to
ISO 15189:20122


Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Tayside

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY FLUIDS AND TISSUES (cont'd)	<u>Microbiological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures based on manufacturer's instructions	
	<u>Molecular Biology</u> (cont'd)		
Plasma	HCV rRNA Viral Load HIV Viral Load HBV Viral Load	Aptima Viral Load assays on the Hologic Panther platform, using Transcription Mediated Amplification (TMA) SOP NV265 HIV, HCV and HBV Viral Load Testing on the Hologic Panther	A
Urine and swabs Vaginal/vulvovaginal, Endocervical, urethral, rectal and throat	Detection of <i>Chlamydia trachomatis</i> and <i>Neisseria gonorrhoeae</i> rRNA	Aptima Combo 2 assay on the Hologic Panther platform using Transcription Mediated Amplification (TMA) SOP NV266 Chlamydia and N gonorrhoeae Testing using the Aptima Combo Assay on the Hologic Panther. NV268 Maintenance of the Hologic panther NV269 Handling samples for CTNG testing on the Hologic panther	A
Nasopharyngeal combined swab	Detection of SARS-CoV-2 virus	RT-PCR for qualitative detection via Seegene Starlet using SOP NV285.	A
Nasopharyngeal combined swab	Detection of SARS-CoV-2 virus	Aptima SARS-CoV-2 assay combines the technologies of target capture, Transcription Mediated Amplification (TMA), and Dual Kinetic Assay (DKA) via Hologic Panther using SOP NV286.	A



issued by

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Issue No: 016 Issue date: 02 April 2025

Testing performed by the Organisation at the locations specified

END