


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 9086 Accredited to ISO/IEC 17025:2017	Scope iT Ltd Issue No: 012 Issue date: 10 June 2025	
	Unit 14 The Quad Atherleigh Business Park Gibfield Park Avenue Manchester M46 0SY	Contact: Raymond Mailey Tel: +44 (0)194 287 9067 E-Mail: r.mailey@scope-it-group.com Website: www.scope-it-group.com

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Unit 14 The Quad Atherleigh Business Park Gibfield Park Avenue Manchester M46 0SY	Local contact Raymond Mailey Tel: +44 (0)194 287 9067 Email: r.mailey@scope-it-group.com	Health and Hygiene Asbestos – All Support Functions	A
Address Office S Prestwich Industrial Estate Coal Pit Lane Atherton M46 0FY	Local contact Raymond Mailey Tel: +44 (0)194 287 9067 Email: r.mailey@scope-it-group.com	Health and Hygiene Asbestos – Bulk Analysis Support Functions	D

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



9086
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Scope iT Ltd

Issue No: 012 Issue date: 10 June 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method SI014 based on HSG 248	B
	Fibre counting	Documented In-House Method SI014, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
	4 Stage Clearance Process	Documented In-House Method SI014, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
ASBESTOS IN BULK MATERIALS including materials and products suspected to contain asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method SI054 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, D
END			