

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>9152</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>S M Gauge Company Ltd</h3> <p>Issue No: 018 Issue date: 16 June 2026</p>	
	<p>Unit 6 Armstrong Business Park Yate Road Yate Bristol BS37 5AA United Kingdom</p>	<p>Contact: Mr Andrew White Tel: +44 (0)1179 654615 E-Mail: andy@pressuregauge.co.uk Website: www.pressuregauge.co.uk</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
<p>PRESSURE</p> <p>Gas Pressure (Gauge)</p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>-100 kPa to -50 kPa -50 kPa to -25 kPa -25 kPa to -15 kPa -15 kPa to 15 kPa 15 kPa to 25 kPa 25 kPa to 50 kPa 50 kPa to 100 kPa 100 kPa to 400 kPa 400 kPa to 1 MPa 1 MPa to 3.5 MPa 3.5 MPa to 7 MPa 7 MPa to 10 MPa 10 MPa to 16 MPa 16 MPa to 20 MPa</p>	<p>18 Pa 11 Pa 4.4 Pa 2.9 Pa 4.4 Pa 11 Pa 18 Pa 51 Pa 170 Pa 480 Pa 910 Pa 6.5 kPa 5.8 kPa 16 kPa</p>	<p>Calibration of instruments with an electrical output can also be undertaken Methods consistent with EURAMET CG17</p> <p>Calibration by comparison with indication from a reference instrument.</p> <p>Absolute pressure calibrations can be undertaken using associated barometric pressure measurement correction. The uncertainties quoted will be increased by 22 Pa</p>
<p>Hydraulic Pressure (Gauge)</p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>0 kPa to 100 kPa 100 kPa to 250 kPa 250 kPa to 400 kPa 400 kPa to 700 kPa 700 kPa to 1 MPa 1 MPa to 1.6 MPa 1.6 MPa to 2.5 MPa 2.5 MPa to 10 MPa 10 MPa to 300 MPa 300 MPa to 500 MPa</p>	<p>90 Pa 118 Pa 180 Pa 414 Pa 500 Pa 830 Pa 1.1 kPa Q [0.059 %, 1.4 kPa] 0.026 % 0.036 %</p>	<p>Calibration by comparison with indication from a reference instrument up to 2.5 MPa.</p> <p>Above 2.5 MPa pressure is generated by dead weight tester.</p>
<p>Calibration of Oxygen-duty and oil free pressure instruments</p>	<p>0 Pa to 1.6 MPa 1.6 MPa to 4 MPa 4 MPa to 16 MPa 16 MPa to 25 MPa 25 MPa to 40 MPa 40 MPa to 70 MPa</p>	<p>0.97 kPa 2.6 kPa 12 kPa 17 kPa 28 kPa 44 kPa</p>	
<p>END</p>			



9152
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

S M Gauge Company Ltd
Issue No: 018 **Issue date:** 16 June 2026

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$