


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS MEDICAL 9344</p> <p>Accredited to ISO 15189:2022</p>	<p>Unilabs Ltd</p> <p>Issue No: 012 Issue Date: 14 January 2026</p>	
	<p>24-32 Stephenson Way London NW1 2HD</p>	<p>Contact: Ms Jayne Holloway Tel: +44 (0)207 299 4490 Fax: +44 (0)207 299 4491 E-Mail: jayne.holloway@unilabs.com Website: http://www.unilabs.co.uk</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
<p>Address 24-32 Stephenson Way London NW1 2HD</p>	<p>Local contact Dhilum Alacon dhilum.alacon@unilabs.com</p>	<p>Histopathology: Special Stains Interpretive microscopy IHC Diagnostic Cytology (Non-Gynae): Interpretive microscopy Digital Pathology Cytology (Gynae) Interpretive microscopy HPV Testing</p> <p>Storage of Records</p>	STEP
<p>Address 100a New Cavendish Street London W1W 6NR</p>	<p>Local contact As above</p>	<p>Biochemistry Haematology Coagulation</p>	CAVE



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 Issue date: 14 January 2026

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>HUMAN BODY TISSUE</p> <p>Fixed, fresh and frozen tissue; excisional and incisional biopsies and surgical resection specimens</p> <p>Formalin fixed tissue Tissue samples (including breast; gastrointestinal; skin; urological; bronchial/lung/thoracic; endocrine including thyroid, parathyroid, adrenal; lymphoreticular including lymph nodes, spleen, and bone marrow trephine biopsies, liver; bone and soft tissues; gynaecological; head and neck.</p>	<p><u>Histopathological examination activities for the purposes of clinical diagnosis</u></p> <p><u>Routine Histopathology</u></p> <p>Examination of tissues to identify or exclude morphological and cytological abnormalities for the purposes of diagnosis</p>	<p>In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified:</p> <p><u>Specimen dissection</u> Manual method using: Thermo printmate Dragon Dictation system/LIS Balances –Ohaus & Denver XP3000 SOP H49, H55, H56, D007</p>	<p>STEP</p>



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Histopathological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	
Formalin fixed tissues samples (as above)		<u>Tissue Processing</u> Thermo Excelsior As processor SOP H23 Tissue processing	STEP
Processed tissue samples (as above)		<u>Tissue embedding</u> Thermo Histostar embedding centres SOP 51	STEP
Paraffin waxed tissue samples (as above)		<u>Tissue sectioning (microtomy)</u> Cutting of 3u sections using: Thermo Finesse microtomes Thermo HM325 rotary microtome Thermo HM255S automatic microtome SOP 51	STEP
Fresh Tissue		<u>Cryotomy</u> Leica Cryostat CM1510 & Leica SOPH 53	STEP



9344
Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Histopathological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	
Frozen Section Tissue on glass slides	H&E staining for tissue architecture and nuclear detail/ Identification of basophilic and eosinophilic structures	<u>H&E staining</u> <u>Manual method</u>	
Formalin Fixed Paraffin Embedded Tissue (FFPE) sections on glass slides		<u>Automated staining</u> Leica Auto Stainer and CV5030 coverslipper Thermo Hyperclean TRUAIR Hood SOP H51	STEP STEP
Formalin Fixed Paraffin Embedded Tissue (FFPE) sections on glass slides	<u>Special staining</u> Examination of tissues to identify or exclude morphological and cytological abnormalities for the purposes of diagnosis (cont'd)	Documented in-house procedures for manual tinctorial hand staining	
	Acid Mucopolysaccharides	SOP H52 Using stains: Alcian blue	STEP
	Acid and Neutral Mucopolysaccharides	Alcian Blue/Periodic Acid Schiff	STEP
	Amyloid	Congo Red	STEP
	Elastic tissue fibres-black Muscle-yellow Collagen-red	EVG - Miller's Elastic Tissue Stain	STEP



9344
Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Histopathological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	
Formalin Fixed Paraffin Embedded Tissue (FFPE) sections on glass slides	<u>Special staining</u> (cont'd)	Documented in-house procedures for manual tinctorial hand staining (cont'd)	
	Gram +/- organisms	Grams stain	STEP
	Fungi and Pneumocystis carinii	Grocott- Hexamine Silver	STEP
	Connective tissue	Haematoxylin Van Gieson	STEP
	Melanin pigment	Masson Fontana	STEP
	Connective tissue	Masson Trichrome	STEP
	Fibrin	MSB	STEP
	Hepatitis B surface antigen	Shikata Orcein	STEP
	PAS +ve substances e.g. fungi, glycogen (and cells of pituitary)	Periodic Acid Schiff with and without diastase	STEP
	Haemosiderin and ferric iron	Perls Prussian Blue	STEP
	Reticulin	Reticulin method – Gordon & Sweets	STEP
	Metachromatic substances	Toluidine Blue	STEP
	Calcium	Von Kossa	STEP
	Tubercle bacilli (TB)	Ziehl-Neelsen	STEP
Formalin Fixed Paraffin Embedded Tissue (FFPE) sections on glass slides	Morphological assessment and interpretation/diagnosis	Microscopy (qualitative analysis) In-house procedures using light microscopy	
		Olympus BX45 microscope SOP H31	STEP
FFPE slides	Scanning and subsequent Morphological assessment and interpretation/diagnosis	Using Digital Pathology- Scanner Philips Intellisite Ultra Fast Scanners SOP H58	STEP



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>HUMAN BODY TISSUE (cont'd)</p> <p>Formalin Fixed Paraffin Embedded (FFPE) human tissue and cell block sections on adhesive glass slides</p>	<p><u>Histological examinations for the purpose of clinical diagnosis</u></p> <p>Examination of tissues to identify or exclude morphological and cytological abnormalities for the purpose of clinical diagnosis</p> <p>Basal cells and Alpha methacyl CoA racemase</p> <p>B cells of the mantle zone and interfollicular T cells</p> <p>B-lymphocytes, mostly located in germinal centers</p> <p>Epithelial Cells</p> <p>Smooth muscle tissue and myoepithelium</p> <p>Smooth muscle</p> <p>Neural tissues, normal and reactive mesothelium, eccrine glands of skin, Sertoli cells of the testis, ovarian stromal cells and adrenal cortical cells</p> <p>Early lymphoid cells, and various nonlymphoid tissues including breast myoepithelial cells, endometrium, bile canaliculi, fibroblasts, kidney tubular brush borders and small intestine epithelium</p>	<p>Documented in-house procedures SOP's IC01 and IC02 in conjunction with manufacturer's instructions for antibodies and using Ventana Benchmark Ultra:</p> <p>Basal Cell (p63+34be12) + Racemase</p> <p>BCL-2</p> <p>BCL-6</p> <p>BEREP4 (Ep-Cam)</p> <p>Calponin-1</p> <p>Caldesmon</p> <p>Calretinin</p> <p>CD10</p>	<p>STEP</p>



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 Issue date: 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Histological examinations for the purpose of clinical diagnosis (cont'd)</u>		
Formalin Fixed Paraffin Embedded (FFPE) human tissue and cell block sections on adhesive glass slides (cont'd)	Examination of tissues to identify or exclude morphological and cytological abnormalities for the purpose of clinical diagnosis(cont'd)	Documented in-house procedures SOP's IC01 and IC02 in conjunction with manufacturer's instructions for antibodies and using Ventana Benchmark Ultra:	STEP
	Hematopoietic stem cells, melanocytes, mast cells, Cajal cells, germ cells, basal cells of skin, and mammary ductal epithelia	CD117	
	Plasma cells	CD138	
	Reed-Sternberg cells, granulocytes and granulocyte precursors, monocytes, a subset of tissue macrophages, and activated T-lymphocytes	CD15	
	B-cells	CD20	
	Follicular dendritic cell meshwork	CD21	
	Follicular dendritic cell meshwork	CD23	
	T-cells	CD3	
	Reed-Sternberg cells, anaplastic large cell lymphomas.	CD30	
	Endothelial cells	CD31	
	Endothelial cells	CD34	
	T-helper cells	CD4	
	Lymphocytes	CD45	



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Histological examinations for the purpose of clinical diagnosis (cont'd)</u>		
Formalin Fixed Paraffin Embedded (FFPE) human tissue and cell block sections on adhesive glass slides (cont'd)	Examination of tissues to identify or exclude morphological and cytological abnormalities for the purpose of clinical diagnosis(cont'd)	Documented in-house procedures SOP's IC01 and IC02 in conjunction with manufacturer's instructions for antibodies and using Ventana Benchmark Ultra:	STEP
	T-cells	CD5	
	Natural killer cells, neuroendocrine glands, neural tissue	CD56	
	Macrophages and some myeloid elements	CD68	
	B-cells	CD79a	
	T-cytotoxic cells	CD8	
	Gastrointestinal epithelium	CDX2	
	Adenocarcinomas from the lung, colon, stomach, oesophagus, pancreas, gallbladder, urachus, salivary gland, ovary, and endocervix	CEA (monoclonal)	
	Endocrine cells, tumours of neuroendocrine origin	Chromogranin A	
	Epithelial Cells	CK AE1/AE3	
	Epithelial Cells	CK CAM 5.2	
	Epithelial Cells	CK MNF116	
	Epithelial Cells	CK14	
	Epithelial Cells	CK19	



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Histological examinations for the purpose of clinical diagnosis (cont'd)</u>		
Formalin Fixed Paraffin Embedded (FFPE) human tissue and cell block sections on adhesive glass slides (cont'd)	Examination of tissues to identify or exclude morphological and cytological abnormalities for the purpose of clinical diagnosis(cont'd)	Documented in-house procedures SOP's IC01 and IC02 in conjunction with manufacturer's instructions for antibodies and using Ventana Benchmark Ultra :	STEP
	Epithelial Cells	CK20	
	Epithelial Cells	CK5/6	
	Epithelial Cells	CK7	
	Cytomegalovirus	Cytomegalovirus	
	Proliferating cells, mantle cell lymphoma	Cyclin D1	
	Myocytic cells, muscle cells	Desmin	
	Cajal cells, gastrointestinal stromal tumour	DOG-1	
	Epithelial tissue	E-CAD	
	Glandular epithelia	EMA	
	Oestrogen Receptor	ER	
	Endothelial cells	ERG	
	Breast carcinoma and urothelial carcinoma	GATA-3	
	Helicobacter Pylori	Helicobacter Pylori	
	c-erbB-2 oncoprotein	HER-2	
	Human herpesvirus 8 (Kaposi sarcoma-associated herpes virus)	HHV8 (KSHV)	



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Histological examinations for the purpose of clinical diagnosis (cont'd)</u>		
Formalin Fixed Paraffin Embedded (FFPE) human tissue and cell block sections on adhesive glass slides (cont'd)	Examination of tissues to identify or exclude morphological and cytological abnormalities for the purpose of clinical diagnosis(cont'd)	Documented in-house procedures SOP's IC01 and IC02 in conjunction with manufacturer's instructions for antibodies and using Ventana Benchmark Ultra :	STEP
	Melanocytes, melanoma	HMB45	
	Clonal restriction of B cell and plasma cell populations	Kappa	
	Proliferating cells	Ki-67 (MIB-1)	
	Clonal restriction of B cell and plasma cell populations	Lambda	
	Melanocytes, melanoma	Melan-A	
	Mismatch repair protein deficiency	MLH1	
	Mismatch repair protein deficiency	MSH2	
	Mismatch repair protein deficiency	MSH6	
	B-cell lymphomas	MUM-1	
	Gonadal and extra-gonadal seminomas, dysgerminomas and embryonal carcinomas	OCT-4	
	Surrogate marker for HPV	p16	
	Basal cells	p40	
	Malignant cells of the colon, bladder, ovary, breast and mesothelium.	p53	



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd) Formalin Fixed Paraffin Embedded (FFPE) human tissue and cell block sections on adhesive glass slides (cont'd)	<u>Histological examinations for the purpose of clinical diagnosis</u> (cont'd) Examination of tissues to identify or exclude morphological and cytological abnormalities for the purpose of clinical diagnosis(cont'd) Basal cells B-cells Ovarian serous carcinoma, clear cell renal cell carcinoma, and papillary thyroid carcinoma Mismatch repair protein deficiency Lymphatic endothelium Progesterone Receptor Prostate specific antigen Glial cells of the central and peripheral nervous system, melanocytes, chondrocytes, and adipocytes Smooth muscle cells, myofibroblasts and myoepithelial cells Smooth muscle cells, myoepithelial cells Melanocytes, melanoma Endocrine cells, tumours of neuroendocrine origin	Documented in-house procedures SOP's IC01 and IC02 in conjunction with manufacturer's instructions for antibodies and using Ventana Benchmark Ultra : p63 PAX-5 PAX-8 PMS2 Podoplanin (D2-40) PR PSA S100 SMA SMMHC (Myosin) SOX-10 Synaptophysin	STEP



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd) Formalin Fixed Paraffin Embedded (FFPE) human tissue and cell block sections on adhesive glass slides (cont'd)	<u>Histological examinations for the purpose of clinical diagnosis (cont'd)</u> Examination of tissues to identify or exclude morphological and cytological abnormalities for the purpose of clinical diagnosis(cont'd) Thyroid and lung tumours Fibroblasts, smooth muscle cells and lymphoid cells Wilms' tumour 1 protein Anaplastic lymphomas Nuclear accumulation of Beta-Catenin protein. T-cells T-cells Megakaryocytes Dermatofibroma Astrocytes, Schwann cells, satellite cells, enteric glial cells Plasma cells Sex cord stromal tumours and adrenal cortical tumours Neutrophil granulocytes and monocytes Neurons of the central and peripheral nervous system	Documented in-house procedures SOP's IC01 and IC02 in conjunction with manufacturer's instructions for antibodies and using Ventana Benchmark Ultra : TTF-1 Vimentin WT-1 ALK-1 (lymphoma) Beta-Catenin CD2 CD7 CD61 Factor XIIIa GFAP IgG4 Inhibin, alpha Myeloperoxidase Neurofilament Protein (NFP)	STEP



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>HUMAN BODY TISSUE (cont'd)</p> <p>Formalin Fixed Paraffin Embedded (FFPE) human tissue and cell block sections on adhesive glass slides (cont'd)</p>	<p><u>Histological examinations for the purpose of clinical diagnosis (cont'd)</u></p> <p>Examination of tissues to identify or exclude morphological and cytological abnormalities for the purpose of clinical diagnosis(cont'd)</p> <p>Lung adenocarcinoma</p> <p>Lymphoblastic lymphomas</p>	<p>Documented in-house procedures SOP's IC01 and IC02 in conjunction with manufacturer's instructions for antibodies and using Ventana Benchmark Ultra :</p> <p>Napsin A</p> <p>TdT</p>	<p>STEP</p>



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>HUMAN BODY TISSUE (cont'd)</p> <p>Unfixed body fluids cell brushings, cell washings, Ascitic, pleural, pericardial, and peritoneal Cell Blocks CSF Urine Cyst fluids FNA; - Lymph Node Salivary Gland Submandibular Parotid Liver Kidney Thorax Abdomen Pancreatic brushings Synovial Fluids Bronchial specimens, including bronchial brushings, bronchial washings, sputum, transbronchial FNA</p> <p>Unstained slides prepared from above</p>	<p><u>Cytopathological examination activities for the purposes of clinical diagnosis</u></p> <p><u>Non-Gynaecological Cytology</u></p> <p>Examination of cellular material in order to identify or exclude morphological and cytological abnormalities</p>	<p>In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified</p> <p>Diagnostic cytology (non-gynae) in-house documented manual procedures for preparation using: Thermo Cytospin Hologic ThinPrep 2000 processor SOP CC1 SOP CC2 SOP CC4 SOP CC6 SOP CC8 SOP CC11 SOP CP13</p> <p><u>Romanowsky staining</u> Manual method SOP CC5</p> <p><u>Papanicolaou staining</u> Automated staining using Sakura Tissue Tek Prisma Automated Stainer and ST5030 coverslipper SOP CP12</p>	<p>STEP</p> <p>STEP</p>



9344
Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Cytopathological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	
Formalin Fixed Paraffin Embedded Tissue (FFPE) sections on glass slides	Morphological assessment and interpretation/diagnosis	Microscopy (qualitative analysis) In-house procedures using light microscopy Olympus BX45 microscope SOP H31	STEP
Joint fluids (synovial fluids)	Assessment of birefringent crystals of monosodium urate and calcium pyrophosphate crystals	Microscopic examination: Olympus BX45 microscope with polarising filter SOP CC1	STEP
Liquid based cytology samples	<u>Gynaecological Cytology</u>		
Liquid based cytology samples containing cervical/vaginal cells	Examination of tissues to identify or exclude morphological and cytological abnormalities for the purposes of diagnosis	<u>Preparation of slides for staining</u> Documented in-house procedures in conjunction with manufacturer's instructions using: Hologic Thinprep 5000 benchtop processor SOP CS1 SOP CS2 SOP CP1	STEP



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Cytopathological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	
Unstained slides prepared from above		<u>Papanicolaou staining</u> Automated staining using Sakura Prisma Automated Stainer and ST5030 coverslipper SOP CP 12	
Formalin Fixed Paraffin Embedded Tissue (FFPE) sections on glass slides	Morphological assessment and interpretation/ diagnosis	Microscopy (qualitative analysis) In-house procedures using light microscopy Olympus microscopes BX45 Olympus microscopes BX51 SOP H31	STEP
Liquid based cytology samples containing cervical/ vaginal cells	<u>High-Risk Human Papilloma Virus (HR-HPV) testing</u> Detection of high-risk HPV genotypes 16 and 18 and 12 others pooled together	Detection of ribosomal RNA from: HPV by Nucleic acid amplification technique (NAAT) using Roche COBAS 480 plus analyser SOP CP14 HPV	STEP
	<u>Serological examination activities for the purposes of clinical diagnosis</u>	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	
	<u>Molecular detection for the purposes of clinical diagnosis</u>	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	



9344
Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 Issue date: 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Haematological examination activities for the purposes of clinical diagnosis</u>	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	CAVE
Blood	Cell Morphology	Manual staining and microscopy using SOP HAE003	CAVE
Blood	ESR	Diesse CUBE 30 method using the Modified Westergren SOP HAE026	CAVE
Blood	FBC	Siemens Advia 560 using Impedance, light scatter and light absorbance measurements	CAVE
	RBC	SOP HAE001	
	Differential WCC HGB PLT		
Plasma	PT INR APTT FIB D Dimer	Stago Compact Max using Chronometry measurement SOP HAE024 Using photometry	CAVE



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Biochemical examination activities for the purposes of clinical diagnosis</u>	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	CAVE
Serum unless otherwise indicated		Siemens Atellica using Photometry SOPBIO001	
Serum/urine	ACE Albumin Amylase Total Bilirubin Calcium Adjusted Calcium CK Bicarbonate GGT ALT AST ASTPLc Glucose Direct HDL EGFR TIBC Iron LDH Magnesium Microalbumin Inorganic Phosphorus		
Plasma	Phosphate Protein Transferrin Urea IgA IgG IgM Total Cholesterol Creatinine Total Protein Triglyceride Calculated LDL Uric Acid/ Urate		
Serum/urine	HS CRP Ferritin		
Serum/urine			
Urine			
Urine			
Serum/urine			
Serum/urine			
Serum			



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 Issue date: 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Biochemical examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	CAVE
Serum unless otherwise indicated (cont'd)	Folate FT4 TSH Troponin I HS Vitamin B12 Vitamin D NT Pro BNP FT3 FSH LH Progesterone Testosterone Oestradiol Prolactin PTH HCG E2 PSA SHBG aTPO CA125 CA19-9 CEA	Siemens Atellica using Photometry SOPBIO001	
Serum/urine	Sodium	Siemens Atellica using Indirect ISE SOPBIO001	CAVE
Serum/urine	Potassium Chloride		



9344

Accredited to
ISO 15189:2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Unilabs Ltd

Issue No: 012 **Issue date:** 14 January 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN BODY TISSUE (cont'd)	<u>Biochemical examination activities for the purposes of clinical diagnosis</u> (cont'd)	In house documented procedures (listed) in conjunction with manufacturer's instructions for equipment as specified	
Urine	Calculated Ratios: 24hr Calcium 24hr Creatinine 24hr Uric acid 24hr Phosphate 24hr Magnesium 24hr Microalbumin 24hr Protein Albumin/Creatine Protein/Creatinine Calcium Creatinine Creatinine Clearance	Siemens Atellica using photometry	CAVE
Serum	Cortisol TSI aTG	Siemens Immulite using Chemiluminescent enzyme immunoassay SOP BIO009	CAVE
Faeces	iFOBT: Immunochemical Faecal Occult Blood Test	QuickRead Go using photometry SOP INF 003	STEP
END			