Schedule of Accreditation

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



9574

Accredited to ISO 15189:2022

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Department of Clinical Biochemistry Gloucestershire Royal Hospital Great Western Road E-Mail: linford.rees@nhs.net

Gloucester Gloucestershire

GL1 3NN

Contact: Linford Rees Tel: +44 (0)3004 225 385

Website: www.gloshospitals.nhs.uk

Testing performed by the Organization at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details: | | Activity: | Location code |
|---|--------------------------|-----------------------|---------------|
| Address: Department of Clinical Biochemistry Gloucestershire Royal Hospital Great Western Road Gloucester Gloucestershire GL1 3NN | Contact: Linford Rees | Clinical Biochemistry | A |
| Address: Department of Clinical Biochemistry Cheltenham General Hospital Sandford Road Cheltenham Gloucestershire GL53 7AN | Contact: Linford Rees | Clinical Biochemistry | В |

Assessment Manager: ML2 Page 1 of 11



ISO 15189:2022

Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|---|---|--|------------------|
| HUMAN BODY FLUIDS | Biochemistry examination activities for the purpose of clinical diagnosis | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum or plasma unless otherwise stated | Quantitation of: | Roche Cobas ISE/c702/c702 and e801(long line) and ISE/ c702 and e801 (short line) | A |
| | | Roche Cobas c501/c501/e601 and c501 | В |
| | Albumin | CHSOPR010 Colourimetric | A, B |
| Urine | Albumin | CHSOPR041 Immunoturbidimetric | В |
| | Alkaline phosphatase | CHSOPR011 Colourimetric | A, B |
| | Alanine transaminase | CHSOPR012 Catalytic activity | A, B |
| | Ammonia | CHSOPR035 Enzymatic method with glutamate dehydrogenase | A |
| | Amylase | CHSOPR019 Enzymatic colourimetric | A, B |
| Urine | Amylase | CHSOPR019 Enzymatic colourimetric | А |
| | Aspartate Aminotransferase | CHSOPR022 Catalytic activity/enzymatic | A, B |
| | Bicarbonate | CHSOPR038 Colourimetric | A, B |
| | | | |
| | | | |

Assessment Manager: ML2 Page 2 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|--|--|--|------------------|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum or plasma unless otherwise stated (cont'd) | Quantitation of: | Roche Cobas ISE/c702/c702 and e801(long line) and ISE/ c702 and e801 (short line) | A |
| | | Roche Cobas c501/c501/e601 and c501 | В |
| | Bile acids | CHSOPR040 Enzymatic | А |
| | Calcium | CHSOPR006 Colourimetric | A, B |
| | Adjusted Calcium | CHSOPR006 Calculated parameter | A, B |
| Urine | Calcium | CHSOPR006 Colourimetric | А |
| | Chloride | CHSOPR002 Indirect ISE | A, B |
| | Cholesterol | CHSOPR015 Enzymatic colourimetric | А |
| | Creatine kinase | CHSOPR021 UV-test | A, B |
| | Creatinine | CHSOPR004 Enzymatic | A, B |
| Urine | Creatinine | CHSOPR004/045 Enzymatic | A, B |
| | AKI | CHMN027 Calculated parameter | A, B |
| | | | |
| | | | |

Assessment Manager: ML2 Page 3 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|--|--|--|------------------|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum or plasma unless otherwise stated (cont'd) | Quantitation of: (cont'd) | Roche Cobas ISE/c702/c702 and e801(long line) and ISE/ c702 and e801 (short line) | A |
| | | Roche Cobas c501/c501/e601 and c501 | В |
| | eGFR | CHMN027 Calculated parameter CKD-EPI | A, B |
| | Conjugated Bilirubin | CHSOPR013 Diazo method | А |
| | C reactive protein | CHSOPR020 Particle enhanced Immunoturbidimetric assay | А, В |
| | Ethanol | CHSOPR033 Enzymatic using alcohol dehydrogenase | А, В |
| | Gamma Glutamyl transferase (GGT) | CHSOPR037 Enzymatic assay | A, B |
| | Glucose | CHSOPR005 Hexokinase method | A, B |
| CSF | Glucose | CHSOPR039 Hexokinase method | A |
| | High density lipoprotein (HDL) | CHSOPR017 Homogeneous enzymatic colourimetric method | A |
| | Insulin Like Growth Factor-1 (IGF-1) | CHSPOR-111 Sandwich ECLIA | A |
| | | | |
| | | | |

Assessment Manager: ML2 Page 4 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|--|--|--|------------------|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum or plasma unless otherwise stated (cont'd) | Quantitation of: (cont'd) | Roche Cobas ISE/c702/c702 and e801(long line) and ISE/ c702 and e801 (short line) | A |
| | | Roche Cobas c501/c501/e601 and c501 | В |
| | Iron | CHSOPR036 Ferrozine method | A |
| | Lactate | CHSOPR034 Enzymatic colourimetric | A, B |
| CSF | Lactate | CHSOPR034 Colourimetric assay | А |
| | Lactate dehydrogenase | CHSOPR023 Optimised standard method acc IFCC | A, B |
| | Lithium | CHSOPR032 Colourimetric method | А |
| | Magnesium | CHSOPR007 Colourimetric endpoint | A, B |
| Urine | Magnesium | CHSOPR007 Colourimetric endpoint | А |
| | Salicylate | CHSOPR028 Enzymatic assay | A, B |
| | Paracetamol | CHSOPR029 Colourimetric assay | A, B |
| | Phosphate | CHSOPR008 Molybdate UV | A, B |
| | | | |

Assessment Manager: ML2 Page 5 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|--|--|--|------------------|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum or plasma unless otherwise stated (cont'd) | Quantitation of: (cont'd) | Roche Cobas ISE/c702/c702 and e801(long line) and ISE/ c702 and e801 (short line) | А |
| | | Roche Cobas c501/c501/e601 and c501 | В |
| Urine | Phosphate | CHSOPR008 Molybdate UV | A |
| | Potassium | CHSOPR002 Indirect ISE | A, B |
| Urine | Potassium | CHSOPR045 Indirect ISE | A, B |
| Urine | Protein | CHSOPR042 Turbidimetric | A |
| CSF | Protein | CHSOPR039 Turbidimetry | A |
| | Sodium | CHSOPR002 Indirect ISE | A, B |
| Urine | Sodium | CHSOPR045 Indirect ISE | A, B |
| | Soluble fms-like tyrosine kinase-1 (sFlt-1) | CHSOPR-112 Sandwich ECLIA | A |
| | Placental Growth Factor (PIGF) | CHSOPR-112 Sandwich ECLIA | A |
| | PIGF/sFlt-1 Ratio | CHSOPR-112 Calculated Test | A |
| Whole blood lysed | Tacrolimus | CHGN128 Competitive ECLIA | А |
| | Thyroglobulin (Tg) | CHSOPR-113 Sandwich ECLIA | В |
| | Anti-Thyroglobulin (TgAb) | CHSOPR-113 Sandwich ECLIA | В |

Assessment Manager: ML2 Page 6 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|--|--|--|------------------|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum or plasma unless otherwise stated (cont'd) | Quantitation of: (cont'd) | Roche Cobas ISE/c702/c702 and e801(long line) and ISE/ c702 and e801 (short line) | A |
| | | Roche Cobas c501/c501/e601 and c501 | В |
| | Elecsys NT-proBNP | CHSOPR-100 Sandwich ECLIA | А |
| | Transferrin | CHSOPR036 Immunoturbidimetric | A |
| | Total Protein | CHSOPR009 Colourimetric | A, B |
| | Total Bilirubin | CHSOPR013 Colourimetric Diazo method | A, B |
| | Triglyceride | CHSOPR016 Enzymatic colourimetric | A |
| | Urate | CHSOPR018 Enzymatic colourimetric | A, B |
| Urine | Urate | CHSOPR018 Enzymatic colourimetric | A |
| | Urea | CHSOPR003 Kinetic assay | A, B |
| Urine | Urea | CHSOPR045 Kinetic analysis | A |
| | Gentamicin | CHSOPR026 Homogeneous enzyme immunoassay (CEDIA) | A, B |
| | Tobramycin | CHSOPR027 Homogeneous enzyme immunoassay (CEDIA) | A |

Assessment Manager: ML2 Page 7 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|--|--|--|------------------|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum or plasma unless otherwise stated (cont'd) | Quantitation of: (cont'd) | Roche Cobas ISE/c702/c702 and e801(long line) and ISE/ c702 and e801 (short line) | A |
| | | Roche Cobas c501/c501/e601 and c501 | В |
| | Vancomycin | CHSOPR068 Homogeneous enzyme immunoassay (CEDIA) | A |
| | Carbamazepine | CHSOPR025 Homogeneous microparticle agglutination assay | A |
| | Phenytoin | CHSOPR024 Kinetic interaction of microparticles in solution (KIMS) | А |
| | Theophylline | CHSOPR031 Kinetic interaction of microparticles in solution (KIMS) | A |
| Serum | Quantitation of: | Roche Cobas e801 (Gloucester) and e601 (Cheltenham) electrochemiluminescence using in-house procedures: | |
| | Active B12 | CHSOPR-116 Sandwich Principle | А |
| | Cortisol | CHSOPR047 Competition principle | А |
| | Follicle Stimulating hormone | CHSOPR053 Sandwich principle | А |
| | Free T3 | CHSOPR050 Competition principle | А |
| | Free T4 | CHSOPR050 Competition principle | А |

Assessment Manager: ML2 Page 8 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|----------------------------|--|--|------------------|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum (cont'd) | Quantitation of: (cont'd) | Roche Cobas e801 (Gloucester) and e601 (Cheltenham) electrochemiluminescence using in-house procedures: | |
| | Growth hormone | CHSOPR080 Sandwich principle | В |
| | Luteinizing hormone | CHSOPR053 Sandwich principle | А |
| | Oestradiol | CHSOPR057 Competition principle | В |
| | Progesterone | CHSOPR052 Competition principle | А |
| | Prolactin | CHSOPR054 Sandwich principle | А |
| | Parathyroid Hormone | CHSOPR059 Sandwich principle | А |
| | Sex Hormone Binding globulin | CHSOPR058 Sandwich principle | А |
| | Testosterone | CHSOPR056 Competition principle | А |
| | Thyroid stimulating hormone | CHSOPR050 Sandwich principle | А |
| | Digoxin | CHSOPR030 Competition principle | А |
| | Procalcitonin | CHSOPR078 Sandwich principle | А |
| | Troponin T | CHSOPR046 Sandwich principle | A, B |

Assessment Manager: ML2 Page 9 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|----------------------------|--|--|------------------|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction with in-house procedures: | |
| Serum (cont'd) | Quantitation of: (cont'd) | Roche Cobas e801 (Gloucester) and e601 (Cheltenham) electrochemiluminescence using in-house procedures: | |
| | Vitamin D | CHSOPR088 Competition principle | A |
| | Alpha feto protein | CHSOPR063 Sandwich principle | В |
| | CA125 | CHSOPR060 Sandwich principle | В |
| | CA-15-3 | CHSOPR075 Sandwich principle | В |
| | CA 19-9 | CHSOPR061 Sandwich principle | В |
| | Carcino-embryonic antigen (CEA) | CHSOPR062 Sandwich Principle | В |
| | Total Prostate specific antigen | CHSOPR055 Sandwich principle | А |
| | Beta HCG | CHSOPR048 Sandwich principle | А |
| | Quantification of: | Advanced Micro Osmometer OsmoPro freezing point depression analysis and in-house procedure: | |
| Serum & Urine | Osmolality | CHMN001 | A |
| | | | |
| | | | |

Assessment Manager: ML2 Page 10 of 11



Schedule of Accreditation issued by

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gloucestershire Hospitals NHS Foundation Trust

Issue No: 010 Issue date: 24 October 2025

Testing performed by the Organisation at the locations specified

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code | |
|----------------------------|--|---|------------------|--|
| HUMAN BODY FLUIDS (cont'd) | Biochemistry examination activities for the purpose of clinical diagnosis (cont'd) | Manufacturer's instructions for equipment and kits where relevant in conjunction within house procedures: | | |
| | Detection and quantification of: | | | |
| Whole blood (EDTA) | HbA1c | TOSOH HLC-723 G11 High Performance Liquid Chromatography | A | |
| | | CHSOPM-1 | | |
| | | Dynex DS2 Enzyme linked Immunosorbent assay (ELISA) and in-house procedure | A | |
| Faeces | Calprotectin | CHMN033 | | |
| Sweat | Sweat chloride | Wescor Macroduct sweat collection system and Sherwood 926S chloride analyser coulometric titration using in-house procedure: | A | |
| | | CHMN011 | | |
| CSF | CSF Xanthochromia | Northstar Bio-UV UVIKON XL Spectrophotometer- Spectrophotometry | A | |
| | | CHGN027 | | |
| END | | | | |

Assessment Manager: ML2 Page 11 of 11