

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



9730

Accredited to
ISO/IEC 17025:2017

DNKA LTD.

Issue No: 006 Issue date: 18 July 2025

75 Loanbank Quadrant
Govan
Glasgow
G51 3HZ

Contact: Mr D Coll
Tel: +44 (0)141 425 1191
Fax: +44 (0)141 445 0333
E-Mail: info@dnka.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 75 Loanbank Quadrant Govan Glasgow G51 3HZ	Local contact Mr D Coll	Health and Hygiene Asbestos – All Support Functions
		A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and hygiene	B
Mobile laboratories	Health and hygiene	C



9730
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DNKA LTD.

Issue No: 006 Issue date: 18 July 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method SOP01 based on HSG 248	B
	Fibre counting	Documented In-House Method SOP01 Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method SOP01, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for asbestos identification	Documented In-House Method SOP03 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method SOP02 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
ASBESTOS IN SOILS – The Identification of Asbestos fibres in bulk samples of Soil, <i>specifically: Soil</i>	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method SOP04 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248.	A



9730
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DNKA LTD.

Issue No: 006 **Issue date:** 18 July 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN SOILS – The Identification and Quantification of Asbestos fibres in bulk samples of Soil, <i>specifically:</i> <i>Soil</i>	Health and Hygiene (cont'd) Free Fibre Identification and Quantification of Asbestos content of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021 Documented In-House Method SOP04 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			