


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>9854</b> Accredited to ISO/IEC 17025:2017	<b>Pendragon Scientific Ltd</b>	
	<b>Issue No: 007    Issue date: 05 June 2025</b>	
	<b>44 Winslow Road Granborough Buckinghamshire MK18 3NQ</b>	<b>Contact: Mr R Johns Tel: +44 (0)845 496 0613 E-Mail: info@pendragonscientific.com Website: www.pendragonscientific.com</b>
<b>Calibration performed by the Organisation at the locations specified</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
44 Winslow Road Granborough Buckinghamshire MK18 3NQ  Richard Johns	Temperature calibration of Incubators  Dynamic validation of PCR Machines and thermal cyclers  Temperature in a centrifuge  Rotational speed (Centrifuges)  Time (Centrifuges)  Carbon Dioxide content in an incubator	P

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Temperature calibration of Incubators  Dynamic validation of PCR Machines and thermal cyclers  Temperature in a centrifuge  Rotational speed (Centrifuges)  Time (Centrifuges)  Carbon Dioxide content in an incubator	S



9854

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Pendragon Scientific Ltd**

**Issue No: 007    Issue date: 05 June 2025**

**Calibration performed by the Organisation at the locations specified**

**Calibration and Measurement Capability (CMC)**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
TEMPERATURE			By comparison with a reference instrument	
Temperature controlled incubators, refrigerators, media preparators and chambers	0 °C to 70 °C	0.21 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	P & S
Dynamic validation of PCR <sup>®</sup> machines and thermal cyclers Accuracy and uniformity	10 °C to 30 °C 30 °C to 90 °C 90 °C to 100 °C	0.60 °C 0.37 °C 0.41 °C	The PCR <sup>®</sup> (Polymerase Chain Reaction), process is covered by patents owned by Hoffman-LaRoche Inc. Multiple sensor system	P & S
Temperature in a centrifuge	4 °C to 40 °C	0.50 °C		P & S
ROTATIONAL SPEED				
Centrifuges	30 rpm to 20000 rpm 20000 rpm to 60000 rpm	4.0 rpm 18 rpm	By comparison with a reference instrument	P & S
TIME INTERVAL				
Timers associated with centrifuges	0 s to 3600 s	0.35 s	By comparison with a reference instrument	P & S
GAS CONTENT				
Carbon Dioxide	5%	0.06%	Gas Analyser By comparison with a reference gas mixture	P & S
Carbon Dioxide	5%	0.26%	Single point calibration in the centre of the incubator by comparison with a reference instrument	P & S
END				



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Pendragon Scientific Ltd**

**Issue No: 007    Issue date: 05 June 2025**

**Calibration performed by the Organisation at the locations specified**

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$