


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|--|---|
|  10391 Accredited to ISO/IEC 17025:2017 | Climatic Services Limited | |
| | Issue No: 004 Issue date: 06 December 2021 | |
| | 41 Cavendish Way Southfield Estate Glenrothes Fife KY6 2SB | Contact: Tony Cox Tel: +44 (0)1592 771254 Fax: +44 (0)1592 775626 E-Mail: info@climaticservices.co.uk Website: www.climaticservices.co.uk |
| Calibration performed by the Organisation at the locations specified | | |

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details | Activity | Location code |
|--|-------------|---------------|
| Address 41 Cavendish Way Southfield Estate Glenrothes Fife KY6 2SB Local contact Tony Cox Tel: +44 (0)1592 771254 Fax: +44 (0)1592 775626 E-Mail: info@climaticservices.co.uk | Temperature | P |

Site activities performed away from the locations listed above:

| Location details | Activity | Location code |
|---|-------------|---------------|
| The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. Local contact Tony Cox Tel: +44 (0)1592 771254 Fax: +44 (0)1592 775626 E-Mail: info@climaticservices.co.uk | Temperature | S |



10391

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Climatic Services Limited

Issue No: 004 **Issue date:** 06 December 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|---|------------------|
| Temperature indicators and recorders, with temperature sensor(s) | -90 °C to +125 °C +125 °C to +500 °C | 0.15 °C 0.20 °C | Calibration performed within Metal Block Baths | P & S |
| Block Calibrators | -90 °C to +125 °C | 0.30 °C | By comparison with reference standards | P & S |
| Temperature controlled fridges, freezers, ovens and environmental chambers | -90 °C to +200 °C | 1.35 °C | Multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping | P & S |
| END | | | | |



10391

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Climatic Services Limited

Issue No: 004 Issue date: 06 December 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$