


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0043</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>Roxspur Measurement and Control Ltd</b>	
	Issue No: 066 Issue date: 04 April 2022	
	<b>2 Downgate Drive</b> <b>Sheffield</b> <b>South Yorkshire</b> <b>S4 8BT</b>	<b>Contact: Mr M Donnelly</b> <b>Tel: +44 (0)114 244 2521</b> <b>Fax: +44 (0)114 243 4838</b> <b>E-Mail: <a href="mailto:Mark.Donnelly@ttelectronics.com">Mark.Donnelly@ttelectronics.com</a></b> <b>Website: <a href="http://www.roxspur.com">www.roxspur.com</a></b>

**Calibration performed by the Organisations at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b> 2 Downgate Drive Sheffield South Yorkshire S4 8BT	<b>Local contact</b> Mr Mark Donnelly  Tel: +44 (0)114 244 2521 Fax: +44 (0)114 243 4838 Email: <a href="mailto:Mark.Donnelly@ttelectronics.com">Mark.Donnelly@ttelectronics.com</a> Website: <a href="http://www.roxspur.com">www.roxspur.com</a>	<a href="#">Electrical</a> <a href="#">Flow</a> <a href="#">Pressure</a> <a href="#">Temperature</a>	P1

#### Site activities performed away from the locations listed above:

Location details		Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	<b>Local contact</b> Mr Mark Donnelly  Tel: +44 (0)114-244 2521 Fax: +44 (0)114-243 4838 Email: <a href="mailto:mark.donnelly@ttelectronics.com">mark.donnelly@ttelectronics.com</a> Website: <a href="http://www.roxspur.com">www.roxspur.com</a>	<a href="#">Electrical</a> <a href="#">Pressure</a> <a href="#">Temperature</a>	S



0043

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Roxspur Measurement and Control Calibration Laboratory**

**Issue No: 066 Issue date: 04 April 2022**

Calibration performed by the Organisation at the locations specified

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>ELECTRICAL</b>				
DC Voltage	0 V to 200 mV	15 $\mu$ V/V + 0.70 $\mu$ V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	200 mV to 2 V	10 $\mu$ V/V + 0.60 $\mu$ V		
DC Current	2 V to 20 V	10 $\mu$ V/V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	S
	20 V to 200 V	15 $\mu$ V/V		
DC Current	200 V to 1000 V	15 $\mu$ V/V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	0 V to 30 V	5.0 mV		
DC Current	0 $\mu$ A to 200 $\mu$ A	50 $\mu$ A/A + 2.0 nA	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	S
	200 $\mu$ A to 2 mA	75 $\mu$ A/A + 15 nA		
Generation	2 mA to 20 mA	75 $\mu$ A/A	For the calibration of clampmeters only, using multi-turn coil method.	P1
	20 mA to 200 mA	100 $\mu$ A/A		
Generation	200 mA to 2 A	200 $\mu$ A/A	For the calibration of clampmeters only, using multi-turn coil method.	P1
	0 mA to 100 mA	0.010 mA		
DC Resistance	320 mA to 3.2 A	550 $\mu$ A/A + 150 $\mu$ A	For the calibration of clampmeters only, using multi-turn coil method.	P1
	3.2 A to 10 A	500 $\mu$ A/A + 1.1 mA		
DC Resistance	10 A to 20 A	0.11 % + 5.2 mA	For the calibration of clampmeters only, using multi-turn coil method.	P1
	20 A to 1000 A	0.37 %		
Measurement	0 $\Omega$ to 20 $\Omega$	100 $\mu\Omega/\Omega$ + 2.0 m $\Omega$	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	20 $\Omega$ to 2 k $\Omega$	20 $\mu\Omega/\Omega$ + 2.0 m $\Omega$		
AC Voltage	2 k $\Omega$ to 20 k $\Omega$	20 $\mu\Omega/\Omega$	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	20 k $\Omega$ to 200 k $\Omega$	30 $\mu\Omega/\Omega$		
AC Voltage	200 k $\Omega$ to 2 M $\Omega$	50 $\mu\Omega/\Omega$	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	2 M $\Omega$ to 20 M $\Omega$	200 $\mu\Omega/\Omega$		
AC Voltage	20 M $\Omega$ to 200 M $\Omega$	700 $\mu\Omega/\Omega$	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	200 M $\Omega$ to 2 G $\Omega$	0.10 %		
AC Voltage	10 mV to 200 mV	280 $\mu$ V/V + 8.0 $\mu$ V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	20 Hz to 10 kHz	550 $\mu$ V/V + 10 $\mu$ V		
AC Voltage	10 kHz to 100 kHz	400 $\mu$ V/V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	200 mV to 200 V	480 $\mu$ V/V		
AC Voltage	20 Hz to 30 kHz	450 $\mu$ V/V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	30 kHz to 100 kHz	510 $\mu$ V/V		
AC Voltage	200 V to 1100 V	450 $\mu$ V/V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	45 Hz to 10 kHz	510 $\mu$ V/V		
AC Voltage	10 kHz to 30 kHz	450 $\mu$ V/V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	10 kHz to 30 kHz	510 $\mu$ V/V		



0043

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Roxspur Measurement and Control Calibration Laboratory**

**Issue No: 066 Issue date: 04 April 2022**

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>ELECTRICAL</b> (continued)				
AC Current	20 Hz to 1 kHz 10 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 200 mA 200 mA to 2 A	250 $\mu$ A/A + 41 nA 350 $\mu$ A/A 600 $\mu$ A/A	Measured using digital multimeter. The capability includes generation and measurement of this quantity	P1
	1 kHz to 5 kHz 10 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 200 mA 200 mA to 2 A	800 $\mu$ A/A + 30 nA 700 $\mu$ A/A 0.13 %		P1
Generation	10 Hz to 3 kHz 2 A to 20 A	0.25 % + 7.0 mA	For the calibration of clampmeters only, using multi-turn coil method.	
	40 Hz to 100 Hz 20 A to 1000 A	0.50 %		
Frequency	0.1 Hz to 120 MHz	0.70 $\mu$ Hz/Hz	Using counter timer.	P1
Time interval	30 s to 8 hrs 30 s to 8 hrs	0.15 s 0.25 s	Using counter timer.	P1 S
<b>Electrical calibration of temperature indicators, controllers and recorders for the following sensors:</b>				
Noble metal thermocouples			Using millivolt injection or measurement.	
Types R and S	0 °C to 200 °C 200 °C to 1700 °C	0.50 °C 0.40 °C	with cold junction Compensation.	P1
Type B	600 °C to 1700 °C	0.40 °C	with cold junction compensation.	P1
Type R and S	0 °C to 200 °C 200 °C to 1700 °C	0.60 °C 0.60 °C	with cold junction compensation.	S
Type B	600 °C to 1700 °C	1.0 °C	with cold junction compensation.	S
Base metal thermocouples	-200 °C to 0.0 °C 0 °C to 1372 °C	0.26 °C 0.20 °C	with cold junction compensation.	P1
	-200 °C to 0.0 °C 0 °C to 1372 °C	0.80 °C 0.60 °C	with cold junction compensation.	S



0043

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Roxspur Measurement and Control Calibration Laboratory**

**Issue No: 066 Issue date: 04 April 2022**

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>Electrical calibration of temperature indicators, controllers and recorders for the following sensors: (continued)</b>				
Pt 100	-200 °C to +850 °C	0.050 °C	Using DC Resistance measurement or injection.	P1
	-200 °C to +850 °C	0.50 °C		S
<b>FLOW</b>				
Flow rate of calibration gas: Compressed Air Oxygen Argon Carbon Dioxide Nitrous Oxide Helium Nitrogen	0.005 l/min to 0.25 l/min 0.25 l/min to 2.5 l/min 2.5 l/min to 50 l/min	0.60 % 0.50 % 0.60 %	Calibration of flowmeters measuring volumetric or mass flowrates under ambient conditions.	P1
<b>PRESSURE</b>				
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges	3.5 kPa to 120 kPa 120 kPa to 200 kPa 200 kPa to 2.6 MPa	1.2 % + 5.8 Pa 1.4 % + 9.0 Pa 1.4 % + 5.8 Pa	Methods consistent with EURAMET CG17.  Calibration of pressure devices with an electrical output may be undertaken.	P1
	3 kPa to 96 kPa 96 kPa to 105 kPa 105 kPa to 2.1 MPa	3.9 kPa 530 Pa 3.9 kPa		S
Gas pressure (gauge)				
Calibration of pressure indicating instruments and gauges	-90 kPa to -2.5 kPa -2.5 kPa to 2.5 kPa 2.5 kPa to 100 kPa 100 kPa to 2.5 MPa	1.17 % + 25 Pa 0.24 % + 0.058 Pa 0.18 % + 6.0 Pa 0.011 %		P1
	-90 kPa to +2.0 MPa	3.8 kPa		S
Hydraulic pressure (absolute)				
Calibration of pressure indicating instruments and gauges	700 kPa to 6.1 MPa 6.1 MPa to 120.1 MPa	1.4 % + 5.8 Pa 1.4 % + 5.8 Pa		P1
	10 kPa to 60 MPa	260 kPa		S
Hydraulic pressure (gauge)				
Calibration of pressure indicating instruments and gauges	600 kPa to 6 MPa 6 MPa to 120 MPa	0.013 % 0.014 %		P1
	0 to 70 MPa	0.26 MPa		S



0043

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Roxspur Measurement and Control Calibration Laboratory**

**Issue No: 066 Issue date: 04 April 2022**

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>TEMPERATURE</b>				
Resistance thermometers	-196 °C 0 °C to -80 °C Ice Point (0 °C) Triple Point of water (0.01 °C) Gallium Melt Point (29.7646 °C) 0 °C to 300 °C	0.040 °C 0.050 °C 0.031 °C 0.012 °C 0.010 °C 0.050 °C	Calibration performed within Liquid Baths.	P1
Platinum thermocouples	0 °C to 1100 °C 1100 °C to 1600 °C  1064.18 °C 1553.5 °C  0 °C to 1100 °C 1100 °C to 1600 °C  Fixed point calibrations  Gallium Melt Point (29.7646 °C) Tin (231.9 °C) Zinc (419.5 °C)	0.50 °C 1.8 °C  0.50 °C 1.8 °C  2.0 °C 4.0 °C  0.35 °C 0.50 °C 0.50 °C	Calibration performed within Tube Furnace or Metrology Well. Au and Pd wire bridge measurements.	P1          P1
Other thermocouples	-196 °C 0 °C to -80 °C Gallium Melt Point (29.7646 °C) 0 °C to 300 °C 300 °C to 650 °C 650 °C to 1100 °C 1100 °C to 1300 °C  -80 °C to +200 °C 200 °C to 1000 °C 1000 °C to 1300 °C	0.26 °C 0.15 °C 0.10 °C 0.15 °C 0.25 °C 1.0 °C 2.5 °C  1.0 °C 3.0 °C 5.0 °C	Calibration performed within Liquid Bath, Metrology Well or Furnace.	P1          S
Compensating and extension cables for Noble metal thermocouples Base metal thermocouples	0 °C to 40 °C 0 °C to 40 °C	0.26 °C 0.15 °C	Calibration performed within Liquid Baths	P1
Liquid-in-glass thermometers	-80 °C to -40 °C -40 °C to 0 °C Ice point (0 °C) 0 °C to 100 °C 100 °C to 300 °C	0.11 °C 0.050 °C 0.050 °C 0.050 °C 0.050 °C		P1
Electronic thermometers with sensors	Range as per sensors	As for sensors		P1
Electronic thermometers with sensors Analogue Digital	0 °C to 200 °C 200 °C to 600 °C 600 °C to 1100 °C 1100 °C to 1300 °C	0.50 °C 1.3 °C 2.0 °C 3.4 °C		S



0043

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Roxspur Measurement and Control Calibration Laboratory**

**Issue No: 066 Issue date: 04 April 2022**

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>TEMPERATURE</b> (continued)				
Metal block calibrators	-100 °C to +300 °C 300 °C to 1100 °C	0.050 °C 1.0 °C		P1
Temperature surveys				
Autoclaves, incubators, fridges/refrigerators and freezers	-80 °C to +200 °C	1.0 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping.	S
Ovens, furnaces, fridges/refrigerators and environmental chambers	0 °C to 600 °C 600 °C to 1100 °C 1100 °C to 1600 °C	1.1 °C 1.9 °C 3.5 °C		
END				



0043

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Roxspur Measurement and Control Calibration Laboratory**

**Issue No: 066 Issue date: 04 April 2022**

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \cdot 0.01 \cdot q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$