

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0143 Accredited to ISO/IEC 17025:2017	Calmet Laboratory Services, a division of Lazgill Ltd	
	Issue No: 038 Issue date: 12 March 2020	
	11b Upper Teddington Road Kingston-upon-Thames Surrey KT1 4DL	Contact: Mr D Howell Tel: +44 (0)20 8977 8455 Fax: +44 (0)20 8614 8048 E-Mail: sales@calmet.co.uk Website: www.calmet.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 11b Upper Teddington Road Kingston-upon-Thames Surrey KT1 4DL Contact: Mr D Howell Tel: +44 (0)20 8977 8455 Fax: +44 (0)20 8614 8048 E-Mail: sales@calmet.co.uk	Dimensional, Electrical, Pressure and Torque	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At customers premises Mr D Howell	Dimensional and Electrical	B



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
DIMENSIONAL			All linear calibrations may be given in inch units.	
LENGTH				
Plain plug gauges (parallel) cylindrical setting standards and rollers	1 to 50 diameter	1.0	Using length measuring machine and end standards.	A
	50 to 100	2.0		
	100 to 200	3.0		
	200 to 300	4.0		
Plain ring gauges (parallel)	1 to 10 diameter	2.0	Using length measuring machine and end standards.	A
	10 to 50	1.5		
	50 to 100	2.0		
	100 to 200	3.5		
	200 to 300	5.0		
Screw plug gauges (parallel) including check and setting plugs	1 to 100 diameter	4.0	Single and multi-start, symmetrical thread forms only, using length measuring machine.	A
	100 to 200	5.0		
	200 to 300	6.0		
Screw ring gauges (parallel)	1 to 50	5.0	Single and multi-start, symmetrical thread forms only, using length measuring machine. The 1 mm to 12 mm diameter range relates to functional test of size using check plugs.	A
	50 to 150	6.0		
	150 to 300	8.0		
Screw thread pitch	0.2 to 8	1.5	Using length measuring machine	A
Screw thread flank angles	0° to 52°	5.0 minutes of arc	Using a projector	A
Length gauges, flat and spherical ended	25 to 600	1.0 + (8.0 x length in m)	Using end standards	A
Plain gap gauges (parallel)	0.5 to 100 100 to 200	3.0	Using gauge blocks	A
		5.0		
Parallels	As BS 906:1972 5 to 50 x 100 x 400	From 1.5 up to 5.0		A
Vee blocks	As BS 3731:1987 20 to 150 diameter, vee capacity	From 2.5 up to 7.0		A



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
ANGLE			The uncertainties quoted for angle and form are for the departure from flatness, straightness, parallelism or squareness, ie the distance separating the parallel planes which just enclose the surface under consideration.	
Square Blade type	As BS 939:2007 50 to 300 300 to 600	3.0 on squareness 5.0		A
Angle plates and box angle plates	As BS 5535:1978 50 to 600	Squareness: 3.0 + (1.0 per 100 mm) Parallelism: 1.0 + (1.0 per 100 mm)		A
FORM				
Straight edges Cast iron	As BS 5204:Part 1:1975 300 to 8000))) 1.0 +) (2.0 x length in m))		A
Steel, Granite	As BS 5204:Part 2:1977 300 to 2000))) 1.0 +) (2.0 x length in m))		A & B
Surface plates Granite Cast iron	As BS 817:2008 160 x 100 to 4000 x 6000	1.5 + (0.80 x diagonal in m)	A & B	
MEASURING INSTRUMENTS AND MACHINES				
Micrometers External	As BS 870:2008 0 to 900))) Heads 2.0 between) any two points.		A
Internal	As BS 959:2008 0 to 900) Setting and extension) rods 1.0 +) (8.0 x length in m)		
Depth	As BS 6468:2008 0 to 300))		
Three point bore micrometers	3 to 100 100 to 150	Overall performance 5.0 Overall performance 8.0	Using setting rings	A
Bore indicators	2 to 100 100 to 150	Overall performance 5.0 Overall performance 8.0	Using setting rings or length measuring machine	A
Micrometer heads	As BS 1734:1951 0 to 100	1.0		A



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
MEASURING INSTRUMENTS AND MACHINES (cont'd)				
Vernier / Digital / Dial gauges Calliper	BS 887:2008 0 to 1000	Overall performance 10 + (30 x length in m)		A
Height	ISO13225:2012 and BS 1643:2008 0 to 1000	Overall performance 10 + (30 x length in m)		
Depth	BS 6365:2008 0 to 600	Overall performance 10 + (30 x length in m)		
Dial gauges and dial test indicators	As BS 907:2008 and BS 2795:1981 0 to 50	1.0		A
Electronic height gauges (including setting masters)	0 to 1000	1.0 + (5.0 x length in m)	Using end standards	A
Profile projectors	10 to 100 magnifications	125 at the screen 5.0 linear 3.0 minutes of arc	Using glass scales	A & B
Feeler gauges	As BS 957:2008 0.02 to 1.00	3.0		A
Spirit levels	As BS 3509:1962 and BS 958:1968 5 seconds of arc to 60 minutes of arc nominal sensitivity	Mean sensitivity 10 % of nominal Minimum of 0.50 seconds of arc		A
Electronic indicating levels	0 to 10 minutes of arc	1.0 % of range Minimum 0.50 seconds of arc	Using small angle generator	A
TORQUE				
Torque Wrenches and Torque Drivers	As BS EN ISO 6789-2:2017 0.1 N·m to 3000 N·m	1.0 % See Notes 3 and 6	Calibration results may also be given in units of lbf·in and lbf·ft. The uncertainties quoted are for both the application of the calibration torque and the characteristics of the device being calibrated	A
Torque Wrenches and Torque Drivers	As BS EN ISO 6789:2003 (Withdrawn & superseded) 0.1 N·m to 3000 N·m	1.0 % See Notes 3 and 6		A



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
ELECTRICAL				
Electrical measurements are carried out by direct comparison unless otherwise defined in the remarks column.				
DC Resistance				
Spot values	100 $\mu\Omega$ 1 m Ω 10 m Ω 100 m Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1.0 G Ω	60 ppm 35 ppm 15 ppm 12 ppm 6.0 ppm 5.0 ppm 4.0 ppm 4.0 ppm 3.0 ppm 3.0 ppm 6.0 ppm 10 ppm 0.60 % 0.65 %	For the calibration of measuring instruments	A
Range values	0 Ω to 12 Ω 12 Ω to 120 Ω 120 Ω to 1.2 k Ω 1.2 k Ω to 12 k Ω 12 k Ω to 120 k Ω 120 k Ω to 1.2 M Ω 1.2 M Ω to 12 M Ω 12 M Ω to 120 M Ω 120 M Ω to 1 G Ω 0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 120 M Ω	6.7 ppm + 35 $\mu\Omega$ 4.4 ppm + 350 $\mu\Omega$ 3.6 ppm + 230 $\mu\Omega$ 4.4 ppm + 2.3 m Ω 3.6 ppm + 23 m Ω 14 ppm + 1.2 Ω 61 ppm + 120 Ω 650 ppm + 1.2 k Ω 0.65 % + 12 k Ω 130 ppm + 3.5 m Ω 120 ppm + 4.6 m Ω 220 ppm 580 ppm 1.2 %	For the calibration of resistance sources For the calibration of measuring instruments	A B
Spot values	1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	96 ppm 24 ppm 11 ppm 6.8 ppm 6.8 ppm 8.8 ppm 15 ppm 41 ppm 120 ppm	For the calibration of measuring instruments	B



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
DC Voltage Standard Cell Values Other Values	10 V	0.40 ppm		A
	0 V to 100 mV	0.80 ppm + 0.20 μ V	For the calibration of measuring and sourcing instruments	A
	100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1 kV 1 kV to 12 kV	1.6 ppm + 0.20 μ V 0.80 ppm 2.0 ppm 1.0 % 0.50 %		
	0 mV to 120 mV	45 ppm + 4.5 μ V		B
	120 mV to 1.2 V 1.2 V to 12 V 12 V to 120 V 120 V to 1050 V 1 kV to 12 kV	100 ppm 80 ppm 110 ppm 600 ppm 0.50 %	Measurement of sources	
	0 V to 220 mV	9.0 ppm + 0.80 μ V		B
	220 mV to 2.2 V 2.2 V to 11 V 11 V to 22 V 22 V to 220 V 220 V to 1100 V	5.5 ppm + 1.0 μ V 3.8 ppm + 2.6 μ V 3.8 ppm + 4.2 μ V 5.2 ppm + 40 μ V 6.8 ppm + 400 μ V	For the calibration of measuring instruments	
DC Current	1 μ A to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A 3 A to 10 A 10 A to 20 A 20 A to 40 A 40 A to 100 A	15 ppm 20 ppm 25 ppm 25 ppm 45 ppm 50 ppm 85 ppm 380 ppm	For the calibration of sourcing instruments	A
	10 A to 550 A 550 A to 1000 A	0.55 % 0.55 %	For the Calibration of Current Clamps	A&B A
	0 A to 120 μ A	600 ppm + 31 nA	For the calibration of sourcing instruments	B
	120 μ A to 1.2 mA 1.2 mA to 12 mA 12 mA to 120 mA 120 mA to 400 mA 400 mA to 1.2 A 1.2 A to 3 A 3 A to 10 A 10 A to 20 A	600 ppm + 60 nA 600 ppm + 2.4 μ A 600 ppm + 12 μ A 600 ppm + 26 μ A 600 ppm + 240 μ A 0.12 % + 700 μ A 0.18 % + 930 μ A 75 ppm + 75 μ A		



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
DC Current Continued	0 A to 220 μ A 220 μ A to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 10 A	43 ppm + 6.0 nA 37 ppm + 7.0 nA 37 ppm + 40 nA 48 ppm + 700 nA 84 ppm + 12 μ A 480 ppm + 260 μ A	For the calibration of measuring instruments	A&B
AC Voltage	1 mV to 2.2 mV 40 Hz to 20 kHz 2.2 mV to 7 mV 40 Hz to 20 kHz 7 mV to 22 mV 40 Hz to 20 kHz 22 mV to 70 mV 40 Hz to 20 kHz 70 mV to 220 mV 40 Hz to 20 kHz 220 mV to 700 mV 20 Hz to 40 Hz 40Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz 100kHz to 300kHz 700 mV to 22 V 20 Hz to 40 Hz 40Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz 100kHz to 300kHz	680 ppm + 1.0 μ V 250 ppm + 1.0 μ V 110 ppm + 1.0 μ V 100 ppm + 1.0 μ V 33 ppm + 1.0 μ V 55 ppm + 1.0 μ V 29 ppm + 1.0 μ V 39 ppm + 1.4 μ V 57 ppm + 1.7 μ V 150 ppm + 2.7 μ V 49 ppm 24 ppm 37 ppm 58 ppm 150 ppm	For the calibration of sourcing instruments	A
AC Voltage	22 V to 70 V 20 Hz to 40 Hz 40Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz 100kHz to 300kHz 70 V to 220 V 20 Hz to 40 Hz 40Hz to 20kHz	49 ppm 28 ppm 43 ppm 76 ppm 150 ppm 51 ppm 29 ppm	For the calibration of sourcing instruments	A



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC Voltage Continued	220 V to 700 V 50 Hz to 20 kHz	35 ppm	For the calibration of sourcing instruments	A
	700 V to 1000 V 50 Hz to 20 kHz	33 ppm		
	5 mV to 120 mV 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.085 % + 47 μ V 0.15 % + 58 μ V 0.70 % + 95 μ V	For the calibration of sourcing instruments	B
	0.12 V to 1.2 V 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.080 % + 350 μ V 0.15 % + 580 μ V 0.70 % + 950 μ V		
	1.2 V to 12 V 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.080 % + 3.5 mV 0.15 % + 5.8 mV 0.70 % + 9.5 mV		
	12 V to 120 V 10Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz	0.080 % + 35 mV 0.15 % + 58 mV 0.70 % + 95 mV	For the calibration of measuring instruments	A&B
	120 V to 1000 V 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.080 % + 260 mV 0.15 % + 440 mV 0.70 % + 700 mV		
	200 μ V to 2.2 mV 10 Hz to 2 0Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	420 ppm + 4.0 μ V 350 ppm + 4.0 μ V 350 ppm + 4.0 μ V 470 ppm + 4.0 μ V 710 ppm + 5.0 μ V 0.13 % + 10 μ V 0.17 % + 20 μ V 0.40 % + 20 μ V		
	2.2 mV to 22 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	260 ppm + 4.0 μ V 120 ppm + 4.0 μ V 110 ppm + 4.0 μ V 220 ppm + 4.0 μ V 530 ppm + 5.0 μ V 0.11 % + 10 μ V 0.15 % + 20 μ V 0.29 % + 45 μ V		



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC Voltage Continued	22 mV to 220 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	250 ppm + 12 μ V 100 ppm + 7.0 μ V 62 ppm + 7.0 μ V 130 ppm + 7.0 μ V 320 ppm + 17 μ V 670 ppm + 20 μ V 0.15 % + 25 μ V 0.29 % + 45 μ V	For the calibration of measuring instruments	A&B
	220 mV to 2.2 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	250 ppm + 40 μ V 100 ppm + 15 μ V 46 ppm + 8.0 μ V 70 ppm + 10 μ V 88 ppm + 30 μ V 350 ppm + 80 μ V 0.11 % + 200 μ V 0.18 % + 300 μ V		
	2.2 V to 22 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	250 ppm + 400 μ V 92 ppm + 150 μ V 46 ppm + 50 μ V 70 ppm + 100 μ V 86 ppm + 200 μ V 260 ppm + 600 μ V 0.11 % + 2.0 mV 0.16 % + 3.2 mV		
	22 V to 220 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	250 ppm + 4.0 mV 92 ppm + 1.5 mV 56 ppm + 600 μ V 82 ppm + 1.0 mV 150 ppm + 2.5 mV 910 ppm + 16 mV 0.45 % + 40 mV 0.92 % + 80 mV		
	220 V to 1100 V 15 Hz to 50 Hz 50 Hz to 1 kHz	310 ppm + 16 mV 74 ppm + 4.0 mV		
AC Current	6 μ A to 120 μ A 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.17 % + 35 nA 750 ppm + 35 nA 750 ppm + 35 nA	For the calibration of sourcing instruments	A



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC Current Continued	120 μ A to 1.2 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.50 % + 240 nA 0.20 % + 240 nA 730 ppm + 240 nA 400 ppm + 240 nA 730 ppm + 240 nA 0.50 % + 470 nA 0.66 % + 1.8 μ A	For the calibration of sourcing instruments	A
	1.2mA to 12mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.48 % + 2.5 μ A 0.20 % + 2.5 μ A 700 ppm + 2.5 μ A 380 ppm + 2.5 μ A 730 ppm + 2.5 μ A 0.50 % + 4.8 μ A 0.65 % + 24 μ A		
	12mA to 120mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.48 % + 24 μ A 0.18 % + 24 μ A 730 ppm + 24 μ A 380 ppm + 24 μ A 730 ppm + 2.5 μ A 0.50 % + 48 μ A 0.65 % + 180 μ A		
	120 mA to 1 A 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz	0.48 % + 240 μ A 0.20 % + 240 μ A 0.10 % + 240 μ A 0.14 % + 240 μ A 0.38 % + 240 μ A 1.2 % + 470 μ A		
	1 A to 20 A 40 Hz to 1 kHz	750 ppm		
	5 μ A to 10 Hz to 5 kHz 5 kHz to 10 kHz	0.18 % + 70 nA 0.41 % + 810 nA		
	0.12 mA to 1.2 mA 10 Hz to 5 kHz 5 kHz to 10 kHz	0.18 % + 480 nA 0.24 % + 2.9 μ A		
	1.2 mA to 12mA 10 Hz to 5 kHz 5 kHz to 10 kHz	0.18 % + 7.0 μ A 0.41 % + 81 μ A		
	12 mA to 120 mA 10 Hz to 5 kHz 5 kHz to 10 kHz	0.12 % + 47 μ A 0.23 % + 290 μ A		



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC Current Continued	120 mA to 400 mA 10 Hz to 1 kHz 1 kHz to 10 kHz 400 mA to 1.2 A 10 Hz to 5 kHz 5 kHz to 10 kHz 1.2 A to 3 A 10 Hz to 5 kHz 5 kHz to 10 kHz 3 A to 10 A 10 Hz to 5 kHz 5 kHz to 10 kHz	0.12 % + 460 μ A 0.24 % + 3.3 mA 0.12 % + 460 μ A 0.41 % + 8.1 mA 0.19 % + 7.0 mA 0.41 % + 24 mA 0.19 % + 7.0 mA 0.42 % + 81 mA	For the calibration of sourcing instruments	B
AC Current	9 μ A to 220 μ A 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 220 μ A to 2.2 mA 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 2.2 mA to 22 mA 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 22 mA to 220 mA 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 220 mA to 2.2 A 20 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	260 ppm + 16 nA 170 ppm + 10 nA 120 ppm + 8.0 nA 310 ppm + 12 nA 0.12 % + 65 nA 260 ppm + 40 nA 170 ppm + 35 nA 110 ppm + 35 nA 220 ppm + 110 nA 0.12 % + 650 nA 260 ppm + 400 nA 170 ppm + 350 nA 110 ppm + 350 nA 220 ppm + 550 nA 0.12 % + 5.0 μ A 260 ppm + 4.0 μ A 170 ppm + 3.5 μ A 110 ppm + 2.5 μ A 220 ppm + 3.5 μ A 0.12 % + 10 μ A 260 ppm 35 μ A 410 ppm + 80 μ A 0.71 % + 160 μ A	For the calibration of measuring instruments	A&B



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC Current Continued	2.2 A to 11 A 45 Hz to 65 Hz 65 Hz to 500 Hz 500 Hz to 1 kHz	520 ppm + 1.6 mA 810 ppm + 1.6 mA 0.26 % + 1.6 mA	For the calibration of measuring instruments	A&B
	45 Hz to 60 Hz 10 A to 100 A 100 A to 500 A	0.15 % 0.55 %	Current simulation using multi turn coil for the calibration of clamp meters	A&B
	500 A to 1000 A	0.60 %		A
AC Resistance	40 Hz to 1 kHz 0.05 Ω 0.1 Ω 0.2 Ω 0.5 Ω 1 Ω 2 Ω 5 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω	120 ppm 90 ppm 80 ppm 75 ppm 70 ppm 70 ppm 70 ppm 45 ppm 45 ppm 45 ppm 50 ppm 65 ppm		A
AC/DC Voltage Transfer	250 mV to 16 V 20 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz 100 kHz 16 V to 125 V 20 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz 100 kHz 250 V 20 Hz to 10 kHz 20 kHz 50 kHz 500 V to 1 kV 20 Hz to 10 kHz 20 kHz	40 ppm 40 ppm 30 ppm 35 ppm 35 ppm 25 ppm 25 ppm 25 ppm 60 ppm 35 ppm 40 ppm 85 ppm 120 ppm		A



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC Power	47 Hz to 63 Hz Voltages 60 V to 240 V Currents 0.5 A to 5 A 15 W to 1200 W Voltages 60 V to 240 V Currents 5 A to 100 A 150 W to 24 kW Combination of specific voltage and current values: V = 75, 100, 150, 300 V I = 0.5, 1, 2, 5, 10, 20 A	90 ppm 150 ppm 0.035 %))) Unity to 0.5 power) factor lagging or) leading))) Unity to 0.5 power) factor lagging or) leading. Calibrations at) lower power factors) can be carried out to) greater uncertainties.	A
Frequency	0.2 Hz to 1 kHz 1 kHz to 1000 MHz	2 in 10 ⁸ 2 in 10 ⁸	Multi-period measurement Frequency measurement	A
Elapsed time	100 ms to 1 day	20 ms	Stop watches	
Rise/Fall Time	160 ps to 500 ns	5.0 % + 20 ps	For the calibration of	A
Bandwidth	50 kHz to 1.0 GHz	5.0 %	Oscilloscopes	A
Capacitance	At 1 kHz 1 nF to 10 μF	770 ppm		
Temperature simulation Ambient in support of reference junction	17 °C to 23 °C	0.10 °C		
Temperature indicators, calibration by electrical simulation	Simulated temperature	Uncertainty of simulated temperature	including cold junction compensation	A
Type K	-200 °C to 1370 °C	0.20 °C		
Type J	-200 °C to 750 °C	0.20 °C		
Type N	-200 °C to 400 °C	0.20 °C		
Type T	-200 °C to 1300 °C	0.20 °C		
Type R	0 °C to 1700 °C	0.20 °C		
Type S	0 °C to 1700 °C	0.20 °C		



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
Resistance thermometer (Pt 100)	-200°C to 800°C	0.030 °C		B
Type K	-200 °C to 1370 °C	0.50 °C		
Type J	-200 °C to 750 °C	0.50 °C		
Type N	-200 °C to 400 °C	0.80 °C		
Type T	-200 °C to 1300 °C	0.50 °C		
Resistance thermometer (Pt 100)	-200°C to 800°C	0.10 °C		
Measurements to support 17th edition type test equipment				A & B
Insulation Resistance	100 kΩ 1 MΩ 5 MΩ 10 MΩ 100 MΩ 1 GΩ	500 Ω 80 kΩ 80 kΩ 500 kΩ 5.0 MΩ 50 MΩ		
Insulation Voltage	50 V to 250 V 250 V to 500 V 500 V to 1 kV	5.0 V 10 V 15 V		
RCD Trip Current	10 mA to 15 mA 15 mA to 50 mA 100 mA to 100 mA 100 mA to 150 mA 150 mA to 500 mA 500 mA to 1 A	500 μA 1.0 mA 3.0 mA 5.0 mA 10 mA 25 mA		
RCD Trip Time	0 s to 100 ms	2.0 ms		
Continuity Resistance	10 mΩ to 1 Ω 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ	100 mΩ 500 mΩ 3.0 Ω 50 Ω 300 Ω		
Loop impedance	100 mΩ to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ	100 mΩ 1.0 Ω 15 Ω		
PAT test Voltage	50 V to 100 V nominal 50 Hz 100 V to 400 V nominal 50 Hz	1.0 V 1.5 V		
Earth Bond resistance	5 mΩ to 50 mΩ 50 mΩ to 5 Ω	5.0 mΩ 10 mΩ		



0143
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd
Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
Measurements to support 17th edition type test equipment Continued				A&B
Earth bond current	10 mA to 500 mA	10 mA		
	500 mA to 10 A	20 mA		
	10 A to 25 A	65 mA		
PAT Leakage current	50 μ A to 7.7 mA	50 μ A		
Flash Test	1 kV to 3 kV	100 V		
	3 kV to 12 kV	200 V		
Flash Current	1 kV to 3 kV nominal 50 Hz	100 V		
	3 kV to 12 kV nominal 50 Hz	250 V		
	2 μ A to 200 μ A	2.0 μ A		
	200 μ A to 2 mA	30 μ A		
	2 mA to 20 mA	200 μ A		
	2 μ A to 200 μ A nominal 50 Hz	3.0 μ A		
	200 μ A to 2 mA nominal 50 Hz	30 μ A		
	2 mA to 20 mA nominal 50 Hz	300 μ A		
Load for PAT	0.13 kW	1.0 % + 1.5 Ω	At nominal UK mains supply voltage	
PRESSURE			Calibration of devices with an electrical output may be undertaken.	A
Gas pressure (gauge)				A
Calibration of pressure indicators and gauges	-95 kPa to 200 kPa	150 Pa	Calibration by comparison with digital pressure standards.	A
	200 kPa to 2 MPa	650 Pa		
	2 MPa to 10 MPa	3.0 kPa		
Gas pressure (absolute)				A
Calibration of pressure indicators and gauges	10 kPa to 300 kPa	350 Pa	Calibration by comparison with digital pressure standards.	A
	300 kPa to 2.1 MPa	850 Pa		
	2.1 MPa to 10 MPa	3.2 kPa		
Hydraulic pressure (gauge)				A
Calibration of pressure indicators and gauges	550 kPa to 69 MPa	0.037 %	Calibration by comparison with deadweight tester.	A
Calibrations using water	0 MPa to 100 MPa	30 kPa	Calibration by comparison with digital pressure standards.	A
END				



0143

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services, a division of Lazgill Ltd

Issue No: 38 Issue date: 12 March 2020

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are not mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 % · V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 % · p + (0.12 · 10⁻⁶ · p · 10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.