


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0143 Accredited to ISO/IEC 17025:2005	Calmet Laboratory Services	
	Issue No: 033 Issue date: 17 March 2017	
	11b Upper Teddington Road Kingston-upon-Thames Surrey KT1 4DL	Contact: Mr R Griffiths Tel: +44 (0)20 8977 8455 Fax: +44 (0)20 8614 8048 E-Mail: sales@calmet.co.uk Website: www.calmet.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 11b Upper Teddington Road Kingston-upon-Thames Surrey KT1 4DL	Contact: Mr R Griffiths Tel: +44 (0)20 8977 8455 Fax: +44 (0)20 8614 8048 E-Mail: sales@calmet.co.uk	Dimensional, Electrical, Pressure and Torque A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At customers premises	Mr R Griffiths	Dimensional and Electrical B



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
DIMENSIONAL			NOTES	
LENGTH				
Plain plug gauges (parallel) cylindrical setting standards and rollers	1 to 50 diameter 50 to 100 100 to 200 200 to 300	1.0 2.0 3.0 4.0	1. All linear calibrations may be given in inch units.	A
Plain ring gauges (parallel)	1 to 10 diameter 10 to 50 50 to 100 100 to 200 200 to 300	2.0 1.5 2.0 3.5 5.0	2 The uncertainty quoted is for the departure from flatness, straightness, parallelism or squareness, ie the distance separating the parallel planes which just enclose the surface under consideration.	A
Screw plug gauges (parallel) including check and setting plugs See Note 4	1 to 100 diameter 100 to 200 200 to 300	4.0 5.0 6.0	3 Calibration results may also be given in units of lbf.in and lbf.ft.	A
Screw ring gauges (parallel) See Notes 4 and 5	1 to 50 50 to 150 150 to 300	5.0 6.0 8.0	4 Single and multi-start, symmetrical thread forms only.	A
Screw thread pitch	0.2 to 8	1.5	5. 1 mm to 12 mm diameter range relates to functional test of size using check plugs.	A
Screw thread flank angles	0° to 52°	5.0 minutes of arc		A
Length gauges, flat and spherical ended	25 to 600	1.0 + (8.0 x length in m)		A
Plain gap gauges (parallel)	0.5 to 100 100 to 200	3.0 5.0		A
Parallels	As BS 906:1972 5 to 50 x 100 x 400	From 1.5 up to 5.0		A
Vee blocks	As BS 3731:1987 20 to 150 diameter, vee capacity	From 2.5 up to 7.0		A



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
ANGLE				
Square Blade type	As BS 939:2007 50 to 300 300 to 600	3.0 on squareness 5.0 (See Note 2)		A
Angle plates and box angle plates	As BS 5535:1978 50 to 600	Squareness: 3.0 + (1.0 per 100 mm) Parallelism: 1.0 + (1.0 per 100 mm) See note 2		A
FORM				
Straight edges Cast iron	As BS 5204:Part 1:1975 300 to 8000))) 1.0 +) (2.0 x length in m)		A
Steel, Granite	As BS 5204:Part 2:1977 300 to 2000) See note 2))		
Surface plates Granite Cast iron] As BS 817:2008] 160 x 100 to 4000 x 6000	1.5 + (0.80 x diagonal in m) See Note 2		A & B
MEASURING INSTRUMENTS AND MACHINES				
Micrometers				
External	As BS 870:2008 0 to 900))) Heads 2.0 between) any two points.		A
Internal	As BS 959:2008 0 to 900) Setting and extension) rods 1.0 +) (8.0 x length in m)		
Depth	As BS 6468:2008 0 to 300))		
Three point bore micrometers	3 to 100 100 to 150	Overall performance 5.0 Overall performance 8.0		A
Bore indicators	2 to 100 100 to 150	Overall performance 5.0 Overall performance 8.0		A
Micrometer heads	As BS 1734:1951 0 to 100	1.0		A



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
MEASURING INSTRUMENTS AND MACHINES (cont'd)				
Vernier / Digital / Dial gauges Calliper	BS 887:2008 0 to 1000	Overall performance $10 + (30 \times \text{length in m})$		A
Height	ISO13225:2012 and BS 1643:2008 0 to 1000	Overall performance $8.0 + (27 \times \text{length in m})$		
Depth	BS 6365:2008 0 to 600	Overall performance $10 + (30 \times \text{length in m})$		
Dial gauges and dial test indicators	As BS 907:2008 and BS 2795:1981 0 to 50	1.0		A
Electronic height gauges (including setting masters)	0 to 1000	$1.0 + (5.0 \times \text{length in m})$		A
Profile projectors	10 to 100 magnifications	125 at the screen 5.0 linear 3.0 minutes of arc		A & B
Feeler gauges	As BS 957:2008 0.02 to 1.00	3.0		A
Spirit levels	As BS 3509:1962 and BS 958:1968 5 seconds of arc to 60 minutes of arc nominal sensitivity	Mean sensitivity 10 % of nominal Minimum of 0.50 seconds of arc		A
Electronic indicating levels	0 to 10 minutes of arc	1.0 % of range Minimum 0.50 seconds of arc		A
TORQUE				
Hand torque tools	As BS EN ISO 6789:2003 0.1 N·m to 3000 N·m	1.0 % of applied torque See Note 3		A



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
ELECTRICAL				
DC Resistance				
Sourcing	100 $\mu\Omega$ 1 m Ω 10 m Ω 100 m Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1.0 G Ω	70 ppm 35 ppm 15 ppm 12 ppm 6.0 ppm 5.0 ppm 4.0 ppm 4.0 ppm 3.0 ppm 3.0 ppm 6.0 ppm 10 ppm 0.60 % 0.65 %		A
Measurement	0 Ω to 100 Ω 100 Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω	10 ppm + 0.20 m Ω 10 ppm 15 ppm 20 ppm		A
	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω	25 ppm + 1.0 m Ω 15 ppm 10 ppm 45 ppm 60 ppm		B
Generation	0.001 Ω to 10 Ω 10 Ω to 10 k Ω 10 k Ω to 1 M Ω 1 M Ω to 10 M Ω	100 ppm + 1.0 m Ω 25 ppm 120 ppm 250 ppm		B
DC Voltage				
Standard Cell Values	10 V	0.80 ppm		A
Other Values	1 μ V to 1.1 V 1.1 V to 11 V 11 V to 100 V 100 V to 1 kV 1 kV to 12 kV	0.80 ppm + 0.20 μ V 2.4 ppm + 0.20 μ V 2.5 ppm 2.5 ppm 0.50 %		A
Measurement	0 mV to 100 mV 100 mV to 100 V 100 V to 1000 V 1 kV to 12 kV	25 ppm + 2.0 μ V 20 ppm 20 ppm 0.50 %		B



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
DC Voltage Continued Generation	0 V to 1 V 1 V to 1000 V	25 ppm + 1.0 μ V 35 ppm		B
DC Voltage Ratio	0 to unity	0.10 ppm of input + 0.40 μ V	For input voltages in the range 1 to 100 V	A
DC Current Measurement and sourcing	1 μ A to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A 3 A to 10 A 20 A 50 A and 100 A 100 A to 1000 A	15 ppm 20 ppm 25 ppm 25 ppm 45 ppm 80 ppm 0.020 % 0.55 %	For the Calibration of Current Clamps	A
Generation	1.0 μ A to 100 mA 100 mA to 10 A	0.010 % 0.035 %		B
AC Voltage	10 mV to 250 mV 40 Hz to 20 kHz	0.014 %		A
	250 mV to 16 V 20 Hz to 50 kHz 50 kHz 100 kHz	40 ppm 30 ppm 35 ppm		
	16 V to 125 V 20 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz 100 kHz	30 ppm 25 ppm 25 ppm 25 ppm		
	125 V to 1000 V 20 Hz to 10 kHz 50 kHz	60 ppm 50 ppm		
	250 V 50 kHz	40 ppm		
	1 kV to 10 kV 50 Hz to 60 Hz	1.0 %		
Generation	40 Hz to 100 kHz 10 mV to 10 V 10 V to 1000 V	0.020 % 0.030 %		B



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC Current				A
Specific Values	<i>40 Hz to 1 kHz</i> 1 μ A and 10 μ A 100 μ A 1 mA 10 mA 100 mA 200 mA, 500 mA and 1 A 2 A 5 A 10 A 20 A <i>1 kHz to 5 kHz</i> 1 μ A and 10 μ A 100 μ A <i>5 kHz to 10 kHz</i> 100 μ A <i>1 kHz to 10 kHz</i> 1 mA 10 mA 100 mA 200 mA, 500 mA and 1 A 2 A 5 A 10 A 20 A	40 ppm 20 ppm 10 ppm 15 ppm 10 ppm 55 ppm 60 ppm 65 ppm 75 ppm 100 ppm 130 ppm 25 ppm 30 ppm 20 ppm 15 ppm 10 ppm 65 ppm 80 ppm 80 ppm 70 ppm 120 ppm		
Other Values	<i>40 Hz to 1 kHz</i> 1 μ A to 10 μ A 10 μ A to 100 mA 100 mA to 10 A 10 A to 20 A <i>45 Hz to 60 Hz</i> 50 A to 100 A 100 A to 500 A 500 A to 1000 A	40 ppm 20 ppm 75 ppm 100 ppm 150 ppm 0.55 % 0.60 %	For the Calibration of Current Clamps	
Generation	<i>40 Hz to 10 kHz</i> 10 μ A to 10 A	0.030 %		B



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC Resistance	40 Hz to 1 kHz			A
	0.05 Ω	120 ppm		
	0.1 Ω	90 ppm		
	0.2 Ω	80 ppm		
	0.5 Ω	75 ppm		
	1 Ω	70 ppm		
	2 Ω	70 ppm		
	5 Ω	70 ppm		
	10 Ω	45 ppm		
	100 Ω	45 ppm		
	1 k Ω	45 ppm		
	10 k Ω	50 ppm		
100 k Ω	65 ppm			
AC/DC Voltage Transfer	250 mV to 16 V			A
	20 Hz to 20 kHz	40 ppm		
	20 kHz to 50 kHz	40 ppm		
	50 kHz	30 ppm		
	100 kHz	35 ppm		
	16 V to 125 V			
	20 Hz to 20 kHz	35 ppm		
	20 kHz to 50 kHz	25 ppm		
	50 kHz	25 ppm		
	100 kHz	25 ppm		
	250 V			
	20 Hz to 10 kHz	60 ppm		
20 kHz	35 ppm			
50 kHz	40 ppm			
500 V to 1 kV				
20 Hz to 10 kHz	85 ppm			
20 kHz	120 ppm			
AC Power	47 Hz to 63 Hz			A
	Voltages 60 V to 240 V Currents 0.5 A to 5 A 15 W to 1200 W	90 ppm))) Unity to 0.5 power) factor lagging or) leading	
	Voltages 60 V to 240 V Currents 5 A to 100 A 150 W to 24 kW	150 ppm)))	
	Combination of specific voltage and current values: V = 75, 100, 150, 300 V I = 0.5, 1, 2, 5, 10, 20 A	0.035 %) Unity to 0.5 power) factor lagging or) leading. Calibrations at) lower power factors) can be carried out to) greater uncertainties.	



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
Frequency	0.2 Hz to 1 kHz 1 kHz to 1000 MHz	2 in 10^8 2 in 10^8	Multi-period measurement Frequency measurement	A
Elapsed time	100 ms to 1 day	20 ms	Stop watches	
Rise/Fall Time	160 ps to 500 ns	5.0 % + 20 ps	For the calibration of	A
Bandwidth	50 kHz to 1.0 GHz	5.0 %	Oscilloscopes	A
Capacitance	<i>At 1 kHz</i> 1 nF to 10 μ F	0.077 %		
Temperature simulation				
Ambient in support of reference junction	17 °C to 23 °C	0.10 °C		
Temperature indicators, calibration by electrical simulation	Simulated temperature	Uncertainty of simulated temperature	including cold junction compensation	A
Type K	-200 °C to 1370 °C	0.20 °C		
Type J	-200 °C to 750 °C	0.20 °C		
Type N	-200 °C to 400 °C	0.20 °C		
Type T	-200 °C to 1300 °C	0.20 °C		
Type R	0 °C to 1700 °C	0.20 °C		
Type S	0 °C to 1700 °C	0.20 °C		
Resistance thermometer (Pt 100)	-200°C to 800°C	0.030 °C		
Type K	-200 °C to 1370 °C	0.50 °C		B
Type J	-200 °C to 750 °C	0.50 °C		
Type N	-200 °C to 400 °C	0.80 °C		
Type T	-200 °C to 1300 °C	0.50 °C		
Resistance thermometer (Pt 100)	-200°C to 800°C	0.10 °C		



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
Measurements to support 17 th edition type test equipment				A & B
Insulation Resistance	100 kΩ 1 MΩ 5 MΩ 10 MΩ 100 MΩ 1 GΩ	500 Ω 80 kΩ 80 kΩ 500 kΩ 5.0 MΩ 50 MΩ		
Insulation Voltage	50 V to 250 V 250 V to 500 V 500 V to 1 kV	5.0 V 10 V 15 V		
RCD Trip Current	10 mA to 15 mA 15 mA to 50 mA 100 mA to 100 mA 100 mA to 150 mA 150 mA to 500 mA 500 mA to 1 A	500 μA 1.0 mA 3.0 mA 5.0 mA 10 mA 25 mA		
RCD Trip Time	0 s to 100 ms	2.0 ms		
Continuity Resistance	10 mΩ to 1 Ω 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ	100 mΩ 500 mΩ 3.0 Ω 50 Ω 300 Ω		
Loop impedance	100 mΩ to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ	100 mΩ 1.0 Ω 15 Ω		
PAT test Voltage	50 V to 100 V nominal 50 Hz 100 V to 400 V nominal 50 Hz	1.0 V 1.5 V		
Earth Bond resistance	5 mΩ to 50 mΩ 50 mΩ to 5 Ω	5.0 mΩ 10 mΩ		
Earth bond current	10 mA to 500 mA 500 mA to 10 A 10 A to 25 A	10 mA 20 mA 65 mA		
PAT Leakage current	50 μA to 7.7 mA	50 μA		
Flash Test	1 kV to 3 kV 3 kV to 12 kV 1 kV to 3 kV nominal 50 Hz 3 kV to 12 kV nominal 50 Hz	100 V 200 V 100 V 250 V		



0143
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calmet Laboratory Services
Issue No: 033 Issue date: 17 March 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
Measurements to support 17 th edition type test equipment continued				A & B
Flash Current	2 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA	2.0 μ A 30 μ A 200 μ A		
Load for PAT	2 μ A to 200 μ A nominal 50 Hz 200 μ A to 2 mA nominal 50 Hz 2 mA to 20 mA nominal 50 Hz 0.13 kW	3.0 μ A 30 μ A 300 μ A 1.0 % + 1.5 Ω	At nominal UK mains supply voltage	
PRESSURE			Calibration of devices with an electrical output may be undertaken.	
Gas pressure (gauge)				A
Calibration of pressure indicators and gauges	-95 kPa to 200 kPa 200 kPa to 2 MPa 2 MPa to 10 MPa	150 Pa 650 Pa 3.0 kPa		A
Gas pressure (absolute)				A
Calibration of pressure indicators and gauges	10 kPa to 300 kPa 300 kPa to 2.1 MPa 2.1 MPa to 10 MPa	350 Pa 850 Pa 3.2 kPa		A
Hydraulic pressure (gauge)				A
Calibration of pressure indicators and gauges	550 kPa to 69 MPa	0.043 %		A
Calibrations using water	0 MPa to 100 MPa	30 kPa		A
END				