

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0221 Accredited to ISO/IEC 17025:2017	Druck Ltd Issue No: 068 Issue date: 09 April 2021	
	Fir Tree Lane Groby Leicester LE6 0FH	Contact: Mr S Berdej Tel: +44 (0)116-231 7100 Fax: +44 (0)116-231 7101 E-Mail: sensing.grobyukas@bakerhughes.com Website: www.druck.com

Calibration performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Fir Tree Lane Groby Leicester LE6 0FH Contact: Mr S Berdej Tel: +44 (0)116-231 7100 Fax: +44 (0)116-231 7101 Email: sensing.grobyukas@bakerhughes.com	Electrical Pressure Temperature Mass Humidity	UK

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Pressure	Site



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
<p>PRESSURE</p> <p><u>Gas pressure (absolute)</u></p> <p>Calibration of pressure measuring instruments and gauges and "Pressure equivalent" calibration of Dead Weight Testers (pressure balances supplied with an associated mass set) and Effective area calibration of Dead Weight Testers</p> <p><u>Gas pressure (gauge)</u></p> <p>Calibration of pressure measuring instruments and gauges and "Pressure equivalent" calibration of Dead Weight Testers (pressure balances supplied with an associated mass set) and Effective area calibration of Dead Weight Testers</p> <p><u>Hydraulic pressure (gauge)</u></p> <p>Calibration of pressure measuring instruments and gauges. "Pressure equivalent" calibration of Dead Weight Testers (Pressure balance with associated mass set). Effective area calibration of Dead Weight Testers.</p>	<p>0 kPa to 450 kPa 450 kPa kPa to 3.1 MPa 3.1 MPa to 21.1 MPa 21.1 MPa to 40.1 MPa</p> <p>- 100 kPa to - 3.5 kPa - 3.5 kPa to 3.5 kPa 3.5 kPa to 350 kPa 350 kPa to 3 MPa 3 MPa to 21 MPa 21 MPa to 40 MPa</p> <p>0.50 MPa to 140 MPa 140 MPa to 500 MPa</p>	<p>0.0025 % + 0.40 Pa 0.0028 % + 0.40 Pa 0.0032 % + 0.40 Pa 0.0039 % + 11 Pa</p> <p>0.0025 % + 0.60 Pa 0.55 Pa 0.0025 % 0.0028 % 0.0032 % 0.0039 %</p> <p>0.0036 % + 0.12 ppm/MPa</p> <p>0.010 % + 0.30 ppm/MPa</p>	<p>Methods consistent with EURAMET CG3 and CG17</p> <p>Calibration of pressure measuring devices with an electrical output may be undertaken.</p> <p>Absolute pressure calibrations can be undertaken using gauge pressure generation and the associated barometric pressure with the additional uncertainty of 11 Pa</p>	<p>UK and Site</p> <p>UK and Site</p> <p>UK and Site</p>



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL			Calibration by comparison with a reference standard	UK
DC Resistance measurement (At low current)	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω	17 ppm + 5.3 $\mu\Omega$ 7.0 ppm + 19 $\mu\Omega$ 6.8 ppm + 280 $\mu\Omega$ 5.9 ppm + 2.3 m Ω		
	2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω	6.2 ppm + 22 m Ω 5.8 ppm + 140 m Ω 12 ppm + 1.3 Ω 72 ppm + 120 Ω		
	20 M Ω to 200 M Ω 200 M Ω to 2 G Ω	750 ppm + 120 k Ω 750 ppm + 1.2 M Ω		
DC Resistance measurement (At higher current)	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω	18 ppm + 5.9 $\mu\Omega$ 4.7 ppm + 20 $\mu\Omega$ 4.3 ppm + 100 $\mu\Omega$ 5.9 ppm + 1.1 m Ω		
	2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω	6.2 ppm + 8.8 m Ω 5.8 ppm + 95 m Ω 10 ppm + 1.3 Ω 72 ppm + 120 Ω		
	20 M Ω to 200 M Ω 200 M Ω to 2 G Ω	750 ppm + 12 k Ω 750 ppm + 1.2 M Ω		
DC Resistance Generation Specific values	1 m Ω 10 m Ω 100 m Ω 1 Ω 1.9 Ω	0.545% 0.065 % 245 ppm 91 ppm 55 ppm		
	10 Ω 19 Ω 25 Ω 50 Ω 100 Ω 190 Ω	16 ppm 18 ppm 18 ppm 17 ppm 6 ppm 16 ppm		
	250 Ω 300 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω	18 ppm 8.5 ppm 4.3 ppm 7.5 ppm 7.5 ppm 7.5 ppm		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
DC Resistance Generation Specific values continued	100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω	7.5 ppm 10 ppm 10 ppm 18 ppm 15 ppm		UK
	19 M Ω 100 M Ω 1.0G Ω	35 ppm 50 ppm 250 ppm		
	1.0 k Ω 100 k Ω 1.0 M Ω 10 M Ω 100 M Ω	0.25 % 0.25 % 0.25 % 0.25 % 0.25 %	With applied voltages of 50 V and 100 V	
	1.0 G Ω 10 G Ω	0.50 % 1.0 %		
	1.0 M Ω 10 M Ω 100 M Ω 1.0 G Ω 10 G Ω	0.30 % 0.30 % 0.30 % 0.30 % 0.30 %	With applied voltages of 250 V and 500 V	
	100 G Ω 1.0 T Ω	2.0 % 2.5 %	With an applied voltage of 500 V	
	100 M Ω 1.0 G Ω 10 G Ω 100 G Ω 1.0 T Ω	0.5 % 0.5 % 0.5 % 2.0 % 2.0 %	With an applied voltage of 1000 V	
	1.0 G Ω 10 G Ω 100 G Ω 1.0 T Ω	1.1 % 1.0 % 2.0 % 2.0 %	With an applied voltage of 5000 V	
Other values	0.1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	2.5 m Ω 36 ppm 16 ppm 10 ppm 10 ppm 50 ppm 0.10 % 0.15 % 0.50 %		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
DC Voltage				UK
Voltage Reference Values	1.018 V 10 V	3.2 ppm 1.3 ppm		
Measurement	0 V to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	5.7 ppm + 1.2 μ V 6.1 ppm + 1.3 μ V 2.1 ppm + 5.8 μ V 3.2 ppm + 61 μ V 6.7 ppm + 680 μ V		
DC High Voltage Measurement	0 V to 5 kV 5 kV to 10 kV 10 kV to 15 kV 15 kV to 20 kV 20 kV to 25 kV 25 kV to 30 kV 30 kV to 35 kV 35 kV to 40 kV	1.4 % 0.58 % 0.49 % 0.87 % 0.96 % 0.84 % 0.77 % 0.69 %		
Generation	0 mV to 200 mV 0.2 V to 2 V 2 V to 11 V 11 V to 20 V 20 V to 200 V 200 V to 1100 V	5.9 ppm + 0.70 μ V 3.2 ppm + 1.1 μ V 1.8 ppm + 3.1 μ V 2.6 ppm + 11 μ V 2.4 ppm + 120 μ V 3.0 ppm + 700 μ V		
DC Current				UK
Measurement	0 A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A	8.5 ppm + 1.4 nA 8.8 ppm + 5.3 nA 9.5 ppm + 52 nA 22 ppm + 950 nA 35 ppm + 19 μ A 89 ppm + 460 μ A		
Generation	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A 20 A to 60 A	7.0 ppm + 1.2 nA 10.5 ppm + 3.0 nA 9.2 ppm + 20 nA 10.5 ppm + 120 nA 25 ppm + 2.0 μ A 2 mA 10 mA 40 mA		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
AC Voltage Measurement	50 μ V to 200 mV 20 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 200 mV to 2 V 20 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz 2 V to 20 V 20 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz 20 V to 200 V 20 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz At 100 kHz	8.0 μ V 10 μ V 8.0 μ V 7.2 μ V 9.5 μ V 20 μ V 21 μ V 40 μ V 45 μ V 45 μ V 45 μ V 50 μ V 61 μ V 90 μ V 290 μ V 330 μ V 2.9 mV 24 mV 450 μ V 450 μ V 460 μ V 490 μ V 730 μ V 790 μ V 1.8 mV 2.9 mV 40 mV 270 mV 6.0 mV 6.0 mV 6.0 mV 6.0 mV 7.5 mV 15 mV 29 mV 33 mV 240 mV		UK



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
AC Voltage Measurement Continued	200 V to 1 kV 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz	75 mV 80 mV 90 mV 100 mV 130 mV		UK
	At 30 kHz	310 mV		
High Voltage AC Measurement Nom 50 Hz	2 kV 3 kV 4 kV 5 kV 6 kV	1.7 % 1.2 % 0.97 % 2.3 % 0.96 %	Claims for intermediate values more than 10 % away from the nominal value listed will be the highest of the adjacent points	
	7 kV 8 kV 10 kV	0.88 % 0.81 % 3.7 %		
	15 kV 20 kV 25 kV 28 kV	3.4 % 3.2 % 2.0 % 2.4 %		
Generation	10 Hz to 40 Hz 0.1 mV to 2 mV 2 mV to 20 mV 20 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V	7.0 μ V 9.0 μ V 25 μ V 120 μ V 2.4 mV 24 mV		
	40 Hz to 500 Hz 0.1 mV to 2 mV 2 mV to 20 mV 20 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V	7.0 μ V 8.0 μ V 14 μ V 50 μ V 760 μ V 7.0 mV		
	500 Hz to 1 kHz 0.1 mV to 2 mV 2 mV to 20 mV 20 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V	7.0 μ V 8.0 μ V 12 μ V 24 μ V 400 μ V 4.0 mV		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
AC Voltage Generation Continued	<p><i>1 kHz to 10 kHz</i> 0.1 mV to 2 mV 2 mV to 20 mV 20 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V</p> <p><i>10 kHz</i> 2 mV 20 mV 100 mV 1 V 10 V 100 V</p> <p>1000 V <i>55 Hz to 500 Hz</i> <i>500 Hz to 1 kHz</i> <i>1 kHz</i></p>	<p>7.0 μV 9.0 μV 12 μV 24 μV 400 μV 4.0 mV</p> <p>7.0 μV 7.0 μV 12 μV 26 μV 410 μV 4.0 mV</p> <p>58 mV 61 mV 65 mV</p>		UK
AC Current Measurement	<p>30 nA to 200 μA <i>55 Hz to 305 Hz</i> <i>305 Hz to 1 kHz</i> <i>1 kHz to 3 kHz</i> <i>3 kHz to 10 kHz</i> <i>At 10 kHz</i></p> <p>200 μA to 2 mA <i>55 Hz to 305 Hz</i> <i>305 Hz to 1 kHz</i> <i>1 kHz to 3 kHz</i> <i>3 kHz to 10 kHz</i> <i>At 10 kHz</i></p> <p>2 mA to 20 mA <i>55 Hz to 305 Hz</i> <i>305 Hz to 1 kHz</i> <i>1 kHz to 3 kHz</i> <i>3 kHz to 10 kHz</i> <i>At 10 kHz</i></p> <p>20 mA to 200 mA <i>55 Hz to 305 Hz</i> <i>305 Hz to 1 kHz</i> <i>1 kHz to 3 kHz</i> <i>3 kHz to 10 kHz</i> <i>At 10 kHz</i></p>	<p>27 nA 25 nA 30 nA 40 nA 62 nA</p> <p>250 nA 250 nA 280 nA 280 nA 330 nA</p> <p>2.6 μA 2.4 μA 2.7 μA 4.0 μA 4.0 μA</p> <p>25 μA 25 μA 25 μA 25 μA 30 μA</p>		UK



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd

Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
AC Current Measurement Continued	200 mA to 2 A 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz At 10 kHz	260 μ A 280 μ A 460 μ A 600 μ A 590 μ A		UK
AC Current Generation	2 A to 20 A 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz At 10 kHz	12 mA 12 mA 15 mA 16 mA 18 mA		
	55 Hz to 400 Hz 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	20 nA 150 nA 580 nA 5.0 μ A 75 μ A		
	60 Hz 1 A to 2 A 2 A to 10 A 10 A to 20 A 20 A to 50 A	0.15 % 0.80 % 2.8 % 4.7 %		
	400 Hz to 1 kHz 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 200 mA to 1 A	20 nA 180 nA 700 nA 5.2 μ A 75 μ A		
	1 kHz to 5 kHz 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	30 nA 200 nA 1.1 μ A 15 μ A 300 μ A		
	5 kHz 100 μ A 1 mA 10 mA 100 mA 1 A	100 nA 1.0 μ A 2.3 μ A 25 μ A 500 μ A		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
Frequency	10 MHz 1 Hz to 1 GHz	7.0 in 10^{11} 0.20 ppm	Measurement and generation of repetitive waveforms.	UK
Temperature indicators, calibration by electrical simulation				UK
Base metal thermocouples	- 210 °C to + 1360 °C	0.05 °C	Excluding cold junction compensation	
Nobel metal thermocouples	- 50 °C to + 2300 °C	0.17 °C	Excluding cold junction compensation	
Cold junction compensation	Ambient temperature 18 °C to 30 °C At 0 °C	0.56 °C 0.21 °C		
Resistance thermometers	- 200 °C to + 840 °C	0.050 °C		
Temperature simulators, calibration by electrical simulation				
Base metal thermocouples	- 210 °C to + 1360 °C	0.070 °C	Excluding cold junction compensation	
Noble metal thermocouples	- 50 °C to + 2300 °C	0.15 °C	Excluding cold junction compensation	
Base metal thermocouples	- 210 °C to + 1360 °C	0.11 °C	Including cold junction compensation	
Noble metal thermocouples	- 50 °C to + 2300 °C	0.19 °C	Including cold junction compensation	
Resistance thermometers	- 200 °C to + 840 °C	0.002 °C		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
MASS	Nominal value (grams) 0.001 to 0.02 0.05 to 0.2 0.5 1 2 5 10 20 50 100 200 500 1 000 2 000 5 000 10 000 20 000 30 000	(mg) 0.012 0.015 0.018 0.020 0.024 0.030 0.040 0.050 0.060 0.10 0.20 0.50 1.0 2.0 5.0 10 20 30	Intermediate values under 100g can be calibrated with an uncertainty equal to the uncertainty of the next higher nominal value. Intermediate values over 100 g can be calibrated with an uncertainty of 1 part per million (1 mg per kg). Substitution Method	UK
TEMPERATURE Resistance thermometers and electronic thermometers with PRT, thermocouple or thermistor sensors	-75 °C to -40 °C -40 °C to 0 °C 0 °C (ice point) 0.01 °C (Triple Point of Water) 0 °C to 60 °C 60 °C to 150 °C 150 °C to 250 °C	0.048 °C 0.038 °C 0.022 °C 0.017 °C 0.022 °C 0.030 °C 0.045 °C	Calibration performed in liquid baths	UK
HUMIDITY Relative humidity meters	11 %rh 33 %rh 54 %rh 75 %rh 90 %rh For the temperature range 21 °C ± 3 °C	2.2 %rh 2.2 %rh 2.7 %rh 2.7 %rh 3.2 %rh	Calibrations by comparison with dew- point hygrometer and Platinum Resistance Thermometers	UK
END				



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Ltd
Issue No: 068 Issue date: 09 April 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10⁻⁶·p·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.