


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |  |  |
|--|--|--|
|  <p><b>0232</b><br/>Accredited to<br/>ISO/IEC 17025:2017</p>            | <b>Indentec Hardness Testing Machines Ltd</b>  |  |
|  | <b>Issue No: 039    Issue date: 27 August 2021</b>   |  |
| <b>Unit 30 Navigation Drive</b><br><b>Hurst Business Park</b><br><b>Brierley Hill</b><br><b>West Midlands</b><br><b>DY5 1UT</b><br><b>United Kingdom</b> | <b>Contact: Mr. M. Wood</b><br><b>Tel: +44 (0)1384-484070</b><br><b>Fax: +44 (0)1384-481074</b><br><b>E-Mail: sales@indentec.com</b><br><b>Website: www.indentec.com</b> |  |
| <b>Calibration performed by the Organisations at the locations specified below</b>   |  |  |

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details   |                                   | Activity  | Location code |
|--|-----------------------------------|---|---------------|
| <b>Address</b><br>Unit 30 Navigation Drive<br>Hurst Business Park<br>Brierley Hill<br>West Midlands<br>DY5 1UT<br>United Kingdom | <b>Local contact</b><br>Mr M.Wood | Hardness<br>Indentors, ball<br>Ball Indentor holders<br>Test blocks, Brinell<br>Test blocks, Rockwell<br>Test blocks, Vickers<br>Test blocks, Knoop<br>Calibration machines, Rockwell, direct<br>Testing machines, Rockwell, direct<br>Calibration machines, Rockwell, Indirect<br>Testing machines, Rockwell, indirect<br>Calibration machines, Vickers, and Knoop direct<br>Testing machines, Vickers, and Knoop direct<br>Calibration machines, Vickers, and Knoop Indirect<br>Testing machines, Vickers, and Knoop indirect<br>Calibration machines, Brinell, direct<br>Testing machines, Brinell, direct<br>Calibration machines, Brinell, indirect<br>Testing machines, Brinell, indirect | P             |

#### Site activities performed away from the locations listed above:

| Location details             |  | Activity  | Location code |
|------------------------------|--|---|---------------|
| <b>At Customers Premises</b> |  | Hardness<br>Calibration machines, Rockwell, direct<br>Testing machines, Rockwell, direct<br>Calibration machines, Rockwell, Indirect<br>Testing machines, Rockwell, indirect<br>Calibration machines, Vickers and Knoop direct<br>Testing machines, Vickers, and Knoop direct<br>Calibration machines, Vickers, and Knoop Indirect<br>Testing machines, Vickers, and Knoop indirect<br>Calibration machines, Brinell, direct<br>Testing machines, Brinell, direct<br>Calibration machines, Brinell, indirect<br>Testing machines, Brinell, indirect | S             |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity<br>Instrument or Gauge  | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ )  | Remarks   | Location<br>Code |
|---|---|--|---|------------------|
| VERIFICATION OF HARDNESS<br>MACHINES IN SERVICE                                   |   |  |   |                  |
| Direct verification of Rockwell<br>Hardness Calibration and Testing<br>Machines   | Rockwell scales:<br>A, B, C, D, E, F, G, H, K, L,<br>M, P, R, S, V, N, T, W, X<br>and Y<br>Force  | See Note 1<br><br>0.12 % force   | Note 1 The calibration /<br>verification shall be in<br>accordance with the<br>requirements of<br>BS EN ISO 6508-<br>2:2015, BS EN ISO<br>6508-3:2015 and / or<br>ASTM E18-20<br>Accreditation is limited<br>to machines<br>manufactured by<br>Indentec Hardness<br>Testing Machines Ltd &<br>Zwick Roell.<br>Accreditation excludes<br>Rockwell portable<br>hardness testing<br>machines | P & S            |
|   | Time  | 0.10 second  |   |                  |
|   | Length  | 0.10 $\mu$ m   |   |                  |
| Indirect verification of Rockwell<br>Hardness Calibration and Testing<br>Machines | Rockwell scales:<br>HRA Scale<br>80 to 85<br>70 to 80<br>60 to 70<br><br>HRBW Scale<br>80<br>50 to 80<br>10 to 50<br><br>HRC Scale<br>60 to 70<br>40 to 60<br>20 to 40<br><br>HRD Scale<br>70 to 80<br>50 to 70<br>40 to 50<br><br>HREW Scale<br>89<br>75 to 89<br>65 to 75<br><br>HRFW Scale<br>87<br>70 to 87<br>40 to 70<br><br>HRGW Scale<br>80<br>40 to 80<br>10 to 40 | See Note 1<br><br>0.15 HRA<br>0.16 HRA<br>0.28 HRA<br><br>0.42 HRBW<br>0.87 HRBW<br>1.36 HRBW<br><br>0.31 HRC<br>0.32 HRC<br>0.37 HRC<br><br>0.17 HRD<br>0.25 HRD<br>0.27 HRD<br><br>0.54 HREW<br>0.54 HREW<br>0.54 HREW<br><br>0.40 HRFW<br>0.40 HRFW<br>0.54 HRFW<br><br>0.30 HRGW<br>0.30 HRGW<br>0.76 HRGW |   | P & S            |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge   | Range            | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|--|------------------|---|---------|------------------|
| Indirect verification of Rockwell<br>Hardness Calibration and Testing<br>Machines (cont'd) | Rockwell Scales  | See Note 1                                      |         | P & S            |
|  | HRHW Scale       |   |         |                  |
|  | 90               | 0.40 HRHW                                       |         |                  |
|  | 80 to 90         | 0.40 HRHW                                       |         |                  |
|  | 60 to 80         | 0.68 HRHW                                       |         |                  |
|  | HRKW Scale       |   |         |                  |
|  | 70               | 0.40 HRKW                                       |         |                  |
|  | 30 to 70         | 0.40 HRKW                                       |         |                  |
|  | 10 to 30         | 0.64 HRKW                                       |         |                  |
|  | HRLW Scale       |   |         |                  |
|  | 115              | 0.35 HRLW                                       |         |                  |
|  | 90 to 115        | 0.35 HRLW                                       |         |                  |
|  | HRMW Scale       |   |         |                  |
|  | 100              | 0.56 HRMW                                       |         |                  |
|  | 70 to 100        | 0.56 HRMW                                       |         |                  |
|  | Rockwell Scales: | See Note 1                                      |         |                  |
|  | HRPW Scale       |   |         |                  |
|  | 85               | 0.65 HRPW                                       |         |                  |
|  | 40 to 85         | 0.91 HRPW                                       |         |                  |
|  | HRRW Scale       |   |         |                  |
| 120  | 0.23 HRRW        |   |         |                  |
| 100 to 120   | 0.40 HRRW        |   |         |                  |
| HRSW Scale   |                  |   |         |                  |
| 112  | 0.19 HRSW        |   |         |                  |
| 110 to 112   | 0.91 HRSW        |   |         |                  |
| HRVW Scale   |                  |   |         |                  |
| 104 to 120   | 0.20 HRVW        |   |         |                  |
| 80 to 104  | 0.61 HRVW        |   |         |                  |
| HR15N Scale  |                  |   |         |                  |
| 90 to 95   | 0.18 HR15N       |   |         |                  |
| 80 to 90   | 0.18 HR15N       |   |         |                  |
| 40 to 80   | 0.39 HR15N       |   |         |                  |
| HR15TW Scale   |                  |   |         |                  |
| 88 to 100  | 0.21 HR15TW      |   |         |                  |
| 80 to 88   | 0.21 HT15TW      |   |         |                  |
| 20 to 80   | 0.37 HR15TW      |   |         |                  |
| HR15WW Scale   |                  |   |         |                  |
| 89 to 100  | 0.53 HR15WW      |   |         |                  |
| 80 to 89   | 0.44 HR15WW      |   |         |                  |
| HR15XW Scale   |                  |   |         |                  |
| 88 to 100  | 0.33 HR15XW      |   |         |                  |
| 80 to 88   | 0.62 HR15XW      |   |         |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|--|---|---|---------|------------------|
| Indirect verification of Rockwell<br>Hardness Calibration and Testing<br>Machines (cont'd) | Rockwell scales:<br>HR15YW Scale<br>94 to 100<br>85 to 94 | See Note 1<br>0.63 HR15YW<br>1.30 HR15YW        |         | P & S            |
|  | HR30N Scale<br>77 to 85<br>60 to 77<br>40 to 60           | 0.27 HR30N<br>0.27 HR30N<br>0.55 HR30N          |         |                  |
|  | HR30TW Scale<br>57 to 85<br>50 to 57<br>20 to 50          | 0.39 HR30TW<br>0.66 HR30TW<br>0.90 HR30TW       |         |                  |
|  | HR30WW Scale<br>65 to 100<br>40 to 65                     | 0.76 HR30WW<br>0.90 HR30WW                      |         |                  |
|  | HR30XW Scale<br>79 to 100<br>60 to 79                     | 0.15 HR30XW<br>0.99 HR30XW                      |         |                  |
|  | HR30YW Scale<br>88 to 100<br>60 to 88                     | 0.37 HR30YW<br>0.82 HR30YW                      |         |                  |
|  | HR45N Scale<br>67 to 75<br>50 to 67<br>10 to 50           | 0.18 HR45N<br>0.21 HR45N<br>0.43 HR45N          |         |                  |
|  | HR45TW Scale<br>50 to 75<br>40 to 50<br>10 to 40          | 0.40 HR45TW<br>0.40 HR45TW<br>0.73 HR45TW       |         |                  |
|  | HR45WW Scale<br>49 to 100<br>10 to 49                     | 0.12 HR45WW<br>0.29 HR45WW                      |         |                  |
|  | HR45XW Scale<br>69 to 100<br>40 to 69                     | 0.34 HR45XW<br>0.81 HR45XW                      |         |                  |
|  | HR45YW Scale<br>82 to 100<br>60 to 82                     | 0.29 HR45YW<br>0.94 HR45YW                      |         |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge   | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |       |
|--|---|---|--|------------------|-------|
| Direct verification of Vickers and Knoop hardness Calibration and testing machines & Indentation Measuring Devices | Vickers scales:<br>HV 50 to HV 0.05<br>HK 1.0 to HK 0.05  | See Note 2                                      | Note 2 The calibration / verification shall be in accordance with the requirements of BS EN ISO 6507-2:2018, BS EN ISO 6507-3:2018, BS EN ISO 4545-2:2017, BS EN ISO 4545-3:2017, ASTM E384-17 and ASTM E92-17 | P & S            |       |
|  | Force   | 0.12 % force                                    |  |                  |       |
|  | Time  | 0.10 second                                     |  |                  |       |
|  | Length  | 1.0 $\mu$ m                                     |  |                  |       |
| Indirect verification of Vickers hardness Calibration and testing machines   | Vickers scales:<br>HV 100 200<br>HV 100 400<br>HV 100 700 | See Note 2<br>1.2 HV<br>3.4 HV<br>4.1 HV        |  |                  | P & S |
|  | HV 50 200<br>HV 50 400<br>HV 50 700                       | 1.9 HV<br>3.5 HV<br>6.3 HV                      |  |                  |       |
|  | HV 30 200<br>HV 30 400<br>HV 30 700                       | 2.0 HV<br>4.4 HV<br>9.3 HV                      |  |                  |       |
|  | HV 20 200<br>HV 20 400<br>HV 20 700                       | 2.5 HV<br>6.2 HV<br>11.0 HV                     |  |                  |       |
|  | HV 10 200<br>HV 10 400<br>HV 10 700                       | 3.1 HV<br>7.7 HV<br>14.9 HV                     |  |                  |       |
|  | HV5 200<br>HV5 400<br>HV5 700                             | 3.9 HV<br>11.0 HV<br>19.7 HV                    |  |                  |       |
|  | HV3 200<br>HV3 400<br>HV3 700                             | 6.9 HV<br>16.3 HV<br>31.0 HV                    |  |                  |       |
|  | HV1 200<br>HV1 400<br>HV1 700                             | 8.7 HV<br>21.4 HV<br>44.0 HV                    |  |                  |       |
|  | HV 0.5 200<br>HV 0.5 400<br>HV 0.5 700                    | 5.0 HV<br>15.0 HV<br>17.0 HV                    |  |                  |       |
|  | HV 0.3 200<br>HV 0.3 400<br>HV 0.3 700                    | 6.0 HV<br>16.0 HV<br>19.0 HV                    |  |                  |       |
|  | HV 0.2 200<br>HV 0.2 400<br>HV 0.2 700                    | 7.0 HV<br>17.0 HV<br>20.0 HV                    |  |                  |       |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge  | Range  | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |
|---|--|---|--|------------------|
| Indirect verification of Vickers<br>hardness Calibration and testing<br>machines (cont'd) | HV 0.1 200   | 10.0 HV   |  | P & S            |
|   | HV 0.1 400   | 30.0 HV   |  |                  |
|   | HV 0.1 700   | 40.0 HV   |  |                  |
| Indirect verification of Knoop<br>hardness Calibration and testing<br>machines            | Knoop Scales   | See Note 2                                      |  | P & S            |
|   | HK 0.05 200  | 8.5 HK  |  |                  |
|   | HK 0.05 400  | 19.0 HK   |  |                  |
|   | HK 0.05 700  | 27.0 HK   |  |                  |
|   | Knoop Scales:  | See Note 2                                      |  |                  |
|   | HK 0.1 200   | 8.0 HK  |  |                  |
|   | HK 0.1 400   | 18.0 HK   |  |                  |
|   | HK 0.1 700   | 25.0 HK   |  |                  |
|   | HK 0.2 200   | 7.0 HK  |  |                  |
|   | HK 0.2 400   | 17.0 HK   |  |                  |
|   | HK 0.2 700   | 20.0 HK   |  |                  |
|   | HK 0.3 200   | 6.0 HK  |  |                  |
|   | HK 0.3 400   | 16.0 HK   |  |                  |
|   | HK 0.3 700   | 19.0 HK   |  |                  |
|   | HK 0.5 200   | 5.0 HK  |  |                  |
| HK 0.5 400  | 15.0 HK  |   |  |                  |
| HK 0.5 700  | 17.0 HK  |   |  |                  |
| HK1   | 200  | 8.7 HK  |  |                  |
|   | 400  | 21.4 HK   |  |                  |
|   | 700  | 44.0 HK   |  |                  |
|   |  |   |  |                  |
|   |  |   |  |                  |
| Direct verification of Brinell<br>Hardness Calibration and Testing<br>Machines            | Brinell scales:<br>From HBW 10/3000 to<br>HBW 1/1<br>Force | See Note 3                                      | 3 The calibration /<br>verification shall be in<br>accordance with the<br>requirements of<br>BS EN ISO 6506-<br>2:2018 BS EN ISO<br>6506-3:2014 and /or<br>ASTM E10-18 | P & S            |
|   | Time   | 0.24 %  |  |                  |
|   | Length   | 0.10 second                                     |  |                  |
|   |  | 10 $\mu$ m                                      |  |                  |
| Indirect verification of Brinell<br>Hardness Calibration and Testing<br>Machines          | Brinell scales:<br>Scale 10/3000<br>600HBW to 140 HBW      | See Notes 3                                     |  | P & S            |
|   | Scale 10/1500<br>299 HBW to 55 HBW                         | 8.0 HBW to 2.2 HBW                              |  |                  |
|   | Scale 10/1000<br>169 HBW to 55 HBW                         | 4.1 HBW to 1.2 HBW                              |  |                  |
|   | Scale 5/750<br>600 HBW to 140 HBW                          | 2.3 HBW to 1.2 HBW                              |  |                  |
|   | Scale 5/250<br>169 HBW to 55 HBW                           | 9.8 HBW to 2.4 HBW                              |  |                  |
|   |  | 2.7 HBW to 1.3 HBW                              |  |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge  | Range                                 | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks   | Location<br>Code |
|---|---------------------------------------|---|---|------------------|
| Indirect verification of Brinell<br>Hardness Calibration and Testing<br>Machines (cont'd) | Scale 2.5/187.5<br>600 HBW to 140 HBW | 16 HBW to 2.9 HBW                               |   | P & S            |
|   | Scale 1/1<br>21.8 HBW to 3.18 HBW     | 1.04 HBW to 0.09 HBW                            |   |                  |
| Calibration of indenter holders   | 1/16" to 1/2" ball holders            | See Note 4                                      | Note 4 The calibration /<br>verification shall be in<br>accordance with the<br>requirements of ASTM<br>E18-20. The hardness<br>of the ball holder can<br>also be made using the<br>Vickers hardness scale<br>and the results<br>converted to Rockwell<br>hardness values.<br>Note 5 The calibration /<br>verification shall be in<br>accordance with the<br>requirements of<br>BS EN ISO 6508 and<br>ASTM E18-20 for<br>Rockwell and<br>BS EN ISO 6506 and<br>ASTM E10-18 for<br>Brinell, | P                |
|   | Ball Protrusion                       | 3.5 $\mu$ m                                     |   |                  |
|   | Ball holder hardness                  | 0.37 HRC  |   |                  |
| Calibration of indentor balls   | 1 mm to 12.7 mm (1/2 inch)            | See Note 5                                      |   |                  |
| Calibration of Rockwell<br>Standardised Hardness Blocks                                   | Rockwell scales:                      | See Note 6                                      | Note 6 The calibration<br>shall be in accordance<br>with the requirements of<br>BS EN ISO 6508-<br>3:2015 and / or ASTM<br>E18-20   | P                |
|   | HRA Scale                             |   |   |                  |
|   | 87 to 92                              | 0.10 HRA  |   |                  |
|   | 80 to 87                              | 0.15 HRA  |   |                  |
|   | 70 to 80                              | 0.16 HRA  |   |                  |
|   | 60 to 70.0                            | 0.28 HRA  |   |                  |
|   | HRBW Scale                            |   |   |                  |
|   | 80 to 100                             | 0.42 HRBW                                       |   |                  |
|   | 50 to 80                              | 0.87 HRBW                                       |   |                  |
|   | 10 to 50                              | 1.0 HRBW  |   |                  |
|   | HRC Scale                             |   |   |                  |
|   | 60 to 72                              | 0.31 HRC  |   |                  |
|   | 40 to 60                              | 0.32 HRC  |   |                  |
|   | 10 to 40                              | 0.37 HRC  |   |                  |
|   | HRD Scale                             |   |   |                  |
|   | 70 to 80                              | 0.17 HRD  |   |                  |
|   | 50 to 79                              | 0.25 HRD  |   |                  |
|   | 40 to 50                              | 0.27 HRD  |   |                  |
|   | HREW Scale                            |   |   |                  |
|   | 89 to 100                             | 0.54 HREW                                       |   |                  |
|   | 75 to 89                              | 0.54 HREW                                       |   |                  |
|   | 65 to 75                              | 0.54 HREW                                       |   |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge                            | Range                    | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|---|--------------------------|---|---------|------------------|
| Calibration of Rockwell<br>Standardised Hardness Blocks<br>(cont'd) | HRFW Scale<br>87 to 100  | 0.40 HRFW                                       |         | P                |
|   | 70 to 87                 | 0.40 HRFW                                       |         |                  |
|   | 40 to 70                 | 0.54 HRFW                                       |         |                  |
|   | HRGW Scale<br>80 to 83   | 0.30 HRGW                                       |         |                  |
|   | 40 to 80                 | 0.30 HRGW                                       |         |                  |
|   | 10 to 40                 | 0.76 HRGW                                       |         |                  |
|   | HRHW Scale<br>90 to 100  | 0.40 HRHW                                       |         |                  |
|   | 80 to 90                 | 0.40 HRHW                                       |         |                  |
|   | 60 to 80                 | 0.68 HRHW                                       |         |                  |
|   | HRKW Scale<br>70 to 100  | 0.40 HRKW                                       |         |                  |
|   | 30 to 70                 | 0.40 HRKW                                       |         |                  |
|   | HRLW Scale<br>114 to 123 | 0.35 HRLW                                       |         |                  |
|   | 90 to 114                | 0.35 HRLW                                       |         |                  |
|   | HRMW Scale<br>100 to 118 | 0.56 HRMW                                       |         |                  |
|   | 68 to 100                | 0.56 HRMW                                       |         |                  |
| HRPW Scale<br>85 to 112   | 0.65 HRPW                |   |         |                  |
| Rockwell Scales<br>HRRW Scale<br>120 to 123                         | See Note 1<br>0.23 HRRW  |   |         |                  |
| 86 to 120   | 0.40 HRRW                |   |         |                  |
| HRSW Scale<br>112 to 123  | 0.19 HRSW                |   |         |                  |
| 107 to 112  | 0.91 HRSW                |   |         |                  |
| HRVW Scale<br>104 to 120  | 0.20 HRVW                |   |         |                  |
| 80 to 104   | 0.61 HRVW                |   |         |                  |
| HR15N Scale<br>90 to 95   | 0.18 HR15N               |   |         |                  |
| 80 to 90  | 0.18 HR15N               |   |         |                  |
| 40 to 80  | 0.39 HR15N               |   |         |                  |
| HR15TW Scale<br>88 to 100   | 0.21 HR15TW              |   |         |                  |
| 80 to 88  | 0.21 HT15TW              |   |         |                  |
| 20 to 80  | 0.37 HR15TW              |   |         |                  |





0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge                            | Range  | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|---|--|---|---------|------------------|
| Calibration of Rockwell<br>Standardised Hardness Blocks<br>(cont'd) | HR15WW Scale<br>89 to 100<br>80 to 89            | 0.53 HR15WW<br>0.44 HR15WW                      |         | P                |
|   | HR15XW Scale<br>88 to 100<br>80 to 88            | 0.33 HR15XW<br>0.62 HR15XW                      |         |                  |
|   | HR15YW Scale<br>88 to 98                         | 0.63 HR15YW                                     |         |                  |
|   | HR30N Scale<br>77 to 87<br>60 to 77<br>40 to 60  | 0.27 HR30N<br>0.27 HR30N<br>0.55 HR30N          |         |                  |
|   | HR30TW Scale<br>57 to 85<br>50 to 57<br>20 to 50 | 0.39 HR30TW<br>0.66 HR30TW<br>0.90 HR30TW       |         |                  |
|   | HR30WW Scale<br>65 to 100                        | 0.76 HR30WW                                     |         |                  |
|   | HR30XW Scale<br>79 to 100                        | 0.15 HR30XW                                     |         |                  |
|   | HR30YW Scale<br>88 to 100                        | 0.37 HR30YW                                     |         |                  |
|   | HR45N Scale<br>67 to 80<br>50 to 67<br>19 to 50  | 0.18 HR45N<br>0.21 HR45N<br>0.43 HR45N          |         |                  |
|   | HR45TW Scale<br>50 to 75<br>40 to 50<br>10 to 40 | 0.40 HR45TW<br>0.40 HR45TW<br>0.73 HR45TW       |         |                  |
|   | HR45WW Scale<br>49 to 100                        | 0.12 HR45WW                                     |         |                  |
|   | HR45XW Scale<br>69 to 100                        | 0.34 HR45XW                                     |         |                  |
|   | HR45YW Scale<br>82 to 100                        | 0.29 HR45YW                                     |         |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge            | Range              | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |
|---|--------------------|---|--|------------------|
| Calibration of Vickers Reference<br>Hardness Blocks | Vickers scales:    | See Note 7                                      | Note 7 The calibration shall be in accordance with the requirements of BS EN ISO 6507-3:2018, ASTM E384-17 and ASTM E92-17 | P                |
|   | HV0.010 35 to 45   | 4.8 HV  |  |                  |
|   | HV0.025 35 to 116  | 11.9 HV   |  |                  |
|   | HV0.050 35 to 100  | 6.9 HV  |  |                  |
|   | HV0.050 100 to 200 | 19.2 HV   |  |                  |
|   | HV0.050 200 to 232 | 23.9 HV   |  |                  |
|   | HV0.1 35 to 100    | 4.9 HV  |  |                  |
|   | HV0.1 100 to 200   | 13.6 HV   |  |                  |
|   | HV0.1 200 to 300   | 24.8 HV   |  |                  |
|   | HV0.1 300 to 400   | 38.5 HV   |  |                  |
|   | HV0.1 400 to 464   | 47.7 HV   |  |                  |
|   | HV0.2 35 to 100    | 3.5 HV  |  |                  |
|   | HV0.2 100 to 200   | 9.8 HV  |  |                  |
|   | HV0.2 200 to 300   | 17.8 HV   |  |                  |
|   | HV0.2 300 to 400   | 27.2 HV   |  |                  |
|   | HV0.2 400 to 500   | 37.8 HV   |  |                  |
|   | HV0.2 500 to 600   | 49.6 HV   |  |                  |
|   | HV0.2 600 to 700   | 62.9 HV   |  |                  |
|   | HV0.2 700 to 800   | 76.9 HV   |  |                  |
|   | HV0.2 800 to 900   | 91.3 HV   |  |                  |
|   | HV0.2 900 to 927   | 95.4 HV   |  |                  |
|   | HV0.3 35 to 100    | 2.9 HV  |  |                  |
|   | HV0.3 100 to 200   | 8.1 HV  |  |                  |
|   | HV0.3 200 to 300   | 14.6 HV   |  |                  |
|   | HV0.3 300 to 400   | 22.4 HV   |  |                  |
|   | HV0.3 400 to 500   | 31.1 HV   |  |                  |
|   | HV0.3 500 to 600   | 40.7 HV   |  |                  |
|   | HV0.3 600 to 700   | 51.4 HV   |  |                  |
| HV0.3 700 to 800                                    | 62.6 HV            |   |  |                  |
| HV0.3 800 to 900                                    | 74.5 HV            |   |  |                  |
| HV0.3 900 to 1000                                   | 87.4 HV            |   |  |                  |
| HV0.3 1000 to 1100                                  | 100.7 HV           |   |  |                  |
| HV0.3 1100 to 1200                                  | 115.4 HV           |   |  |                  |
| HV0.3 1200 to 1300                                  | 129.2 HV           |   |  |                  |
| HV0.3 1300 to 1391                                  | 143.2 HV           |   |  |                  |
| HV0.5 35 to 100                                     | 2.3 HV             |   |  |                  |
| HV0.5 100 to 200                                    | 6.4 HV             |   |  |                  |
| HV0.5 200 to 300                                    | 11.5 HV            |   |  |                  |
| HV0.5 300 to 400                                    | 17.5 HV            |   |  |                  |
| HV0.5 400 to 500                                    | 24.4 HV            |   |  |                  |
| HV0.5 500 to 600                                    | 32.0 HV            |   |  |                  |
| HV0.5 600 to 700                                    | 40.2 HV            |   |  |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge                     | Range              | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|--|--------------------|---|---------|------------------|
| Calibration of Vickers Reference<br>Hardness Blocks (cont'd) | Vickers Scales:    | See Note 7                                      |         | P                |
|  | HV0.5 700 to 800   | 48.8 HV   |         |                  |
|  | HV0.5 800 to 900   | 58.3 HV   |         |                  |
|  | HV0.5 900 to 1000  | 67.9 HV   |         |                  |
|  | HV0.5 1000 to 1100 | 78.1 HV   |         |                  |
|  | HV0.5 1100 to 1200 | 89.4 HV   |         |                  |
|  | HV0.5 1200 to 1300 | 100.8 HV  |         |                  |
|  | HV0.5 1300 to 1400 | 111.7 HV  |         |                  |
|  | HV0.5 1400 to 1500 | 124.1 HV  |         |                  |
|  | HV0.5 1500 to 1600 | 136.8 HV  |         |                  |
|  | HV0.5 1600 to 1700 | 149.3 HV  |         |                  |
|  | HV0.5 1700 to 1800 | 163.5 HV  |         |                  |
|  | HV0.5 1800 to 1900 | 177.1 HV  |         |                  |
|  | HV0.5 1900 to 2000 | 192.3 HV  |         |                  |
|  | HV1 35 to 100      | 1.7 HV  |         |                  |
|  | HV1 100 to 200     | 4.6 HV  |         |                  |
|  | HV1 200 to 300     | 8.4 HV  |         |                  |
|  | HV1 300 to 400     | 12.7 HV   |         |                  |
|  | HV1 400 to 500     | 18.0 HV   |         |                  |
|  | HV1 500 to 600     | 23.0 HV   |         |                  |
|  | HV1 600 to 700     | 28.8 HV   |         |                  |
|  | HV1 700 to 800     | 35.1 HV   |         |                  |
|  | HV1 800 to 900     | 41.8 HV   |         |                  |
|  | HV1 900 to 1000    | 48.8 HV   |         |                  |
|  | HV1 1000 to 1100   | 56.1 HV   |         |                  |
|  | HV1 1100 to 1200   | 64.1 HV   |         |                  |
|  | HV1 1200 to 1300   | 71.9 HV   |         |                  |
|  | HV1 1300 to 1400   | 80.4 HV   |         |                  |
|  | HV1 1400 to 1500   | 88.8 HV   |         |                  |
|  | HV1 1500 to 1600   | 97.5 HV   |         |                  |
|  | HV1 1600 to 1700   | 107.5 HV  |         |                  |
|  | HV1 1700 to 1800   | 116.7 HV  |         |                  |
|  | HV1 1800 to 1900   | 125.7 HV  |         |                  |
|  | HV1 1900 to 2000   | 135.8 HV  |         |                  |
|  | HV2 35 to 100      | 1.3 HV  |         |                  |
|  | HV2 100 to 200     | 3.4 HV  |         |                  |
|  | HV2 200 to 300     | 6.1 HV  |         |                  |
|  | HV2 300 to 400     | 9.3 HV  |         |                  |
|  | HV2 400 to 500     | 12.8 HV   |         |                  |
|  | HV2 500 to 600     | 16.7 HV   |         |                  |
|  | HV2 600 to 700     | 20.9 HV   |         |                  |
|  | HV2 700 to 800     | 25.4 HV   |         |                  |
|  | HV2 800 to 900     | 30.2 HV   |         |                  |
|  | HV2 900 to 1000    | 35.3 HV   |         |                  |
|  | HV2 1000 to 1100   | 40.5 HV   |         |                  |
|  | HV2 1100 to 1200   | 46.1 HV   |         |                  |
|  | HV2 1200 to 1300   | 51.9 HV   |         |                  |
| HV2 1300 to 1400   | 57.7 HV            |   |         |                  |
| HV2 1400 to 1500   | 63.7 HV            |   |         |                  |
| HV2 1500 to 1600   | 70.1 HV            |   |         |                  |
| HV2 1600 to 1700   | 77.0 HV            |   |         |                  |
| HV2 1700 to 1800   | 83.7 HV            |   |         |                  |
| HV2 1800 to 1900   | 90.5 HV            |   |         |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge                     | Range                              | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|--|------------------------------------|---|---------|------------------|
| Calibration of Vickers Reference<br>Hardness Blocks (cont'd) | Vickers Scales:<br>HV2.5 35 to 100 | See Note7<br>1.2 HV<br>3.1 HV                   |         | P                |
|  | HV2.5 100 to 200                   | 5.5 HV  |         |                  |
|  | HV2.5 200 to 300                   | 8.3 HV  |         |                  |
|  | HV2.5 300 to 400                   | 11.5 HV   |         |                  |
|  | HV2.5 400 to 500                   | 15.0 HV   |         |                  |
|  | HV2.5 500 to 600                   | 18.8 HV   |         |                  |
|  | HV2.5 600 to 700                   | 23.0 HV   |         |                  |
|  | HV2.5 700 to 800                   | 27.0 HV   |         |                  |
|  | HV2.5 800 to 900                   | 31.6 HV   |         |                  |
|  | HV2.5 900 to 1000                  | 36.4 HV   |         |                  |
|  | HV2.5 1000 to 1100                 | 41.4 HV   |         |                  |
|  | HV2.5 1100 to 1200                 | 46.5 HV   |         |                  |
|  | HV2.5 1200 to 1300                 | 51.9 HV   |         |                  |
|  | HV2.5 1300 to 1400                 | 57.6 HV   |         |                  |
|  | HV2.5 1400 to 1500                 | 63.4 HV   |         |                  |
|  | HV2.5 1500 to 1600                 | 69.3 HV   |         |                  |
|  | HV2.5 1600 to 1700                 | 75.1 HV   |         |                  |
|  | HV2.5 1700 to 1800                 | 81.5 HV   |         |                  |
|  | HV2.5 1800 to 1900                 |   |         |                  |
|  | HV3 35 to 100                      | 1.1 HV<br>2.9 HV                                |         |                  |
|  | HV3 100 to 200                     | 5.1 HV  |         |                  |
|  | HV3 200 to 300                     | 7.7 HV  |         |                  |
|  | HV3 300 to 400                     | 10.7 HV   |         |                  |
|  | HV3 400 to 500                     | 13.9 HV   |         |                  |
|  | HV3 500 to 600                     | 17.4 HV   |         |                  |
|  | HV3 600 to 700                     | 21.1 HV   |         |                  |
|  | HV3 700 to 800                     | 25.1 HV   |         |                  |
|  | HV3 800 to 900                     | 29.2 HV   |         |                  |
|  | HV3 900 to 1000                    | 33.5 HV   |         |                  |
|  | HV3 1000 to 1100                   | 38.1 HV   |         |                  |
|  | HV3 1100 to 1200                   | 42.9 HV   |         |                  |
|  | HV3 1200 to 1300                   | 47.9 HV   |         |                  |
|  | HV3 1300 to 1400                   | 52.9 HV   |         |                  |
|  | HV3 1400 to 1500                   | 58.0 HV   |         |                  |
|  | HV3 1500 to 1600                   | 63.6 HV   |         |                  |
|  | HV3 1600 to 1700                   | 69.1 HV   |         |                  |
|  | HV3 1700 to 1800                   | 74.9 HV   |         |                  |
|  | HV3 1800 to 1900                   |   |         |                  |
|  | HV5 35 to 100                      | 0.9 HV  |         |                  |
|  | HV5 100.1 to 200                   | 2.3 HV  |         |                  |
|  | HV5 200.1 to 300                   | 4.1 HV  |         |                  |
|  | HV5 300.1 to 400                   | 6.2 HV  |         |                  |
| HV5 400.1 to 500   | 8.6 HV                             |   |         |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge                     | Range  | Expanded Measurement<br>Uncertainty ( $k = 2$ )  | Remarks | Location<br>Code |
|--|--|--|---------|------------------|
| Calibration of Vickers Reference<br>Hardness Blocks (cont'd) | Vickers Scales:<br>HV5 500 to 600<br>HV5 600 to 700<br>HV5 700 to 800<br>HV5 800 to 900<br>HV5 900 to 1000<br>HV5 1000 to 1100<br>HV5 1100 to 1200<br>HV5 1200 to 1300<br>HV5 1300 to 1400<br>HV5 1400 to 1500<br>HV5 1500 to 1600<br>HV5 1600 to 1700<br>HV5 1700 to 1800<br>HV5 1800 to 1900<br><br>HV10 35 to 100<br><br>HV10 100 to 200<br>HV10 200 to 300<br>HV10 300 to 400<br>HV10 400 to 500<br>HV10 500 to 600<br>HV10 600 to 700<br>HV10 700 to 800<br>HV10 800 to 900<br>HV10 900 to 1000<br>HV10 1000 to 1100<br>HV10 1100 to 1200<br>HV10 1200 to 1300<br>HV10 1300 to 1400<br>HV10 1400 to 1500<br>HV10 1500 to 1600<br>HV10 1600 to 1700<br>HV10 1700 to 1800<br>HV10 1800 to 1900<br><br>HV20 35 to 100<br><br>HV20 100 to 200<br>HV20 200 to 300<br>HV20 300 to 400<br>HV20 400 to 500<br>HV20 500 to 600<br>HV20 600 to 700<br>HV20 700 to 800<br>HV20 800 to 900<br>HV20 900 to 1000<br>HV20 1000 to 1100<br>HV20 1100 to 1200<br>HV20 1200 to 1300<br>HV20 1300 to 1400<br>HV20 1400 to 1500<br>HV20 1500 to 1600<br>HV20 1600 to 1700<br>HV20 1700 to 1800<br>HV20 1800 to 1900 | See Note 2<br>11.2 HV<br>13.9 HV<br>16.6 HV<br>19.9 HV<br>23.0 HV<br>26.4 HV<br>30.1 HV<br>33.8 HV<br>37.6 HV<br>41.2 HV<br>46.1 HV<br>49.8 HV<br>54.0 HV<br>58.6 HV<br><br>0.7 HV<br>1.8 HV<br>3.1 HV<br>4.7 HV<br>6.3 HV<br>8.2 HV<br>10.2 HV<br>12.5 HV<br>14.7 HV<br>17.1 HV<br>19.5 HV<br>22.1 HV<br>24.8 HV<br>27.7 HV<br>30.7 HV<br>33.7 HV<br>36.6 HV<br>39.8 HV<br>41.5 HV<br><br>0.6 HV<br>1.4 HV<br>2.5 HV<br>3.8 HV<br>4.7 HV<br>6.1 HV<br>8.3 HV<br>9.5 HV<br>11.5 HV<br>14.3 HV<br>16.2 HV<br>19.6 HV<br>20.5 HV<br>22.4 HV<br>25.6 HV<br>27.5 HV<br>30.4 HV<br>34.8 HV<br>35.5 HV |         | P                |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge                     | Range                             | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|--|-----------------------------------|---|---------|------------------|
| Calibration of Vickers Reference<br>Hardness Blocks (cont'd) | Vickers Scales:<br>HV30 35 to 100 | See Note 7<br>0.5 HV<br>1.2 HV                  |         | P                |
|  | HV30 100 to 200                   | 2.1 HV  |         |                  |
|  | HV30 200 to 300                   | 3.1 HV  |         |                  |
|  | HV30 300 to 400                   | 4.2 HV  |         |                  |
|  | HV30 400 to 500                   | 5.3 HV  |         |                  |
|  | HV30 500 to 600                   | 6.6 HV  |         |                  |
|  | HV30 600 to 700                   | 8.0 HV  |         |                  |
|  | HV30 700 to 800                   | 9.4 HV  |         |                  |
|  | HV30 800 to 900                   | 10.8 HV   |         |                  |
|  | HV30 900 to 1000                  | 12.4 HV   |         |                  |
|  | HV30 1000 to 1100                 | 14.8 HV   |         |                  |
|  | HV30 1100 to 1200                 | 15.6 HV   |         |                  |
|  | HV30 1200 to 1300                 | 17.3 HV   |         |                  |
|  | HV30 1300 to 1400                 | 19.2 HV   |         |                  |
|  | HV30 1400 to 1500                 |   |         |                  |
|  | HV30 1500 to 1600                 | 21.0 HV   |         |                  |
|  | HV30 1600 to 1700                 | 22.8 HV   |         |                  |
|  | HV30 1700 to 1800                 | 24.7 HV   |         |                  |
|  | HV30 1800 to 1900                 | 26.8 HV   |         |                  |
|  | HV50 35 to 100                    | 0.4 HV<br>1.1 HV                                |         |                  |
|  | HV50 100 to 200                   | 1.8 HV  |         |                  |
|  | HV50 200 to 300                   | 2.6 HV  |         |                  |
|  | HV50 300 to 400                   | 3.5 HV  |         |                  |
|  | HV50 400 to 500                   | 4.5 HV  |         |                  |
|  | HV50 500 to 600                   | 5.5 HV  |         |                  |
|  | HV50 600 to 700                   | 6.6 HV  |         |                  |
|  | HV50 700 to 800                   | 7.5 HV  |         |                  |
|  | HV50 800 to 900                   | 8.9 HV  |         |                  |
|  | HV50 900 to 1000                  | 10.1 HV   |         |                  |
|  | HV50 1000 to 1100                 | 11.5 HV   |         |                  |
|  | HV50 1100 to 1200                 | 12.8 HV   |         |                  |
|  | HV50 1200 to 1300                 | 14.3 HV   |         |                  |
|  | HV50 1300 to 1400                 | 15.7 HV   |         |                  |
|  | HV50 1400 to 1500                 | 17.1 HV   |         |                  |
|  | HV50 1500 to 1600                 | 18.6 HV   |         |                  |
|  | HV50 1600 to 1700                 | 20.1 HV   |         |                  |
|  | HV50 1700 to 1800                 | 21.7 HV   |         |                  |
|  | HV50 1800 to 1900                 |   |         |                  |
|  | HV100 35 to 100                   | 0.4 HV<br>0.9 HV                                |         |                  |
|  | HV100 100 to 200                  | 1.5 HV  |         |                  |
|  | HV100 200 to 300                  | 2.1 HV  |         |                  |
|  | HV100 300 to 400                  | 2.8 HV  |         |                  |
|  | HV100 400 to 500                  | 3.6 HV  |         |                  |
|  | HV100 500 to 600                  | 4.4 HV  |         |                  |
|  | HV100 600 to 700                  | 5.2 HV  |         |                  |
|  | HV100 700 to 800                  |   |         |                  |
|  | HV100 800 to 900                  | 6.1 HV  |         |                  |
| HV100 900 to 1000  | 7.0 HV                            |   |         |                  |
| HV100 1000 to 1100   | 8.0 HV                            |   |         |                  |
| HV100 1100 to 1200   | 9.0 HV                            |   |         |                  |
| HV100 1200 to 1300   | 10.0 HV                           |   |         |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge  | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|---|---|--|--|------------------|
| Calibration of Vickers Reference<br>Hardness Blocks (cont'd)                            | Vickers Scales:<br>HV100 1300 to 1400<br>HV100 1400 to 1500<br>HV100 1500 to 1600<br>HV100 1600 to 1700<br>HV100 1700 to 1800<br>HV100 1800 to 1900   | See Note 7<br>11.0 HV<br>12.1 HV<br>13.2 HV<br>14.3 HV<br>15.5 HV<br>16.7 HV   |  | P                |
| Certification of reference Vickers<br>hardness measurements & Vickers<br>Reading Blocks | All ranges See note 2   | 1.0 $\mu$ m  |  | P                |
| Calibration of Knoop hardness<br>reference blocks                                       | Knoop Scales<br>HK0.005 35 to 100<br><br>HK0.005 100 to 178<br><br>HK0.010 35 to 100<br><br>HK0.010 100 to 200<br>HK0.010 200 to 300<br>HK0.010 300 to 356<br><br>HK0.025 35 to 100<br><br>HK0.025 100 to 200<br>HK0.025 200 to 300<br>HK0.025 300 to 400<br>HK0.025 400 to 500<br>HK0.025 500 to 600<br>HK0.025 600 to 700<br>HK0.025 700 to 800<br>HK0.025 800 to 889<br><br>HK 0.050 35 to 100<br><br>HK 0.050 100 to 200<br>HK 0.050 200 to 300<br>HK 0.050 300 to 400<br>HK 0.050 400 to 500<br>HK 0.050 500 to 600<br>HK 0.050 600 to 700<br>HK 0.050 700 to 800<br>HK 0.050 800 to 900<br>HK 0.050 900 to 1000<br><br>HK 0.1 35 to 100<br><br>HK 0.1 100 to 200<br>HK 0.1 200 to 300<br>HK 0.1 300 to 400<br>HK 0.1 400 to 500<br>HK 0.1 500 to 600<br>HK 0.1 600 to 700<br>HK 0.1 700 to 800<br>HK 0.1 800 to 900<br>HK 0.1 900 to 1000 | See note 8<br>9.6 HK<br>21.2 HK<br><br>7.1 HK<br>18.5 HK<br>33.2 HK<br>42.4 HK<br><br>5.2 HK<br>13.2 HK<br>23.1 HK<br>34.4 HK<br>46.3 HK<br>60.3 HK<br>74.9 HK<br>92.3 HK<br>106.1 HK<br><br>4.3 HK<br>10.5 HK<br>17.8 HK<br>26.4 HK<br>36.2 HK<br>46.2 HK<br>57.5 HK<br>68.9 HK<br>80.3 HK<br>94.6 HK<br><br>3.6 HK<br>8.5 HK<br>14.3 HK<br>20.9 HK<br>28.0 HK<br>35.6 HK<br>43.8 HK<br>52.8 HK<br>61.9 HK<br>71.3 HK | Note 8 The calibration<br>shall be in accordance<br>with the requirements of<br>BS EN ISO 4545-<br>3:2017, ASTM E384-17<br>and ASTM E92-17 | P                |



**0232**  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge  | Range  | Expanded Measurement<br>Uncertainty ( $k = 2$ )  | Remarks | Location<br>Code |
|---|--|--|---------|------------------|
| Calibration of Knoop hardness<br>reference blocks (cont'd)                          | Knoop Scales<br>HK 0.2 35 to 100   | See Note 8<br>3.1 HK<br>7.1 HK<br>11.7 HK<br>17.0 HK<br>22.6 HK<br>28.7 HK<br>34.9 HK<br>41.8 HK<br>48.9 HK<br>56.0 HK |         | P                |
|   | HK 0.2 100 to 200<br>HK 0.2 200 to 300<br>HK 0.2 300 to 400<br>HK 0.2 400 to 500<br>HK 0.2 500 to 600<br>HK 0.2 600 to 700<br>HK 0.2 700 to 800<br>HK 0.2 800 to 900<br>HK 0.2 900 to 1000 | 3.1 HK<br>6.5 HK<br>10.6 HK<br>15.2 HK<br>20.1 HK<br>25.4 HK<br>31.0 HK  |         |                  |
|   | HK 0.3 35 to 100   | 3.1 HK<br>6.5 HK<br>10.6 HK<br>15.2 HK<br>20.1 HK<br>25.4 HK<br>31.0 HK  |         |                  |
|   | HK 0.3 100 to 200<br>HK 0.3 200 to 300<br>HK 0.3 300 to 400<br>HK 0.3 400 to 500<br>HK 0.3 500 to 600<br>HK 0.3 600 to 700<br>HK 0.3 700 to 800<br>HK 0.3 800 to 900<br>HK 0.3 900 to 1000 | 36.9 HK<br>43.0 HK<br>49.5 HK  |         |                  |
|   | HK 0.5 35 to 100   | 2.6 HK<br>5.8 HK<br>9.5 HK<br>13.5 HK<br>17.8 HK<br>22.1 HK<br>26.8 HK<br>31.7 HK<br>36.9 HK<br>42.8 HK                |         |                  |
|   | HK 0.5 100 to 200<br>HK 0.5 200 to 300<br>HK 0.5 300 to 400<br>HK 0.5 400 to 500<br>HK 0.5 500 to 600<br>HK 0.5 600 to 700<br>HK 0.5 700 to 800<br>HK 0.5 800 to 900<br>HK 0.5 900 to 1000 |  |         |                  |
|   | HK 1 35 to 100   | 2.4 HK<br>5.2 HK<br>8.3 HK<br>11.6 HK<br>15.2 HK<br>18.9 HK<br>22.7 HK<br>26.6 HK<br>30.9 HK<br>35.6 HK                |         |                  |
|   | HK 1 100 to 200<br>HK 1 200 to 300<br>HK 1 300 to 400<br>HK 1 400 to 500<br>HK 1 500 to 600<br>HK 1 600 to 700<br>HK 1 700 to 800<br>HK 1 800 to 900<br>HK 1 900 to 1000                   |  |         |                  |
| Certification of reference Knoop<br>hardness measurements & Knoop<br>Reading Blocks | All ranges See note 8  | 1.0 $\mu\text{m}$  |         | P                |





0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge            | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ )  | Remarks   | Location<br>Code |
|---|---|--|---|------------------|
| Calibration of Brinell Reference<br>Hardness Blocks | Brinell Scales:<br>Force diameter index<br>$(F/D^2) = 30$<br><br>10/3000 600HBW to 650<br>HBW<br>10/3000 500HBW to 600<br>HBW<br>10/3000 400HBW to 500<br>HBW<br>10/3000 300HBW to 400<br>HBW<br>10/3000 200HBW to 300<br>HBW<br>10/3000 95HBW to 200<br>HBW<br><br>5/750 600HBW to 650<br>HBW<br>5/750 500HBW to 600<br>HBW<br>5/750 400HBW to 500<br>HBW<br>5/750 300HBW to 400<br>HBW<br>5/750 200HBW to 300<br>HBW<br>5/750 95HBW to 200<br>HBW<br><br>2.5/187.5 600HBW to 650<br>HBW<br>2.5/187.5 500HBW to 600<br>HBW<br>2.5/187.5 400HBW to 500<br>HBW<br>2.5/187.5 300HBW to 400<br>HBW<br>2.5/187.5 200HBW to 300<br>HBW<br>2.5/187.5 95HBW to 200<br>HBW<br><br>1/30 600HBW to 650 HBW<br>1/30 500HBW to 600 HBW<br>1/30 400HBW to 500 HBW<br>1/30 300HBW to 400 HBW<br>1/30 200HBW to 300 HBW<br>1/30 95HBW to 200 HBW | See Note 9<br><br>4.3 HBW<br>4.0 HBW<br>3.3 HBW<br>2.7 HBW<br>2.0 HBW<br>1.4 HBW<br><br>5.3 HBW<br>4.9 HBW<br>4.1 HBW<br>3.3 HBW<br>2.5 HBW<br>1.7 HBW<br><br>5.3 HBW<br>4.9 HBW<br>4.1 HBW<br>3.3 HBW<br>2.1 HBW<br>1.7 HBW<br><br>6.0 HBW<br>5.5 HBW<br>4.4 HBW<br>3.4 HBW<br>2.5 HBW<br>1.7 HBW | 9 The calibration /<br>verification shall be in<br>accordance with the<br>requirements of<br>BS EN ISO 6506-<br>3:2014 and /or ASTM<br>E10-18 | P                |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ )                                      | Remarks | Location<br>Code |
|--|---|--|---------|------------------|
|  | Force diameter index<br>( $F/D^2$ ) = 15<br>10/1500 270HBW to<br>300HBW<br>10/1500 230HBW to<br>270HBW<br>10/1500 200HBW to<br>230HBW<br>10/1500 170HBW to<br>200HBW<br>10/1500 140HBW to<br>170HBW<br>10/1500 110HBW to<br>140HBW<br>10/1500 70HBW to<br>110HBW<br>10/1500 55HBW to<br>70HBW | 2.0 HBW<br>1.8 HBW<br>1.5 HBW<br>1.3 HBW<br>1.1 HBW<br>1.1 HBW<br>1.1 HBW<br>1.1 HBW |         |                  |
|  | Force diameter index<br>( $F/D^2$ ) = 10<br>10/1000 200HBW to 218<br>HBW<br>10/1000 170HBW to 200<br>HBW<br>10/1000 140 HBW to 170<br>HBW<br>10/1000 110HBW to<br>140HBW<br>10/1000 90HBW to<br>110HBW<br>10/1000 55HBW to<br>90HBW   | 1.4 HBW<br>1.3 HBW<br>1.1 HBW<br>1.1 HBW<br>1.1 HBW<br>1.1 HBW                       |         |                  |
|  | 5/250 200HBW to<br>218HBW<br>5/250 170HBW to<br>200HBW<br>5/250 140 HBW to<br>170HBW<br>5/250 110HBW to<br>140HBW<br>5/250 90HBW to<br>110HBW<br>5/250 55HBW to<br>90HBW  | 1.9 HBW<br>1.7 HBW<br>1.5 HBW<br>1.2 HBW<br>1.1 HBW<br>1.1 HBW                       |         |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge                     | Range  | Expanded Measurement<br>Uncertainty ( $k = 2$ )   | Remarks | Location<br>Code |
|--|--|---|---------|------------------|
| Calibration of Brinell Reference<br>Hardness Blocks (cont'd) | Brinell Scales:<br>Force diameter index<br>$(F/D^2) = 30$<br>2.5/62.5 200 HBW to 218<br>HBW<br>2.5/62.5 100 HBW to 200<br>HBW<br>2.5/62.5 47 HBW to 100<br>HBW<br><br>1/10 200 HBW to 218<br>HBW<br>1/10 100 HBW to 200<br>HBW<br>1/10 47 HBW to 100<br>HBW<br><br>Force diameter index $(F/D^2)$<br>$= 5$<br>10/500 100 HBW to 109<br>HBW<br>10/500 70 HBW to 100<br>HBW<br>10/500 47 HBW to 70<br>HBW<br><br>5/125 100 HBW to 109<br>HBW<br>5/125 70 HBW to 100<br>HBW<br>5/125 47 HBW to 70<br>HBW<br><br>2.5/31.25 100 HBW to 109<br>HBW<br>2.5/31.25 47 HBW to 100<br>HBW<br><br>1/5 100 HBW to 109 HBW<br>1/5 47 HBW to 100 HBW<br><br>Force diameter index<br>$(F/D^2) = 2.5$<br>10/250 20HBW to 55HBW<br><br>5/62.5 20.0HBW to<br>55.0HBW<br><br>2.5/15.625 20.0HBW to<br>55.0HBW<br><br>1/2.5 20.0HBW to<br>55.0HBW<br><br>Force diameter index<br>$(F/D^2) = 1.25$ 10/125<br>20HBW | See Note 9<br><br>1.8 HBW<br>1.7 HBW<br>1.1 HBW<br><br>2.0 HBW<br>1.8 HBW<br>1.1 HBW<br><br>1.1 HBW<br>1.1 HBW<br>1.1 HBW<br><br>1.1 HBW<br>1.1 HBW<br>1.1 HBW<br><br>1.1 HBW<br>1.1 HBW<br><br>1.1 HBW<br>1.1 HBW<br><br>1.1 HBW<br><br>1.1 HBW<br><br>1.1 HBW |         | P                |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge  | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ )                | Remarks | Location<br>Code |
|---|---|--|---------|------------------|
|   | 5/31.25 20HBW<br>2.5/7.8125 20HBW<br>Force diameter index<br>( $F/D^2$ ) = 1<br>10/100 20HBW<br>5/25 20HBW<br>2.5/6.25 20HBW<br>1/1 20HBW | 1.1 HBW<br>1.1 HBW<br>1.1 HBW<br>1.1 HBW<br>1.1 HBW<br>1.1 HBW |         |                  |
| Certification of reference Brinell<br>hardness measurements & Brinell<br>Reading Blocks | All ranges See note 3   | 1.0 $\mu\text{m}$  |         | P                |
| END   |   |  |         |                  |



0232  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Indentec Hardness Testing Machines Ltd**  
**Issue No: 039 Issue date: 27 August 2021**

Calibration performed by the Organisation at the locations specified

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$