

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0246

Accredited to  
ISO/IEC 17025:2017

### Defence Electronics & Components Agency

Issue No: 044 Issue date: 19 March 2021

DECA Sealand  
Welsh Road  
Deeside  
Flintshire  
CH5 2LS

Contact: Mr S Knight  
Tel: +44 (0)1244 847091  
Fax: +44 (0)1244 847058  
E-Mail: [tepl.accreditation@deca.mod.uk](mailto:tepl.accreditation@deca.mod.uk)  
Website: [www.gov.uk/deca](http://www.gov.uk/deca)

Calibration performed at the above address only

#### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<b>ELECTRICAL</b>			All electrical calibrations are performed as a comparison against a reference standard unless otherwise stated
FREQUENCY			
Specific Values	100 kHz 1 MHz 5 MHz 10 MHz	2.3 in $10^{12}$ 2.3 in $10^{12}$ 2.3 in $10^{12}$ 2.3 in $10^{12}$	For the calibration of Frequency equipment, timers / counters
RESISTANCE	0 $\Omega$ to 2 $\Omega$ 2 $\Omega$ to 20 $\Omega$ 20 $\Omega$ to 200 $\Omega$ 200 $\Omega$ to 2 k $\Omega$ 2 k $\Omega$ to 20 k $\Omega$ 20 k $\Omega$ to 200 k $\Omega$ 200 k $\Omega$ to 2 M $\Omega$ 2 M $\Omega$ to 20 M $\Omega$ 20 M $\Omega$ to 200 M $\Omega$	18 ppm + 1.0 $\mu\Omega$ 11 ppm + 7.0 $\mu\Omega$ 10 ppm + 43 $\mu\Omega$ 10 ppm + 200 $\mu\Omega$ 10 ppm + 2.0 m $\Omega$ 10 ppm + 20 m $\Omega$ 12 ppm + 200 m $\Omega$ 27 ppm + 4.0 $\Omega$ 210 ppm + 200 $\Omega$	For measurement of instrument outputs
<b>DC/LF MULTIFUNCTION TRANSFER STANDARD SYSTEM</b>			
DC RESISTANCE	10 $\Omega$ 100 $\Omega$ 1 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$ 1 M $\Omega$ 10 M $\Omega$ 100 M $\Omega$	12 ppm 10 ppm 7.5 ppm 5.8 ppm 9.9 ppm 17 ppm 67 ppm 250 ppm	Sourcing resistance for measuring instruments
DC VOLTAGE	100 mV 1 V 10 V 19 V 100 V 1000 V	11 ppm 4.5 ppm 4.3 ppm 4.3 ppm 7.0 ppm 6.1 ppm	Sourcing and measurement capability for the calibration of voltage instruments



**0246**  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Defence Electronics & Components Agency**  
Issue No: 044 Issue date: 19 March 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<b>DC/LF MULTIFUNCTION TRANSFER STANDARD SYSTEM (cont'd)</b>			
DC CURRENT	100 $\mu$ A	39 ppm	Sourcing and measurement capability for the calibration of current instruments
	1 mA	35 ppm	
	10 mA	30 ppm	
	100 mA	36 ppm	
	1 A	51 ppm	
	10 A	73 ppm	
AC VOLTAGE	1 mV		Sourcing and measurement capability for the calibration of voltage instruments
	20 Hz	0.21 %	
	30 Hz	0.21 %	
	40 Hz	0.21 %	
	55 Hz	0.22 %	
	300 Hz	0.21 %	
	1 kHz	0.21 %	
	10 kHz	0.21 %	
	20 kHz	0.21 %	
	30 kHz	0.21 %	
	50 kHz	0.21 %	
	100 kHz	0.32 %	
	10 mV		
	20 Hz	320 ppm	
	30 Hz	310 ppm	
	40 Hz	310 ppm	
	55 Hz	310 ppm	
	300 Hz	300 ppm	
	1 kHz	310 ppm	
	10 kHz	300 ppm	
	20 kHz	310 ppm	
	30 kHz	360 ppm	
	50 kHz	340 ppm	
	100 kHz	530 ppm	
	100 mV		
	20 Hz	130 ppm	
	30 Hz	120 ppm	
	40 Hz	130 ppm	
	55 Hz	160 ppm	
	300 Hz	120 ppm	
	1 kHz	110 ppm	
	10 kHz	110 ppm	
	20 kHz	130 ppm	
30 kHz	180 ppm		
50 kHz	190 ppm		
100 kHz	370 ppm		



0246  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Defence Electronics & Components Agency**  
Issue No: 044 Issue date: 19 March 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks	
AC VOLTAGE (cont'd)	1 V			
	10 Hz	49 ppm		
	20 Hz	51 ppm		
	30 Hz	46 ppm		
	40 Hz	41 ppm		
	55 Hz	41 ppm		
	300 Hz	39 ppm		
	1 kHz	45 ppm		
	10 kHz	41 ppm		
	20 kHz	40 ppm		
	30 kHz	40 ppm		
	50 kHz	47 ppm		
	100 kHz	78 ppm		
	300 kHz	180 ppm		
	500 kHz	470 ppm		
	1 MHz	0.16 %		
	10 V			
	10 Hz	55 ppm		
	20 Hz	48 ppm		
	30 Hz	53 ppm		
	40 Hz	41 ppm		
	55 Hz	40 ppm		
	300 Hz	40 ppm		
	1 kHz	45 ppm		
	19 V			
	1 kHz	39 ppm		
	10 kHz	36 ppm		
	20 kHz	38 ppm		
	30 kHz	46 ppm		
	50 kHz	57 ppm		
	100 kHz	91 ppm		
	300 kHz	160 ppm		
	500 kHz	710 ppm		
	1 MHz	0.14 %		
	100 V			
	10 Hz	55 ppm		
	20 Hz	51 ppm		
	30 Hz	50 ppm		
	40 Hz	52 ppm		
	55 Hz	48 ppm		
	300 Hz	44 ppm		
	1 kHz	42 ppm		
	10 kHz	43 ppm		
	20 kHz	44 ppm		
	30 kHz	45 ppm		
	50 kHz	51 ppm		
	100 kHz	100 ppm		
700 V				
50 kHz	190 ppm			
100 kHz	480 ppm			



0246  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Defence Electronics & Components Agency**  
Issue No: 044 Issue date: 19 March 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC VOLTAGE (cont'd)	1000 V	53 ppm	Sourcing and measurement capability for the calibration of current instruments
	40 Hz	51 ppm	
	55 Hz	52 ppm	
	300 Hz	49 ppm	
	1 kHz	60 ppm	
	10 kHz	63 ppm	
	20 kHz	89 ppm	
	30 kHz		
AC CURRENT	100 $\mu$ A	190 ppm	
	10 Hz	190 ppm	
	20 Hz	180 ppm	
	30 Hz	180 ppm	
	40 Hz	190 ppm	
	55 Hz	180 ppm	
	300 Hz	170 ppm	
	1 kHz	390 ppm	
	5 kHz		
	1 mA	140 ppm	
	10 Hz	150 ppm	
	20 Hz	130 ppm	
	30 Hz	130 ppm	
	40 Hz	110 ppm	
	55 Hz	110 ppm	
	300 Hz	110 ppm	
	1 kHz	110 ppm	
	5 kHz	170 ppm	
	10 mA	140 ppm	
	10 Hz	130 ppm	
	20 Hz	130 ppm	
	30 Hz	110 ppm	
	40 Hz	110 ppm	
	55 Hz	110 ppm	
	300 Hz	110 ppm	
	1 kHz	110 ppm	
	5 kHz	170 ppm	
	100 mA	140 ppm	
	10 Hz	130 ppm	
	20 Hz	130 ppm	
	30 Hz	110 ppm	
	40 Hz	110 ppm	
	55 Hz	110 ppm	
	300 Hz	110 ppm	
	1 kHz	110 ppm	
	5 kHz	170 ppm	
	1 A	180 ppm	
	10 Hz	180 ppm	
	20 Hz	170 ppm	
	30 Hz	140 ppm	
	40 Hz	140 ppm	
	55 Hz	130 ppm	
	300 Hz	180 ppm	
	1 kHz	300 ppm	
	5 kHz		



0246  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Defence Electronics & Components Agency**  
Issue No: 044 Issue date: 19 March 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC CURRENT (cont'd)	10 A 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz 10 kHz	370 ppm 380 ppm 380 ppm 360 ppm 360 ppm 370 ppm 390 ppm 640 ppm 0.12 %	
<b>DC/LF AUTOMATED CALIBRATION SYSTEM</b>			
DC VOLTAGE	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	10 ppm + 0.60 $\mu$ V 6.5 ppm + 1.7 $\mu$ V 4.5 ppm + 5.2 $\mu$ V 6.5 ppm + 130 $\mu$ V 9.0 ppm + 0.58 mV	Sourcing and measurement capability for the calibration of voltage instruments
DC CURRENT	0 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	120 ppm + 2.0 nA 50 ppm + 12 nA 50 ppm + 120 nA 50 ppm + 1.2 $\mu$ A 120 ppm + 24 $\mu$ A	Sourcing and measurement capability for the calibration of current instruments
DC RESISTANCE	0 $\Omega$ 10 $\Omega$ 100 $\Omega$ 1 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$ 1 M $\Omega$ 10 M $\Omega$ 100 M $\Omega$	10 m $\Omega$ (2-wire configuration) 35 ppm 15 ppm 15 ppm 15 ppm 16 ppm 35 ppm 65 ppm 250 ppm	The zero uncertainty for 4-wire configurations will be lower than that for 2-wire configurations and will largely depend on the resolution of the instrument being calibrated.
AC VOLTAGE	2 mV to 200 mV 20 Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 40 kHz 40 kHz to 100 kHz	190 ppm + 11 $\mu$ V 140 ppm + 11 $\mu$ V 150 ppm + 11 $\mu$ V 400 ppm + 11 $\mu$ V 530 ppm + 11 $\mu$ V	Sourcing and measurement capability for the calibration of voltage instruments
	200 mV to 2 V 20 Hz to 300 Hz 300 Hz to 500 Hz 500 Hz to 30 kHz 30 kHz to 40 kHz 40 kHz to 100 kHz	110 ppm + 47 $\mu$ V 65 ppm + 47 $\mu$ V 55 ppm + 47 $\mu$ V 100 ppm + 47 $\mu$ V 110 ppm + 47 $\mu$ V	
	2 V to 20 V 20 Hz to 300 Hz 300 Hz to 500 Hz 500 Hz to 30 kHz 30 kHz to 100 kHz	110 ppm + 0.47 mV 65 ppm + 0.47 mV 55 ppm + 0.47 mV 100 ppm + 0.47 mV	



0246  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Defence Electronics & Components Agency**  
Issue No: 044 Issue date: 19 March 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC VOLTAGE (cont'd)	20 V to 200 V 20 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 30 kHz 30 kHz to 40 kHz 40 kHz to 100 kHz	130 ppm + 4.7 mV 76 ppm + 4.7 mV 76 ppm + 4.7 mV 78 ppm + 4.7 mV 150 ppm + 4.7 mV 160 ppm + 4.7 mV	
AC CURRENT	200 V to 1000 V 500 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 30 kHz	130 ppm + 47 mV 170 ppm + 47 mV 180 ppm + 47 mV	
	2 $\mu$ A to 200 $\mu$ A 500 Hz	210 ppm + 12 nA	Sourcing and measurement capability for the calibration of voltage instruments
	200 $\mu$ A to 2 mA 500 Hz	160 ppm + 120 nA	
	2 mA to 20 mA 500 Hz	160 ppm + 1.2 $\mu$ A	
	20 mA to 200 mA 40 Hz to 500 Hz 500 Hz to 5 kHz	160 ppm + 12 $\mu$ A 280 ppm + 12 $\mu$ A	
	200 mA to 2 A 40 Hz to 500 Hz 500 Hz to 5 kHz	370 ppm + 120 $\mu$ A 570 ppm + 120 $\mu$ A	
<b>PRESSURE</b>			Methods consistent with EURAMET CG17.
Gas pressure (absolute)			
Calibration of pressure measuring instruments and gauges	3.5 kPa to 175 kPa 175 kPa to 700 kPa	50 ppm + 2 Pa 46 ppm + 2 Pa	Calibration against a deadweight tester standard
Gas pressure (gauge)			
Calibration of pressure measuring instruments and gauges	-95 kPa to 0 kPa 3.5 kPa to 175 kPa 175 kPa to 700 kPa 700 kPa to 3.5 MPa	70 Pa 43 ppm 43 ppm 57 ppm	Calibration against a deadweight tester standard
Hydraulic pressure (gauge)			
Calibration of pressure measuring instruments and gauges	0.35 MPa to 0.69 MPa 0.69 MPa to 6.9 MPa 6.9 MPa to 69 MPa	108 ppm 85 ppm 58 ppm	Calibration against a deadweight tester standard



0246  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Defence Electronics & Components Agency**  
Issue No: 044 Issue date: 19 March 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<b>FORCE</b> Calibration of push strength testers in compression	5 N to 500 N	0.33% of applied force	Calibrated using alignment guides
<b>TEMPERATURE</b> Temperature Indicators and/or recorders with temperature sensors	5 °C to 60 °C 60 °C to 120 °C	0.045 °C 0.064 °C	Calibration against a reference standard
<b>ROTATIONAL SPEED</b> Optical Tachometers	45 RPM to 1000 RPM 1000 RPM to 100,000 RPM	110 ppm + 0.050 RPM 110 ppm + 0.50 RPM	Calibration against a reference standard
END			



0246  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Defence Electronics & Components Agency**  
Issue No: 044 Issue date: 19 March 2021

Calibration performed at main address only

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

### Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %· $p$  + (0.12·10<sup>-6</sup>· $p$ ·10<sup>6</sup>) + 4.0 Pa, where  $p$  is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 ·  $i$ , where  $i$  is the instrument indication.