


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0247</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Poulten Selfe and Lee Ltd</p> <p>Issue No: 027 Issue date: 19 February 2020</p>	
	<p>PSL Calibration Laboratory Russell House Burnham Business Park Burnham-on-Crouch Essex CM0 8TE</p>	<p>Contact: Mr S J Gosling Tel: +44 (0)1621 787100 Fax: +44 (0)1621 787175 E-Mail: sales@psl-rheotek.com Website: www.psl-rheotek.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
TEMPERATURE			
Temperature indicators and recorders with sensors	-40 °C to 0 °C Ice Point (0 °C) 0 °C to 150 °C	0.010 °C 0.010 °C 0.010 °C	Digital thermometers for use according to ASTM D445, IP71 part1 and ISO3104 Different immersion characteristics may increase the uncertainty value
VISCOSITY			
Capillary viscometer calibration			
Viscosity constant 'C'	'C' (mm ² s ⁻¹)s ⁻¹		
Master viscometers - Ubbelohde	0.001 0.003 0.01 0.03 0.05 0.1 0.3 0.5 1 3 5 15 30 100	0.11 % 0.11 % 0.11 % 0.12 % 0.13 % 0.14 % 0.15 % 0.17 % 0.19 % 0.20 % 0.22 % 0.24 % 0.26 % 0.28 %	Calibrated to ASTM D2162-17
Reference viscometers	'C' (mm ² s ⁻¹)s ⁻¹		



0247
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Poulten Selfe and Lee Ltd
Issue No: 027 Issue date: 19 February 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks
Direct flow - PSL suspended level type	0.001 0.003 0.01 0.03 0.1 0.3 1 3 10 30 100	0.12 % 0.12 % 0.13 % 0.15 % 0.17 % 0.19 % 0.22 % 0.23 % 0.27 % 0.45 % 0.45 %	Calibrated to ASTM D446-12(2017) and ISO 3105:1994 in direct comparison with master viscometers calibrated to ASTM D2162-17
VISCOSITY (cont'd)			
Routine viscometers	'C' (mm ² s ⁻¹)s ⁻¹	Direct flow - Ubbelohde Suspended Level	Reverse flow
	0.001 0.003 0.01 0.03 0.1 0.3 1 3 10 30 100	0.13 % 0.13 % 0.14 % 0.17 % 0.20 % 0.22 % 0.24 % 0.29 % 0.29 % 0.46 % 0.46 %	0.19 % 0.20 % 0.20 % 0.23 % 0.25 % 0.27 % 0.31 % 0.35 % 0.35 % 0.50 % 0.50 %
Viscosity Reference Standards and Viscosity Measurement	mm ² s ⁻¹ 0.4 to 6 6 to 19 19 to 100 100 to 240 240 to 465 465 to 1130 1130 to 1740 1740 to 3980 3980 to 12500 12500 to 34000 34000 to 150000	Kinematic Dynamic 0.11 % 0.11 % 0.12 % 0.12 % 0.14 % 0.14 % 0.15 % 0.15 % 0.17 % 0.17 % 0.19 % 0.19 % 0.20 % 0.20 % 0.22 % 0.22 % 0.24 % 0.24 % 0.26 % 0.26 % 0.37 % 0.37 %	Calibrated to ASTM D446-12 (2017) and ISO 3105:1994 in direct comparison with reference viscometers or using viscosity standards calibrated to ASTM D2162-17. Temperature range 20 °C to 150 °C.
Viscosity Reference Standards and Viscosity Measurement	mm ² s ⁻¹ 0.4 to 6 6 to 19 19 to 100 100 to 240 240 to 465 465 to 1130 1130 to 1740 1740 to 150000	Kinematic 0.18 % 0.21 % 0.28 % 0.33 % 0.38 % 0.50 % 0.50 % 0.73 %	Viscosity reference standards calibrated to ASTM D2162-17 using reference viscometers. Temperature range -40 °C to 0 °C.



0247
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Poulten Selfe and Lee Ltd
Issue No: 027 Issue date: 19 February 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
Viscosity Reference Standards and Viscosity Measurement	mm ² s ⁻¹	Kinematic Dynamic	Viscosity reference standards calibrated to ASTM D2162-17 using reference viscometers. Temperature range 20 °C to 150 °C.
	0.4 to 6	0.18 % 0.18 %	
	6 to 19	0.18 % 0.18 %	
	19 to 100	0.24 % 0.24 %	
	100 to 240	0.27 % 0.27 %	
	240 to 465	0.38 % 0.38 %	
	465 to 1130	0.45 % 0.45 %	
	1130 to 1740	0.45 % 0.45 %	
	1740 to 3980	0.47 % 0.47 %	
	3980 to 12500	0.56 % 0.56 %	
	12500 to 34000	0.69 % 0.69 %	
34000 to 150000	0.69 % 0.69 %		
DENSITY			
Liquids	0.65 g/ml to 0.93 g/ml	0.040 %	Calibrated to IP 189/190-2005 Temperature range 20 °C to 150 °C.
END			



0247

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Poulten Selfe and Lee Ltd
Issue No: 027 Issue date: 19 February 2020

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
- As an explicit function of the measurand or of a parameter (see below).
- As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
- As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %· p + (0.12·10⁻⁶· p ·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i , where i is the instrument indication.