


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0251</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>Lloyd Instruments</b> <b>a trading division of Ametek (GB) Limited</b>  Issue No: 021    Issue date: 01 September 2021	
	Steyning Way Bognor Regis West Sussex PO22 9ST	Contact: Mr Wayne Shaw Tel: +44 (0) 1243 833389 Fax: +44 (0)1243 833401 E-Mail: wayne.shaw@ametek.co.uk Website: www.ametektest.com

**Calibration performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customers' sites or premises  The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer	Force	S



0251  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Lloyd Instruments a trading division of Ametek (GB) Limited**  
**Issue No: 021 Issue date: 01 Spetember 2021**

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
FORCE  UNIVERSAL MATERIALS TESTING MACHINES  Verification and calibration of the force measuring system by force proving instruments in tension  Verification and calibration of the force measuring system by force proving instruments in compression  Verification and calibration of the force measuring system by calibrated masses in tension  Verification and calibration of the force measuring system by calibrated masses in compression	0.1 kN to 300 kN for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018  0.1 kN to 300 kN for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018  0.1 N to 100 N for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018  0.1 N to 100 N for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018	0.22 %  0.22 %  0.10 %  0.10 %		S
END				



0251  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Lloyd Instruments a trading division of Ametek (GB) Limited**  
**Issue No: 021 Issue date: 01 Spetember 2021**

Calibration performed by the Organisation at the locations specified

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$