


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0254</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Devonport Royal Dockyard Limited trading as Devonport Electrical and Nucleonic Calibration Facility</p> <p>Issue No: 038 Issue date: 29 June 2020</p>	
	<p>Babcock International Group PC1409, Devonport Royal Dockyard Plymouth Devon PL1 4SG</p>	<p>Contact: Mr C Burrow Tel: +44 (0)1752 324739 Fax: +44 (0)1752 324737 E-Mail: Chris.Burrow@babcockinternational.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL CALIBRATION			Electrical calibrations are as a comparison against a reference standard
DC Resistance Generation Specific Values	1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	61 ppm 58 ppm 32 ppm 22 ppm 21 ppm 22 ppm 24 ppm 29 ppm 110 ppm	
Measurement	0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 2 G Ω 2 G Ω to 100 G Ω	20 ppm + 0.20 m Ω 20 ppm + 0.60 m Ω 20 ppm + 0.60 m Ω 15 ppm + 7.0 m Ω 15 ppm + 65 m Ω 25 ppm + 1.5 Ω 150 ppm + 80 Ω 200 ppm + 9.0 k Ω 500 ppm + 1.0 M Ω 2.0 %	
DC Voltage Generation	0 mV to 220 mV 220 mV to 2.2 V 2.2 V to 11 V 11 V to 22 V 22 V to 220 V 220 V to 1100 V	13 ppm + 2.0 μ V 5.0 ppm + 2.0 μ V 6.0 ppm + 4.0 μ V 6.0 ppm + 8.0 μ V 6.0 ppm + 0.10 mV 6.0 ppm + 0.60 mV	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 038 Issue date: 29 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
DC Voltage (continued)			
Measurement	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	10 ppm + 2.0 μ V 10 ppm + 2.0 μ V 10 ppm + 3.0 μ V 10 ppm + 40 μ V 10 ppm + 100 μ V	
DC Current			
Generation	0 μ A to 220 μ A 220 μ A to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A	52 ppm + 10 nA 15 ppm + 10 nA 30 ppm + 100 nA 37 ppm + 1.0 μ A 100 ppm + 30 μ A 190 ppm + 480 μ A	
Measurement	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A 1 A to 10 A	50 ppm + 1.0 nA 30 ppm + 5.0 nA 40 ppm + 40 nA 40 ppm + 1.0 μ A 52 ppm + 10 μ A 50 ppm	
AC Voltage			
Generation	40 Hz to 20 kHz 10 mV to 220 mV 220 mV to 2.2 V 2.2 V to 22 V 22 V to 220 V 220 V to 750 V 750 V to 1100 V	56 ppm + 10 μ V 24 ppm + 20 μ V 22 ppm + 70 μ V 63 ppm + 2.0 mV 110 ppm + 6.0 mV 270 ppm + 11 mV	
	20 kHz to 50 kHz 10 mV to 220 mV 220 mV to 2.2 V 2.2 V to 22 V 22 V to 220 V 220 V to 750 V	82 ppm + 10 μ V 27 ppm + 80 μ V 22 ppm + 200 μ V 91 ppm + 4.0 mV 270 ppm + 11 mV	
	50 kHz to 100 kHz 10 mV to 220 mV 220 mV to 2.2 V 2.2 V to 22 V 22 V to 220 V	160 ppm + 30 μ V 56 ppm + 80 μ V 52 ppm + 400 μ V 200 ppm + 10 mV	
	100 kHz to 1 MHz 220 mV to 2.2 V 2.2 V to 22 V	0.17 % + 1.0 mV 0.37 % + 9.0 mV	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility
Issue No: 038 Issue date: 29 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
AC Voltage (continued) Measurement	<p><i>20 Hz to 10 kHz</i> 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V</p> <p><i>10 kHz to 100 kHz</i> 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V</p> <p><i>55 Hz to 3 kHz</i> 200 V to 1000 V</p> <p><i>3 kHz to 30 kHz</i> 200 V to 1000 V</p>	<p>270 ppm + 4.0 μV 680 ppm + 20 μV 260 ppm + 200 μV 230 ppm + 2.0 mV</p> <p>690 ppm + 20 μV 600 ppm + 200 μV 550 ppm + 2.0 mV 700 ppm + 20 mV</p> <p>580 ppm + 10 mV</p> <p>940 ppm + 20 mV</p>	
AC Current Generation	<p><i>50 Hz to 400 Hz</i> 10 μA to 200 μA 0.2 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A</p> <p><i>400 Hz to 1 kHz</i> 10 μA to 200 μA 0.2 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A</p> <p><i>1 kHz to 5 kHz</i> 10 μA to 200 μA 0.2 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A</p> <p><i>5 kHz to 10 kHz</i> 10 μA to 200 μA 0.2 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A</p>	<p>190 ppm + 20 nA 150 ppm + 40 nA 58 ppm + 0.40 μA 48 ppm + 4.0 μA 110 ppm + 40 μA 190 ppm + 170 μA</p> <p>190 ppm + 50 nA 150 ppm + 0.50 μA 58 ppm + 5.0 μA 54 ppm + 50 μA 110 ppm + 100 μA 250 ppm + 380 μA</p> <p>700 ppm + 0.10 μA 350 ppm + 1.0 μA 530 ppm + 10 μA 640 ppm + 100 μA 410 ppm + 200 μA 390 ppm + 750 μA</p> <p>0.18 % + 0.10 μA 0.12 % + 1.0 μA 0.18 % + 10 μA 0.22 % + 100 μA 0.090 % + 200 μA 0.060 % + 750 μA</p>	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**
Issue No: 038 Issue date: 29 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
AC Current (continued) Measurement	<i>55 Hz to 1 kHz</i> 20 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A <i>1 kHz to 5 kHz</i> 20 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A	850 ppm + 20 nA 850 ppm + 200 nA 850 ppm + 2.0 μ A 800 ppm + 20 μ A 800 ppm + 200 μ A 0.11 % + 20 nA 0.11 % + 200 nA 0.11 % + 2.0 μ A 0.10 % + 20 μ A 0.18 % + 400 μ A	
Capacitance	<i>1 kHz</i> 100 pF to 1 μ F	0.070 %	2 terminal measurements
Frequency			Averaged over 100 s
Generation			
Specific Values	100 kHz, 1 MHz, 5 MHz & 10 MHz	4.0 in 10^{12}	
	1 PPS	4.0 in 10^{11}	
Range values	1 Hz to 20 GHz	1.0 in 10^{11}	
Measurement	1 Hz to 20 GHz	1.0 in 10^{11} + 10 mHz	
Voltage Reflection Coefficient	<i>10 MHz to 18 GHz</i> 0 to 0.33	0.040	The CMC applies to 7 mm 50 Ω coaxial line fitted with Type N connectors; for other connector types the quoted uncertainties may be increased. Results may also be quoted in terms of VSWR or Return Loss with the uncertainty being converted to the corresponding units.
RF Attenuation	<i>10 MHz to 1 GHz</i> 0 dB to 10dB 10 dB to 20 dB 20 dB to 30 dB 30 dB to 40 dB 40 dB to 50 dB 50 dB to 60 dB 60 dB to 70 dB 70 dB to 80 dB	0.080 dB 0.10 dB 0.15 dB 0.18 dB 0.22 dB 0.25 dB 0.30 dB 0.35 dB	The CMCs apply to devices with input and output VSWR not exceeding 1.1: 1. For devices that exhibit higher values of VSWR the quoted uncertainties may be increased.



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**
Issue No: 038 Issue date: 29 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks
RF Power			
Specific value	50 MHz 1 mW	0.70 %	The CMC for measurement of 50MHz 1mW 50 Ω Reference sources with an output VSWR not exceeding 1.1 : 1, For devices that exhibit higher values of VSWR the quoted uncertainties may be increased.
Other values	-20 dBm to +20 dBm		
	10 MHz	2.5%	The CMCs for RF Power (Diode Sensors) apply to the measurement of the output power of a 50 Ω source with an output VSWR not exceeding 1.5:1. For devices that exhibit higher values of VSWR the quoted uncertainties may be increased.
	30 MHz	2.5%	
	50 MHz	2.3%	
	100 MHz	2.2%	
	200 MHz	2.2%	
	300 MHz	2.2%	
	400 MHz	2.2%	
	800 MHz	2.2%	
	1 GHz	2.2%	
	2 GHz	2.2%	
	4 GHz	2.2%	
	5 GHz	2.2%	
	6 GHz	2.2%	
	7 GHz	2.2%	
	8 GHz	2.5%	
	9 GHz	2.7%	
	10 GHz	2.5%	
	11 GHz	2.5%	
	12 GHz	2.5%	
	13 GHz	2.7%	
	14 GHz	3.0%	
	15 GHz	2.7%	
	16 GHz	2.7%	
	17 GHz	3.5%	
	18 GHz	4.5%	
	50 dBm to -20 dBm		
	10 MHz to 30 MHz	8.0 %	
	30 MHz to 8 GHz	3.5%	
	8 GHz to 18 GHz	4.5 %	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 038 Issue date: 29 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
RF Calibration Factor			
Nominal 1 mW	10 MHz and 30 MHz	1.6 %	The stated CMCs are for the measurement of RF Calibration Factor of power sensors in a 50 Ω coaxial system at a nominal level of 1 mW, with respect to a reference frequency of 50 MHz.
	100 MHz, 200 MHz, 300 MHz, 400 MHz, 800 MHz and 1000 MHz	1.4 %	
	2 GHz, 3 GHz and 4 GHz	1.7 %	
	5 GHz, 6 GHz, 7 GHz and 8 GHz	3.8 %	
	9 GHz, 10 GHz and 11 GHz	6.0 %	
	12 GHz, 13 GHz and 14 GHz	5.5 %	
	15 GHz, 16 GHz, 17 GHz and 18 GHz	5.3 %	
Nominal 1 μ W	10 MHz	1.7 %	The stated CMCs are for the measurement of RF Calibration Factor of power sensors in a 50 Ω coaxial system at a nominal level of 1 μ W, with respect to a reference frequency of 50 MHz.
	30 MHz, 100 MHz, 200 MHz, 300 MHz, 400 MHz, 800 MHz, 1 GHz, 2 GHz, 3 GHz and 4 GHz	1.5 %	
	5 GHz	1.6 %	
	6 GHz, 7 GHz and 8 GHz	3.0 %	
	9 GHz to 18 GHz	3.4 %	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**
Issue No: 038 Issue date: 29 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks
RADIOLOGICAL CALIBRATION			
Air Kerma Rate	¹³⁷ Cs: 2.0 μGyh ⁻¹ to 1.4 Gyh ⁻¹	3.5 %	Calibration of air kerma/air kerma rate monitors using air kerma rates traceable to national standards through a secondary standard dosimeter
	⁶⁰ Co: 270.4 μGyh ⁻¹ to 15.0 mGyh ⁻¹	3.0 %	
	²⁴¹ Am: 4.2 μGyh ⁻¹ to 4.6 mGyh ⁻¹	5.0 %	
Ambient Dose Equivalent Rate	¹³⁷ Cs: 2 μSvh ⁻¹ to 1.713 Sv h ⁻¹	3.5 %	Calibration of ambient dose equivalent/dose rate monitors using air kerma rates traceable to national standards through a secondary standard dosimeter and using appropriate coefficients given in ISO Standards for H*(10).
	⁶⁰ Co: 311 μSvh ⁻¹ to 18.0mSvh ⁻¹	3.0 %	
	²⁴¹ Am: 7.2 μSvh ⁻¹ to 8 mSvh ⁻¹	5.0 %	
Personal Dose Equivalent Rate	¹³⁷ Cs: 2 μSvh ⁻¹ to 1.713 Sv h ⁻¹	3.5 %	Calibration of electronic personal dosimeters using air kerma rates traceable to national standards through a secondary standard dosimeter, and using appropriate coefficients given in ISO Standards for Hp(10)
	⁶⁰ Co: 292 μSvh ⁻¹ to 18.0 mSvh ⁻¹	3.0 %	
	²⁴¹ Am: 7.8 μSvh ⁻¹ to 8.7 mSvh ⁻¹	5.0 %	
Surface Contamination			
Monitor Response:			
Statutory tests, including: Test before first use; Periodic tests			
Alpha (α) Contamination	Alpha-emitting nuclides: ²⁴¹ Am, ²³⁸ Pu, ²³⁸ U	5.0 % to 20 % depending upon monitor type	Calibration against large area radioactive sources In with GPG 14
Beta (β) Contamination	Beta-emitting nuclides: ⁹⁰ Sr/ ⁹⁰ Y, ¹⁴⁷ Pm, ³⁶ Cl, ¹⁴ C, ⁶⁰ Co, ¹³⁷ Cs	5.0 % to 20 % depending upon monitor type	Calibration against large area radioactive sources In with GPG 14
END			



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 038 Issue date: 29 June 2020

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %· p + (0.12·10⁻⁶· p ·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i , where i is the instrument indication.