


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0280</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Central Calibration Services Limited</p> <p>Issue No: 035 Issue date: 01 September 2021</p>	
	<p>7 Everdon Park Heartlands Business Park Daventry Northamptonshire NN11 8YJ</p>	<p>Contact: Mr J Tyrrell Tel: +44 (0) 1327 300648 E-Mail: john@ccs0280.com</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address 7 Everdon Park Heartlands Business Park Daventry Northamptonshire NN11 8YJ</p> <p>Local contact Mr J Tyrrell</p>	Force	P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>Customers' sites or premises</p> <p>The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer</p>	Force	S



0280
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Central Calibration Services Limited
Issue No: 035 Issue date: 01 September 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
FORCE FORCE MEASURING DEVICES				
Calibration of force measuring devices (e.g. strain gauged load cells and load measuring rings) but excluding proving devices in tension by the use of calibrated masses or elastic reference devices	0.2 N to 260 N 0.2 kN to 500 kN	0.10 % 0.36 %		P
Calibration of force measuring devices (e.g. strain gauged load cells and load measuring rings) but excluding proving devices in compression by the use of calibrated masses or elastic reference devices	0.2 N to 260 N 0.2 kN to 500 kN 0.2 kN to 3 MN	0.10 % 0.36 % 0.48 %		
Calibration of force measuring devices used in soil testing machines in compression and tension by the use of elastic reference devices	0.2 kN to 500 kN to BS 1377:Part 1:2016	0.36 %		S
END				



0280
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Central Calibration Services Limited
Issue No: 035 **Issue date:** 01 September 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$