

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0292</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Tyne and Wear Trading Standards</h3> <p>Issue No: 018 Issue date: 27 October 2016</p>	
	<p>Joint Committee Metrology Laboratory Saltmeadows Road Gateshead Tyne and Wear NE8 3AH</p>	<p>Contact: Mr A Hayward Tel: +44 (0)191 478 4550 Fax: +44 (0)191 478 4550 E-Mail: metrologylaboratory@gateshead.gov.uk Website: www.legalmetrology.info</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
MASS	(g)	(mg)	1. Intermediate values can be calibrated to an uncertainty interpolated from the next higher and lower values in the table.
	25 000	12.5	
	20 000	10	
	10 000	5.0	
	5 000	2.5	
	2 000	1.0	
	1 000	0.50	
	500	0.25	
	200	0.10	
	100	0.050	
	50	0.033	
	20	0.026	
	10	0.020	
	5	0.016	
	2	0.013	
	1	0.010	
	0.5	0.0080	
	0.2	0.0066	
	0.1	0.0050	
	0.05	0.0040	
	0.02	0.0030	
	0.01	0.0026	
	0.005	0.0020	
	0.002	0.0020	
	0.001	0.0020	
	0.000 5	0.0020	
END			