

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0292 Accredited to ISO/IEC 17025:2017	Tyne and Wear Trading Standards Issue No: 022 Issue date: 13 August 2021	
	Joint Committee Metrology Laboratory Saltmeadows Road Gateshead Tyne and Wear NE8 3AH	Contact: Mr A Hayward Tel: +44 (0)191 478 4550 Fax: +44 (0)191 478 4550 E-Mail: metrologylaboratory@gateshead.gov.uk Website: www.legalmetrology.info

Calibration performed by the Organisation at the locations specified

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Metrology Laboratory Saltmeadows Road Gateshead Tyne and Wear NE8 3AH	Local contact Mr A Hayward	Mass Volume P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer's sites or premises The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer	Contact Mr A Hayward	Mass (Weighing machines) S



0292

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tyne and Wear Trading Standards

Issue No: 022 **Issue date:** 13 August 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
MASS	(g) 50.000 25 000 20 000 10 000 5 000 2 000 1 000 500 200 100 50 20 10 5 2 1 0.5 0.2 0.1 0.05 0.02 0.01 0.005 0.002 0.001 0.000 5	(mg) 260 12.5 10 5.0 2.5 1.0 0.50 0.25 0.10 0.050 0.033 0.026 0.020 0.016 0.013 0.010 0.0080 0.0066 0.0050 0.0040 0.0030 0.0026 0.0020 0.0020 0.0020 0.0020	1. Intermediate values can be calibrated to an uncertainty interpolated from the next higher and lower values in the table. Performed by Borda's method of substitution	P
VOLUMETRIC GLASSWARE Pipettes	(ml) 25 10 5 2 1	(ml) 0.028 0.016 0.010 0.005 2 0.004 0	Calibration for all volumetric glassware is performed gravimetrically and is based on ISO 4787	P



0292

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tyne and Wear Trading Standards

Issue No: 022 Issue date: 13 August 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Volumetric flasks	(ml)	(ml)		P
	5 000	2.1		
	2 000	0.85		
	1 000	0.43		
	500	0.22		
	250	0.11		
	200	0.093		
	100	0.050		
	50	0.029		
	25	0.017		
	20	0.015		
	10	0.008 9		
	5	0.006 6		
Measuring cylinders	(ml)	(ml)		P
	2 000	2.0		
	1 000	1.2		
	500	0.77		
	250	0.41		
	100	0.21		
	50	0.16		
	25	0.083		
	10	0.046		
	5	0.032		
	Non-automatic Weighing Instruments	(kg)		
150		380		
100		150		
50		76		
35		27		
20		15		
10		7.6		
5		3.8		
2		1.5		
1		0.76		
(g)		(mg)		
500		0.38		
200		0.15		
100		0.076		
50		0.050		
20		0.040		
10		0.031		
5		0.024		
2		0.020		
1		0.015		
0.5	0.012			
0.2	0.010			
END				



0292

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tyne and Wear Trading Standards

Issue No: 022 Issue date: 13 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$