


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|---|--|
|  <p>0310</p> <p>Accredited to ISO/IEC 17025:2017</p> | <p>RS Calibration a trading name of RS Components Ltd</p> <p>Issue No: 062 Issue date: 23 September 2024</p> | |
| | <p>DPN 175 Birchington Road Corby Northamptonshire NN17 5JF</p> | <p>Contact: Sean Smith Tel: +44 (0)1536 405545 E-Mail: calibration.uk@rs-components.com Website: https://uk.rs-online.com/web/</p> |
| Calibration performed at the above address only | | |

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---|---|---|
| ELECTRICAL MEASUREMENTS | | | |
| DC RESISTANCE | | | |
| Specific values Generation | 1 mΩ 10 mΩ 100 mΩ 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ 1 GΩ | 130 μΩ/Ω 52 μΩ/Ω 52 μΩ/Ω 6.0 μΩ/Ω 12 μΩ/Ω 5.0 μΩ/Ω 5.8 μΩ/Ω 11 μΩ/Ω 53 μΩ/Ω 53 μΩ/Ω 53 μΩ/Ω 54 μΩ/Ω 300 μΩ/Ω | Known values of resistance for application to resistance measuring instruments. Specific values are those which fall within $\pm 0.5\%$ of the stated values. |
| Measurement | 0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 kΩ 2 kΩ to 20 kΩ 20 kΩ to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 2 GΩ | 20 μΩ/Ω + 4.0 μΩ 11 μΩ/Ω + 14 μΩ 10 μΩ/Ω + 50 μΩ 10 μΩ/Ω + 0.50 mΩ 10 μΩ/Ω + 5.0 mΩ 10 μΩ/Ω + 50 mΩ 11 μΩ/Ω + 1.0 Ω 24 μΩ/Ω + 100 Ω 140 μΩ/Ω + 10 kΩ 0.18 % + 1.0 MΩ | Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased. |
| DC VOLTAGE | | | |
| Measurement | 0 V to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1 kV | 6.0 μV/V + 0.10 μV 5.0 μV/V + 0.40 μV 5.0 μV/V + 4.0 μV 7.0 μV/V + 40 μV 7.0 μV/V + 500 μV | Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased. |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---|---|---|
| DC CURRENT | | | |
| Generation | 20 A to 500 A 500 A to 2500 A | 590 μ A/A 490 μ A/A | Calibration of clamp-on ammeters and similar devices using multi-turn technique. |
| Measurement | 0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A 2 A to 10 A 10 A to 20 A 20 A to 100 A | 14 μ A/A + 0.40 nA 14 μ A/A + 4.0 nA 17 μ A/A + 40 nA 56 μ A/A + 800 nA 220 μ A/A + 16 μ A 470 μ A/A + 400 μ A 320 μ A/A 330 μ A/A 75 μ A/A | Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased. Using digital multimeter and shunt. Generation of these values may also be undertaken however the uncertainties may be increased. |
| AC VOLTAGE Specific Values | | | Using AC Measurement Standard. |

Expanded uncertainty ($k = 2$) expressed in terms of (μ V/V + μ V)

Generation of these values may also be undertaken, with increased uncertainties and with the following limitations:

600 μ V to 220 V, 10 Hz to 1 MHz: A maximum volt-Hertz product of 2.2×10^7 applies.

220 V to 750 V: The applied frequency will be in the range 40 Hz to 100 kHz.

750 V to 1000 V: The applied frequency will be in the range 40 Hz to 30 kHz.

| Voltage Range | 10 Hz to 20 Hz | 20 Hz to 40 Hz | 40 Hz to 20 kHz | 20 kHz to 50 kHz | 50 kHz to 100 kHz | 100 kHz to 300 kHz | 300 kHz to 500 kHz | 500 kHz to 1 MHz |
|-----------------------|-------------------|-------------------|--------------------|---------------------|----------------------|-----------------------|-----------------------|---------------------|
| 600 μ V to 2.2 mV | 1500 + 1.3 | 850 + 1.3 | 710 + 1.3 | 890 + 2.0 | 1200 + 2.5 | 1900 + 4.0 | 2100 + 8.0 | 3500 + 8.0 |
| 2.2 mV to 7 mV | 710 + 1.3 | 390 + 1.3 | 270 + 1.3 | 370 + 2.0 | 510 + 2.5 | 990 + 4.0 | 1200 + 8.0 | 2800 + 8.0 |
| 7 mV to 22 mV | 260 + 1.3 | 190 + 1.3 | 140 + 1.3 | 200 + 2.0 | 270 + 2.5 | 660 + 4.0 | 870 + 8.0 | 2500 + 8.0 |
| 22 mV to 70 mV | 210 + 1.5 | 130 + 1.5 | 100 + 1.5 | 130 + 2.0 | 220 + 2.5 | 440 + 4.0 | 740 + 8.0 | 2300 + 8.0 |
| 70 mV to 220 mV | 180 + 1.5 | 100 + 1.5 | 80 + 1.5 | 90 + 2.0 | 150 + 2.5 | 270 + 4.0 | 600 + 8.0 | 2300 + 8.0 |
| 220 mV to 700 mV | 180 + 1.5 | 100 + 1.5 | 80 + 1.5 | 90 + 2.0 | 100 + 2.5 | 230 + 4.0 | 570 + 8.0 | 750 + 8.0 |
| 700 mV to 2.2 V | 160 + 0 | 60 + 0 | 30 + 0 | 40 + 0 | 60 + 0 | 130 + 0 | 210 + 0 | 700 + 0 |
| 2.2 V to 7 V | 160 + 0 | 60 + 0 | 30 + 0 | 40 + 0 | 70 + 0 | 150 + 0 | 320 + 0 | 940 + 0 |
| 7 V to 22 V | 160 + 0 | 60 + 0 | 30 + 0 | 40 + 0 | 70 + 0 | 150 + 0 | 320 + 0 | 940 + 0 |
| 22 V to 70 V | 160 + 0 | 60 + 0 | 30 + 0 | 50 + 0 | 80 + 0 | | | |
| 70 V to 220 V | 160 + 0 | 60 + 0 | 30 + 0 | 60 + 0 | 90 + 0 | | | |
| 220 V to 700 V | 160 + 0 | 80 + 0 | 40 + 0 | 110 + 0 | 390 + 0 | | | |
| 700 V to 1000 V | 160 + 0 | 80 + 0 | 40 + 0 | 110 + 0 | 390 + 0 | | | |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|---|--|--|--|
| AC VOLTAGE (continued) Other values Measurement | <i>10 Hz to 40 Hz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V <i>40 Hz to 100 Hz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V <i>100 Hz to 2 kHz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V <i>2 kHz to 10 kHz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V <i>10 kHz to 30 kHz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V <i>30 kHz to 100 kHz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V <i>100 kHz to 300 kHz</i> 200 mV to 2 V 2 V to 20 V 20 V to 200 V <i>300 kHz to 1 MHz</i> 200 mV to 2 V 2 V to 20 V | 170 $\mu\text{V/V} + 4.0 \mu\text{V}$ 140 $\mu\text{V/V} + 20 \mu\text{V}$ 140 $\mu\text{V/V} + 200 \mu\text{V}$ 140 $\mu\text{V/V} + 2.0 \text{ mV}$ 140 $\mu\text{V/V} + 4.0 \mu\text{V}$ 110 $\mu\text{V/V} + 20 \mu\text{V}$ 110 $\mu\text{V/V} + 200 \mu\text{V}$ 110 $\mu\text{V/V} + 2.0 \text{ mV}$ 140 $\mu\text{V/V} + 20 \text{ mV}$ 130 $\mu\text{V/V} + 2.0 \mu\text{V}$ 90 $\mu\text{V/V} + 20 \mu\text{V}$ 90 $\mu\text{V/V} + 200 \mu\text{V}$ 87 $\mu\text{V/V} + 2.0 \text{ mV}$ 140 $\mu\text{V/V} + 20 \text{ mV}$ 160 $\mu\text{V/V} + 4.0 \mu\text{V}$ 130 $\mu\text{V/V} + 20 \mu\text{V}$ 130 $\mu\text{V/V} + 200 \mu\text{V}$ 130 $\mu\text{V/V} + 2.0 \text{ mV}$ 140 $\mu\text{V/V} + 20 \text{ mV}$ 400 $\mu\text{V/V} + 8.0 \mu\text{V}$ 260 $\mu\text{V/V} + 40 \mu\text{V}$ 260 $\mu\text{V/V} + 400 \mu\text{V}$ 260 $\mu\text{V/V} + 4.0 \text{ mV}$ 260 $\mu\text{V/V} + 40 \text{ mV}$ 900 $\mu\text{V/V} + 20 \mu\text{V}$ 660 $\mu\text{V/V} + 200 \mu\text{V}$ 660 $\mu\text{V/V} + 2.0 \text{ mV}$ 660 $\mu\text{V/V} + 20 \text{ mV}$ 670 $\mu\text{V/V} + 200 \text{ mV}$ 0.35 % + 2.0 mV 0.35 % + 20 mV 0.35 % + 200 mV 1.2 % + 20 mV 1.2 % + 200 mV | Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased. |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---|---|---|
| AC CURRENT Generation | 2 A to 250 A 50 Hz to 800 Hz | 0.090 % | Calibration of clamp-on ammeters and similar devices using multi-turn technique. |
| | 250 A to 1750 A 50 Hz to 100 Hz | 0.090 % | |
| | 1750 A to 2500 A 50 Hz to 60 Hz | 0.090 % | |
| Measurement | 10 Hz to 2 kHz 1 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A | 580 μ A/A + 20 nA 350 μ A/A + 200 nA 350 μ A/A + 2.0 μ A 340 μ A/A + 20 μ A 720 μ A/A + 200 μ A 950 μ A/A + 2.0 mA | Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased. |
| | 2 kHz to 10 kHz 1 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A | 580 μ A/A + 20 nA 350 μ A/A + 200 nA 350 μ A/A + 2.0 μ A 340 μ A/A + 20 μ A 850 μ A/A + 200 μ A | |
| | 50 Hz to 800 Hz 2 A to 20 A 10 A to 20 A 20 A to 50 A 50 A to 100 A | 590 μ A/A 580 μ A/A 0.11 % 0.11 % | Using digital multimeter and shunt. Generation of these values may also be undertaken however the uncertainties may be increased. |
| | 50 Hz to 60 Hz 50A to 100A | 0.11 % | |
| INDUCTANCE Generation | At 1 kHz: 1 mH 10 mH 100 mH 1 H | 490 μ H/H 500 μ H/H 470 μ H/H 470 μ H/H | Known inductance values for the calibration of inductance measuring instruments |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---|---|---|
| CAPACITANCE | | | |
| Generation | At 1 kHz: 1 nF 2 nF 3 nF 4 nF 5 nF 6 nF to 9 nF 10 nF 20 nF to 80 nF 90 nF 100 nF to 900 nF 1 μ F 2 μ F 3 μ F and 4 μ F 5 μ F and 9 μ F 10 μ F to 30 μ F 40 μ F and 50 μ F 60 μ F and 80 μ F 90 μ F and 100 μ F | 0.092 % 0.064 % 0.058 % 0.055 % 0.053 % 0.052 % 0.050 % 0.037 % 0.037 % 0.043 % 0.12 % 0.12 % 0.098 % 0.12 % 0.19 % 0.19 % 0.19 % 0.19 % | Known capacitance values for the calibration of capacitance measuring instruments |
| FREQUENCY | | | |
| Measurement | 10 Hz to 100 kHz 100 kHz to 1 MHz 1 MHz to 6 GHz | 3.0 in 10^9 2.0 in 10^{10} 5.0 in 10^{10} | Using counter timer and off-air standard. |
| Timer and stopwatch calibrations | 5 s to 99 999s | 0.10 s | Manual calibration. |
| OSCILLOSCOPE CALIBRATION | | | |
| Horizontal deflection coefficients | 500 ps to 10 ms | 0.29 μ s/s | Using time markers. The uncertainty quoted will be particularly dependent on the horizontal resolution of the oscilloscope being calibrated. |
| Vertical deflection coefficients | 6 mV to 60 mV 60 mV to 600 mV 600 mV to 60 V 60 V to 100 V 100 V to 120 V | 0.33 % 0.15 % 0.12 % 0.12 % 0.59 % | Using chopped waveforms of known peak to peak amplitude. The uncertainty quoted will be particularly dependent on the vertical resolution of the oscilloscope being calibrated. |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---|--|--|
| Bandwidth | <i>Input VSWR up to 1.2:1</i> 50 kHz to 550 MHz 550 MHz to 1 GHz <i>Input VSWR from 1.2:1 to 1.6:1</i> 50 kHz to 550 MHz 550 MHz to 1 GHz | 0.29 dB 0.39 dB 0.44 dB 0.54 dB | Relative to a low frequency reference. The uncertainty quoted will be dependent on the vertical resolution of the oscilloscope being calibrated. |
| Rise and fall times | Using 150 ps edge Using 500 ps edge | 21 ps 50 ps | Using fast rise pulses. The uncertainty quoted will be dependent on the vertical and horizontal resolution of the oscilloscope being calibrated. |
| Input resistance (DC) | 50 Ω 1 M Ω | 0.13 % 0.13 % | For values within 20% of the nominal values shown. |
| CALIBRATION OF MULTI-FUNCTION CALIBRATORS | | | |
| DC RESISTANCE | | | Using automated system. Generation of these parameters up to and including 100 kHz may also be undertaken but the uncertainties may be increased |
| Specific Values | 0 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω | 70 $\mu\Omega$ 70 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 13 $\mu\Omega/\Omega$ 12 $\mu\Omega/\Omega$ 12 $\mu\Omega/\Omega$ 19 $\mu\Omega/\Omega$ 32 $\mu\Omega/\Omega$ 49 $\mu\Omega/\Omega$ 640 $\mu\Omega/\Omega$ | |
| DC VOLTAGE | | | |
| Specific Values | 0 V 100 mV 1 V 10 V 19 V 100 V 1000 V | 1.1 μV 12 $\mu\text{V/V}$ 7.0 $\mu\text{V/V}$ 6.0 $\mu\text{V/V}$ 6.0 $\mu\text{V/V}$ 8.0 $\mu\text{V/V}$ 8.0 $\mu\text{V/V}$ | |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---|--|--|
| DC CURRENT Specific Values | 0 μ A 100 μ A 1 mA 10 mA 100 mA 1 A 10 A | 5.0 nA 48 μ A/A 45 μ A/A 45 μ A/A 46 μ A/A 70 μ A/A 100 μ A/A | |
| AC VOLTAGE | At 10 Hz, 20 Hz, 30 Hz, 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz, 1 kHz, 10 kHz and 20 kHz: 1 mV 10 mV 100 mV At 30 kHz and 50 kHz: 1 mV 10 mV 100 mV At 100 kHz: 1 mV 10 mV 100 mV 1 V and 10 V: 10 Hz, 20 Hz and 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz, 1 kHz, 10 kHz, 20 kHz and 30 kHz 50 kHz 100 kHz 300 kHz 500 kHz 1 MHz 19 V: 1 kHz 100 V: 10 Hz, 20 Hz and 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz, 1 kHz, 10 kHz, 20 kHz and 30 kHz 50 kHz 100 kHz 1000 V: 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz | 0.80 % 800 μ V/V 170 μ V/V 0.80 % 810 μ V/V 230 μ V/V 1.1 % 0.13 % 440 μ V/V 49 μ V/V 44 μ V/V 80 μ V/V 110 μ V/V 270 μ V/V 480 μ V/V 0.11 % 44 μ V/V 55 μ V/V 49 μ V/V 80 μ V/V 130 μ V/V 70 μ V/V | 1 mV and 10 mV are not available at 10 Hz |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|---|---|---|--|
| AC CURRENT | 100 μ A: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 1 mA: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 10 mA: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 100 mA: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 1 A: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 10 A: 40 Hz, 55 Hz, 300 Hz and 1 kHz | 230 μ A/A 220 μ A/A 210 μ A/A 380 μ A/A 190 μ A/A 190 μ A/A 180 μ A/A 290 μ A/A 190 μ A/A 190 μ A/A 180 μ A/A 280 μ A/A 190 μ A/A 190 μ A/A 180 μ A/A 280 μ A/A 230 μ A/A 220 μ A/A 190 μ A/A 411 μ A/A 800 μ A/A | |
| ELECTRICAL SIMULATION OF TEMPERATURE (Base metal thermocouple indicators & PT100) | | | Application or measurement of DC voltages equivalent to those for the thermocouple types indicated, with cold junction compensation enabled. |
| K type | -200 °C to -190 °C -190 °C to -100 °C -100 °C to +1300 °C | 0.18 °C 0.11 °C 0.086 °C | |
| T Type | -150 °C to -100 °C -100 °C to 0 °C 0 °C to 400 °C | 0.11 °C 0.082 °C 0.065 °C | |
| J Type | -100 °C to 0 °C 0 °C to 1000 °C | 0.083 °C 0.076 °C | |
| PT100 | -200 °C to -50 °C -50 °C to 200 °C 200 °C to 300 °C 300 °C to 500 °C 500 °C to 600 °C 600 °C to 800 °C | 1.4 m°C 2.5 m°C 5.0 m°C 7.4 m°C 10 m°C 11 m°C | By Resistance simulation. |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|--------------------------------|---|-------------------------------------|
| CALIBRATION OF 16TH/17TH EDITION TEST EQUIPMENT | | | Using dedicated calibration system. |
| Insulation Resistance | 10 k Ω to 5 M Ω | 0.16 % | |
| | 5 M Ω to 100 M Ω | 1.2 % | |
| | 100 M Ω to 1 G Ω | 1.4 % | |
| Continuity Resistance | 900 m Ω to 2 Ω | 3.5 % | |
| | 2 Ω to 6 Ω | 1.1 % | |
| | 6 Ω to 20 Ω | 0.67 % | |
| | 100 Ω | 0.32 % | |
| | 1 k Ω | 1.2 % | |
| Continuity Current | 100 mA | 2.2 % | |
| | 200 mA | 1.9 % | |
| | 300 mA | 1.7 % | |
| Insulation Voltage | 50 V | 3.0 % | |
| | 100 V | 2.1 % | |
| | 250 V | 1.5 % | |
| | 500 V | 1.3 % | |
| | 1000 V | 1.3 % | |
| Current on Insulation resistance function | | | |
| | | | |
| 1000 V range | 0.5 mA | 3.1 % | |
| | 1.0 mA | 2.1 % | |
| 500 V range | 0.5 mA | 3.1 % | |
| | 1.0 mA | 2.1 % | |
| Loop Resistance at 50 Hz | 0.33 Ω | 10 % | |
| | 0.5 Ω | 6.7 % | |
| | 1 Ω | 3.5 % | |
| | 5 Ω | 1.2 % | |
| | 10 Ω | 1.0 % | |
| | 100 Ω | 0.87 % | |
| | 1000 Ω | 0.87 % | |
| | | | |
| RCD Current at 50 Hz | 10 mA to 30 mA | 2.1 % | |
| | 30 mA to 300 mA | 1.9 % | |
| | 300 mA to 2 A | 1.6 % | |
| RCD Trip Time | 20 ms to 40 ms | 4.8 % | |
| | 40 ms to 200 ms | 2.4 % | |
| | 200 ms to 390 ms | 0.48 % | |
| | 390 ms to 900 ms | 0.90 % | |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|--|--|--|
| PAT Insulation Resistance | 1 M Ω to 4 M Ω 4 M Ω to 10 M Ω | 0.16 % 1.2 % | |
| PAT Earth Bond Resistance at 50 Hz | 0.1 Ω 0.22 Ω 0.33 Ω 0.5 Ω 1 Ω 5 Ω 10 Ω 100 Ω 1000 Ω | 8.0 % 4.1 % 3.0 % 2.3 % 1.6 % 1.0 % 0.94 % 0.87 % 0.87 % | |
| PAT Earth Bond Current at 50 Hz | 0 mA to 100 mA 100 mA to 10 A 10 A to 30 A | 2.1 % + 6 mA 1.7 % + 60 mA 1.7 % + 60 mA | |
| PAT Leakage Current Test | 2.7 mA at 240 V 4.7 mA at 240 V 7.7 mA at 240 V | 1.9 % 1.8 % 1.8 % | |
| PAT Flash Voltage Test At 50 Hz | 1000 V (Class 1) 1500 V (Class 1) 1000 V (Class 2) 3000 V (Class 2) | 5.8 % 5.4 % 5.8 % 5.0 % | |
| PAT Flash Current Test At 50 Hz | 0.67 mA at 1000 V (Class 1) 1.00 mA at 1500 V (Class 1) 0.34 mA at 1000 V (Class 2) 1.00 mA at 3000 V (Class 2) | 6.0 % 5.9 % 6.8 % 5.9 % | |
| TEMPERATURE CALIBRATION | | | Unless otherwise stated the calibration is performed by comparison with reference standards. Other units other than Celsius can be reported. |
| Temperature indicators and recorders with temperature sensor(s) – resistance | -20 °C to -10 °C -10 °C to +50 °C 50 °C to 100 °C 100 °C to 200 °C | 0.042 °C 0.037 °C 0.055 °C 0.051 °C | Calibration performed within Liquid Baths. |
| Temperature indicators and recorders with temperature sensor(s) - thermocouple | -20 °C to -10 °C -10 °C to +50 °C 50 °C to 100 °C 100 °C to 200 °C | 0.046 °C 0.040 °C 0.12 °C 0.21 °C | Calibration performed within Liquid Baths. |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|---|---|--|---|
| DEW POINT and RELATIVE HUMIDITY Dew-point Temperature probes in air and Temperature probes associated with hygrometers Relative humidity instruments | 1°C to 40°C 10 °C to 20 °C 20 °C to 25 °C 25 °C to 50 °C Example conditions At 10 °C 53 %rh to 85 %rh At 23 °C 24 %rh to 85 %rh At 30 °C 16 %rh to 85 %rh At 60 °C 10 %rh to 37 %rh | 0.33°C to 0.39°C 0.30 °C 0.30 °C 0.30 °C Corresponding to above dew- point and temperature uncertainties 1.5 %rh to 2.3 %rh 0.63 %rh to 1.8 %rh 0.49 %rh to 1.8 %rh 0.41 %rh to 0.87 %rh | Calibration by comparison with a reference dew point hygrometer and reference PRTs. Calibration performed within an air chamber. |
| DIMENSIONAL MEASUREMENTS Unless otherwise stated, the ranges are presented in millimetres (mm) and the uncertainties in micrometres (µm). MEASURING INSTRUMENTS AND MACHINES Micrometers External Internal Depth | BS 870:2008 0 to 600 Heads: (Zero) Setting, 0 to 25: (Zero) Setting, 25 to 600: Flatness of anvils: Parallelism of anvils: BS 959:2008 0 to 600 Heads: (Zero) Setting and extension rods: BS 6468:2008 0 to 300 Heads: (Zero) Setting: Base Flatness: Rod Flatness: Parallelism: Rod axis of rotation: Squareness of Face to spindle / rod axis: Rod axis to datum face: | 2.0 between any two points 1.0 1.0 + (8.0 x length in m) 0.50 1.2 2.0 between any two points 1.0 + (8.0 x length in m) 2.0 between any two points 1.0 + (8.0 x length in m) 0.42 0.42 2.7 5.0 1.6 10 | |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---|--|--|
| Vernier, Digital and Dial Gauges | | | |
| Calipers | Based on BS 887:2008 0 to 600 Overall performance: Flatness: Parallelism, External: Parallelism, Internal: Squareness: Co-Planer error of Jaws: Width of internal jaws: | $10 + (30 \times \text{length in m})$ 4.7 4.4 2.7 6.2 10 2.7 | |
| Height gauges | Based on BS 1643:2008 0 to 600 Overall performance: | $10 + (30 \times \text{length in m})$ | |
| Depth gauges | Based on BS 6365:2008 0 to 300 Overall performance: Flatness / Straightness: Parallelism: | $10 + (30 \times \text{length in m})$ 2.8 2.7 | |
| Internal and external dial caliper and thickness gauges | Scale accuracy: 0 to 50 50 to 200 Parallelism: | 1.6 10 1.6 | Using procedures MLCP 12 or MLCP 15 |
| Dial gauges and dial test indicators | BS 907:2008 and BS 2795:1981 0 to 100 | 1.0 | |
| Road measuring wheels Derived calibration factor Diameter Circumference Counter distance | 0.95 to 1.05 200 to 350 600 to 1050 29.9 to 30.1 revolutions | 0.0014 0.12 mm 0.37 mm 0.030 m | Using procedure MLCP 52 |
| LENGTH | | | |
| Length gauges, flat and spherical ended (excluding length bars) | BS 870:2008 0 to 600 | $1.0 + (8.0 \times \text{length in m})$ | |
| Feeler gauges | BS 957:2008 0.03 to 1 | 1.8 | |
| Steel rules, engineers | 0 to 1200 | $8.0 + (10 \times \text{length in metres})$ | Using procedure MLCP 57 |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|--|---|---|
| PRESSURE CALIBRATION | | | Methods consistent with EURAMET CG17 |
| Gas pressure (gauge) | | | |
| Calibration of pressure indicating instruments and gauges | -90 kPa to 100 kPa 100 kPa to 250 kPa 250 kPa to 2 MPa 2 MPa to 10 MPa | 14 Pa 33 Pa 0.33 kPa 1.3 kPa | Using deadweight tester. Comparison with pressure controllers and indicators. |
| Hydraulic pressure (gauge) | | | |
| Calibration of pressure indicating instruments and gauges | 0.6 MPa to 6 MPa 6 MPa to 16 MPa 16 MPa to 40 MPa 40 MPa to 70 MPa 70 MPa to 100 MPa | 0.020 % 6.8 kPa 61 kPa 0.020 % 78 kPa | Using deadweight tester. Comparison with pressure indicator. |
| END | | | |



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 062 Issue date 23 September 2024

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$