


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0318</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>CoMech Metrology Limited</h3> <p>Issue No: 038 Issue date: 01 June 2020</p>	
	<p>Metrology Division Calibration House Castings Road Derby DE23 8YL</p>	<p>Contact: Mr K Pallett Tel: +44 (0)1332 867 700 E-Mail: sales@comech.co.uk Website: www.comech.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
<p>RANGE IN MILLIMETRES AND UNDERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED</p>			
<p>MEASURING INSTRUMENTS AND MACHINES</p>			<p>NOTES</p>
<p>Micrometers External Internal Depth</p>	<p>BS 870:2008, 0 to 600 BS 959:2008, 0 to 600 BS 6468:2008, 0 to 300</p>	<p>Heads 2.0 Setting and extension rods 1.0 + (5.0 x length in m)</p>	<p>Calibrations may be made in metric or imperial units.</p>
<p>Vernier gauges Caliper Height Depth Dial gauge type Digital type</p>	<p>BS 887:2008 0 to 600 BS 1643:2008, 0 600 BS 6365:2008, 0 to 300 0 to 300 0 to 300</p>	<p>Overall performance: 10 + (30 x length in m)</p>	
<p>Dial gauges and dial test indicators</p>	<p>0 to 50 BS 907:2008 and BS 2795:1981</p>	<p>3.6</p>	
<p>Surface texture of Gauges (excluding surface texture standards)</p>	<p>As BS 1134:Part 1:1988</p>	<p>7.0 % (minimum 1.0 μm Ra)</p>	
<p>Length gauges, flat and spherical ended (excluding length bars)</p>	<p>25 to 1200</p>	<p>1.0 + (5.0 x length in m)</p>	<p>Using a length measuring Machine or by comparison with End standards</p>
<p>Plain plug gauges parallel</p>	<p>0 to 100</p>	<p>2.1</p>	<p>Using a length measuring machine</p>
<p>Receiver and position gauges, jigs, fixtures</p>	<p>Track Grinding Gauges (HB221) 0 to 1440</p>	<p>29</p>	<p>Procedure CM-C-686</p>
	<p>Autocoupler Pin Checking Gauges (HB035) 0 to 1.0</p>	<p>3.0</p>	<p>Procedure CM-C-706</p>
	<p>Brake Pad Wear Gauge (HB100) 0 to 10.0</p>	<p>2.7</p>	<p>Procedure CM-C-705</p>
	<p>Back to Back Gauges (HB018) 0 to 1360</p>	<p>8.1</p>	<p>Procedure CM-C-704</p>



0318
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CoMech Metrology Limited
Issue No: 038 Issue date: 01 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
PRESSURE <u>Gas pressure (gauge)</u> Calibration of pressure measuring instruments and gauges. <u>Gas pressure (absolute)</u> Calibration of pressure measuring instruments and gauges. <u>Hydraulic pressure (gauge)</u> Calibration of pressure measuring instruments and gauges.	- 80 kPa to 2 MPa 70 kPa to 2 MPa 300 kPa to 5.5 MPa 5 MPa to 110 MPa	0.48 kPa 0.48 kPa 0.044 % 0.030 %	Methods consistent with EURAMET CG3
TORQUE Hand torque tools (excluding torque screwdrivers)	As BS EN ISO 6789 :2017 5 N-m to 1356 N-m	1.0 %	Calibrations may also be given in lbf.in and lbf.ft.
ELECTRICAL MEASUREMENTS			

Electrical values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. The method used is by direct comparison against laboratory standards unless otherwise stated in the remarks column

DC Voltage High Voltage DC Current	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1.1 kV 1 kV to 40 kV 0 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 10.5 A 10.5 A to 20 A 20 A to 105 A 105 A to 525 A 525 A to 1000 A	10 ppm + 860 nV 9.2 ppm + 3.5 µV 9.2 ppm + 25 µV 12 ppm + 340 µV 18 ppm + 3.4 mV 0.70 % + 15 V 23 ppm + 3.3 nA 23 ppm + 21 nA 23 ppm + 200 nA 40 ppm + 2.3 µA 130 ppm + 43 µA 740 ppm + 1.2 mA 0.45 % + 26 mA 750 ppm + 12 mA 750 ppm + 60 mA 0.50 % + 1.3 A	Output values above 10 A -Simulation using a multi - turn Coil
--	---	---	--



0318
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CoMech Metrology Limited
Issue No: 038 Issue date: 01 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
ELECTRICAL MEASUREMENTS (cont'd)			
Resistance	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 100 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1.1 G Ω 1 G Ω to 2 G Ω	17 ppm + 180 $\mu\Omega$ 14 ppm + 1.0 m Ω 12 ppm + 58 m Ω 12 ppm + 84 m Ω 12 ppm + 1.0 Ω 17 ppm + 61 Ω 58 ppm + 320 Ω 580 ppm + 32 k Ω 0.58 % + 720 k Ω 21 M Ω	
AC Voltage	1 mV to 100 mV 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz 100 mV to 1 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz 1 V to 10 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz 10 V to 100 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz 100 V to 750 V 40 Hz to 1 kHz 1 kHz to 10 kHz 750 V to 1050 V 10 Hz to 3 kHz 3 kHz to 10 kHz 10 kHz to 20 kHz	700 ppm + 17 μV 240 ppm + 16 μV 0.16 % + 67 μV 710 ppm + 130 μV 250 ppm + 130 μV 0.18 % + 1.5 mV 700 ppm + 1.3 mV 250 ppm + 1.3 mV 0.18 % + 4.9 mV 700 ppm + 13 mV 360 ppm + 12 mV 0.18 % + 50 mV 940 ppm + 240 mV 0.18 % + 460 mV 940 ppm + 160 mV 9304ppm + 250 mV 0.14 % + 370 mV	
High Voltage	1 kV to 40 kV 50 Hz to 60 Hz	1.5 % + 45 V	



0318
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CoMech Metrology Limited
Issue No: 038 Issue date: 01 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
ELECTRICAL MEASUREMENTS (cont'd)			
AC Current	1 μ A to 100 μ A 40 Hz to 1 kHz	0.18 % + 38 nA	
	100 μ A to 1 mA 40 Hz to 1 kHz	0.46 % + 270 nA	
	1 mA to 10 mA 40 Hz to 1 kHz 1 kHz to 5 kHz	0.47 % + 2.8 μ A 0.035 % + 5.3 μ A	
	10 mA to 100 mA 40 Hz to 1 kHz	0.46 % + 15 μ A	
	100 mA to 1 A 40 Hz to 1 kHz	0.46 % + 340 μ A	
	1 A to 3.2 A 10 Hz to 3 kHz 3 kHz to 10 kHz	0.12 % + 560 μ A 0.13 % + 3.0 mA	
AC Current	3.2 A to 10.5 A 10 Hz to 3 kHz 3 kHz to 10 kHz	0.23 % + 4.4 mA 0.58 % + 12 mA	
	10.5 A to 105 A 10 Hz to 400 Hz	0.28 % + 44 mA	Output values above 10 A
	105 A to 525 A 10 Hz to 400 Hz	0.28 % + 220 mA	-Simulation using a multi -turn coil
Capacitance 1 kHz	0.5 nF to 4 nF 4 nF to 40 nF 40 nF to 400 nF 400 nF to 4 μ F 4 μ F to 40 μ F 40 μ F to 400 μ F 400 μ F to 4 mF 4 mF to 40 mF	0.50 % + 18 pF 0.40 % + 36 pF 0.40 % + 190 pF 0.50 % + 580 pF 0.60 % + 20 nF 0.60 % + 190 nF 0.60 % + 59 μ F 0.20 % + 70 μ F	Simulated values for the Calibration of capacitance Meters.
17th Edition capability			
Insulation Resistance	10 k Ω to 40 k Ω 40 k Ω to 200 k Ω 200 k Ω to 10 M Ω 10 M Ω to 1 G Ω 1 G Ω to 2 G Ω	200 ppm + 15 Ω 250 ppm + 15 Ω 500 ppm + 590 Ω 0.22 % + 120 k Ω 0.52 % + 120 k Ω	
Insulation Resistance: test current	0 mA to 9.9 mA	76 μ A	



0318
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CoMech Metrology Limited
Issue No: 038 Issue date: 01 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
ELECTRICAL MEASUREMENTS (cont'd)			
Continuity resistance	100 mΩ to 400 mΩ	1.0 % + 1.5 mΩ	
	400 mΩ to 5 Ω	0.80 % + 8.3 mΩ	
	5 Ω to 30 Ω	0.12 %	
	30 Ω to 200 Ω	0.11 %	
	200 Ω to 2 kΩ	0.11 %	
	2 kΩ to 10 kΩ	0.17 %	
Continuity resistance current	50 mA to 400 mA	21 %	
Voltage Output	3 V to 600 V	0.15 %	
	3 V 50 Hz to 600 V 50 Hz	0.18 %	
Voltage measurement	180 mV to 1 kV	0.060 % + 180 mV	
	130 mV 50 Hz to 1 kV 50 Hz	0.10 % + 130 mV	
High Voltage 50 Hz	1 kV to 7 kV	0.25 % + 24 V	
	10 kV to 25 kV	0.25 % + 120 V	
High Voltage DC	1 kV to 10 kV	0.060 % + 40 V	
	10 kV to 25 kV	0.060 % + 120 V	
High Voltage Current 50 Hz	300 μA	0.36 % + 3.4 μA	
	3 mA	0.24 % + 8.5 μA	
	30 mA	0.58 % + 27 μA	
	300 mA	0.23 % + 180 μA	
High Voltage Current DC	300 μA	0.36 % + 3.4 μA	
	3 mA	0.24 % + 8.5 μA	
	30 mA	0.58 % + 25 μA	
	300 mA	0.23 % + 180 μA	
Loop impedance 50 Hz	10 mΩ to 90 mΩ	3.7 % + 18 mΩ	
	90 mΩ to 320 mΩ	1.8 % + 18 mΩ	
	320 mΩ to 490 mΩ	0.90 % + 23 mΩ	
	490 mΩ to 1 Ω	0.60 % + 24 mΩ	
	1 Ω to 5 Ω	0.30 % + 43 mΩ	
	5 Ω to 500 Ω	0.10 % + 200 mΩ	
	500 Ω to 1.8 kΩ	0.10 % + 24 Ω	
RCD Trip Current 50 Hz	1 mA to 30 mA	0.20 % + 38 μA	
	30 mA to 300 mA	0.20 % + 1.4 mA	
	300 mA to 3 A	0.20 % + 11 mA	
RCD Trip time	0 ms to 5 s	0.10 % + 4.7 ms	



0318
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CoMech Metrology Limited
Issue No: 038 Issue date: 01 June 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
ELECTRICAL MEASUREMENTS (cont'd)			
PAT Insulation resistance	10 kΩ to 40 kΩ 40 kΩ to 200 kΩ 200 kΩ to 10 MΩ 10 MΩ to 1 GΩ 1 GΩ to 2 GΩ	200 ppm + 15 Ω 250 ppm + 15 Ω 500 ppm + 590 Ω 0.22 % + 120 kΩ 0.52 % + 120 kΩ	
Earth Bond resistance	0 Ω to 10 Ω 10 Ω to 1 kΩ	0.30 % + 40 mΩ 0.11 % + 200 mΩ	
Earth bond current	100 μA to 100 mA 100 mA to 20 A	4.0 mA 0.53 % + 32 mA	
Leakage current	1 μA to 30 mA	0.52 % + 20 μA	
Temperature simulation Thermocouple type			
K	-180 °C to 0 °C 0 °C to 1300 °C	0.16 °C 0.14 °C	Excluding internal reference junction compensation
J	-200 °C to 0 °C 0 °C to 1190 °C	0.15 °C 0.13 °C	
K	-180 °C to 0 °C 0 °C to 1300 °C	0.64 °C 0.63 °C	Including internal reference junction compensation
J	-200 °C to 0 °C 0 °C to 1190 °C	0.63 °C 0.63 °C	
Resistance thermometer simulation			
PT 100	-200 °C to 0 °C 0 °C to 850 °C	0.21 °C 0.21 °C	
END			



0318

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CoMech Metrology Limited
Issue No: 038 Issue date: 01 June 2020

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10⁻⁶·p·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.