


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0318</b> Accredited to ISO/IEC 17025:2017	<b>CoMech Metrology Limited</b>	
	Issue No: 041    Issue date: 04 April 2022	
	<b>Metrology Division</b> Calibration House Castings Road Derby DE23 8YL	<b>Contact: Mr K Pallett</b> Tel: +44 (0)1332 867 700 E-Mail: sales@comech.co.uk Website: www.comech.co.uk
<b>Calibration performed by the Organisation at the locations specified</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Castings Road Derby DE23 8YL	<b>Local contact</b>  <a href="#">Dimensional Torque</a>	A
<b>Address</b> 1 Pride Park View, Victoria Way, Pride Park, Derby DE24 8AN	<b>Local contact</b>  <a href="#">Electrical Pressure</a>	B



0318

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**CoMech Metrology Limited**

Issue No: 041 Issue date: 04 April 2022

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
RANGE IN MILLIMETRES AND UNDERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
<b>DIMENSIONAL</b>				
Micrometers External Internal Depth	BS 870:2008, 0 to 600 BS 959:2008, 0 to 600 BS 6468:2008, 0 to 300	Heads 2.0 Setting and extension rods 1.0 + (5.0 x length in m)	Calibrations may be made in metric or imperial units.	A
Vernier gauges Caliper Height Depth Dial gauge type Digital type	BS 887:2008 0 to 600 BS 1643:2008, 0 600 BS 6365:2008, 0 to 300 0 to 300 0 to 300	Overall performance: 10 + (30 x length in m)		A
Dial gauges and dial test indicators	0 to 50 BS 907:2008 and BS 2795:1981	3.6		A
Surface texture of Gauges (excluding surface texture standards)	As BS 1134:Part 1:1988	7.0 % (minimum 1.0 $\mu$ m Ra)		A
Length gauges, flat and spherical ended (excluding length bars	25 to 1200	1.0 + (5.0 x length in m)	Using a length measuring Machine or by comparison with End standards	A
Plain plug gauges parallel	0 to 100	2.1	Using a length measuring machine	A
Receiver and position gauges, jigs, fixtures	Track Grinding Gauges (HB221) 0 to 1440	29	Procedure CM-C-686	A
	Autocoupler Pin Checking Gauges (HB035) 0 to 1.0	3.0	Procedure CM-C-706	
	Brake Pad Wear Gauge (HB100) 0 to 10.0	2.7	Procedure CM-C-705	
	Back to Back Gauges (HB018) 0 to 1360	8.1	Procedure CM-C-704	



0318

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**CoMech Metrology Limited**

Issue No: 041 Issue date: 04 April 2022

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>PRESSURE</b>  <u>Gas pressure (gauge)</u>  Calibration of pressure measuring instruments and gauges.  Gas pressure (absolute) Calibration of pressure measuring instruments and gauges.  Hydraulic pressure (gauge)  Calibration of pressure measuring instruments and gauges.	- 80 kPa to 2 MPa  70 kPa to 2 MPa  500 kPa to 5.5 MPa 5 MPa to 110 MPa	0.48 kPa  0.48 kPa  0.039 % 0.030 %	Methods consistent with EURAMET CG17	B
<b>TORQUE</b>  Hand torque tools (excluding torque screwdrivers)	As BS EN ISO 6789 :2017 5 N·m to 1356 N·m	1.0 %	Calibrations may also be given in lbf.in and lbf.ft.	A
<b>ELECTRICAL</b>				B
Electrical values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. The method used is by direct comparison against laboratory standards unless otherwise stated in the remarks column.				
<b>DC Voltage</b>	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1.1 kV	10 $\mu$ V/V + 860 nV 9.2 $\mu$ V/V + 3.5 $\mu$ V 9.2 $\mu$ V/V + 25 $\mu$ V 12 $\mu$ V/V + 340 $\mu$ V 18 $\mu$ V/V + 3.4 mV		B
<b>High Voltage</b>	1 kV to 40 kV	0.70 % + 15 V	These output values can be measured.	B
<b>DC Current</b>	0 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A  1 A to 2 A 2 A to 20 A 20 A to 120 A	23 $\mu$ A/A + 3.3 nA 23 $\mu$ A/A + 21 nA 23 $\mu$ A/A + 200 nA 40 $\mu$ A/A + 2.3 $\mu$ A 130 $\mu$ A/A + 43 $\mu$ A  82 $\mu$ A/A + 370 $\mu$ A 76 $\mu$ A/A + 9.8 mA 50 $\mu$ A/A +	Values available for direct measurement	B
<b>DC Current</b>	1 A to 10.5 A 10.5 A to 20 A 20 A to 105 A 105 A to 525 A 525 A to 1000 A 1000 A to 2500 A	0.074 % + 1.2 mA 0.45 % + 26 mA 0.075 % + 12 mA 0.075 % + 60 mA 0.50 % + 1.3 A 0.069 % + 520 mA	Output values above 10 A – simulation Using a multiturn coil	B



0318

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**CoMech Metrology Limited**

Issue No: 041 Issue date: 04 April 2022

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
ELECTRICAL continued				
Resistance	0 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1 k $\Omega$ 1 k $\Omega$ to 10 k $\Omega$ 10 k $\Omega$ to 100 k $\Omega$ 100 k $\Omega$ to 1 M $\Omega$ 1 M $\Omega$ to 100 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$ 100 M $\Omega$ to 1.1 G $\Omega$ 1 G $\Omega$ to 2 G $\Omega$	17 $\mu\Omega/\Omega + 180 \mu\Omega$ 14 $\mu\Omega/\Omega + 1.0 \text{ m}\Omega$ 12 $\mu\Omega/\Omega + 58 \text{ m}\Omega$ 12 $\mu\Omega/\Omega + 84 \text{ m}\Omega$ 12 $\mu\Omega/\Omega + 1.0 \Omega$ 17 $\mu\Omega/\Omega + 61 \Omega$ 58 $\mu\Omega/\Omega + 320 \Omega$ 0.058 % + 32 k $\Omega$ 0.58 % + 720 k $\Omega$ 21 M $\Omega$		B
AC Voltage	1 mV to 100 mV 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz  100 mV to 1 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz  1 V to 10 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz  10 V to 100 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz  100 V to 750 V 40 Hz to 1 kHz 1 kHz to 10 kHz  750 V to 1050 V 10 Hz to 3 kHz 3 kHz to 10 kHz 10 kHz to 20 kHz	0.070 % + 17 $\mu\text{V}$ 0.024 % + 16 $\mu\text{V}$ 0.16 % + 67 $\mu\text{V}$  0.071 % + 130 $\mu\text{V}$ 0.025 % + 130 $\mu\text{V}$ 0.18 % + 1.5 mV  700 $\mu\text{V}/\text{V} + 1.3 \text{ mV}$ 250 $\mu\text{V}/\text{V} + 1.3 \text{ mV}$ 0.18 % + 4.9 mV  0.070 % + 13 mV 0.036 % + 12 mV 0.18 % + 50 mV  0.094 % + 240 mV 0.18 % + 460 mV  0.094 % + 160 mV 0.094 % + 250 mV 0.14 % + 370 mV		B
High Voltage	1 kV to 40 kV 50 Hz to 60 Hz	1.5 % + 45 V	These output values can be measured.	B
AC Current	1 $\mu\text{A}$ to 100 $\mu\text{A}$ 40 Hz to 1 kHz  100 $\mu\text{A}$ to 1 mA 40 Hz to 1 kHz  1 mA to 10 mA 40 Hz to 1 kHz 1 kHz to 5 kHz  10 mA to 100 mA 40 Hz to 1 kHz	0.18 % + 38 nA  0.46 % + 270 nA  0.47 % + 2.8 $\mu\text{A}$ 0.035 % + 5.3 $\mu\text{A}$  0.46 % + 15 $\mu\text{A}$		B



0318

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**CoMech Metrology Limited**

Issue No: 041 Issue date: 04 April 2022

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
AC Current (continued)	100 mA to 1 A <i>40 Hz to 1 kHz</i>	0.46 % + 340 $\mu$ A	Simulation using a Multiturn coil	B
	1 A to 3.2 A <i>10 Hz to 3 kHz 3 kHz to 10 kHz</i>	0.12 % + 560 $\mu$ A 0.13 % + 3.0 mA		
	3.2 A to 10.5 A <i>10 Hz to 3 kHz 3 kHz to 10 kHz</i>	0.23 % + 4.4 mA 0.58 % + 12 mA		
	10.5 A to 120 A <i>50 Hz 51 Hz to 100 Hz</i>	0.36 % + 140 mA 0.58 % + 110 mA		
	10.5 A to 105 A <i>10 Hz to 400 Hz</i>	0.28 % + 44 mA		
	105 A to 525 A <i>10 Hz to 400 Hz</i>	0.28 % + 220 mA		
	500 A to 1000 A <i>50 Hz 51 Hz to 100 Hz</i>	0.043 % + 130 mA 0.043 % + 310 mA		
Capacitance 1 kHz	1000 A to 2500 A <i>50 Hz to 100 Hz</i>	0.043 % + 1.3 A	Simulated values for the calibration of capacitance meters.	B
	0.5 nF to 4 nF	0.50 % + 18 pF		
	4 nF to 40 nF	0.40 % + 36 pF		
	40 nF to 400 nF	0.40 % + 190 pF		
	400 nF to 4 $\mu$ F	0.50 % + 580 pF		
	4 $\mu$ F to 40 $\mu$ F	0.60 % + 20 nF		
	40 $\mu$ F to 400 $\mu$ F	0.60 % + 190 nF		
	400 $\mu$ F to 4 mF	0.60 % + 59 $\mu$ F		
4 mF to 40 mF	0.20 % + 70 $\mu$ F			
Oscilloscopes				
Vertical deflection	<i>At 1 kHz</i> 10 mV to 100 mV 100 mV to 120 V	0.040 % + 27 $\mu$ V 0.050 %	B	
Horizontal deflection	2 ns to 5 s	0.12 %		
<b>17<sup>th</sup> Edition capability</b>				
Insulation Resistance	10 k $\Omega$ to 40 k $\Omega$ 40 k $\Omega$ to 200 k $\Omega$ 200 k $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 1 G $\Omega$ 1 G $\Omega$ to 2 G $\Omega$	0.020 % + 15 $\Omega$ 0.025 % + 15 $\Omega$ 0.050 % + 590 $\Omega$ 0.22 % + 120 k $\Omega$ 0.52 % + 120 k $\Omega$	B	
Insulation Resistance: test current	0 mA to 9.9 mA	76 $\mu$ A		



0318

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**CoMech Metrology Limited**

**Issue No: 041 Issue date: 04 April 2022**

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
Continuity resistance	100 mΩ to 400 mΩ	1.0 % + 1.5 mΩ		B
	400 mΩ to 5 Ω	0.80 % + 8.3 mΩ		
	5 Ω to 30 Ω	0.12 %		
	30 Ω to 200 Ω	0.11 %		
	200 Ω to 2 kΩ	0.11 %		
	2 kΩ to 10 kΩ	0.17 %		
Continuity resistance current	50 mA to 400 mA	21 %		B
Voltage Output	3 V to 600 V	0.15 %		
	3 V 50 Hz to 600 V 50 Hz	0.18 %		
Voltage measurement	180 mV to 1 kV	0.060 % + 180 mV		B
	130 mV 50 Hz to 1 kV 50 Hz	0.10 % + 130 mV		
High Voltage 50 Hz	1 kV to 7 kV	0.25 % + 24 V		B
	10 kV to 25 kV	0.25 % + 120 V		
High Voltage DC	1 kV to 10 kV	0.060 % + 40 V		B
	10 kV to 25 kV	0.060 % + 120 V		
High Voltage Current 50 Hz	300 μA	0.36 % + 3.4 μA		B
	3 mA	0.24 % + 8.5 μA		
	30 mA	0.58 % + 27 μA		
	300 mA	0.23 % + 180 μA		
High Voltage Current DC	300 μA	0.36 % + 3.4 μA		B
	3 mA	0.24 % + 8.5 μA		
	30 mA	0.58 % + 25 μA		
	300 mA	0.23 % + 180 μA		
Loop impedance 50 Hz	10 mΩ to 90 mΩ	3.7 % + 18 mΩ		B
	90 mΩ to 320 mΩ	1.8 % + 18 mΩ		
	320 mΩ to 490 mΩ	0.90 % + 23 mΩ		
	490 mΩ to 1 Ω	0.60 % + 24 mΩ		
	1 Ω to 5 Ω	0.30 % + 43 mΩ		
	5 Ω to 500 Ω	0.10 % + 200 mΩ		
	500 Ω to 1.8 kΩ	0.10 % + 24 Ω		
RCD Trip Current 50 Hz	1 mA to 30 mA	0.20 % + 38 μA		B
	30 mA to 300 mA	0.20 % + 1.4 mA		
	300 mA to 3 A	0.20 % + 11 mA		
RCD Trip time	0 ms to 5 s	0.10 % + 4.7 ms		B
PAT Insulation resistance	10 kΩ to 40 kΩ	0.020 % + 15 Ω		B
	40 kΩ to 200 kΩ	0.025 % + 15 Ω		
	200 kΩ to 10 MΩ	0.050 % + 590 Ω		
	10 MΩ to 1 GΩ	0.22 % + 120 kΩ		
	1 GΩ to 2 GΩ	0.52 % + 120 kΩ		
Earth Bond resistance	0 Ω to 10 Ω	0.30 % + 40 mΩ		B
	10 Ω to 1 kΩ	0.11 % + 200 mΩ		
Earth bond current	100 μA to 100 mA	4.0 mA		B
	100 mA to 20 A	0.53 % + 32 mA		
Leakage current	1 μA to 30 mA	0.52 % + 20 μA		B



0318

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**CoMech Metrology Limited**

**Issue No: 041 Issue date: 04 April 2022**

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>Temperature simulation</b> Thermocouple type				B
K	-180 °C to 0 °C	0.16 °C	Excluding internal reference junction compensation	
	0 °C to 1300 °C	0.14 °C		
J	-200 °C to 0 °C 0 °C to 1190 °C	0.15 °C 0.13 °C	Including internal reference junction compensation	
K	-180 °C to 0 °C	0.64 °C		
	0 °C to 1300 °C	0.63 °C		
J	-200 °C to 0 °C 0 °C to 1190 °C	0.63 °C 0.63 °C		
Resistance thermometer simulation				
PT 100	-200 °C to 0 °C 0 °C to 850 °C	0.21 °C 0.21 °C		
<b>END</b>				



0318

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**CoMech Metrology Limited**

**Issue No: 041 Issue date: 04 April 2022**

**Calibration performed by the Organisation at the locations specified**

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$