

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0328</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>Control Plant Ltd Trading as Lansbury International</b>	
	Issue No: 016    Issue date: 23 August 2018	
	<b>The Lansbury Estate 102 Lower Guildford Road Knaphill Woking Surrey GU21 2EP</b>	<b>Contact: Mr M J Craven Tel: +44 (0)1483-472571 Fax: +44 (0)1483-475618 E-Mail: sales@lansbury.co.uk Website: www.springhammer.co.uk</b>
<b>Calibration performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
Impact Energy Imparted from Spring Operated Impact Test Apparatus - as specified in BS EN 60068-2-75:1997, BS EN 60068-2-75:2014 and IEC 60068-2-75 1997	0.10 Joule to 1.0 Joule	0.010 Joule	Calibrations can be given in Joule or Newton Metre units.
END			