

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> CALIBRATION</p> <p>0331</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>SGS United Kingdom Limited</b></p> <p>Issue No: 047    Issue date: 20 July 2020</p>	
	<p>2 Martin Close Blenheim Industrial Estate Bulwell Nottingham NG6 8UW</p>	<p>Contact: Mr S Cooke Tel: +44 (0)1159 424748 Fax: +44 (0)1159 424746 E-Mail: sam.cooke@sgs.com Website: www.uk.sgs.com</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b> 2 Martin Close Blenheim Industrial Estate Bulwell Nottingham NG6 8UW</p> <p><b>Local contact</b> Mr S Cooke</p>	<p><a href="#">Dimensional</a> <a href="#">Electrical</a> <a href="#">Pressure</a> <a href="#">Temperature</a> <a href="#">Humidity</a></p>	<p>Nottingham</p>

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p><b>Customers' sites or premises</b></p> <p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p> <p>Mr S Cooke</p>	<p><a href="#">Dimensional</a> <a href="#">Electrical</a> <a href="#">Pressure</a> <a href="#">Temperature</a></p>	<p>Based at Nottingham</p>



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				Nottingham
LENGTH			NOTE	
Plain plug gauges (parallel), cylindrical setting standards and rollers	1 to 50 50 to 100 100 to 150 150 to 200 200 to 300	0.50 0.80 1.0 1.2 1.6	Plain cylindrical limit gauges are calibrated using a single axis measuring machine and reference standards	
Thread measuring cylinders	BS 5590 0.1 to 5	0.50		
Plain ring gauges (parallel) and setting standards	2 to 10 diameter 10 to 25 25 to 50 50 to 100 100 to 150 150 to 250	1.0 0.80 1.0 1.5 2.0 3.0		
Length gauges, flat and spherical ended	0 to 600	1.0 + (8.0 x length in m)		
Plain gap gauges (parallel)	0.5 to 100 100 to 200 200 to 300	3.0 5.0 8.0	Using gauge block	
Feeler Gauges and paint thickness foils	BS 957:2008 0.025 to 1	1.0	Using a single axis measuring machine	
Screw plug gauges (parallel) including check and setting plugs See Note 3	1 to 100 diameter 100 to 150 150 to 200	2.5 5.0 8.0	on pitch diameter	
Screw plug gauges (taper) including check plugs See Note 2	5 to 100 diameter	5.0	on pitch diameter	
Screw ring gauges (parallel) See Notes 3 and 4	3 to 75 diameter 75 to 150 150 to 250	5.0 7.0 10	on pitch diameter	
Screw ring gauges (taper) See Note 2	6 up to 150 diameter	7.0	on pitch diameter	
Screw pitch Screw flank angle	0.2 to 8 0° to 52°	1.5 5.0 minutes of arc		
Parallels	BS906:1972 5 to 50 x 100 x 400	1.5 up to 5.0		



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				Nottingham
LENGTH (cont'd)				
Vee blocks	BS 3731:1987 20 to 150	2.5 to 5.0		
Receiver, position and profile gauges, jigs, fixtures See note 5	Maximum dimensions 0 to 1000 x 750 x 500 (Limited to gauges where a specific procedure and uncertainty budget are available).	Minimum per co-ordinate: 3.0 + (10 x length in m)	Documented in- house methods using a cmm and first principles	
ANGLE				
Squares Blade type	BS 939:2007 50 to 300 300 to 600	3.0 On squareness 5.0 See Note 1		
Angle plates and box angle plates	BS 5535:1978 50 to 600	Squareness: 3.0 + (1.0 per 100 mm) Parallelism: 1.0 + (1.0 per 100 mm) See Note 1		
Sine bars and tables	BS 3064:1978 0 to 500 length	Linear dimensions: 1.0 + (10 x length in m) Overall performance: 3.0 seconds of arc		
FORM				
Surface plates Granite and Cast iron	BS 817:2008 and above 160 x 100 to 1600 x 1000	1.5 + (0.80 diagonal in m) See Note 1		
Steel balls	1 to 25 diameter	0.50 on diameter	Single axis measuring machine and reference standard	



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code	
<b>MEASURING INSTRUMENTS AND MACHINES</b>					
Micrometers External	BS 870:2008 0 to 600	Heads: 2.0			
Internal	BS 959:2008 0 to 900	Setting and extension rods 1.0 + (8.0 x length in m)			
Depth	BS 6468:2008 0 to 300				
Micrometer heads	BS 1734:1951 0 to 50	1.0			
Bench micrometer	NPL MOY/SCMI 22 0 to 100	Overall performance 1.0			
Vernier caliper, height and depth gauges	BS 887:2008 0 to 1000 BS 1643:2008 0 to 1000 BS 6365:2008 0 to 600	Overall performance 10 + (30 x length in m)			
Height gauges - (Simple) including vernier, dial and digital types.	As BS EN ISO 13225:2012 0 to 1000	Length measurement error (E): 2.0 + (5.0 x length in metres)			
Dial gauges and dial test indicators	BS 907:2008 and BS 2795:1981 0 to 50	1.0			
Comparators (external)	BS 1054:1975 250 to 10 000 magnifications	1.0 % of range Minimum 0.20			
Graduated rules	BS 4372:1968 0 to 1000	5.0 + (10 x length in m)			
Bevel protractors	BS 1685:2008 0° to 360°	6.0 minutes of arc			
Electronic height gauges with microprocessor control	0 to 1000	2.0 + (5.0 x length in m)	Documented in- house methods		
Notes for length, angle and form					
1 The uncertainty quoted is for the departure from flatness, straightness, parallelism, or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration.					
2. Single start, symmetrical thread forms only.					
3. Single and multi-start symmetrical and asymmetrical thread forms.					
4. Includes use of check plugs for screw rings from 1 mm to 14 mm diameter					
5. Features and associated parts of these gauges can be measured to the uncertainties given for equivalent items listed in this schedule.					



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
ELECTRICAL				Nottingham
DC Voltage	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V 1kV to 50 kV	10 ppm + 0.80 $\mu$ V 7.0 ppm + 1.0 $\mu$ V 8.0 ppm + 3.5 $\mu$ V 8.0 ppm +60 $\mu$ V 8.0 ppm + 0.60 mV 0.20 % + 7.0 V	Electrical calibrations are performed as a comparison against a reference standard  These values can be generated for the calibration of measuring instruments, Outputs of instruments can be measured directly	
DC Resistance	0 $\Omega$ to 20 $\Omega$ 20 $\Omega$ to 200 $\Omega$ 200 $\Omega$ to 2 k $\Omega$ 2 k $\Omega$ to 20 k $\Omega$ 20 k $\Omega$ to 200 k $\Omega$ 200 k $\Omega$ to 2 M $\Omega$ 2 M $\Omega$ to 20 M $\Omega$ 20 M $\Omega$ to 200 M $\Omega$ 200 M $\Omega$ to 1 G $\Omega$	18 ppm + 24 $\mu$ $\Omega$ 11 ppm + 0.10 m $\Omega$ 9.0 ppm + 0.80 m $\Omega$ 9.0 ppm + 8.0 m $\Omega$ 13 ppm + 80 m $\Omega$ 22 ppm + 1.7 $\Omega$ 38 ppm + 100 $\Omega$ 200 ppm + 11 k $\Omega$ 0.12 % + 0.10 M $\Omega$		
DC Current	0 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A  2 A to 30 A	35 ppm + 0.80 nA 30 ppm + 6.0 nA 35 ppm + 52 nA 46 ppm + 1.2 $\mu$ A 70 ppm + 25 $\mu$ A  400 ppm + 0.30 mA		
AC Voltage	20 Hz to 60 Hz 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V  60 Hz to 3 kHz 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V  3 kHz to 30 kHz 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V  3 kHz to 10 kHz 200 V to 1000 V	0.025 % + 5.0 $\mu$ V 0.024 % + 24 $\mu$ V 0.021 % + 0.23 mV 0.021 % + 2.3 mV 0.024 % + 12 mV  0.022 % + 5.0 $\mu$ V 0.014 % + 24 $\mu$ V 0.014 % + 0.23 mV 0.015 % + 2.3 mV 0.020 % + 12 mV  0.025 % + 10 $\mu$ V 0.015 % + 50 $\mu$ V 0.015 % + 0.50 mV 0.016 % + 5.0 mV  0.040 % + 120 mV		



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
ELECTRICAL (cont'd)				Nottingham
AC Voltage (cont'd)	30 kHz to 100 kHz 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V	0.060 % + 24 $\mu$ V 0.075 % + 0.24 mV 0.030 % + 2.3 mV 0.050 % + 23 mV		
	100 kHz to 500 kHz 200 mV to 2 V 2 V to 20 V	0.20 % + 2.4 mV 0.30 % + 23 mV		
	500 kHz to 1 MHz 200 mV to 2 V 2 V to 20 V	0.25 % + 24 mV 0.25 % + 230 mV		
	50 Hz 1 kV to 50 kV	0.60 % + 60 V		
AC Current	40 Hz to 1 kHz 1 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA	0.040 % + 24 nA 0.040 % + 0.24 $\mu$ A 0.040 % + 2.4 $\mu$ A 0.045 % + 24 $\mu$ A		
	40 Hz to 100 Hz 200 mA to 2 A	0.030 % + 0.50 mA		
	100 Hz to 1 kHz 200 mA to 2 A	0.065 % + 1.0 mA		
	1 kHz to 5 kHz 1 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	0.10 % + 24 nA 0.075 % + 0.24 $\mu$ A 0.075 % + 2.4 $\mu$ A 0.090 % + 24 $\mu$ A 0.14 % + 1.0 mA		
	50 Hz to 60 Hz 2 A to 30 A	0.040 % + 2.5 mA		



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
ELECTRICAL (cont'd)				Nottingham
Electrical calibration of temperature indicators for the following sensors:				
Noble metal thermocouples	0 °C to 1600 °C	0.50 °C	Including cold junction compensation	
	0 °C to 1600 °C	0.15 °C	Cold junction disabled	
Base metal thermocouples	-200 °C to +1300 °C	0.13 °C	Including cold junction compensation	
	-200 °C to +1300 °C	0.060 °C	Cold junction disabled	
Resistance sensors	-200 °C to +250 °C +250 °C to +650 °C	5.0 m°C 30 m°C		
Cold junction compensation of thermocouple indicators and sources	At ambient temperature	0.10 °C		
Capacitance Sourcing	1 nF 10 nF 20 nF 50 nF 100 nF 1 µF 10 µF	4.2 pF 30 pF 48 pF 72 pF 93 pF 2.0 nF 50 nF		
Frequency Measurement	0 Hz to 3 GHz 3 GHz to 26 GHz	4.0 in 10 <sup>12</sup> + 3.0 mHz 3.0 in 10 <sup>9</sup>		
Generation	0 Hz to 6 GHz	4.0 in 10 <sup>12</sup>		
Period	10 ns to 1700 s	1.5 in 10 <sup>9</sup>		
Time interval	1 s to 8 x 10 <sup>5</sup> s	8.0 in 10 <sup>9</sup>		



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code		
ELECTRICAL (cont'd)						
17 <sup>th</sup> and 18 <sup>th</sup> edition electrical test instruments						
Insulation Testers						
Resistance	10 kΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ	0.30 % 0.85 % 1.5 %	Using the calibrator to generate known quantities to simulate measurements made by the Electrical Test Instruments	Nottingham		
Output Voltage DC	50 V to 1000 V	0.40 %				
Output Current DC	0 mA to 1 mA	0.060 % + 6.0 μA				
Continuity Resistance	50 mΩ to 1 Ω 1 Ω to 50 kΩ	0.30 % + 12 mΩ 0.40 %				
Continuity Current DC	0 mA to 320 mA	0.70 mA				
Voltage AC	100 V to 400 V at 50 Hz	0.40 %				
Loop Testers						
Loop Resistance	0 Ω to 5 Ω 9 Ω to 1000 Ω	0.60 % + 24 mΩ 0.60 % + 45 mΩ				
RCD Testers						
Current Measurement AC Trip time	0.1mA to 3000mA at 50 Hz 0 ms to 390 ms 390 ms to 5 s	1.4 % + 0.070 mA 0.70 ms 8.2 ms				
Appliance Testers						
Insulation Resistance Earth Bond Resistance	0.1 MΩ to 10 MΩ 50 mΩ to 10 Ω 10 Ω to 1000 Ω	0.30 % 0.60 % + 8.0 mΩ 0.60 % + 35 mΩ				
Earth Bond Current AC	0.1 mA to 500 mA at 50 Hz	1.75 % + 7.0 mA				
Leakage Current AC	500 mA to 30 A at 50 Hz 1 mA to 8 mA at 50 Hz	1.75 % + 75 mA 1.75 % + 10 μA				
Flash Test (Accessory)						
AC Voltage AC Leakage Current	0.1 kV to 3 kV at 50 Hz 0.1 mA to 3 mA at 50 Hz	2.0 % + 12 V 6.0 % + 17 μA				





0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
ELECTRICAL (cont'd)  Electrical verification of ultrasonic flaw detection equipment	AS BS EN 12668-1:2010  Stability after warm up (height) Stability after warm up (width) Jitter – screen height Jitter – screen width Stability against supply variations (height) Stability against supply variations (width) Pulser Voltage Pulser Rise Time Pulser Reverberation Pulse Duration Frequency Response <i>0.1 MHz to 25 MHz</i> Equivalent input noise Calibrated Attenuator <i>0 MHz to 100 MHz</i> Vertical Linearity Linearity of Timebase	0.30 % of screen height 0.020 % of screen width 0.30 % of screen height 0.30 % of screen width 0.30 % of screen height 0.020 % of screen width 3.6 V 1.7 ns 1.0 % of pulsar voltage 1.7 ns 30 kHz at -3 dB point $2.4 \times 10^{-9} \text{ V}\sqrt{\text{Hz}}$ 0.60 dB 0.70 % of screen height 0.25 % relative to 20 % and 80 % of the horizontal scale		Nottingham
PRESSURE  <u>Gas Pressure Gauge</u>  Calibration of pressure indicating instruments and gauges	-95 kPa to -20 kPa -20 kPa to 20 kPa 20 kPa to 200 kPa 200 kPa to 2 MPa	0.012 % + 66 Pa 0.015 % + 10 Pa 0.012 % + 66 Pa 0.012 % + 590 Pa	Methods consistent with EURAMET CG3	
<u>Gas Pressure Absolute</u>  Calibration of pressure indicating instruments and gauges	3.5 kPa to 120 kPa 120 kPa to 200 kPa	100 Pa 140 Pa		
<u>Hydraulic Pressure Gauge</u>  Calibration of pressure indicating instruments and gauges	0 Pa to 13.5 MPa 13.5 MPa to 70 MPa	0.010 % + 6.0 kPa 0.015 % + 20 kPa		



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
TEMPERATURE			Calibration by Comparison with reference instruments	Nottingham
Resistance thermometers	-20 °C to +100 °C 100 °C to 250 °C 250 °C to 600 °C	0.055 °C 0.075 °C 0.15 °C	Liquid bath  Block calibrator	
Thermocouples	-20 °C to 250 °C 250 °C to 600 °C 600 °C to 1100 °C 1100 °C to 1300 °C	0.50 °C 0.90 °C 1.9 °C 2.8 °C	Liquid bath Block calibrator Spherical furnace	
Temperature indicators with probes	-20 °C to +100 °C 100 °C to 250 °C 250 °C to 600 °C	0.060 °C 0.090 °C 0.15 °C	Liquid bath  Block calibrator	
	Ambient (typically 20 °C)	0.070 °C	Ambient air	
Dry Block Calibrators	-20 °C to +600 °C	0.13 °C		
Radiation thermometers (Pyrometers)	0 °C to +100 °C 100 °C to 200 °C 200 °C to 350 °C 350 °C to 500 °C	1.0 °C 1.5 °C 2.5 °C 3.5 °C	Radiation thermometers working in the wavelength range 8 µm to 14 µm	
Calibration in air chamber	5 °C to 23 °C 23 °C to 60 °C	0.10 °C 0.15 °C	Comparison in air chamber	
HUMIDITY			Calibration by comparison with a reference hygrometer and reference thermometers	
Relative humidity meters	At 5 °C to 12 °C 15 %rh to 90 %rh	0.45 %rh to 1.9 %rh	Comparison in air chamber	
	At 12 °C to 23 °C 10 %rh to 90 %rh	0.50 %rh to 1.9 %rh		
	At 23 °C to 60 °C 10 %rh to 90 %rh	0.50 %rh to 1.6 %rh		



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks	Location Code
<b>DIMENSIONAL</b>				On site calibrations Based at Nottingham
Surface plates Granite and Cast iron	BS 817:2008 and above 160 x 100 to 1600 x 1000	1.5 + (0.80 diagonal in m) See Note 1		
Electronic height gauges with microprocessor control	0 to 1000	2.0 + (5.0 x length in m)		
<b>ELECTRICAL</b>				
Electrical calibration of temperature indicators for the following sensors:				
Noble metal thermocouples	0 °C to 1600 °C	0.90 °C	Including cold junction compensation	
Base metal thermocouples	-200 °C to +1300 °C	0.70 °C	Including cold junction compensation	
Resistance sensors	-200 °C to +850 °C	0.30 °C		
<b>PRESSURE</b>			Methods consistent with EURAMET CG3	
<u>Gas Pressure Gauge</u>				
Calibration of pressure indicating instruments and gauges	-95 kPa to -20 kPa -20 kPa to 20 kPa 20 kPa to 200 kPa 200 kPa to 2 MPa	0.012 % + 66Pa 0.015 % + 10 Pa 0.012 % + 66 Pa 0.012 % + 590 Pa		
<u>Gas Pressure Absolute</u>				
Calibration of pressure indicating instruments and gauges	3.5 kPa to 120 kPa 120 kPa to 200 kPa	100 Pa 140 Pa		
<u>Hydraulic Pressure Gauge</u>				
Calibration of pressure indicating instruments and gauges	0 Pa to 13.5 MPa 13.5 MPa to 70 MPa	0.010 % + 6.0 kPa 0.015 % + 20 kPa		
<b>TEMPERATURE</b>				
Resistance thermometers	50 °C to 300 °C 300 °C to 600 °C	0.50 °C 1.3 °C	Comparison in block calibrator	
Thermocouples	0 °C to 130 °C 130 °C to 350 °C 350 °C to 600 °C	0.60 °C 0.85 °C 1.5 °C		
Temperature indicators with probes	0 °C to 130 °C 130 °C to 350 °C 350 °C to 600 °C	0.50 °C 0.80 °C 1.4 °C		
<b>END</b>				



0331  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SGS United Kingdom Limited**  
Issue No: 047 Issue date: 20 July 2020

Calibration performed by the Organisation at the locations specified

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
  - As an explicit function of the measurand or of a parameter (see below).
  - As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
  - As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

**Expression of CMCs - symbols and units**

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10<sup>-6</sup>·p·10<sup>-6</sup>) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.