# **Schedule of Accreditation**

issued by

# **United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0339

Nuneaton

Accredited to ISO/IEC 17025:2017

## **Harris Gauges Limited**

Issue No: 023 Issue date: 05 September 2025

Unit 2 Contact: Mr P Bexon
Ptarmigan Place Tel: +44 (0)24 76326362

Attleborough Fields Industrial E-Mail: sales@harrisgauges.co.uk Website: www.harrisgauges.co.uk

CV11 6RX

## Calibration performed at the above address only

#### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks			
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED						
LENGTH						
Gauge Blocks		Class See Note 1				
Inch (Steel & Tungsten carbide)	As BS 4311-1:2007 0.01 inch to 0.4 inch. 0.4 inch to 1 inch 2 inch 3 inch 4 inch Variation in length of the	C 3.0 4.0 5.0 μ inches 6.0 7.0	Note 1. Class C uncertainties apply to the measurement of length of gauges by comparison with grade K standards of length of a similar material.			
Millimetre (Steel & Tungsten carbide)	above gauge blocks  As BS EN ISO 3650:1999 0.25 to 10 10 to 25 30, 40, 50 60, 70, 75 80, 90, 100  Variation in length of the above gauge blocks	<u>C</u> 0.080 0.10 0.12 0.15 0.18 0.03	Class C uncertainties apply to new and used grade 0, 1 and 2 gauges to BS 4311-1:2007 and BS EN ISO 3650:1999			
Plain plug gauges (parallel) Including pin gauges	0.1 to 50 diameter 50 to 100 100 to 150	1.0 on diameter 1.5 2.0	By comparision to reference standards using a length measuring machine			
Plain ring gauges (parallel) and setting standards	2 to 10 diameter 10 to 100 100 to 200	1.5 on diameter 1.0 2.0	By comparision to master ring gauges using a length measuring machine			
Length gauges, flat and spherical ended (excluding length bars)	25 to 600	1.0 + 8.0 x (length in m)	By comparision to reference standards			
Plain gap gauges (parallel)	0.5 to 100 100 to 200	3.0 5.0	By comparision to reference standards			
Vee blocks	20 to 150 diameter, vee capacity	2.5 to 5.0	As BS 3731:1987 By comparision datum surfaces			

Assessment Manager: DH Page 1 of 5



# **Schedule of Accreditation** issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

## **Harris Gauges Limited**

**Issue No: 023** Issue date: 05 September 2025

## Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks			
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED						
ANGLE						
Squares, Blade type	50 to 300 300 to 600	3.0 On squareness See Note 3 5.0	As BS 939:2007 By comparision to reference squares and datum surfaces			
Bevel protractors	0 to 360 degrees	6.0 minutes of arc 2.0 Flatness, Straightness and Parallelism: See note 3	As BS 1685:2008 By comparision to reference angles and datum surfaces			
MEASURING INSTRUMENTS ANI	D MACHINES					
Micrometers			By comparision to reference			
External	0 to 1200		standards As BS 870:2008			
Internal	0 to 1200	Heads 2.0 between any two points.	As BS 959:2008			
Depth	0 to 300	Setting and extension rods 1.0 + (8.0 x length in m)	As BS 6468:2008			
Micrometer heads	0 to 100	1.0	As BS 1734:1951 By comparision to reference standards			
Bore micrometers (three point)	3 to 10 diameter 10 to 50 diameter 50 to 100 diameter 100 to 150 diameter	3.0 2.5 3.0 4.0	By comparision to master setting ring gauges			
Bore indicators	2 to 150	Overall performance 5.0	By comparision to reference standards			
Internal and External Calliper Gauges	0 to 150	Overall performance 5.0	By comparision to reference standards			
Feeler Gauges	0.03 to 1.0	3.0	As BS 957:2008 Using a length measuring instrument			
Digital and dial gauges and digital and dial test indicators (Excluding LVDT type)	0 to 50 Discrimination	1.0 0.5	As BS 907:2008 and BS 2795:1981 and in-house method. By comparision to reference standards or length measuring machine			
Comparators (external)	250 to 20 000 magnifications	1.0 % of range Minimum 0.10	As BS 1054:1975 By comparision to reference standards			

Assessment Manager:DH Page 2 of 5



# Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Harris Gauges Limited

Issue No: 023 Issue date: 05 September 2025

## Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks			
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED						
MEASURING INSTRUMENTS AN	ND MACHINES (cont'd)					
Digital, dial and vernier type calliper gauges	0.01 to 50 50 to 200 200 to 400 400 to 600 600 to 800 800 to 1000  3 to 50 Shift error (S) internal jaws	1.0 See note 4 3.0 6.0 10 13 16	As BS EN ISO 13385-1:2019 By comparision to reference standards			
	3 to 50 Shift error (S) depth and step	2.5				
Vernier type gauges including dial and digital						
Calliper	0 to 1200	Overall performance 10 + (30 x length in m)	As BS 887:2008 By comparision to reference standards			
Height	0 to 1200	Overall performance 15 + (10 x length in m)	As ISO13225:2012 and BS 1643:2008 By comparision to reference standards and datum surfaces			
Depth	0 to 300	Overall performance 10 + (30 x length in m)	As BS 6365:2008 By comparision to reference standards and datum surfaces			
ANCILLERY MEASUREMENTS						
Flatness		0.2	Ancillery measurements made for completeness of calibration. Best CMC's are dependent on methodology and range.			
Parallelism		0.4				
Squareness		3.0				
Straightness		2.0				
Angular		30 seconds of arc				

Assessment Manager:DH Page 3 of 5



## Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

## **Harris Gauges Limited**

**Issue No: 023** Issue date: 05 September 2025

## Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks			
Notes						
Class C uncertainties apply to the measurement of length of gauges by comparison with grade K standards of length of a similar material.  Class C uncertainties apply to new and used grade 0, 1 and 2 gauges to BS 4311-1:2007 and BS EN ISO 3650:1999.						
2. All linear calibrations may be given in inch units.						
3. The uncertainty quoted is for the departure from flatness, straightness, or squareness, ie the distance separating the two parallel planes which just enclose the surface under consideration.						
4. The stated uncertainties have been calculated in accordance with the relevant standard method and relate to the test value uncertainty. The uncertainty quoted excludes contributions relating to the instrument under test.						
END						

Assessment Manager:DH Page 4 of 5



## Schedule of Accreditation issued by **United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

## Harris Gauges Limited

Issue No: 023 Issue date: 05 September 2025

#### Calibration performed at main address only

### Appendix - Calibration and Measurement Capabilities

#### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

#### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant nonrepeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

#### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means 1.5  $\times$  0.01  $\times$  q, where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets: Q[a, b] =  $[a^2 + b^2]^{1/2}$ 

Assessment Manager:DH Page 5 of 5