


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|--|---|
|  0375 Accredited to ISO/IEC 17025:2017 | SERCAL Materials Testing Machines Services Ltd | |
| | Issue No: 049 | Issue date: 11 March 2020 |
| | Southern Avenue Leominster Herefordshire HR6 0QH | Contact: Dr N Wrigley Tel: +44 (0)1527 514015 Fax: +44 (0)1527 514016 E-Mail: nigel.wrigley@sercalcalibrations.co.uk Website: www.sercalcalibrations.co.uk |
| Calibration performed by the Organisations at the locations specified below | | |

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details | Activity | Location code |
|---|--------------------------------------|----------------|
| Address Southern Avenue Leominster Herefordshire HR6 0QH | Local contact Dr N Wrigley | Force P |

Site activities performed away from the locations listed above:

| Location details | Activity | Location code |
|--|--------------------------------|----------------------------|
| Customer's sites or premises The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer | Contact Dr N Wrigley | Force Hardness S |



0375

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SERCAL Materials Testing Machines Services Ltd

Issue No: 049 Issue date: 11 March 2020

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|--|---|--|--|------------------|
| FORCE | | | NOTES | |
| UNIVERSAL MATERIALS TESTING MACHINES | | | | S |
| Verification and calibration of the force measuring system by force proving instruments in tension | 25 N to 600 kN for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018 | 0.20 % | 1. Calibration also includes the alignment and restraint of the upper machine platen required by BS EN 12390-4:2000. | |
| | From 50 N up to 2000 kN for Class 1, 2 and 3 machines to BS EN ISO 7500-1:2018 and ASTM E4-16 | 0.32 % | 2 The indirect verification shall be in accordance with the requirements of BS EN ISO 6508-2:2015 and ASTM E18-17. | |
| Verification and calibration of the force measuring system by force proving instruments in compression | 5 N to 600 kN for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018 | 0.20 % | 3 The indirect verification shall be in accordance with the requirements of BS EN ISO 6506-2:2014 ASTM E10-17. | |
| | 5 N to 16.5 MN for Class 1, 2 and 3 machines to BS EN ISO 7500-1:2018 and ASTM E4-16 | 0.32 % | 4 The indirect verification shall be in accordance with the requirements of BS EN ISO 6507-2:2005 and ASTM E92-17. | |
| Verification and calibration of the force measuring system by calibrated masses in tension | 0.01 N to 1000 N for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018 and ASTM E4-16 | 0.10 % | | |
| Verification and calibration of the force measuring system by calibrated masses in compression | 0.01 N to 1000 N for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018 and ASTM E4-16 | 0.10 % | | |
| FORCE MEASURING DEVICES | | | | P |
| Calibration of force measuring devices, eg, strain gauged load cells and load measuring rings (but excluding proving devices in) Tension and Compression | From 0,1 N up to 1 000 N | 0.10 % | | |
| | From 500 N up to 500 kN | 0.41 % | | |
| COMPRESSION TESTING MACHINES FOR CONCRETE | | | | S |
| Verification of concrete testing machines by proving devices in Compression | 100 kN to 16.5 MN for Class 1, 2 and 3 machines to BS EN ISO 7500-1:2018 | 0.32 % See note 1 | | |
| Rate of application of force (Pacer rate) | As BS EN 12390-2:2000 3 kN/min to 1300 kN/min | 2.25 % | | |
| Flatness of platens and spacing blocks | As BS EN 12390-4:2000 40 mm to 300 mm | 0.010 mm | | |



0375

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SERCAL Materials Testing Machines Services Ltd

Issue No: 049 Issue date: 11 March 2020

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|--|---|--|---------|------------------|
| TENSION CREEP TESTING MACHINES | | | | |
| Verification of the applied load using force proving instruments | 25 N to 500 kN for Class 0.5, 1 and 2 machines to BS EN ISO 7500-2:2006 and ASTM E4-16 | 0.20 % | | S |
| Verification of the applied load using masses | 0.01 N to 1000 N for Class 0.5, 1 and 2 machines to BS EN ISO 7500-2:2006 and ASTM E4-16 | 0.10 % | | |
| LENGTH | | | | |
| Extensometers | As BS EN ISO 9513:2012 for the following classes and gauge lengths: Class 0.2 from 25 mm Class 0.5 from 10 mm Class 1 from 5 mm Class 2 from 5 mm As ASTM:E83-16 for the following classes and gauge lengths: B-1 from 20 mm B-2 from 10 mm C from 5 mm Displacements 0.005 mm to 50 mm | | | S |
| Long Travel Extensometry | As BS EN ISO 5893:2019 Displacements 3 mm to 600 mm | 0.015 mm + 0.2 mm/m | | |
| Crosshead Rate | Timed between 30 seconds and 10 minutes | 0.15 seconds | | |
| Testing machine crosshead displacement and actuator displacement | 1 mm to 1200 mm | 0.011 mm + (0.13 mm per metre) | | |
| TORSION TESTING MACHINES | | | | |
| Torque | 4 N.m to 5000 N.m | 0.43 % | | S |
| Angle | 0° to 360° | 0.25° | | |



0375
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SERCAL Materials Testing Machines Services Ltd

Issue No: 049 Issue date: 11 March 2020

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|---|--|--|------------|------------------|
| IMPACT TESTING MACHINES Charpy Izod Plastics | Absorbed Energy (joules) 1 J to 600 J BS EN ISO 148-2:2016 ASTM E23-18 BS 131:Part 4:1972 BS ISO 13802:2015 | 0.70 J 0.11J | | S |
| VERIFICATION OF HARDNESS TESTING MACHINES Indirect verification of Rockwell Hardness Testing Machines | Rockwell scales: A, B, C, D, E, F, G, H, K, N and T HRA Scale 80 to 85 70 to 79 60 to 69 HRB Scale 80 51 to 79 10 to 50 HRC Scale 60 to 70 40 to 59 20 to 39 HRD Scale 70 to 80 50 to 69 40 to 49 HRE Scale 89 75 to 88 65 to 87 HRF Scale 87 70 to 86 40 to 69 HRG Scale 80 40 to 79 10 to 39 HRH Scale 90 80 to 89 60 to 79 | 0.15 HRA 0.16 HRA 0.28 HRA 0.42 HRB 0.87 HRB 1.36 HRB 0.31 HRC 0.32 HRC 0.37 HRC 0.17 HRD 0.25 HRD 0.27 HRD 0.54 HRE 0.54 HRE 0.54 HRE 0.40 HRF 0.40 HRF 0.54 HRF 0.30 HRG 0.30 HRG 0.76 HRG 0.40 HRH 0.40 HRH 0.68 HRH | See Note 2 | S |



0375
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SERCAL Materials Testing Machines Services Ltd

Issue No: 049 Issue date: 11 March 2020

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|--|--|---|-------------------|------------------|
| <p>VERIFICATION OF HARDNESS TESTING MACHINES</p> <p>Indirect verification of Rockwell Hardness Testing Machines (cont'd)</p> | <p>Rockwell scales:</p> <p>HRK Scale 70 30 to 69 10 to 29</p> <p>HR45N Scale 67 to 75 50 to 66 10 to 49</p> <p>HR45T Scale 50 to 75 40 to 49 10 to 39</p> <p>HR30N Scale 77 to 85 60 to 76 40 to 59</p> <p>HR30T Scale 57 to 85 50 to 56 20 to 49</p> <p>HR15N Scale 90 to 95 80 to 89 40 to 79</p> <p>HR15T Scale 88 to 100 80 to 87 20 to 79</p> | <p>0.40 HRK 0.40 HRK 0.64 HRK</p> <p>0.18 HR45N 0.21 HR45N 0.43 HR45N</p> <p>0.40 HR45T 0.40 HR45T 0.73 HR45T</p> <p>0.27 HR30N 0.27 HR30N 0.55 HR30N</p> <p>0.39 HR30T 0.66 HR30T 0.90 HR30T</p> <p>0.18 HR15N 0.18 HR15N 0.39 HR15N</p> <p>0.21 HR15T 0.21 HT15T 0.37 HR15T</p> | <p>See Note 2</p> | |
| <p>Indirect verification of Brinell Hardness Testing and Calibration machines</p> | <p>Brinell scales:</p> <p>Scale 10/3000 600HBW to 140 HBW</p> <p>Scale 10/1500 299 HBW to 55 HBW</p> <p>Scale 10/1000 169 HBW to 55 HBW</p> <p>Scale 5/750 600 HBW to 140 HBW</p> | <p>8.0 HBW to 2.2 HBW</p> <p>4.1 HBW to 1.2 HBW</p> <p>2.3 HBW to 1.2 HBW</p> <p>9.8 HBW to 2.4 HBW</p> | <p>See Note 3</p> | |



0375
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SERCAL Materials Testing Machines Services Ltd

Issue No: 049 Issue date: 11 March 2020

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|---|--|--|------------|------------------|
| VERIFICATION OF HARDNESS TESTING MACHINES | | | | |
| Indirect verification of Brinell Hardness Testing machines (cont'd) | Brinell scales: Scale 5/250 169 HBW to 55 HBW Scale 2.5/187.5 600 HBW to 140 HBW Scale 1/1 21.8 HBW to 3.18 HBW | 2.7 HBW to 1.3 HBW 16 HBW to 2.9 HBW 1.04 HBW to 0.09 HBW | See Note 3 | |
| Indirect verification of Vickers hardness testing machines | Vickers scales: HV 100 200 HV 100 400 HV 100 700 HV 50 200 HV 50 400 HV 50 700 HV 30 200 HV 30 400 HV 30 700 HV 20 200 HV 20 400 HV 20 700 HV 10 200 HV 10 400 HV 10 700 HV5 200 HV5 400 HV5 700 HV3 200 HV3 400 HV3 700 HV1 200 HV1 400 HV1 700 | 1.2 HV 3.4 HV 4.1 HV 1.9 HV 3.5 HV 6.3 HV 2.0 HV 4.4 HV 9.3 HV 2.5 HV 6.2 HV 11.0 HV 3.1 HV 7.7 HV 14.9 HV 3.9 HV 11.0 HV 19.7 HV 6.9 HV 16.3 HV 31.0 HV 8.7 HV 21.4 HV 44.0 HV | See Note 4 | |



0375
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SERCAL Materials Testing Machines Services Ltd
Issue No: 049 Issue date: 11 March 2020

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|---|---|--|------------|------------------|
| VERIFICATION OF HARDNESS TESTING MACHINES | | | | |
| Indirect verification of Vickers hardness testing machines (cont'd) | Vickers Scales: HV 0.5 200 HV 0.5 400 HV 0.5 700 HV 0.3 200 HV 0.3 400 HV 0.3 700 HV 0.2 200 HV 0.2 400 HV 0.2 700 HV 0.1 200 HV 0.1 400 HV 0.1 700 HV 0.05 200 HV 0.05 400 HV 0.05 700 HV 0.025 200 HV 0.025 400 HV 0.025 700 HV 0.01 200 HV 0.01 400 HV 0.01 700 | 5.0 HV 15.0 HV 17.0 HV 6.0 HV 16.0 HV 19.0 HV 7.0 HV 17.0 HV 20.0 HV 10.0 HV 30.0 HV 40.0 HV 8.5 HV 19.0 HV 27.0 HV 9.0 HV 20.0 HV 30.0 HV 10.0 HV 30.0 HV 40.0 HV | See Note 4 | |
| END | | | | |



0375

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SERCAL Materials Testing Machines Services Ltd

Issue No: 049 Issue date: 11 March 2020

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or*
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.*

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples, and an indication of how they are to be interpreted, are shown below.

DC voltage, 100 mV to 1 V: $0.0025 \% + 5.0 \mu\text{V}$:

Over the range 100 mV to 1 V, the CMC is $0.0025 \% \cdot V + 5.0 \mu\text{V}$, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: $0.0036 \% + 0.12 \text{ ppm/MPa} + 4.0 \text{ Pa}$

Over the range 0.5 MPa to 140 MPa, the CMC is $0.0036 \% \cdot p + (0.12 \cdot 10^{-6} \cdot p \cdot 10^{-6}) + 4.0 \text{ Pa}$, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means $1.5 \cdot 0.01 \cdot i$, where i is the instrument indication.