


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|---|--|---|
|  <p>0394</p> <p>Accredited to ISO/IEC 17025:2017</p> | <p>Trescal Limited (Trescal EMS – Rolls-Royce)</p> <p>Issue No: 062 Issue date: 26 May 2022</p> | |
| | <p>Trescal EMS</p> <p>Unit 2, Riverside Road</p> <p>Pride Park</p> <p>Derby</p> <p>DE24 8HY</p> | <p>Contact: Matt Gypps</p> <p>Tel: +44 (0) 1942 761226</p> <p>Fax: +44 (0) 2476 623626</p> <p>E-Mail: matt.gypps@trescal.com</p> <p>Website: www.trescal.com</p> |

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details | Activity | Location code |
|--|---|---------------|
| <p>Address (Pride Park)</p> <p>Trescal EMS Unit 2, Riverside Road Pride Park Derby DE24 8HY</p> <p>Local contact</p> <p>Trevor Smith</p> <p>Tel: +44 (0) 1332 238102 Email: Trevor.smith@trescal.com</p> | <p>Dimensional</p> <p>Electrical</p> <p>Humidity</p> <p>Temperature</p> <p>Torque</p> | Pride Park |
| <p>Address (Ansty)</p> <p>Trescal EMS – Rolls-Royce Standards Room Building 6 Ansty Coventry CV7 9JR</p> <p>Local contact</p> <p>David Williams</p> <p>Tel: +44 (0) 2476 623625 Fax: +44 (0) 2476 623626 Email: David.williams2@rolls-royce.com</p> | <p>Torque</p> <p>Pressure</p> | Ansty |
| <p>Address (Inchinnan)</p> <p>Trescal EMS – Rolls-Royce Inchinnan Drive Inchinnan Renfrewshire PA4 9AF</p> <p>Local contact</p> <p>Robert Simpson</p> <p>Tel: +44 (0) 141 626 8540 Email: Robert.simpson@trescal.com</p> | <p>Dimensional</p> <p>Torque</p> | Inchinnan |
| <p>Address (Washington)</p> <p>Trescal EMS – Rolls-Royce Calibration Laboratory Radial Park Road Washington Tyne and Wear NE38 9DA</p> <p>Local contact</p> <p>Robert Simpson Steve Jones</p> <p>Tel: +44 (0) 191 297 3023 Email: Robert.simpson@trescal.com</p> | <p>Dimensional</p> <p>Torque</p> | Washington |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)

Issue No: 062 **Issue date:** 26 May 2022

Calibration performed by the Organisation at the locations specified

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details | | Activity | Location code |
|--|--|---|---------------|
| Address (Bristol) Trescal EMS – Rolls-Royce Metrology Laboratory (EW6/7) PO Box 3 Filton Bristol BS34 7QE | Local contact Mr M Viney Tel: +44 (0) 117 979 6099 Fax: +44 (0) 117 979 5038 Email: michael.viney@rolls-royce.com | Fuel Flow Torque | Bristol |
| Address (Solihull) Rolls-Royce Derwent Building 5000 Solihull Parkway Birmingham Business Park Birmingham B37 7YP | Local contact Jim Attwooll Tel +44 (0) 121 2732781 Email: jim.attwooll@rolls-royce.com | Electrical DC&LF Dimensional | Solihull |

Site activities performed away from the locations listed above:

| | | | |
|--|---|---|------|
| All Rolls-Royce sites: The site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. | Local contact Trevor Smith Tel: +44 (0) 1332 238102 Email: Trevor.smith@trescal.com | Form Electrical Temperature | Site |
|--|---|---|------|



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|--|------------------|
| LENGTH | | | | |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | | | | |
| Thread measuring cylinders | BS3777:1964 and BS 5590:1978 and specials 0.1 to 5.0 diameter | 0.50 on diameter | NOTES 1 In addition to all items in the first column, other similar items, including parts of measuring instruments and machines, may be calibrated in accordance with the stated CMCs. Where the item or part calibrated is of lower quality due to wear, errors in geometry or form, or poor surface texture, or where any other factor adversely affects the measurement capability, greater uncertainties will be quoted. 2 The uncertainty quoted is for the departure from flatness, straightness, or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration. 3 All linear calibrations may be given in inch units. 4 Single start symmetrical thread forms only. 5 Single start symmetrical thread forms only. 6 By comparison with end standards using a length measuring machine. | Pride Park |
| Plain plug gauges (parallel), cylindrical setting standards, gear measuring cylinders and rollers. See Note 6 | 1 to 50 diameter 50 to 100 100 to 150 150 to 200 200 to 300 | 0.50 0.80 1.0 1.2 1.6 | | |
| Plain ring gauges (parallel) and setting standards | CCP 2.3.2, issue 11 1 to 50 diameter 50 to 100 diameter 100 to 150 diameter 150 to 200 diameter | 0.80 1.2 1.8 2.5 | | |
| Length gauges, flat and spherical ended See Note 6 | 0 m to 3 m | 1.0 + (5.0 x length in m) | | |
| Length bars Inspection and workshop grades 1 and 2 | BS 1790:1961 BS 5317:1976 | 0.45 + (1.1 x length in m) | | |
| Plain gap gauges (parallel) | BS 969:2008 0.5 to 100 100 to 200 200 to 300 | 3.0 5.0 8.0 | | |
| Screw plug gauges (parallel) including check and setting plugs See Notes 5 and 6 | 1 to 100 diameter 100 to 300 diameter | 2.5 5.0 | | |
| Screw ring gauges (parallel) See Notes 4 and 6 | 5 to 75 diameter 100 to 150 diameter 150 to 300 diameter | 4.0 5.0 8.0 | | |
| Screw pitch | 0.2 to 8 | 1.5 | | |
| Screw flank angle | 0° to 50° | 5.0 minutes of arc | | |
| Parallels | BS 906:Parts 1 and 2:1992 5 to (50 x 100 x 400) | 1.5 to 5.0 | Using a length measuring machine. | |
| | | | Using a projector. | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|---|------------------|
| LENGTH (continued) | | | | |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | | | | |
| Gauge blocks | | Class (see note) | Note | |
| Inch (Steel) | BS 4311-1:2007 0 in to 0.4 in 0.4 in to 1 in Size 2 in 3 in 4 in | C D 3.0 μ in 4.0 μ in 4.0 μ in 5.0 μ in 5.0 μ in 6.0 μ in 7.0 μ in | Class C uncertainties apply to the measurement of length by comparison with grade K standards of a similar material. Class D uncertainties apply to the measurement of length by comparison with grade K standards of a dissimilar material. | |
| Millimetre (Steel) | BS EN ISO 3650:1999 0 to 10 10 to 25 Size 30, 40, 50 60, 70, 75 80, 90, 100 | C D 0.080 0.10 0.10 0.13 0.12 0.15 0.18 | The uncertainties apply to new and used grade 0, 1 and 2 gauges to BS EN ISO 3650:1999 and BS 4311-1:2007. | |
| Vee blocks | BS 3731:1987 20 to 150 diameter, Vee capacity | 2.5 to 5.0 | | |
| Receiver, position and profile gauges, jigs, fixtures | 1500 x 750 x 750 1500 x 3200 x 1100 | From first principles: Dependant on size and features Minimum per co-ordinate: 3.0 + (10 x length in m) Using CMM: Dependant on size and features Minimum per co-ordinate: 5.0 + (10 x length in m) | | |
| ANGLE | | | | |
| Squares | | | | |
| Blade type | BS 939:2007, CCP 2.4.17 issue 10 50 to 300 300 to 600 | 3.0 5.0 | | |
| Cylindrical | BS 939:2007, CCP 2.4.17 issue 10 75 to 300 300 to 600 | 2.0 On squareness 4.0 See Note 2 | | |

Pride Park



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty (<i>k</i> = 2) | Remarks | Location Code |
|--|---|--|---|------------------|
| ANGLE (continued) | | | | Pride Park |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | | | | |
| Block | BS 939:2007 50 to 300 300 to 600 | 3.0 5.0 | In-house methods based on MOY/SCMI/18 | |
| Angle gauges | NPL type Other types | 2.0 seconds of arc 3.0 seconds of arc | | |
| Sine bars and tables | BS 3064:1978 100 to 500 length | Linear dimensions: 1.0 + (10 x length in m) Overall performance: 3.0 seconds of arc | | |
| Sine centres | 100 to 500 length or between centres | Linear dimensions: 1.0+ (10 x length in m) Overall performance | In-house methods based on BS 3064:1978 | |
| Compound sine tables | 100 to 500 length | 5.0 seconds of arc | | |
| FORM | | | | |
| Straightedges Cast iron Steel Granite | BS 5204:Part 1:1975 and BS 5204:Part 2:1977 0 m to 2m | 1.0 + (2.0 x length in m) See Note 2 | By comparison with end standards using a length measuring machine. | |
| Roundness External Internal | BS 3730:Part 2:1982 0 to 350 diameter 3 to 350 diameter | 0.050 on radius 0.050 on radius | | |
| Steel balls | 1 to 25 diameter | 0.50 on diameter | | |
| MEASURING INSTRUMENTS AND MACHINES | | | | |
| Micrometers | | | | |
| External | BS 870:2008, CCP 2.4.1 issue 12 0 to 600 | Heads: 2.0 Setting and Extension rods: 1.0 + (5.0 x length in m) | | |
| Internal | BS 959:2008 0 to 300 | | | |
| Depth | BS 6468:2008 0 to 300 | | | |
| Micrometer heads | BS 1734:1951; 0 to 100 | 1.0 | In-house method based on MOY/SCMI/22 | |
| Bench micrometer | 0 to 100 | Overall performance 1.0 | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty (<i>k</i> = 2) | Remarks | Location Code |
|--|---|--|---|------------------|
| MEASURING INSTRUMENTS AND MACHINES (continued) | | | | Pride Park |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | | | | |
| Height setting micrometer | 0 to 300 | Heads 1.0 Stepped column 1.6 Overall performance 2.0 | By comparison with end standards. | |
| Riser Blocks | 150 300 | 1.6 1.7 | By comparison with end standards. | |
| Height gauges - (Simple) including vernier, dial and digital types | BS EN ISO 13225:2012 0 to 300 | 4.0 | | |
| Vernier gauges Caliper Height | BS 887:2008 BS 1643:2008 0 to 1200 | Overall performance: 10 + (30 x length in m) | | |
| Depth | BS 6365:2008 0 to 600 | | | |
| Dial gauges and dial test indicators | BS 907:2008 and BS 2795:1981 0 to 50 | 1.0 | | |
| Spirit levels | BS 958:1968 and BS 3509:1962 Nominal sensitivity 5 seconds of arc to 60 minutes of arc | Mean sensitivity: 10 % of nominal; minimum 0.50 seconds of arc | | |
| Clinometers | 0° to 360° | 10 seconds of arc | In-house method based on MOY/SCMI/36 | |
| Levels, electronic | 0 seconds of arc to 10 minutes of arc | 1.0 % of range minimum 0.50 seconds of arc | The quoted uncertainty will be particularly dependent on the sensitivity of the device. Using small angle generator. | |
| Orifice plates | BS EN ISO 5167-2:2003 (and similar devices) Bore d diameter 1.0 mm to 1 m | 4.0 + (6.0 x length in m) | | |
| TORQUE | | | | |
| Hand torque tools (excluding torque screwdrivers) | BS EN ISO 6789:2017 And BS EN ISO 6789:2003 (withdrawn and superseded) and CCP 3.6.6 Issue 9.0 1.0 N·m to 1000 N·m 2.0 | 1.0 % | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test. | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty (<i>k</i> = 2) | Remarks | Location Code |
|--|---|---|--|------------------|
| ELECTRICAL MEASUREMENTS | | | | |
| DC VOLTAGE | | | | Pride Park |
| Measurement | Up to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1 000 V | 6.5 μV/V + 1.3 μV 5.1 μV/V 6.1 μV/V 9.4 μV/V 9.6 μV/V | | |
| Generation | 0 mV to 2 mV 2 mV to 20 mV 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1100 V | 1.3 μV 1.3 μV 1.3 μV 2.8 μV/V + 0.90 μV 2.2 μV/V + 2.5 μV 3.2 μV/V + 39 μV 5.6 μV/V + 0.39 mV | | |
| DC RESISTANCE | | | | |
| Measurement | 0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 2 GΩ | 30 μΩ/Ω + 20 μΩ 13 μΩ/Ω 14 μΩ/Ω 24 μΩ/Ω 55 μΩ/Ω 450 μΩ/Ω 0.50% | The stated CMCs are for a four-terminal configuration and may be increased if a two-terminal configuration is necessary. | |
| Generation | | | | |
| Four terminal configuration | 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ | 5.7 μΩ/Ω 3.9 μΩ/Ω 3.6 μΩ/Ω 3.2 μΩ/Ω 4.5 μΩ/Ω 10 μΩ/Ω 19 μΩ/Ω 65 μΩ/Ω | | |
| Two terminal configuration | 0 Ω, 10 Ω and 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ | 10 mΩ 79 μΩ/Ω 8.3 μΩ/Ω 4.5 μΩ/Ω 10 μΩ/Ω 19 μΩ/Ω 65 μΩ/Ω | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---------|------------------|
| DC CURRENT | | | | Pride Park |
| Measurement | 10 μ A to 200 μ A 200 μ A to 200 mA 200 mA to 2 A | 100 μ A/A 100 μ A/A 170 μ A/A | | |
| Generation | 10 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A | 22 μ A/A + 1.6 nA 15 μ A/A + 7.8 nA 15 μ A/A + 78 nA 15 μ A/A + 0.78 μ A 26 μ A/A + 16 μ A | | |
| AC VOLTAGE | | | | |
| Measurement | 10 mV to 200 mV 10 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz | 390 μ V/V 640 μ V/V 0.17% | | |
| | 200 mV to 2 V 10 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz | 190 μ V/V 270 μ V/V 870 μ V/V | | |
| | 2 V to 20 V 10 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz | 180 μ V/V 270 μ V/V 870 μ V/V | | |
| | 20 V to 200 V 10 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz | 190 μ V/V 270 μ V/V 870 μ V/V | | |
| | 200 V to 300 V 40 Hz to 10 kHz 10 kHz to 30 kHz | 250 μ V/V 390 μ V/V | | |
| | 300 V to 1 kV 40 Hz to 10 kHz 10 kHz to 30 kHz | 0.11 % 0.12 % | | |
| | 200 V to 1 kV 30 kHz to 50 kHz | 0.20 % | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|---|---------|------------------|
| AC VOLTAGE (continued) | | | | Pride Park |
| Generation | 1 mV to 2 mV 20 Hz to 100 kHz | 0.74% + 4.2 μ V | | |
| | 2 mV to 20 mV 20 Hz to 100 kHz | 0.032% + 4.2 μ V | | |
| | 20 mV to 200 mV 20 Hz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz | 130 μ V/V + 7.0 μ V 0.044% 0.17% 0.83% | | |
| | 200 mV to 2 V 10 Hz to 20 Hz 20 Hz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz | 310 μ V/V 89 μ V/V 180 μ V/V 0.13 % 0.52% | | |
| | 2 V to 20 V 10 Hz to 20 Hz 20 Hz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz | 290 μ V/V 73 μ V/V 240 μ V/V 0.12% 0.52% | | |
| | 20 V to 200 V 10 Hz to 20 Hz 20 Hz to 50 kHz 50 kHz to 100 kHz | 220 μ V/V 100 μ V/V 200 μ V/V | | |
| | 200 V to 1 kV 45 Hz to 33 kHz | 130 μ V/V | | |
| AC CURRENT | | | | |
| Measurement | 40 Hz to 1 kHz: 10 μ A to 200 μ A 200 μ A to 200 mA 200 mA to 2 A | 370 μ A/A + 16 nA 840 μ A/A 660 μ A/A + 310 μ A | | |
| Generation | 40 Hz to 1 kHz: 10 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A | 90 μ A/A + 7.8 nA 85 μ A/A + 78 nA 85 μ A/A + 0.78 μ A 110 μ A/A + 7.8 μ A 370 μ A/A + 78 μ A | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|--|------------------|
| FREQUENCY | | | | Pride Park |
| Specific Values | 1 MHz and 10 MHz | 1.2 parts in 10^9 | For calibrating oscillators | |
| Other Values | 0.1 Hz to 1 Hz 1 Hz to 10 Hz 10 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 60 MHz 60 MHz to 100 MHz 100 MHz to 150 MHz 150 MHz to 500 MHz | 1.5 parts in 10^3 1.5 parts in 10^4 1.5 parts in 10^5 1.5 parts in 10^6 1.5 parts in 10^7 1.7 parts in 10^8 3.9 parts in 10^9 2.5 parts in 10^9 1.2 parts in 10^9 2.4 parts in 10^9 1.4 parts in 10^9 | Measurement capability only above 60 MHz | |
| ELAPSED TIME | | | | |
| Stop watches (mechanical and electronic) | ± 0.5 s error / 24 hours ± 2.0 s error / 24 hours 10 s to 24 hours | 0.062 s 0.090 s 0.41 s | Time reference measurement per 24 hour period per 24 hour period Real time measurement | |
| TEMPERATURE SIMULATION | | | | |
| Temperature indicators and simulators (thermocouple type), calibration by electrical simulation | | | | |
| Base metal thermocouples | Type J, -210 °C to 0 °C Type J, 0 °C to 1200 °C Type K, -270 °C to -200 °C Type K, -200 °C to 0 °C Type K, 0 °C to 1370 °C Type N, -270 °C to -200 °C Type N, -200 °C to 0 °C Type N, 0 °C to 1300 °C Type T, -270 °C to -200 °C Type T, -200 °C to 0 °C Type T, 0 °C to 400 °C | 0.064 °C 0.018 °C 0.23 °C 0.070 °C 0.022 °C 0.62 °C 0.084 °C 0.027 °C 0.19 °C 0.070 °C 0.020 °C | excluding cold junction compensation excluding cold junction compensation excluding cold junction compensation excluding cold junction compensation | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---|------------------|
| Temperature indicators and simulators (thermocouple type), calibration by electrical simulation (continued) | | | | Pride Park |
| Cold junction compensation | At ambient temperature of 20 °C ± 2.0 °C | 0.13 °C | | |
| Base metal thermocouples | Type J, -210 °C to 0 °C | 0.14 °C | | |
| | Type J, 0 °C to 1200 °C | 0.13 °C | including cold junction compensation | |
| | Type K, -270 °C to -200 °C | 0.24 °C | | |
| | Type K, -200 °C to 0 °C | 0.15 °C | including cold junction compensation | |
| | Type K, 0 °C to 1370 °C | 0.13 °C | | |
| | Type N, -270 °C to -200 °C | 0.53 °C | | |
| | Type N, -200 °C to 0 °C | 0.15 °C | including cold junction compensation | |
| | Type N, 0 °C to 1300 °C | 0.13 °C | | |
| | Type T, -270 °C to -200 °C | 0.21 °C | | |
| | Type T, -200 °C to 0 °C | 0.15 °C | including cold junction compensation | |
| | Type T, 0 °C to 400 °C | 0.13 °C | | |
| Noble metal thermocouples | -50 °C to 0 °C | 0.19 °C | | |
| | 0 °C to 250 °C | 0.17 °C | excluding cold junction compensation | |
| | 250 °C to 1760 °C | 0.089 °C | | |
| Cold junction compensation | At ambient temperature of 20 °C ± 2 °C | 0.17 °C | | |
| Temperature indicators and simulators (thermocouple type), calibration by electrical simulation | | | | |
| Noble metal thermocouples | -50 °C to 0 °C | 0.24 °C | | |
| | 0 °C to 250 °C | 0.22 °C | including cold junction compensation | |
| | 250 °C to 1760 °C | 0.18 °C | | |
| PRT simulation (Pt 100) | -200 °C to 0 °C | 0.017 °C | | |
| | 0 °C to 100 °C | 0.018 °C | | |
| | 100 °C to 400 °C | 0.020 °C | | |
| | 400 °C to 630 °C | 0.023 °C | | |
| | 630 °C to 850 °C | 0.026 °C | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|--|------------------|
| TEMPERATURE | | | | Pride Park |
| Thermocouples | | | | |
| Base metal | -20 °C to +200 °C | 0.45 °C | Calibration within both liquid and metal medium | |
| Noble metal | -20 °C to 200 °C | 0.92 °C | Calibration within both liquid and metal medium | |
| Resistance thermometers | -20 °C to +200 °C | 0.070 °C | Calibration within both liquid and metal medium | |
| Electronic thermometers with sensors; analogue or digital | Ranges as per sensor | As per sensor type | Calibration within both liquid and metal medium | |
| HUMIDITY | | | | |
| Dew point | -10 °C to 0 °C 0 °C to 20 °C | 0.12 °C dp 0.10 °C dp | By comparison with dew-point hygrometer and Platinum Resistance Thermometers | |
| Relative Humidity | 5 %rh to 95 %rh | 2.0 %rh | At air temperature 5 °C to 60 °C | Ansty |
| Air Temperature | 5 °C to 60 °C | 0.4 °C | | |
| PRESSURE | | | Methods consistent with EURAMET CG17 | |
| Hydraulic pressure (Gauge) | | | | |
| Pressure indicating instruments and gauges | 600 kPa to 120 MPa | 0.010 % | Calibration of pressure measuring devices with an electrical output may be undertaken. | |
| Pneumatic pressure (Gauge) | | | | |
| Pressure indicating instruments and gauges | 3.70 kPa to 3.5 MPa | 0.010 % | | |
| Pneumatic pressure (Absolute) | | | | |
| Pressure indicating instruments and gauges | 3.70 kPa to 3.5 MPa 75 kPa to 120 kPa | 0.010 % + 5.0 Pa 17 Pa | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|--|------------------|
| TORQUE | CCP 3.6.6 issue 9.0 0.113 N·m to 1356 N·m | 1.0 % | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test. | Ansy |
| Hand torque tools | | | | |
| LENGTH | RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | 0.50 on diameter | By comparison with end standards using a length measuring machine. | Inchman |
| Thread measuring cylinders | | | | |
| Plain plug gauges (parallel), cylindrical setting standards, gear measuring cylinders and rollers | | | | |
| Plain ring gauges (parallel) and setting standards | | | | |
| Length gauges, flat and spherical ended | | | | |
| Plain gap gauges (parallel) | | | | |
| Screw plug gauges (parallel) excluding check and setting plugs | | | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|---------|------------------|
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | | | | |
| MEASURING INSTRUMENTS AND MACHINES | | | | |
| LENGTH | | | | |
| Micrometers | | | | |
| External | BS 870:2008, CCP 2.4.1 issue12 0 to 300 | Heads: 2.0 Setting and Extension rods: 1.0 + (5.0 x length in m) | | |
| Internal | BS 959:2008 0 to 300 | | | |
| Depth | BS 6468:2008 0 to 300 | | | |
| Vernier gauges | | | | |
| Caliper | BS 887:2008 0 to 300 | Overall performance: 10 + (30 x length in m) | | |
| Height | BS 1643:2008 0 to 300 | | | |
| Depth | BS 6365:2008 0 to 300 | | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty (<i>k</i> = 2) | Remarks | Location Code |
|--|---|---|---|------------------|
| MEASURING INSTRUMENTS AND MACHINES (continued) | | | | Inchman |
| LENGTH (continued) | | | | |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | | | | |
| Dial gauges and dial test indicators | BS 907:2008 and BS 2795:1981 0 to 50 | 1.0 <div></div> | | |
| TORQUE | | | | |
| Hand torque tools | CCP 3.6.6 issue 9.0 0.136 N·m to 677.91 N·m | 1.0 % | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test | |
| LENGTH | | | | Washington |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | | | | |
| Micrometers External | BS 870:2008, CCP 2.4.1 issue12 0 to 600 | Heads: 2.0 Setting and Extension rods: 1.0 + (5.0 x length in m) | | |
| Internal Depth | BS 959:2008; 0 to 150 BS 6468:2008; 0 to 150 | | | |
| Vernier gauges Caliper Depth | BS 887:2008; 0 to 600 BS 6365:2008; 0 to 150 | Overall performance: 10 + (30 x length in m) | | |
| Dial gauges and dial test indicators | BS 907:2008 and BS 2795:1981 0 to 50 | 1.5 | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|---|------------------|
| TORQUE | | | | Washington |
| Hand torque tools | CCP 3.6.6 issue 9.0 0.1 N·m to 1000 N·m | 1.0 % | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test | |
| FUEL FLOW | | | Piston prover method | Bristol |
| Flow rate - volume Flow rate - mass | 5 l/hr to 27000 l/hr 4 kg/hr to 21330 kg/hr | 0.10 % 0.20 % | Calibration fluid AVTUR (Aviation fuel) | |
| TORQUE | | | | |
| Hand torque tools | CCP 3.6.6 issue 9.0 0.1 N·m to 1000 N·m | 1.0 % | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test | |
| ELECTRICAL MEASUREMENTS | | | | Southall |
| DC RESISTANCE Measurement | 0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 1 G Ω | 28 $\mu\Omega/\Omega + 25 \mu\Omega$ 16 $\mu\Omega/\Omega + 100 \mu\Omega$ 13 $\mu\Omega/\Omega + 1.0 m\Omega$ 13 $\mu\Omega/\Omega + 10 m\Omega$ 16 $\mu\Omega/\Omega + 100 m\Omega$ 27 $\mu\Omega/\Omega + 2.0 \Omega$ 75 $\mu\Omega/\Omega + 100 \Omega$ 500 $\mu\Omega/\Omega + 12 k\Omega$ 1.0 % + 1.1 M Ω | | |
| DC VOLTAGE Measurement | 0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1 kV | 11 $\mu V/V + 1.2 \mu V$ 8.5 $\mu V/V + 0.9 \mu V$ 8.5 $\mu V/V + 4.0 \mu V$ 13 $\mu V/V + 60 \mu V$ 13 $\mu V/V + 600 \mu V$ | | |
| DC CURRENT Measurement | 0 μA to 200 μA 200 μA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A | 140 $\mu A/A + 0.60 nA$ 130 $\mu A/A + 6.0 nA$ 130 $\mu A/A + 60 nA$ 130 $\mu A/A + 1.3 \mu A$ 240 $\mu A/A + 25 \mu A$ | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location |
|--|---|--|---------|----------|
| AC VOLTAGE Measurement | 10 mV to 200 mV 40 Hz to 10 kHz | 320 $\mu\text{V/V} + 5.0 \mu\text{V}$ | | Solihull |
| | 200 mV to 2 V 40 Hz to 10 kHz | 210 $\mu\text{V/V} + 25 \mu\text{V}$ | | |
| | 2 V to 20 V 40 Hz to 10 kHz | 210 $\mu\text{V/V} + 250 \mu\text{V}$ | | |
| | 20 V to 200 V 40 Hz to 10 kHz | 210 $\mu\text{V/V} + 2.5 \text{ mV}$ | | |
| | 200 V to 1 kV 55 Hz to 1 kHz 1 kHz to 10 kHz | 360 $\mu\text{V/V} + 50 \text{ mV}$ 450 $\mu\text{V/V} + 50 \text{ mV}$ | | |
| AC CURRENT Measurement | 10 μA to 200 μA 55 Hz to 1 kHz | 600 $\mu\text{A/A} + 25 \text{ nA}$ | | |
| | 200 μA to 2 mA 55 Hz to 1 kHz | 400 $\mu\text{A/A} + 250 \text{ nA}$ | | |
| | 2 mA to 20 mA 55 Hz to 1 kHz | 400 $\mu\text{A/A} + 2.5 \mu\text{A}$ | | |
| | 20 mA to 200 mA 55 Hz to 1 kHz | 400 $\mu\text{A/A} + 25 \mu\text{A}$ | | |
| | 200 mA to 2 A 55 Hz to 1 kHz | 900 $\mu\text{A/A} + 500 \mu\text{A}$ | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location |
|--|--|---|---------|----------|
| DC RESISTANCE Generation | | | | Solihull |
| Specific Values | 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω | 35 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 18 $\mu\Omega/\Omega$ 80 $\mu\Omega/\Omega$ 180 $\mu\Omega/\Omega$ | | |
| Other Values | 0 Ω to 11 Ω 11 Ω to 33 Ω 33 Ω to 110 Ω 110 Ω to 330 Ω 330 Ω to 1.1 k Ω 1.1 k Ω to 3.3 k Ω 3.3 k Ω to 11 k Ω 11 k Ω to 33 k Ω 33 k Ω to 110 k Ω 110 k Ω to 330 k Ω 330 k Ω to 1.1 M Ω 1.1 M Ω to 3.3 M Ω 3.3 M Ω to 11 M Ω 11 M Ω to 33 M Ω 33 M Ω to 110 M Ω 110 M Ω to 330 M Ω | 180 $\mu\Omega/\Omega + 11$ m Ω 150 $\mu\Omega/\Omega + 19$ m Ω 110 $\mu\Omega/\Omega + 19$ m Ω 110 $\mu\Omega/\Omega + 19$ m Ω 110 $\mu\Omega/\Omega + 90$ m Ω 110 $\mu\Omega/\Omega + 90$ m Ω 110 $\mu\Omega/\Omega + 900$ m Ω 110 $\mu\Omega/\Omega + 900$ m Ω 140 $\mu\Omega/\Omega + 9.0$ Ω 150 $\mu\Omega/\Omega + 9.0$ Ω 180 $\mu\Omega/\Omega + 80$ Ω 200 $\mu\Omega/\Omega + 80$ Ω 710 $\mu\Omega/\Omega + 800$ Ω 0.14 % + 800 Ω 0.60 % + 8.0 k Ω 0.60 % + 21 k Ω | | |
| DC VOLTAGE Generation | 0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1 kV | 12 $\mu\text{V/V} + 1.0$ μV 7.5 $\mu\text{V/V} + 1.5$ μV 6.0 $\mu\text{V/V} + 5.0$ μV 8.0 $\mu\text{V/V} + 70$ μV 10 $\mu\text{V/V} + 700$ μV | | |
| DC CURRENT Generation | 0 μA to 220 μA 220 μA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A | 70 $\mu\text{A/A} + 10$ nA 60 $\mu\text{A/A} + 12$ nA 60 $\mu\text{A/A} + 120$ nA 70 $\mu\text{A/A} + 1.2$ μA 100 $\mu\text{A/A} + 35$ μA 710 $\mu\text{A/A} + 510$ μA | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty (<i>k</i> = 2) | Remarks | Loca tion |
|--|--|---|---------|--------------|
| AC VOLTAGE Generation | <i>40 Hz to 10 kHz</i> 0.22 mV to 2.2 mV 2.2 mV to 22 mV 22 mV to 220 mV 220 mV to 2.2 V 2.2 V to 22 V 22 V to 220 V <i>55 Hz to 1 kHz</i> 220 V to 1 kV | 700 μV/V + 6.0 μV 230 μV/V + 7.0 μV 140 μV/V + 10 μV 100 μV/V + 14 μV 100 μV/V + 130 μV 110 μV/V + 1.5 mV 120 μV/V + 8.0 mV | | Solihull |
| AC CURRENT Generation | <i>55 Hz to 1 kHz</i> 10 μA to 220 μA 220 μA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A | 260 μA/A + 20 nA 250 μA/A + 55 nA 200 μA/A + 550 nA 200 μA/A + 5.5 μA 800 μA/A + 55 μA | | |
| MEASURING INSTRUMENTS AND MACHINES | | | | |
| Micrometers | | | | |
| External Depth | As BS 870:2008 and above As BS 6468:2008 | Heads: 2.0 between any two points Setting and extension rods: 1.0 + 5.0 x length in m | | |
| Vernier gauges Caliper Height Depth | As BS 887:2008 As BS 1643:2008 As BS 6365:2008 | Overall performance: 10 + (30 x length in m) | | |
| Dial gauges and dial test indicators | As BS 907:2008 and BS 2795:1981 | 1.0 | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---|------------------|
| FORM | | | | |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED | | | | |
| Surface plates Granite Cast iron | As BS 817:2008 160 x 100 to 4 m x 4 m | 1.5 + (0.80 x diagonal in m) See Note 2 | | |
| ELECTRICAL | | | | |
| Temperature indicators and simulators (thermocouple type), calibration by electrical simulation: | | | Internal Reference junction enabled. Ambient temperature range 18 °C to 22 °C (controlled customer environment). | |
| Base metal thermocouple types | Type J, -210 °C to 0 °C Type J, 0 °C to 1200 °C Type K, -270 °C to -200 °C Type K, -200 °C to 0 °C Type K, 0 °C to 1000 °C Type K, 1000 °C to 1370 °C Type N, -270 °C to -200 °C Type N, -200 °C to -100 °C Type N, -100 °C to 0 °C Type N, 0 °C to 800 °C Type N, 800 °C to 1300 °C Type T, -270 °C to -200 °C Type T, -200 °C to 0 °C Type T, 0 °C to 400 °C | 0.36 °C 0.28 °C 4.6 °C 0.37 °C 0.29 °C 0.27 °C 1.9 °C 0.49 °C 0.34 °C 0.26 °C 0.24 °C 0.81 °C 0.36 °C 0.26 °C | | |
| Noble metal thermocouple types | Type R, -50 °C to 0 °C Type R, 0 °C to 150 °C Type R, 150 °C to 400 °C Type R, 400 °C to 1768 °C Type S, -50 °C to 0 °C Type S, 0 °C to 100 °C Type S, 100 °C to 300 °C Type S, 300 °C to 1768 °C | 0.91 °C 0.71 °C 0.51 °C 0.62 °C 0.80 °C 0.66 °C 0.55 °C 0.48 °C | | |
| RTD Pt100 | Up to 0 °C Up to 0 °C 0 °C to 850 °C 0 °C to 850 °C | 0.072 °C 0.042 % + 0.072 °C 0.029 % + 0.075 °C 0.051 % + 0.075 °C | Ambient temperature range 18 °C to 28 °C -10 °C to +50 °C 18 °C to 28 °C -10 °C to +50 °C | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)
Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|--|------------------|
| DC Voltage | 0 V to 150 mV 0 V to 150 mV | 0.023 % + 5.0 μ V 0.048 % + 5.0 μ V | Ambient temperature range 18 °C to 28 °C -10 °C to +50 °C | Site |
| | 0.15 V to 0.25 V 0.15 V to 0.25 V | 0.023 % + 8.4 μ V 0.048 % + 8.4 μ V | 18 °C to 28 °C -10 °C to +50 °C | |
| | 0.25 V to 1 V 0.25 V to 1 V | 0.023 % + 12 μ V 0.048 % + 12 μ V | 18 °C to 28 °C -10 °C to +50 °C | |
| | 1 V to 25 V 1 V to 25 V | 0.023 % + 0.65 mV 0.048 % + 0.65 mV | 18 °C to 28 °C -10 °C to +50 °C | |
| | 25 V to 60 V 25 V to 60 V | 0.023 % + 1.2 mV 0.048 % + 1.2 mV | 18 °C to 28 °C -10 °C to +50 °C | |
| DC Current | 0 to 25 mA 0 to 25 mA | 0.025 % + 1.7 μ A 0.049 % + 1.7 μ A | 18 °C to 28 °C -10 °C to +50 °C | |
| | 25 mA to 100 mA 25 mA to 100 mA | 0.025 % + 2.0 μ A 0.049 % + 2.0 μ A | 18 °C to 28 °C -10 °C to +50 °C | |
| DC Resistance | 0 Ω to 250 Ω 0 Ω to 250 Ω | 0.023 % + 4.3 m Ω 0.048 % + 4.3 m Ω | 18 °C to 28 °C -10 °C to +50 °C | |
| | 250 Ω to 2650 Ω 250 Ω to 2650 Ω | 0.023 % + 11 m Ω 0.048 % + 11 m Ω | 18 °C to 28 °C -10 °C to +50 °C | |
| | 2650 Ω to 4000 Ω 2650 Ω to 4000 Ω | 0.023 % + 100 m Ω 0.048 % + 100 m Ω | 18 °C to 28 °C -10 °C to +50 °C | |
| TEMPERATURE | | | | |
| Temperature controlled, ovens, environmental chambers, fridges and freezers. | -80 °C to 400 °C 400 °C to 1000 °C 1000 °C to 1300 °C | 1.8 °C 2.0 °C 2.3 °C | Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping using procedures: QCR LCP 0020 and 0023 | |
| END | | | | |



0394
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Trescal Limited (Trescal EMS - Rolls-Royce)

Issue No: 062 Issue date: 26 May 2022

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$