


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0417 Accredited to ISO/IEC 17025:2017	Snap-on Diagnostics - a trading division of Snap-on UK Holdings Ltd Issue No: 027 Issue date: 18 August 2021	
	Denney Road King's Lynn Norfolk PE30 4HG	Contact: Mr Nicholas Burnham Tel: +44 (0)1553 692422 E-Mail: Nick.Burnham@snapon.com Website: www.snapon.com/diagnostics/uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Premises, e.g., MoT Test Centres, Garages and Repair/Service Facilities (*including those of Snap-on, King's Lynn)	Vehicle Exhaust Analyser (VEGA/DSM) Calibrations	S



0417
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees , Chertsey Lane, Staines-upon-Thames , TW18 3HR, UK

Snap-on Diagnostics
- a trading division of Snap-on UK Holdings Ltd
Issue No: 027 Issue date: 18 August 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Instruments approved by the Vehicle Inspectorate on behalf of the DfT for the measurement of:-	As per Vehicle Inspectorate requirements:-		NOTE: A list of individual approved signatories and the type of approved instruments they may calibrate is held by the laboratory and by UKAS.	
a) vehicle exhaust emissions	VPB/07/24/20/CAL dated May 1995 (1 st revision August 1995)	2.5 % of reading		S
b) free acceleration smoke	MOT/08/19/01 dated October 2001 (3 rd revision January 2007)	0.10 m ⁻¹	Smoke obscuration coefficient	S
END				



0417
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees , Chertsey Lane, Staines-upon-Thames , TW18 3HR, UK

Snap-on Diagnostics
- a trading division of Snap-on UK Holdings Ltd
Issue No: 027 Issue date: 18 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$