

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0467</b> <b>Accredited to ISO/IEC 17025:2017</b>	<b>Devon County Council trading as Heart of the South West Calibration Services</b>	
	<b>Issue No: 019 Issue date: 05 November 2021</b>	
	<b>Heart of the South West Trading Standards Service</b> County Hall Topsham Road Exeter EX2 4QH	<b>Contact: Victor Wardle</b> <b>Tel: +44 (0) 1392 383316 / 383000</b> <b>Fax: +44 (0) 1392 382732</b> <b>E-Mail: calibration@devon.gov.uk</b> <b>Website: www.devonsomersettradingstandards.gov.uk/ weight-testing-service</b>
<b>Calibration performed by the Organisation at the locations specified</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Heart of the South West Calibration Services Heart of the South West Trading Standards Service County Hall Topsham Road Exeter EX2 4QD	<b>Local contact</b> As above	Mass - Weights and artefacts A
<b>Address</b> Heart of the South West Calibration Services Heart of the South West Trading Standards Service West Buckland Road Chelston Somerset TA21 9HP	<b>Local contact</b> As above	Mass - Weights and artefacts B



0467

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devon County Council trading as Heart of the South West Calibration Services**

**Issue No: 019 Issue date: 05 November 2021**

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
MASS	Nominal value (g)	(mg)		A, B
Weights and artefacts	25 000	75	1. Calibrations can be given in other units as required.  2. Intermediate values can be calibrated with an uncertainty interpolated from the next higher and lower values in the table above.  3. Calibration by comparison using Borda's method of substitution.	
	20 000	20		
	10 000	10		
	5 000	5.0		
	2 000	2.0		
	1 000	1.0		
	500	0.50		
	200	0.10		
	100	0.050		
	50	0.033		
	20	0.027		
	10	0.020		
	5	0.017		
	2	0.013		
	1	0.010		
	0.5	0.0080		
	0.2	0.0070		
	0.1	0.0050		
	0.05	0.0040		
	0.02	0.0030		
	0.01	0.0030		
	0.005	0.0020		
	0.002	0.0020		
	0.001	0.0020		
END				



0467

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devon County Council trading as Heart of the South West Calibration Services**

**Issue No: 019 Issue date: 05 November 2021**

Calibration performed by the Organisation at the locations specified

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$