


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|---|---|--|
|  <p>UKAS CALIBRATION 0478</p> <p>Accredited to ISO/IEC 17025:2005</p> | <p>NPL Management Ltd</p> <p>Issue No: 093 Issue date: 16 November 2018</p> | |
| | <p>Hampton Road Teddington Middlesex TW11 0LW</p> | <p>Contact: Customer Helpline Tel: +44 (0)20 8943 7070 Fax: +44 (0)20 8614 0482 E-Mail: measurement_services@npl.co.uk Website: www.npl.co.uk</p> |
| <p>Calibration performed by the Organisation at the locations specified below</p> | | |

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details | Activity | Location code |
|--|---|---------------|
| <p>Address National Physical Laboratory Hampton Road Teddington Middlesex TW11 0LW</p> <p>Local contact Mr Tahir Maqba, Customer Services Manager Tel: +44 (0)20 8943 6796 Fax: +44 (0)20 8614 0482 Email: tahir.maqba@npl.co.uk</p> | <p><u>Calibration</u></p> <p>Chemical Mass Density Optical Dimensional Pressure Electromagnetic Radiological Fibre optics Temperature Flow Time and Frequency Force Ultrasonics Humidity Underwater Acoustics</p> | Teddington |
| <p>Address Wraysbury Reservoir Coppermill Road Wraysbury Middlesex TW19 5NW</p> <p>Local contact Mr J Ablitt Tel: +44 (0)20 8943 6695 Email: justin.ablitt@npl.co.uk</p> | <p><u>Calibration</u></p> <p>Underwater Acoustics</p> | Wraysbury |

Site activities performed away from the locations listed above:

| Location details | Activity | Location Code |
|--|---|------------------|
| <p>Customer's sites or premises</p> <p>The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p> | <p><u>Calibration</u></p> <p>Time and Frequency Chemical (Environmental air quality monitoring instruments)</p> | Customer's sites |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2) | Remarks | Location Code |
|--|--|---|--|-------------------|
| DC VOLTAGE <i>Service Reference ED01</i> | | | | Teddington |
| Standard cells, not thermostated | 1.018 V nominal | 0.090 ppm | Measured in a thermostated air enclosure at 20 °C | |
| Standard cells in a thermostated enclosure | 1.018 V nominal | 0.090 ppm | | |
| Electronic reference standards | 1.0 V 1.018 V 10 V | 0.14 ppm 0.14 ppm 0.020 ppm | Supplementary data can be supplied showing detailed behaviour of standard cells or electronic devices | |
| DC RESISTANCE <i>Service Reference ED02</i> | | | | |
| | 0.1 mΩ 1 mΩ 10 mΩ 100 mΩ 1 Ω 10 Ω 25 Ω 100 Ω 1 kΩ 10 kΩ | 2.5 ppm 0.85 ppm 0.80 ppm 0.18 ppm 0.060 ppm 0.050 ppm 0.050 ppm 0.050 ppm 0.050 ppm 0.060 ppm | 4 terminal resistors at temperatures between 17 °C and 25 °C and at or less than 1 mW power dissipation | |
| | 100 kΩ 1 MΩ 10 MΩ 100 MΩ | 0.080 ppm 0.12 ppm 0.20 ppm 0.40 ppm | 2-terminal resistors at temperatures between 17 °C and 25 °C and at or less than 1 mW power dissipation. Values >10 MΩ are not measured in oil. | |
| | 1 GΩ | 1.6 ppm | Measured in a 2-terminal configuration, in air, at 20 °C or 23 °C | |
| Temperature Coefficient | α β | 0.0020 ppm K ⁻¹ 0.0010 ppm K ⁻² | Resistance measurements at 4 temperatures in the range 15 °C to 30 °C. Uncertainty dependent on fit to curve and nominal value of resistor | |
| Current Carrying Resistors | 100 μΩ to 10 Ω | 0.50 ppm | For current levels up to 100 A. The uncertainty is dependent on measurement current, nominal value of resistor, elapsed time for which the current is applied and the specific calibration undertaken. | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---|-------------------|
| AC RESISTANCE <i>Service Reference ED02</i> | | | | Teddington |
| Specific values | 1 Ω 40 Hz to 1 kHz 1 kHz to 2 kHz 2 kHz to 3 kHz 3 kHz to 5 kHz 5 kHz to 10 kHz 20 kHz | 5.0 ppm 4.0 ppm 5.0 ppm 6.0 ppm 15 ppm 50 ppm | The uncertainties quoted for AC resistance may depend on the type and construction of the resistor | |
| Other values | 10 Ω to 10 k Ω 40 Hz to 400 Hz 400 Hz to 2 kHz 2 kHz to 3 kHz 3 kHz to 5 kHz 5 kHz to 10 kHz 20 kHz | 1.0 ppm 0.50 ppm 1.2 ppm 1.8 ppm 6.0 ppm 50 ppm | <i>Exceptions:</i> 10 Ω /40 Hz: 1.5 ppm 100 Ω /400 Hz to 1.59 kHz: 0.6 ppm 100 Ω /1.59 kHz to 2 kHz: 0.8 ppm 100 Ω /2 kHz to 3 kHz: 1.5 ppm 10 k Ω /10 kHz: 8.0 ppm | |
| Time constant (τ) | 0 ns to \pm 200 ns | 10 ns | All nominal values and frequencies shown above. | |
| AC CURRENT RATIO <i>Service Reference ED07</i> | | | | |
| <u>Current Transformers</u> | | | | |
| Ratio and phase error | 0.25 A to 0.5 A 50 Hz 5 A to 1000 A 50 Hz to 400 Hz 1000 A to 5000 A 50 Hz to 60 Hz 5 kA to 10 kA 50 Hz | <i>Ratio error</i> 10 ppm 10 ppm 10 ppm 20 ppm <i>Phase error</i> 10 μ rad 10 μ rad 10 μ rad 20 μ rad | The CMCs apply to compensated current transformers only. 1 A or 5 A secondary | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation

issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | | | | | | | | | | |
|--|---|--|--------------------|--------------------|--------|--------------|---------|--------------|---------|---------|---|-------------------|--|--|
| <u>Current Transformers</u> (continued) Ratio and phase error (continued) <u>Current Transducers</u> with output voltage greater than 0.10V AC/DC TRANSFER VOLTAGE Service Reference ED11 | 50 Hz to 400 Hz Class 0.01, 0.02 and 0.03 Class 0.1 and higher 50 Hz | <table style="width: 100%; border: none;"> <tr> <td style="text-align: center;"><i>Ratio error</i></td> <td style="text-align: center;"><i>Phase error</i></td> </tr> <tr> <td style="text-align: center;">10 ppm</td> <td style="text-align: center;">10 μrad</td> </tr> <tr> <td style="text-align: center;">30 ppm</td> <td style="text-align: center;">30 μrad</td> </tr> <tr> <td colspan="2" style="text-align: center;">0.050 %</td> </tr> </table> | <i>Ratio error</i> | <i>Phase error</i> | 10 ppm | 10 μ rad | 30 ppm | 30 μ rad | 0.050 % | | The CMCs apply to measurements carried out on uncompensated current transformers in accordance with BS EN 61869-2:2012 at unity or 0.8 power factor as specified or required. | Teddington | | |
| <i>Ratio error</i> | <i>Phase error</i> | | | | | | | | | | | | | |
| 10 ppm | 10 μ rad | | | | | | | | | | | | | |
| 30 ppm | 30 μ rad | | | | | | | | | | | | | |
| 0.050 % | | | | | | | | | | | | | | |
| CMCs for AC/DC Transfer Voltage, at Specific Values, expressed as an Expanded Uncertainty ($k = 2$) [ppm of value] For intermediate points the uncertainty will be determined using linear interpolation between the adjacent points. | | | | | | | | | | | | | | |
| Voltage | Frequency | | | | | | | | | | | | | |
| | 10 Hz | 20 Hz to 5 kHz | 10 kHz | 20 kHz | 50 kHz | 100 kHz | 200 kHz | 300 kHz | 500 kHz | 700 kHz | 1 MHz | | | |
| 1 mV | 66 | 66 | 66 | 66 | 71 | 85 | 120 | 180 | 260 | 480 | 800 | | | |
| 2 mV | 66 | 57 | 57 | 57 | 59 | 71 | 99 | 140 | 190 | 330 | 510 | | | |
| 5 mV | 28 | 28 | 28 | 28 | 33 | 48 | 83 | 120 | 180 | 290 | 460 | | | |
| 10 mV | 28 | 28 | 28 | 28 | 33 | 46 | 78 | 110 | 170 | 260 | 410 | | | |
| 20 mV | 28 | 28 | 28 | 28 | 31 | 46 | 78 | 110 | 160 | 250 | 370 | | | |
| 70 mV | 26 | 26 | 26 | 26 | 31 | 41 | 76 | 110 | 160 | 250 | 370 | | | |
| 100 mV | 7.0 | 7.0 | 7.0 | 7.0 | 9.0 | 14 | 24 | 36 | 58 | 82 | 120 | | | |
| 200 mV | 7.0 | 7.0 | 7.0 | 7.0 | 9.0 | 14 | 24 | 36 | 58 | 82 | 120 | | | |
| 300 mV | 6.0 | 6.0 | 6.0 | 6.0 | 7.0 | 11 | 20 | 29 | 48 | 68 | 96 | | | |
| 500 mV | 6.0 | 6.0 | 6.0 | 6.0 | 7.0 | 10 | 16 | 23 | 38 | 54 | 76 | | | |
| 1 V | 6.0 | 6.0 | 6.0 | 6.0 | 7.0 | 7.0 | 13 | 17 | 25 | 38 | 51 | | | |
| 2 V | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 7.0 | 10 | 12 | 17 | 27 | 38 | | | |
| 3 V | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 7.0 | 10 | 12 | 17 | 27 | 38 | | | |
| 4 V | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 8.0 | 9.0 | 13 | 21 | 31 | | | |
| 5 V | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 8.0 | 9.0 | 13 | 21 | 31 | | | |
| 10 V | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 8.0 | 10 | 15 | 22 | 32 | | | |
| 20 V | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 7.0 | 9.0 | 11 | 16 | 25 | 34 | | | |
| 30 V | 6.0 | 6.0 | 6.0 | 6.0 | 6.0 | 7.0 | 11 | | | | | | | |
| 50 V | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 10 | 14 | | | | | | | |
| 70 V | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 10 | 14 | | | | | | | |
| 100 V | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 10 | 14 | | | | | | | |
| 200 V | 8.0 | 8.0 | 8.0 | 10 | 13 | 22 | | | | | | | | |
| 300 V | 8.0 | 8.0 | 8.0 | 10 | 13 | 22 | | | | | | | | |
| 500 V | 11 | 9.0 | 10 | 15 | 24 | 42 | | | | | | | | |
| 600 V | 11 | 9.0 | 10 | 19 | 29 | 52 | | | | | | | | |
| 700 V | 11 | 9.0 | 10 | 19 | 29 | 52 | | | | | | | | |
| 1 kV | 11 | 9.0 | 11 | 23 | 33 | 62 | | | | | | | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | | | | | | | |
|---|-----------|--|---------|-------------------|--------|---------|---------|---------|---------|---------|-------|
| AC/DC TRANSFER VOLTAGE (continued) | | | | Teddington | | | | | | | |
| AC VOLTAGE Service Reference ED11 | | | | | | | | | | | |
| CMCs for AC Voltage, at Specific Values, expressed as an Expanded Uncertainty ($k = 2$) [ppm of value] <i>For intermediate points the uncertainty will be determined using linear interpolation between the adjacent points.</i> | | | | | | | | | | | |
| Voltage | Frequency | | | | | | | | | | |
| | 10 Hz | 20 Hz to 5 kHz | 10 kHz | 20 kHz | 50 kHz | 100 kHz | 200 kHz | 300 kHz | 500 kHz | 700 kHz | 1 MHz |
| 500 mV | 8.0 | 8.0 | 8.0 | 8.0 | 10 | 13 | 20 | 28 | 43 | 62 | 86 |
| 1 V | 8.0 | 8.0 | 8.0 | 9.0 | 11 | 13 | 18 | 22 | 30 | 43 | 60 |
| 2 V | 8.0 | 8.0 | 8.0 | 8.0 | 9.0 | 11 | 14 | 17 | 23 | 31 | 43 |
| 3 V | 8.0 | 8.0 | 8.0 | 8.0 | 9.0 | 8.0 | 14 | 15 | 21 | 29 | 41 |
| 4 V | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 9.0 | 11 | 15 | 22 | 32 |
| 5 V | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 9.0 | 11 | 15 | 22 | 32 |
| 10 V | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 8.0 | 9.0 | 11 | 16 | 24 | 34 |
| 20 V | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 9.0 | 10 | 13 | 17 | 27 | 35 |
| 30 V | 7.0 | 7.0 | 7.0 | 7.0 | 7.0 | 9.0 | 11 | | | | |
| 50 V | 8.0 | 8.0 | 8.0 | 8.0 | 8.0 | 12 | 15 | | | | |
| 70 V | 8.0 | 8.0 | 8.0 | 8.0 | 8.0 | 12 | 15 | | | | |
| 100 V | 8.0 | 8.0 | 8.0 | 8.0 | 8.0 | 12 | 15 | | | | |
| 200 V | 9.0 | 9.0 | 9.0 | 10 | 13 | 22 | | | | | |
| 300 V | 9.0 | 9.0 | 9.0 | 10 | 13 | 22 | | | | | |
| 500 V | 12 | 10 | 11 | 15 | 25 | 42 | | | | | |
| 600 V | 12 | 10 | 11 | 19 | 29 | 52 | | | | | |
| 700 V | 12 | 10 | 11 | 19 | 29 | 52 | | | | | |
| 1 kV | 12 | 10 | 12 | 23 | 33 | 62 | | | | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | | | | | | | | | | |
|---|-----------|--|---------|-------------------|--------|--------|-------|-------|-------|--------|--------|--------|--------|---------|
| AC/DC TRANSFER CURRENT <i>Service Reference ED11</i> | | | | Teddington | | | | | | | | | | |
| CMCs for AC/DC Transfer Current, at Specific Values, expressed as an Expanded Uncertainty ($k = 2$) [ppm of value] <i>For intermediate points the uncertainty will be determined using linear interpolation between the adjacent points.</i> | | | | | | | | | | | | | | |
| Current | Frequency | | | | | | | | | | | | | |
| | 10 Hz | 20 Hz | 40 Hz | | 100 Hz | 400 Hz | 1 kHz | 2 kHz | 5 kHz | 10 kHz | 20 kHz | 50 kHz | 70 kHz | 100 kHz |
| 1 mA | 31 | 30 | 30 | | 30 | 30 | 30 | 30 | 30 | 30 | 31 | 31 | 33 | 35 |
| 2 mA | 17 | 17 | 16 | | 16 | 16 | 16 | 16 | 16 | 17 | 18 | 19 | 22 | 25 |
| 3 mA | 12 | 12 | 12 | | 12 | 12 | 12 | 12 | 12 | 13 | 14 | 16 | 19 | 22 |
| 5 mA | 11 | 10 | 10 | | 10 | 10 | 10 | 10 | 10 | 12 | 13 | 15 | 19 | 22 |
| 10 mA | 11 | 10 | 10 | | 10 | 10 | 10 | 10 | 10 | 12 | 13 | 15 | 19 | 22 |
| 20 mA | 11 | 10 | 10 | | 10 | 10 | 10 | 10 | 10 | 12 | 13 | 15 | 19 | 22 |
| 30 mA | 11 | 10 | 10 | | 10 | 10 | 10 | 10 | 10 | 12 | 13 | 15 | 19 | 22 |
| 50 mA | 11 | 10 | 10 | | 10 | 10 | 10 | 10 | 10 | 12 | 13 | 15 | 19 | 22 |
| 0.1 A | 14 | 13 | 12 | | 12 | 12 | 12 | 12 | 12 | 12 | 13 | 20 | 23 | 42 |
| 0.2 A | 23 | 20 | 16 | | 16 | 16 | 16 | 16 | 16 | 16 | 17 | 28 | 33 | 61 |
| 0.25 A | 23 | 20 | 16 | 16 | 16 | 16 | 16 | 16 | 16 | 17 | 28 | 33 | 61 | |
| 0.3 A | 30 | 26 | 16 | 17 | 16 | 18 | 16 | 17 | 15 | 24 | 43 | 52 | 81 | |
| 0.5 A | 30 | 26 | 16 | 17 | 16 | 18 | 16 | 17 | 15 | 24 | 43 | 52 | 81 | |
| 1 A | 38 | 31 | 19 | 19 | 19 | 20 | 17 | 18 | 17 | 33 | 53 | 62 | 100 | |
| 2 A | 47 | 37 | 22 | 20 | 21 | 22 | 20 | 21 | 20 | 43 | 63 | 83 | 120 | |
| 2.5 A | 47 | 37 | 22 | 20 | 21 | 22 | 20 | 21 | 20 | 43 | 63 | 83 | 120 | |
| 3 A | 55 | 43 | 25 | 23 | 24 | 24 | 23 | 21 | 22 | 53 | 83 | 100 | 160 | |
| 5 A | 55 | 43 | 25 | 23 | 24 | 24 | 23 | 21 | 22 | 53 | 83 | 100 | 160 | |
| 10 A | 63 | 49 | 27 | 25 | 26 | 26 | 25 | 26 | 23 | 62 | 100 | 120 | 200 | |
| 20 A | 72 | 56 | 31 | 28 | 30 | 29 | 28 | 29 | 28 | 73 | 120 | 140 | 240 | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | | |
|--|-----------|--|---|------------------|--|------------|
| AC VOLTAGE RATIO (continued) <i>Service Reference ED13</i> <u>Inductive Voltage Dividers</u> Voltage ratio | | | | | | |
| | LF System | | Normal operating range: Minimum voltage: 1 V Maximum voltage: 0.1 x f(Hz) from 40 Hz to 80 Hz; 0.15 x f(Hz) from 100 Hz to 200 Hz; 30 V otherwise. | Teddington | | |
| | 40 Hz | 16 | | | 17 | |
| | 60 Hz | 16 | | | 17 | |
| | 80 Hz | 12 | | | 14 | |
| | 100 Hz | 9.2 | | | 11 | |
| | 120 Hz | 7.1 | | | 8.4 | |
| | 200 Hz | 6.1 | | | 7.7 | |
| | 300 Hz | 6.1 | | | 6.9 | |
| | 400 Hz | 6.1 | | | 6.9 | |
| | 600 Hz | 6.1 | | | 6.9 | |
| | 800 Hz | 6.1 | | | 6.9 | |
| | 1000 Hz | 6.1 | | | 6.9 | |
| | 1300 Hz | 6.1 | | | 6.9 | |
| | 1592 Hz | 6.1 | | | 6.9 | |
| | 2000 Hz | 6.8 | | | 8.0 | |
| | 3000 Hz | 9.1 | | | 9.8 | |
| | 4000 Hz | 14 | | | 14 | |
| | 5000 Hz | 21 | | | 21 | |
| | HF System | | | | Normal operating range: Minimum voltage: 1 V Maximum voltage: 30 V | Teddington |
| | 5 kHz | 21 | | | | |
| | 8 kHz | 30 | 30 | | | |
| | 10 kHz | 38 | 38 | | | |
| | 20 kHz | 72 | 75 | | | |
| | 30 kHz | 120 | 120 | | | |
| | 40 kHz | 180 | 190 | | | |
| | 50 kHz | 280 | 300 | | | |
| | 80 kHz | 630 | 650 | | | |
| | 100 kHz | 990 | 1000 | | | |
| | 120 kHz | 1500 | 1600 | | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---|-------------------|
| AC POWER <i>Service Reference ED06</i> <i>Sinusoidal waveforms</i> | <i>40 Hz to 400 Hz: Current 2 mA to 130 A Voltage 1 V to 1000 V</i> | 40 μ W/VA 25 μ W/VA | 20 °C and 23 °C at unity power factor 20 °C and 23 °C at zero power factor <i>Uncertainties increase at other power factors</i> | Teddington |
| Current Response of Wattmeters | 2 mA to 20 A | 30 μ W/VA | 20 °C and 23 °C | |
| Voltage Response of Wattmeters | 1 V to 1000 V | 25 μ W/VA | 20 °C and 23 °C | |
| Auxiliary DC Voltage | DC, 1 V to 10 V | 5.0 ppm of value | 20 °C and 23 °C | |
| AC REACTIVE VOLT-AMPERES <i>Sinusoidal waveforms</i> | <i>50 Hz to 400 Hz: Current 2 mA to 130 A Voltage 1 V to 1000 V</i> | 40 μ W/VA 25 μ W/VA | 20 °C and 23 °C at zero power factor 20 °C and 23 °C at unity power factor <i>Uncertainties increase at other power factors</i> | |
| CALIBRATION OF EN 61000 HARMONIC AND FLICKER ANALYSERS <i>Service Reference ED17</i> <i>Sinusoidal waveforms</i> | | | | |
| Current accuracy | 100 mA to 20 A, 50 Hz | 40 ppm | | |
| Current frequency response | 100 mA to 20 A 50 Hz to 2 kHz | 150 ppm | | |
| Voltage accuracy | 1 V to 1000 V, 50 Hz | 30 ppm | | |
| Power measurements | Ranges as in <i>AC Power</i> above | 45 ppm of full-scale | At unity power factor | |
| <i>Non-sinusoidal waveforms</i> Harmonic measurements for current waveforms | Peak values 1A to 10 A 50 Hz fundamental; harmonics up to 2 kHz | 200 ppm | Steady-state, burst fluctuating or smoothly fluctuating harmonics | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | |
|---|---|--|---|---------------|--|
| CALIBRATION OF EN 61000 HARMONIC AND FLICKER ANALYSERS (continued) | | | | | |
| Flicker (Pst) | 0.95 Pst to 1.05 Pst Square or sine wave modulated, 230 V 50 Hz sine wave | 0.30 % of Pst reading | See the NPL web site for details about these complex waveforms. | Teddington | |
| | 0.95 Pst to 2.05 Pst Complex waveforms, 230 V 50 Hz sine wave | 0.20 % of Pst reading | | | |
| CAPACITANCE and DISSIPATION FACTOR <i>Service Reference ED04</i> | | | | | |
| Fused-silica dielectric capacitors | <i>At 1 kHz and 1.592 kHz</i> | C D | Capacitance and dissipation factor measurements are normally carried out between 20 °C and 23 °C but may exceptionally be carried out at any temperature between 18 °C and 25 °C. | | |
| | 10 pF 100 pF | 0.70 ppm 6.0 x 10 ⁻⁶ 0.90 ppm 7.0 x 10 ⁻⁶ | | | |
| Other types of capacitor | 1 nF 1 kHz 1.592 kHz 10 kHz 100 kHz 1 MHz | 3.0 ppm 20 x 10 ⁻⁶ 3.0 ppm 20 x 10 ⁻⁶ 5.0 ppm 20 x 10 ⁻⁶ 200 ppm 80 x 10 ⁻⁶ 500 ppm 5.0 x 10 ⁻⁴ | | | |
| | <i>At 1 kHz and 1.592 kHz</i> | | | | |
| | 1 pF to 10 pF 10 pF to 1 nF 1 nF to 100 nF 100 nF to 1 µF | 4.0 ppm 10 x 10 ⁻⁶ 3.0 ppm 7.0 x 10 ⁻⁶ 30 ppm 20 x 10 ⁻⁶ 60 ppm 20 x 10 ⁻⁶ | | | |
| General Radio Type 1417 | 1 µF to 10 mF 100 Hz, 120 Hz and 1 kHz | 0.10 % to 0.50 % 0.0010 to 0.0005 | | | |
| | 100 mF to 1 F 100 Hz and 120 Hz | 0.30 % to 1.0 % 0.0030 to 0.010 | | | |
| Dissipation Factor Standard | 0.0 to 0.001 1 kHz | 20 x 10 ⁻⁶ | | | |
| | 0.0 to 0.005 50 Hz | 20 x 10 ⁻⁶ | | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | | | | |
|---|--|--|---------------------------------------|-------------------|---------|--------------------|---------|---------|
| SELF-INDUCTANCE <i>Service Reference ED05</i> | | | | Teddington | | | | |
| <i>Expanded uncertainty at 95% confidence level ($k = 2$) for the frequencies shown</i> | | | | | | | | |
| Nominal value | 20 Hz | 50 Hz | 100 Hz 400 Hz | | 1 kHz | 1.592 kHz 2 kHz | 5 kHz | 10 kHz |
| 1 μ H | | | 2.0 % | | 0.10 % | 0.10 % | 0.25 % | 0.35 % |
| 2 μ H | | | 1.0 % | | 0.10 % | 0.10 % | 0.22 % | 0.30 % |
| 3 μ H | | | 0.61 % | | 0.10 % | 0.10 % | 0.22 % | 0.26 % |
| 5 μ H | | | 0.35 % | | 600 ppm | 600 ppm | 0.11 % | 0.15 % |
| 10 μ H | 0.35 % | 0.25 % | 0.20 % | | 310 ppm | 350 ppm | 620 ppm | 930 ppm |
| 20 μ H | 0.18 % | 0.13 % | 0.10 % | | 150 ppm | 160 ppm | 320 ppm | 460 ppm |
| 30 μ H | 0.12 % | 840 ppm | 670 ppm | | 110 ppm | 120 ppm | 190 ppm | 260 ppm |
| 50 μ H | 700 ppm | 500 ppm | 400 ppm | | 100 ppm | 100 ppm | 160 ppm | 200 ppm |
| 100 μ H | 300 ppm | 200 ppm | 150 ppm | | 75 ppm | 80 ppm | 120 ppm | 150 ppm |
| 200 μ H | 250 ppm | 180 ppm | 100 ppm | | 75 ppm | 85 ppm | 110 ppm | 150 ppm |
| 300 μ H | 250 ppm | 180 ppm | 100 ppm | | 85 ppm | 85 ppm | 120 ppm | 150 ppm |
| 500 μ H | 220 ppm | 160 ppm | 100 ppm | | 80 ppm | 80 ppm | 100 ppm | 150 ppm |
| 1 mH | 180 ppm | 150 ppm | 95 ppm | | 70 ppm | 75 ppm | 100 ppm | 150 ppm |
| 2 mH | 180 ppm | 150 ppm | 100 ppm | | 75 ppm | 80 ppm | 110 ppm | 150 ppm |
| 3 mH | 180 ppm | 150 ppm | 100 ppm | | 85 ppm | 85 ppm | 120 ppm | 150 ppm |
| 5 mH | 180 ppm | 160 ppm | 100 ppm | | 80 ppm | 80 ppm | 110 ppm | 150 ppm |
| 10 mH | 180 ppm | 150 ppm | 100 ppm | | 70 ppm | 70 ppm | 100 ppm | 130 ppm |
| 20 mH | 180 ppm | 150 ppm | 100 ppm | | 75 ppm | 75 ppm | 110 ppm | 130 ppm |
| 30 mH | 180 ppm | 150 ppm | 100 ppm | | 85 ppm | 85 ppm | 110 ppm | 150 ppm |
| 50 mH | 200 ppm | 160 ppm | 100 ppm | | 80 ppm | 80 ppm | 160 ppm | 200 ppm |
| 100 mH | 190 ppm | 150 ppm | 85 ppm | | 70 ppm | 70 ppm | 140 ppm | 200 ppm |
| 200 mH | 230 ppm | 200 ppm | 90 ppm | | 75 ppm | 75 ppm | 200 ppm | 300 ppm |
| 400 mH | 240 ppm | 200 ppm | 90 ppm | | 75 ppm | 75 ppm | 200 ppm | 380 ppm |
| 500 mH | 240 ppm | 210 ppm | 90 ppm | | 80 ppm | 80 ppm | 200 ppm | 400 ppm |
| 1 H | 140 ppm | 110 ppm | 85 ppm | | 70 ppm | 70 ppm | 200 ppm | 400 ppm |
| 2 H | 140 ppm | 110 ppm | 85 ppm | | 70 ppm | 70 ppm | | |
| 5 H | 140 ppm | 110 ppm | 85 ppm | | 80 ppm | 85 ppm | | |
| 10 H | 140 ppm | 110 ppm | 85 ppm | 80 ppm | 85 ppm | | | |
| MUTUAL INDUCTANCE <i>Service Reference ED05</i> | At 1 kHz: 1 mH 5 mH 10 mH 100 mH | | 100 ppm 80 ppm 70 ppm 70 ppm | | | | | |
| NOTE | | | | | | | | |
| Inductance measurements are normally carried out between 20 °C and 23 °C but may exceptionally be carried out at any temperature between 18 °C and 25 °C. The DC resistance of an inductor can also be reported as an indication of its temperature. Inductance measurements may be made at other frequencies between 20 Hz and 10 kHz, but the uncertainties may be increased. | | | | | | | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|--|--|--|---------|-------------------|
| MAGNETIC QUANTITIES <i>Service Reference MT41</i> | | | | Teddington |
| DC MAGNETIC FIELD STRENGTH AND MAGNETIC FLUX DENSITY | | | | |
| | 0.8 mA/m (1 nT) to 16 A/m (20 μ T) | 0.15 % + 0.4 mA/m (0.5 nT) | | |
| | 16 A/m (20 μ T) to 72 A/m (90 μ T) | 0.0030 % | | |
| | 72 A/m (90 μ T) to 280 A/m (350 μ T) | 0.050 % | | |
| | 280 A/m (350 μ T) to 40 kA/m (50 mT) | 0.20 % | | |
| | 40 kA/m (50 mT) to 10.5 MA/m (13 T) | 0.0015 % | | |
| AC MAGNETIC FIELD STRENGTH AND MAGNETIC FLUX DENSITY | | | | |
| | 8 mA/m (10 nT) to 17.5 kA/m (22 mT) 10 Hz to 60 Hz | 0.25 % | | |
| | 8 mA/m (10 nT) to 80 A/m (100 μ T) 60 Hz to 20 kHz | 0.25 % | | |
| | 8 mA/m (10 nT) to 40 A/m (50 μ T) 20 kHz to 50 kHz | 0.40 % | | |
| | 8 mA/m (10 nT) to 15.9 A/m (20 μ T) 50 kHz to 120 kHz | 0.70 % | | |
| MAGNETIC FIELD STRENGTH TO CURRENT RATIO | | | | |
| Standard solenoids and Helmholtz coils | 1 A/m/A to 20 000 A/m/A DC 12 Hz to 60 Hz 60 Hz to 20 kHz | 0.015 % 0.050 % 0.25 % | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|--|---|--|---|-------------------|
| MAGNETIC QUANTITIES (continued) | | | | Teddington |
| TURN AREA (effective area) | | | | |
| Search coils | 0.0029 m ² to 17 m ² 12 Hz to 60 Hz | 0.090 % | | |
| | 0.0001 m ² to 200 m ² 60 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 120 kHz | 0.25 % 0.40 % 0.70 % | | |
| NORMAL DC MAGNETIZATION CURVES AND HYSTERESIS LOOPS | | | | |
| Ring specimens | H = 0.1 kA/m to 10 kA/m B = 0.05 T to 2.5 T | 0.40 % 0.40 % | In accordance with EN 60404 Part 4: 1997 and IEC 60404 Part 4:2008 | |
| Bar or rod specimens | H = 0.1 kA/m to 200 kA/m B = 0.05 T to 2.5 T | 0.40 % 0.40 % | In accordance with EN 60404 Part 4: 1997 and IEC 60404 Part 4:2008 | |
| DC DEMAGNETIZATION CURVE FOR HARD MAGNETIC MATERIALS | | | | |
| Remanence | B _r = 0.02 T to 2 T | 0.30 % | <div style="border-left: 1px solid black; border-right: 1px solid black; border-bottom: 1px solid black; width: 20px; height: 20px; margin: 0 auto;"></div> | |
| Coercivity | H _{CB} = 0.03 to 1.2 MA/m H _{CJ} = 0.03 to 1.6 MA/m | 0.40 % 0.40 % | | |
| Maximum energy product | (B.H) _{max} = 1 to 400 kJ/m ³ | 0.50 % | | |
| DC RELATIVE MAGNETIC PERMEABILITY, μ_r | | | | |
| For low magnetic Permeability materials | ($\mu_r - 1$) = 0.001 to 1.5 ($\mu_r - 1$) = 0.0002 to 0.001 | 0.20 % 2.2 % | In accordance with BS EN 60404 Part 15: 2012 | |
| Permeability measuring instruments and indicators | ($\mu_r - 1$) = 0.001 to 1.5 | 0.20 % | The uncertainty may be increased depending on the characteristics of the device being calibrated | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (<i>k</i> =2) | Remarks | Location Code |
|--|--|---|--|-------------------|
| MAGNETIC QUANTITIES (continued) | | | | Teddington |
| MAGNETIC DIPOLE MOMENT | 0.06 Am ² to 1000 Am ² | 0.11 % | | |
| SPECIFIC TOTAL POWER LOSS | | | | |
| For Epstein strips and 500 mm sheets | 0.02 W/kg to 120 W/kg | | Method of measurement: | |
| Non-oriented material | <i>At 50 Hz and 60 Hz</i> J = 0.1 T to 1.3 T J = 1.3 T to 1.7 T | 0.65 % 0.75 % | For strips: IEC 60404 Part 2:2008 BS EN 60404 Part 2:1998 | |
| Oriented material | <i>At 50 Hz and 60 Hz</i> J = 0.1 T to 1.7 T J = 1.7 T to 1.8 T | 0.65 % 0.75 % | For sheets: IEC 60404 Part 3:2010 BS EN 10280:2001 | |
| For Epstein strips only | 0.02 W/kg to 120 W/kg | | | |
| Non-oriented material | <i>At 400 Hz</i> J = 0.1 T to 1.3 T J = 1.3 T to 1.7 T | 0.70 % 1.8 % | Measurements can be made at frequencies up to 16 kHz but with an increase in uncertainty using IEC 60404 Part 10:1988 BS EN 10252:1997 | |
| Oriented material | <i>At 400 Hz</i> J = 0.1 T to 1.7 T J = 1.7 T to 1.8 T | 0.70 % 1.8 % | | |
| Soft magnetic materials in ring form only | 0.02 W/kg to 120 W/kg <i>50 Hz to 100 kHz</i> J = 1 mT to 100 mT | 0.65 % | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|--|---|--|---|-------------------|
| MAGNETIC QUANTITIES (continued) | | | | Teddington |
| APPARENT POWER | 0.06 VA/kg to 400 VA/kg | | <div style="border-left: 1px solid black; border-right: 1px solid black; border-bottom: 1px solid black; padding: 5px;"> <p>Method of measurement:</p> <p>For strips: IEC 60404 Part 2:2008 BS EN 60404-2: 1998</p> <p>For sheets: IEC 60404 Part 3:2010 BS EN 10280: 2001</p> </div> | |
| Non-oriented material | At 50 Hz and 60 Hz J = 0.1 T to 1.3 T J = 1.3 T to 1.5 T J = 1.5 T to 1.7 T | 1.1 % 1.5 % 2.6 % | | |
| Oriented material | At 50 Hz and 60 Hz J = 0.1 T to 1.5 T J = 1.5 T to 1.8 T | 1.1 % 2.6 % | | |
| For Epstein strips only | At 400 Hz J = 0.1 T to 1.0 T J = 1.0 T to 1.5 T | 1.1 % 2.6 % | | |
| Soft magnetic materials in ring form only | 50 Hz to 100 kHz J = 1 mT to 100 mT | 1.1 % | | |
| AC PERMEABILITY (rms or peak values) | | | | |
| Oriented and non-oriented materials | $\mu_r = 500$ to 200 000 At 50 Hz and 60 Hz B = 0.5 T to 2.2 T H = 0.5 kA/m to 10 kA/m | 0.45 % | | |
| AC MAGNETIC FIELD STRENGTH (rms or peak values) | | | | |
| Oriented and non-oriented materials | At 50 Hz and 60 Hz H = 0.5 kA/m to 10 kA/m | 0.45 % | | |
| AC CONDUCTIVITY <i>Service Reference MT41</i> | | | | |
| AC conductivity reference materials | 2 MS/m to 60 MS/m (3.45 % _{0IACS} to 103 % _{0IACS}) 60 kHz, 20°C | 0.70 % | | |
| AC conductivity instruments | 2 MS/m to 60 MS/m (3.45 % _{0IACS} to 103 % _{0IACS}) 60 kHz, 20°C | 0.70 % | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|--|-------------------|
| POWER FLUX DENSITY <i>Service Reference EF01</i> | 0.11 nW/cm ² to 170 mW/cm ² <i>10 Hz to 10 kHz</i> | 0.68 dB | <u>TEM Cells</u> The maximum frequency and power flux density level is determined by the size of the probe. | Teddington |
| | 0.11 nW/cm ² to 265 mW/cm ² <i>10 kHz to 300 MHz</i> | 0.68 dB | | |
| | 0.17 nW/cm ² to 120 mW/cm ² <i>180 MHz to 1000 MHz</i> | 0.80 dB | <u>Tapered TEM Cell</u> The maximum frequency and power flux density level is determined by the size of the probe. | |
| | 0.17 nW/cm ² to 64 mW/cm ² <i>1000 MHz to 3000 MHz</i> | 0.80 dB | | |
| | 0.03 nW/cm ² to 38 mW/cm ² <i>240 MHz to 270 MHz</i> | 0.65 dB | <u>Anechoic Chambers</u> The Listed Field levels are derived from the lowest unsaturated maximum power in each range. The actual achievable level may be up to 20 % greater than the stated limit. All probes and small active dipoles | |
| | 0.03 nW/cm ² to 65 mW/cm ² <i>270 MHz to 350 MHz</i> | 0.65 dB | | |
| | 0.03 nW/cm ² to 87 mW/cm ² <i>350 MHz to 500 MHz</i> | 0.65 dB | | |
| | 0.03 nW/cm ² to 37 mW/cm ² <i>450 MHz to 550 MHz</i> | 0.62 dB | | |
| | 0.03 nW/cm ² to 72 mW/cm ² <i>550 MHz to 750 MHz</i> | 0.62 dB | | |
| | 0.03 nW/cm ² to 72 mW/cm ² <i>750 MHz to 950 MHz</i> | 0.62 dB | | |
| | 0.03 nW/cm ² to 38 mW/cm ² <i>950 MHz to 1200 MHz</i> | 0.62 dB | | |
| | 0.03 nW/cm ² to 138 mW/cm ² <i>1100 MHz to 1250 MHz</i> | 0.47 dB | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---|------------------|
| POWER FLUX DENSITY (continued) | 0.03 nW/cm ² to 170 mW/cm ² 1250 MHz to 1700 MHz | 0.47 dB | All probes and small active dipoles | Teddington |
| | 0.03 nW/cm ² to 227 mW/cm ² 1700 MHz to 2600 MHz | 0.47 dB | All probes and small active dipoles | |
| | 0.7 nW/cm ² to 569 mW/cm ² 2.45 GHz to 2.7 GHz | 0.40 dB | All probes and small active dipoles | |
| | 0.7 nW/cm ² to 921 mW/cm ² 2.7 GHz to 8.2 GHz | 0.40 dB | All probes and small active dipoles | |
| | 0.7 nW/cm ² to 694 mW/cm ² 8.2 GHz to 18 GHz | 0.40 dB | All probes and small active dipoles | |
| | 1.7 μW/cm ² to 92 mW/cm ² 18 GHz to 40 GHz | 0.35 dB | All probes and small active dipoles | |
| | 60 μW/cm ² to 2.7 mW/cm ² 43.5 GHz to 45.5 GHz | 0.35 dB | All probes and small active dipoles | |
| FIELD STRENGTH <i>Service Reference EF01</i> Electric Field | 0.02 V/m to 800 V/m 10 Hz to 10 kHz | 0.68 dB | <u>TEM Cells</u> The maximum frequency and field strength level is determined by the size of the probe. | Teddington |
| | 0.02 V/m to 1000 V/m 10 kHz to 300 MHz | 0.68 dB | | |
| | 0.008 V/m to 670 V/m 180 MHz to 1000 MHz | 0.80 dB | <u>Tapered TEM Cell</u> The maximum frequency and field strength level is determined by the size of the probe. | |
| | 0.008 V/m to 490 V/m 1000 MHz to 3000 MHz | 0.80 dB | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---|-------------------|
| FIELD STRENGTH (continued) | | | <u>Anechoic Chambers</u> The Listed Field levels are derived from the lowest unsaturated maximum power in each range. The actual achievable level may be up to 20 % greater than the stated limit. | |
| Electric Field | 0.01 V/m to 380 V/m 240 MHz to 270 MHz | 0.65 dB | All probes and small active dipoles | Teddington |
| | 0.01 V/m to 500 V/m 270 MHz to 350 MHz | 0.65 dB | All probes and small active dipoles | |
| | 0.01 V/m to 575 V/m 350 MHz to 500 MHz | 0.65 dB | All probes and small active dipoles | |
| | 0.01 V/m to 375 V/m 450 MHz to 550 MHz | 0.62 dB | All probes and small active dipoles | |
| | 0.01 V/m to 520 V/m 550 MHz to 750 MHz | 0.62 dB | All probes and small active dipoles | |
| | 0.01 V/m to 520 V/m 750 MHz to 950 MHz | 0.62 dB | All probes and small active dipoles | |
| | 0.01 V/m to 380 V/m 950 MHz to 1200 MHz | 0.62 dB | All probes and small active dipoles | |
| | 0.01 V/m to 720 V/m 1100 MHz to 1250 MHz | 0.47 dB | All probes and small active dipoles | |
| | 0.01 V/m to 800 V/m 1250 MHz to 1700 MHz | 0.47 dB | All probes and small active dipoles | |
| | 0.01 V/m to 920 V/m 1700 MHz to 2600 MHz | 0.47 dB | All probes and small active dipoles | |
| | 0.05 V/m to 1460 V/m 2.45 GHz to 2.7 GHz | 0.40 dB | All probes and small active dipoles | |
| | 0.05 V/m to 1860 V/m 2.7 GHz to 8.2 GHz | 0.40 dB | All probes and small active dipoles | |
| | 0.05 V/m to 1610 V/m 8.2 GHz to 18 GHz | 0.40 dB | All probes and small active dipoles | |
| | 2.5 V/m to 590 V/m 18 GHz to 40 GHz | 0.35 dB | All probes and small active dipoles | |
| | 15 V/m to 100 V/m 43.5 GHz to 45.5 GHz | 0.35 dB | All probes and small active dipoles | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|--|------------------|
| FIELD STRENGTH (continued) Magnetic Field | 0.05 mA/m to 2.1 A/m 10 Hz to 100 Hz | 1.35 dB | <u>TEM Cells</u> The maximum frequency and field strength level is determined by the size of the probe. Electrically small probes | Teddington |
| | 0.05 mA/m to 2.1 A/m 100 Hz to 500 Hz | 0.76 dB | Electrically small probes | |
| | 0.05 mA/m to 2.1 A/m 500 Hz to 10 kHz | 0.68 dB | Electrically small probes | |
| | 0.05 mA/m to 2.1 A/m 10 kHz to 300 MHz | 0.68 dB | Electrically small probes | |
| | 0.021 mA/m to 1.8 A/m 180 MHz to 1000 MHz | 0.80 dB | <u>Tapered TEM Cell</u> The maximum frequency and field strength level is determined by the size of the probe. Electrically small probes | |
| | 0.021 mA/m to 1.3 A/m 1000 MHz to 3000 MHz | 0.80 dB | Electrically small probes | |
| | 0.03 mA/m to 1.0 A/m 240 MHz to 270 MHz | 0.65 dB | <u>Anechoic Chambers</u> The Listed Field levels are derived from the lowest unsaturated maximum power in each range. The actual achievable level may be up to 20 % greater than the stated limit. All probes and small active dipoles | |
| | 0.03 mA/m to 1.3 A/m 270 MHz to 350 MHz | 0.65 dB | All probes and small active dipoles | |
| | 0.03 mA/m to 1.5 A/m 350 MHz to 500 MHz | 0.65 dB | All probes and small active dipoles | |
| | 0.03 mA/m to 1.0 A/m 450 MHz to 550 MHz | 0.62 dB | All probes and small active dipoles | |
| | 0.03 mA/m to 1.4 A/m 550 MHz to 750 MHz | 0.62 dB | All probes and small active dipoles | |
| | 0.03 mA/m to 1.4 A/m 750 MHz to 950 MHz | 0.62 dB | All probes and small active dipoles | |
| | 0.03 mA/m to 1.0 A/m 950 MHz to 1200 MHz | 0.62 dB | All probes and small active dipoles | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|-------------------------------------|-------------------|
| FIELD STRENGTH (continued) Magnetic Field (continued) | 0.03 mA/m to 1.9 A/m <i>1100 MHz to 1250 MHz</i> | 0.47 dB | All probes and small active dipoles | Teddington |
| | 0.03 mA/m to 2.1 A/m <i>1250 MHz to 1700 MHz</i> | 0.47 dB | All probes and small active dipoles | |
| | 0.03 mA/m to 2.4 A/m <i>1700 MHz to 2600 MHz</i> | 0.47 dB | All probes and small active dipoles | |
| <u>PFD and Field Strength Indicating Meters</u> | | | | |
| Injected DC Voltage | 0.15 mV to 1 V | 0.47 % | Meter Sensitivity | |
| Measured DC Voltage | 0.09 V to 3.5 V | 0.44 % | Analogue output voltage for FSD | |
| DC Current | 5 mA to 100 mA | 0.10 % | Battery Charge | |
| AC Voltage | 1 V to 5 V, <i>8 kHz to 12 kHz</i> | 0.60 % | Low frequency source check | |
| Frequency | 8 kHz to 12 kHz | 0.012 % | Low frequency source check | |
| Time Interval | 2 s to 660 s | 0.10 s | Response time and maximum hold | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---|--|
| ANTENNA GAIN <i>Service Reference EF03</i> | | | All measurements are performed at 23 °C | The uncertainties apply to calibrations covering a waveguide bandwidth. Teddington |
| Waveguide Feed | 0 dB to 23 dB <i>2.6 GHz to 3.95 GHz</i> | 0.050 dB | Waveguide No 10 | |
| | 0 dB to 24 dB <i>3.3 GHz to 4.9 GHz</i> | 0.050 dB | Waveguide No 11A | |
| | 0 dB to 25 dB <i>3.95 GHz to 5.85 GHz</i> | 0.050 dB | Waveguide No 12 | |
| | 0 dB to 26 dB <i>5.4 GHz to 8.2 GHz</i> | 0.050 dB | Waveguide No 14 | |
| | 0 dB to 27 dB <i>7.05 GHz to 10.0 GHz</i> | 0.050 dB | Waveguide No 15 | |
| | 0 dB to 28 dB <i>8.2 GHz to 12.4 GHz</i> | 0.050 dB | Waveguide No 16 | |
| | 0 dB to 29 dB <i>10.0 GHz to 15.0 GHz</i> | 0.050 dB | Waveguide No 17 | |
| | 0 dB to 29 dB <i>12.4 GHz to 18.0 GHz</i> | 0.040 dB | Waveguide No 18 | |
| | 0 dB to 31 dB <i>18.0 GHz to 26.5 GHz</i> | 0.040 dB | Waveguide No 20 | |
| | 0 dB to 33 dB <i>26.5 GHz to 40.0 GHz</i> | 0.040 dB | Waveguide No 22 | |
| | 0 dB to 34 dB <i>33 GHz to 50 GHz</i> <i>0 dB to 35 dB</i> | 0.060 dB | Waveguide No 23 | |
| | <i>40 GHz to 60 GHz</i> | 0.10 dB | Waveguide No 24 | |
| | <i>0 dB to 36 dB</i> <i>50 GHz to 75 GHz</i> | 0.10 dB | Waveguide No 25 | |
| | <i>0 dB to 37 dB</i> <i>75 GHz to 110 GHz</i> | 0.10 dB | Waveguide No 27 | |
| Coaxial Feed | 0 dB to 28 dB <i>1 GHz to 18 GHz</i> | 0.050 dB | 50 Ω APC-7 or Type N connectors | |
| | 0 dB to 28 dB <i>1 GHz to 26.5 GHz</i> | 0.050 dB | 50 Ω 3.5 mm connector | |
| | <i>0 dB to 28 dB</i> <i>1 GHz to 40 GHz</i> | 0.09 dB | 50 Ω 2.92 mm connector | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---------------------------------------|--|--|-------------------|
| Coaxial Feed (cont'd) | 0 dB to 28 dB 2.6 GHz to 50 GHz | 0.10 dB | 50 Ω 2.4 mm connector Devices fitted with coaxial connectors other than those listed may be calibrated but the uncertainties may be increased | Teddington |
| EMC ANTENNA CALIBRATION <i>Service Reference EF04</i> | | | | |
| Waveguide Feed | 0 dB to 21 dB 2.6 GHz to 3.95 GHz | 0.70 dB | Waveguide No 10 | |
| | 0 dB to 22 dB 3.3 GHz to 4.9 GHz | 0.70 dB | Waveguide No 11A | |
| | 0 dB to 23 dB 3.95 GHz to 5.85 GHz | 0.70 dB | Waveguide No 12 | |
| | 0 dB to 24 dB 5.4 GHz to 8.2 GHz | 0.70 dB | Waveguide No 14 | |
| | 0 dB to 25 dB 7.05 GHz to 10.0 GHz | 0.70 dB | Waveguide No 15 | |
| | 0 dB to 26 dB 8.2 GHz to 12.4 GHz | 0.70 dB | Waveguide No 16 | |
| | 0 dB to 27 dB 10.0 GHz to 15.0 GHz | 0.70 dB | Waveguide No 17 | |
| | 0 dB to 28 dB 12.4 GHz to 18.0 GHz | 0.70 dB | Waveguide No 18 | |
| | 0 dB to 30 dB 18.0 GHz to 26.5 GHz | 0.70 dB | Waveguide No 20 | |
| | 0 dB to 31 dB 26.5 GHz to 40.0 GHz | 0.70 dB | Waveguide No 22 | |
| | 0 dB to 31 dB 43.5 GHz to 45.5 GHz | 0.70 dB | Waveguide No 23 | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 **Issue date:** 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks |
|--|---|--|--|
| Coaxial Feed | 0 dB to 28 dB <i>1 GHz to 18 GHz</i> | 0.80 dB (0.60 dB for conical log spiral antennas) | 50 Ω APC-7 or Type N connectors 50 Ω 3.5 mm connectors 50 Ω 2.92 mm connectors For coaxially fed antennas the antenna factor is calculated from the antenna gain. Devices fitted with coaxial connectors other than those listed above may be calibrated but the uncertainties may be increased. |
| | 0 dB to 30 dB <i>1 GHz to 26.5 GHz</i> | 0.80 dB (0.60 dB for conical log spiral antennas) | |
| | 0 dB to 30 dB <i>18 GHz to 40 GHz</i> | 0.80 dB | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---|-------------------|
| ANTENNA COMPLEX REFLECTION COEFFICIENT | 0 to 0.5, real and imaginary parts, higher reflections with increased uncertainty. | | The uncertainties for complex reflection coefficient apply to both real and imaginary parts. All measurements are performed at 23 °C | Teddington |
| Waveguide feed | 2.6 GHz to 3.95 GHz 3.3 GHz to 4.9 GHz 3.95 GHz to 5.85 GHz 5.4 GHz to 8.2 GHz 7.05 GHz to 10.0 GHz 8.2 GHz to 12.4 GHz 10.0 GHz to 15.0 GHz 12.4 GHz to 18.0 GHz 18.0 GHz to 26.5 GHz 26.5 GHz to 40.0 GHz 33.0 GHz to 50.0 GHz 40 GHz to 60 GHz 50 GHz to 75 GHz 75 GHz to 110 GHz | 0.0080 0.0080 0.0080 0.0080 0.0080 0.0080 0.0080 0.0080 0.0080 0.0080 0.013 0.015 0.015 0.015 | Waveguide No 10 Waveguide No 11A Waveguide No 12 Waveguide No 14 Waveguide No 15 Waveguide No 16 Waveguide No 17 Waveguide No 18 Waveguide No 20 Waveguide No 22 Waveguide No 23 Waveguide No 24 Waveguide No 25 Waveguide No 27 | |
| 7 mm coaxial feed | 1.0 GHz to 1.5 GHz 1.5 GHz to 18.0 GHz 1.0 GHz to 8.2 GHz 8.2 GHz to 18.0 GHz | 0.015 0.011 0.013 0.018 | 50 Ω Type N connectors 50 Ω Type N connectors 50 Ω GPC-7 connectors 50 Ω GPC-7 connectors | |
| 3.5 mm coaxial feed | 1.0 GHz to 8.2 GHz 8.2 GHz to 18 GHz 18.0 GHz to 26.5 GHz | 0.010 0.020 0.029 | 50 Ω GPC-3.5 connectors 50 Ω GPC-3.5 connectors 50 Ω GPC-3.5 connectors | |
| 2.92 mm coaxial feed | 1 GHz to 26.5 GHz 26.5 GHz to 40 GHz | 0.028 0.043 | 50 Ω 2.92 mm connectors 50 Ω 2.92 mm connectors | |
| 2.4 mm coaxial feed | 1 GHz to 26.5 GHz 26.5 GHz to 40 GHz 40 GHz to 50 GHz | 0.021 0.041 0.056 | 50 Ω 2.4 mm connectors 50 Ω 2.4 mm connectors 50 Ω 2.4 mm connectors Devices fitted with coaxial connectors other than those listed above may be calibrated but the uncertainties may be increased. | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|---|-------------------|
| ANTENNA FACTOR <i>Service Reference: EF06</i> | -30 dB/m to +80 dB/m | | Calibrations to meet the requirements of ARP 958, ANSI C63.5 (2006 & 2017), CISPR 16-1-6. | Teddington |
| Linear dipole | 20 MHz to 500 MHz 500 MHz to 1000 MHz | 0.35 dB 0.50 dB | Defined height, tuned element | |
| Linear dipole | 20 MHz to 40 MHz 40 MHz to 1000 MHz | 0.70 dB 0.50 dB | Free-space, tuned element | |
| Biconical antenna | 20 MHz to 300 MHz | 0.50 dB | Free-space or defined height (vs. SRDs) | |
| Mini-Biconical antenna | 300 MHz to 6 GHz | 0.50 dB | Free-space | |
| LPDA antenna | 80 MHz to 200 MHz | 0.70 dB | Free-space | |
| LPDA antenna | 200 MHz to 6 GHz | 0.50 dB | Free-space | |
| Biconical, Hybrid and LPDA | 30 MHz to 1 GHz | 1.0 dB | Standard Site method, horizontal (ANSI C63.5 2006 & 2017) | |
| Hybrid antenna | 20 MHz to 6 GHz | 0.70 dB | Free-space | |
| Spiral antenna | 100 MHz to 1 GHz | 1.0 dB | Free-space | |
| Horn antenna | 200 MHz to 2 GHz | 1.0 dB | Free-space | |
| Rod antenna <i>Service Reference: EF11</i> | 100 Hz to 100 MHz | 1.0 dB 1.2 dB | Plane wave E-field ECSM | |
| DUAL ANTENNA FACTOR <i>Service Reference: EF06</i> | | | For use in NSA measurements | |
| Biconical, LPDA and hybrid antennas | 30 MHz to 1000 MHz | 1.0 dB | Standard Site method, horizontal polarisation | |
| Biconical, LPDA and hybrid antennas | 30 MHz to 1000 MHz | 1.5 dB | Standard Site method, vertical polarisation | |
| RADIATION PATTERNS <i>Service Reference: EF13</i> | Gain < +10 dBi | | | |
| | 500 MHz to 18 GHz | 0.35 dB | From 0 dB to -6 dB, relative to maximum level. | |
| | 500 MHz to 18 GHz | 1.0 dB | From -6 dB to -15 dB, relative to maximum level. | |
| | Note: other parameters may be derived from pattern data, such as Directivity and Efficiency. The uncertainty is related to the change in gain relative to its maximum value. | | | |
| REFLECTION COEFFICIENT <i>Service Reference: EF06</i> | Gamma: 0 to 1 0.3 MHz to 6 GHz | 0.05 | 50 Ω Type N connectors. AUTs with other coaxial connectors can be calibrated but the uncertainty may be increased. | |
| Antenna Balance (Symmetry) | 30 MHz to 300 MHz For values within ± 2 dB | 0.25 dB | ANSI C 63.5 (2006 & 2017) and CISPR 16-1-4 | |
| VSWR (Derived from S11) | 0.3 MHz to 6 GHz For VSWR value = 1.2 to 3 | 0.02 to 0.2 | Uncertainty will be increased for VSWR > 3 | |
| E-field emitters | 30 MHz to 6 GHz 10 kHz to 6 GHz | 1.5 dB 1.0 dB | Radiated, depends on SNR Conducted | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|--|------------------|
| ATTENUATION <i>Service Reference EG03</i> | | | | |
| Coaxial Line | 0 dB to 100 dB <i>0.5 MHz to 18 GHz</i> | (0.00060 dB per 10 dB) + 0.00060 dB | 50 Ω 14 mm Coaxial Line (GR-900 connector) up to 8.0 GHz. | Teddington |
| | 100 dB to 120 dB <i>0.5 MHz to 100 MHz</i> | 0.00080 dB per 10 dB | 50 Ω 7 mm Coaxial Line: Standard N-Type connector up to 12.4 GHz; Precision N-type and GPC-7 connectors to 18 GHz. | |
| | 120 dB to 130 dB <i>0.5 MHz to 100 MHz</i> | (0.0010 dB per 10 dB) + 0.010 dB | 3.5 mm Coaxial Line up to 26.5 GHz (GPC-3.5 connectors) | |
| | 0 dB to 80 dB <i>18 GHz to 40 GHz</i> | (0.00060 dB per 10 dB) + 0.00060 dB | 2.92 mm Coaxial Line up to 40 GHz (K connector). | |
| Waveguide | 0 dB to 90 dB | | 2.4 mm Coaxial Line up to 40 GHz (2.4 mm connector). | |
| | 2.6 GHz to 3.95 GHz 3.3 GHz to 4.9 GHz 3.95 GHz to 5.85 GHz 5.85 GHz to 8.2 GHz 7.05 GHz to 10.0 GHz 8.2 GHz to 12.4 GHz 10.0 GHz to 15.0 GHz 12.4 GHz to 18.0 GHz | (0.00060 dB per 10 dB) + 0.00060 dB | Waveguide No 10 Waveguide No 11A Waveguide No 12 Waveguide No 14 Waveguide No 15 Waveguide No 16 Waveguide No 17 Waveguide No 18 | |
| | 0 dB to 80 dB <i>18 GHz to 26.5 GHz 26.5 GHz to 40 GHz</i> | (0.00060 dB per 10 dB) + 0.00060 dB | Waveguide No 20 Waveguide No 22 | |
| | | | NOTE The uncertainties for attenuation apply to the measurement of a device that is well matched to the ideal characteristic impedance of the transmission line system. The quoted uncertainty will be increased for other devices to account for mismatch and repeatability, when these contributions exceed that which has been allowed for in this schedule. | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation

issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|--|-------------------|
| RF POWER <i>Service Reference EG04</i> | | | | Teddington |
| Calibration factor and effective efficiency - guided wave systems | 0.1 mW to 10 mW 8.2 GHz to 12.4 GHz 12.4 GHz to 18.0 GHz 18.0 GHz to 26.5 GHz 26.5 GHz to 40.0 GHz 40 GHz to 50 GHz 50.0 GHz to 75 GHz 75 GHz to 110.0 GHz | 0.40 % 0.40 % 0.50 % 0.50 % 0.90 % 1.2 % 1.6 % | Waveguide No 16 Waveguide No 18 Waveguide No 20 Waveguide No 22 Waveguide No 23 Waveguide No 25 Waveguide No 27 | |
| RF POWER Calibration factor and effective efficiency - coaxial line systems | Nominal power range 0.01 mW to 10 mW 10 MHz to 100 MHz 100 MHz to 4 GHz 4 GHz to 8 GHz 8 GHz to 12 GHz 12 GHz to 15 GHz 15 GHz to 18 GHz | 0.25 % 0.40 % 0.45 % 0.55 % 0.60 % 0.70 % | Calibration of 7mm power sensors and thermistor mounts against the NPL 7 mm calorimeter. The uncertainties apply to devices with type N connectors with VRC less than 0.01 in a 50 Ω coaxial system. The uncertainties may be increased for devices with a higher VRC or fitted with other connector types. | |
| COMPLEX REFLECTION COEFFICIENT (in support of attenuation and Power calibrations) | | | | |
| Magnitude | 0 to 1.0 2.6 GHz to 3.95 GHz 3.3 GHz to 4.9 GHz 3.95 GHz to 5.85 GHz 5.85 GHz to 8.2 GHz 7.05 GHz to 10.0 GHz 8.2 GHz to 12.4 GHz 10.0 GHz to 15.0 GHz 12.4 GHz to 18.0 GHz 18.0 GHz to 26.5 GHz 26.5 GHz to 40.0 GHz 40.0 GHz to 50 GHz 60.0 GHz to 62 GHz 75 GHz to 110 GHz 10 MHz to 18 GHz 10 MHz to 26.5 GHz | 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 0.0040 | Waveguide No 10 Waveguide No 11A Waveguide No 12 Waveguide No 14 Waveguide No 15 Waveguide No 16 Waveguide No 17 Waveguide No 18 Waveguide No 20 Waveguide No 22 Waveguide No 23 Waveguide No 25 Waveguide No27 50 Ω APC-7 or Type N Connectors 50 Ω 3.5 mm connectors | |
| Phase | -180° to +180° <i>Frequency range as for magnitude</i> | $\sin^{-1} \frac{(\text{magnitude uncertainty})^p}{\text{magnitude}}$ | If the magnitude is less than its uncertainty, then the phase uncertainty is 180° | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|---|------------------|
| PULSE CHARACTERISTICS | | | | |
| <i>Service Reference EU01</i> | | | | |
| Transition Duration | 11 ps to 60 ps 60 ps to 10 ns | 0.90 ps 0.50 % + 1.4 ps | Calibration of pulse generators | Teddington |
| Pulse Amplitude | 50 mV to 800 mV 800 mV to 1 V | 0.30 % 0.50 % | | |
| Pulse Aberrations (e.g. pre-shoot, overshoot, pulse settling error) | Pulse settling error Others | 0.050 % of amplitude 0.10 % of amplitude | | |
| TIME AND FREQUENCY | | | | |
| <i>Service Reference TT02/TT04</i> | | | | |
| Characterisation of GPS disciplined oscillators and frequency standards | | | | |
| Time offset | From UTC (NPL) | 2.0 ns | Calibration of frequency standards with a 1 pps output can also be undertaken. | Customers' sites |
| Time offset | From UTC | 10 ns | | |
| Frequency | 5 MHz and 10 MHz | 5.0×10^{-14} <i>Minimum measurement period 24 hours.</i> | | |
| Time delay (coaxial cables) | 0 ns to 300 ns | 1.0 ns | | |
| Remote characterisation of GPS disciplined oscillators and frequency standards | | | | |
| Time offset | Weekly values relative to UTC (NPL) | 20 ns | The capability relates to a remote common-view service where NPL-supplied software gathers data and returns it to NPL for processing. The user is supplied with instructions for the setting up of the equipment and the antenna. | Customers' sites |
| Time offset | Weekly values relative to estimated UTC | 40 ns | | |
| Time offset | Post-processed values relative to corrected UTC data | 10 ns | | |
| Frequency | 5 MHz and 10 MHz | 1.0×10^{-13} <i>Minimum measurement period 24 hours.</i> | Calibration of frequency standards with a 1 pps output can also be undertaken. | Customers' sites |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | | | | | | | | | | | | | | |
|---|------------|--|---------|---------------|--------|--------|--------|--------|--------|--------|--------|--------|--------|----------|------------|------------|-----------|------------|
| Primary Impedance Measurement System (PIMMS) Service Reference EG02 / EG06 | | | | | | | | | | | | | | | | | | |
| NOTE | | | | | | | | | | | | | | | | | | |
| <p>For the linear voltage reflection and transmission coefficient measurands (i.e. complex-valued S-parameters) described in this section of the Schedule, the uncertainty is shown as an interval of values, where a selected value within the interval represents an expanded uncertainty at a level of confidence of approximately 95%. Furthermore, a selected value within the interval will represent the uncertainty applied equally and simultaneously to <i>both</i> the Real and Imaginary parts of the S-parameter. The uncertainty value therefore defines a circular region of uncertainty, in the appropriate complex S-parameter plane, centred on the measured, quoted, mean value with radius equal to the stated expanded uncertainty. The corresponding k value will not be less than 2.5.</p> <p>For Voltage Reflection Coefficients (VRCs), the stated uncertainty is assumed here to be independent of the nominal VRC , so a single interval is presented applicable for all VRC in the range $0 \leq VRC \leq 1$. For Voltage Transmission Coefficients (VTCs), the stated uncertainty is dependent on the nominal VTC , so uncertainty intervals are presented for selected, representative, values of VTC in the range, $0 \leq VTC \leq 1$.</p> | | | | | | | | | | | | | | | | | | |
| Voltage Reflection Coefficient Magnitude (VRC) in 50 Ω coaxial systems | | | | | | | | | | | | | | | | | | |
| Connector Type | VRC | Frequency (GHz) | | | | | | | | | | | | | | Teddington | | |
| | | 0.01 | 0.02 | 0.04 | 0.045 | 0.05 | 0.07 | 0.09 | 0.1 | 0.2 | 0.3 | 0.5 | 1.0 | 1 to 7.5 | 7.5 to 8.5 | | 8.5 to 18 | 18 to 26.5 |
| 7-16 | 0 to 0.5 | | | | 0.002 | 0.002 | 0.002 | 0.002 | 0.002 | 0.002 | 0.0015 | 0.0015 | 0.001 | 0.001 | | | | |
| 7-16 | 0.5 to 0.7 | | | | 0.0025 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | 0.002 | 0.002 | 0.002 | 0.002 | | | | |
| 7-16 | 0.8 | | | | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | | | | |
| 7-16 | 0.9 | | | | 0.0035 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | 0.003 | 0.003 | 0.003 | 0.003 | | | | |
| 7-16 | 1.0 | | | | 0.004 | 0.004 | 0.004 | 0.004 | 0.004 | 0.004 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | | | | |
| GR900 | 0 to 0.5 | | | | 0.002 | 0.002 | 0.002 | 0.002 | 0.002 | 0.002 | 0.0015 | 0.0015 | 0.001 | 0.001 | 0.001 | | | |
| GR900 | 0.5 to 0.7 | | | | 0.0025 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | 0.002 | 0.002 | 0.002 | 0.002 | 0.002 | | | |
| GR900 | 0.8 | | | | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | | | |
| GR900 | 0.9 | | | | 0.0035 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | | | |
| GR900 | 1.0 | | | | 0.004 | 0.004 | 0.004 | 0.004 | 0.004 | 0.004 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | | | |
| GPC-7 | 0 to 0.6 | 0.0055 | 0.0045 | 0.004 | 0.004 | 0.004 | 0.004 | 0.0035 | 0.0035 | 0.0035 | 0.003 | 0.003 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | | |
| GPC-7 | 0.6 to 0.8 | 0.0055 | 0.005 | 0.0045 | 0.0045 | 0.0045 | 0.0045 | 0.004 | 0.004 | 0.004 | 0.0035 | 0.0035 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | |
| GPC-7 | 0.9 | 0.006 | 0.0055 | 0.005 | 0.005 | 0.005 | 0.005 | 0.0045 | 0.0045 | 0.0045 | 0.004 | 0.004 | 0.004 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | |
| GPC-7 | 1.0 | 0.0065 | 0.0055 | 0.005 | 0.005 | 0.005 | 0.005 | 0.005 | 0.005 | 0.005 | 0.0045 | 0.0045 | 0.0045 | 0.004 | 0.004 | 0.004 | 0.004 | |
| Type-N | 0 to 0.6 | 0.0055 | 0.0045 | 0.004 | 0.004 | 0.004 | 0.004 | 0.0035 | 0.0035 | 0.0035 | 0.003 | 0.003 | 0.0025 | 0.0025 | 0.0025 | 0.0025 | | |
| Type-N | 0.6 to 0.8 | 0.0055 | 0.005 | 0.0045 | 0.0045 | 0.0045 | 0.0045 | 0.004 | 0.004 | 0.004 | 0.0035 | 0.0035 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | |
| Type-N | 0.9 | 0.006 | 0.0055 | 0.005 | 0.005 | 0.005 | 0.005 | 0.0045 | 0.0045 | 0.0045 | 0.004 | 0.004 | 0.004 | 0.0035 | 0.0035 | 0.0035 | 0.0035 | |
| Type-N | 1.0 | 0.0065 | 0.0055 | 0.005 | 0.005 | 0.005 | 0.005 | 0.005 | 0.005 | 0.005 | 0.0045 | 0.0045 | 0.0045 | 0.004 | 0.004 | 0.004 | 0.004 | |
| 3.5 mm | 0 to 0.8 | | | | 0.01 | 0.0095 | 0.009 | 0.0085 | 0.008 | 0.007 | 0.007 | 0.0065 | 0.006 | 0.005 | 0.005 | 0.005 | 0.005 | 0.005 |
| 3.5 mm | 0.8 to 1.0 | | | | 0.01 | 0.01 | 0.0095 | 0.009 | 0.0085 | 0.0075 | 0.0075 | 0.007 | 0.0065 | 0.0055 | 0.0055 | 0.0055 | 0.0055 | 0.0055 |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---------------------------|---------------|
| Primary Impedance Measurement System (PIMMS) (continued) | | | | |
| Voltage Reflection Coefficient Magnitude (VRC) in waveguide systems | | | | |
| Waveguide size | Frequency (GHz) | | | |
| | 5.4 to 8.2 | 8.2 to 12.4 | 12.4 to 18 | 18 to 26.5 |
| R70 | 0.001 | | | |
| R100 | | 0.001 | | |
| R140 | | | 0.0015 | |
| R220 | | | | 0.0015 |
| R320 | | | | 0.003 |
| Voltage Transmission Coefficient Magnitude (VTC) in 50 Ω coaxial systems | | | | |
| Connector Type | VTC and corresponding insertion loss (dB) | Frequency | Minimum uncertainty (VTC) | Teddington |
| 7-16 | 1 (0 dB) | 10 MHz to 7.5 GHz | 0.00040 | |
| 7-16 | 0.316 (10 dB) | 10 MHz to 7.5 GHz | 0.00035 | |
| 7-16 | 0.1 (20 dB) | 10 MHz to 7.5 GHz | 0.00020 | |
| 7-16 | 0.0316 (30 dB) | 10 MHz to 7.5 GHz | 0.00010 | |
| 7-16 | 0.01 (40 dB) | 10 MHz 100 MHz | 0.00010 | |
| 7-16 | 0.01 (40 dB) | 100 MHz to 7.5 GHz | 0.00005 | |
| 7-16 | 0.00316 (50 dB) | 10 MHz to 100 MHz | 0.00010 | |
| 7-16 | 0.00316 (50 dB) | 100 MHz to 7.5 GHz | 0.00004 | |
| 14 mm | 1 (0 dB) | 45 MHz to 8.5 GHz | 0.00040 | |
| 14 mm | 0.316 (10 dB) | 45 MHz to 8.5 GHz | 0.00035 | |
| 14 mm | 0.1 (20 dB) | 45 MHz to 8.5 GHz | 0.00020 | |
| 14 mm | 0.0316 (30 dB) | 45 MHz to 8.5 GHz | 0.00010 | |
| 14 mm | 0.01 (40 dB) | 45 MHz 100 MHz | 0.00010 | |
| 14 mm | 0.01 (40 dB) | 100 MHz to 8.5 GHz | 0.00005 | |
| 14 mm | 0.00316 (50 dB) | 45 MHz to 100 MHz | 0.00010 | |
| 14 mm | 0.00316 (50 dB) | 100 MHz to 8.5 GHz | 0.00004 | |
| Type-N | 1 (0 dB) | 10 MHz to 18 GHz | 0.00040 | |
| Type-N | 0.316 (10 dB) | 10 MHz to 18 GHz | 0.00035 | |
| Type-N | 0.1 (20 dB) | 10 MHz to 18 GHz | 0.00020 | |
| Type-N | 0.0316 (30 dB) | 10 MHz to 18 GHz | 0.00010 | |
| Type-N | 0.01 (40 dB) | 10 MHz to 100 MHz | 0.00010 | |
| Type-N | 0.01 (40 dB) | 100 MHz to 18 GHz | 0.00005 | |
| Type-N | 0.00316 (50 dB) | 10 MHz to 100 MHz | 0.00010 | |
| Type-N | 0.00316 (50 dB) | 100 MHz to 18 GHz | 0.00004 | |
| 3.5 mm | 1 (0 dB) | 45 MHz to 26.5 GHz | 0.0010 | |
| 3.5 mm | 0.316 (10 dB) | 45 MHz to 26.5 GHz | 0.00040 | |
| 3.5 mm | 0.1 (20 dB) | 45 MHz to 100 MHz | 0.00025 | |
| 3.5 mm | 0.1 (20 dB) | 100 MHz to 26.5 GHz | 0.00020 | |
| 3.5 mm | 0.0316 (30 dB) | 45 MHz to 100 MHz | 0.00015 | |
| 3.5 mm | 0.0316 (30 dB) | 100 MHz to 26.5 GHz | 0.00010 | |
| 3.5 mm | 0.01 (40 dB) | 45 MHz to 100 MHz | 0.00006 | |
| 3.5 mm | 0.01 (40 dB) | 100 MHz to 26.5 GHz | 0.00005 | |
| 3.5 mm | 0.00316 (50 dB) | 45 MHz to 100 MHz | 0.00010 | |
| 3.5 mm | 0.00316 (50 dB) | 100 MHz to 26.5 GHz | 0.00004 | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|--|------------------|
| Primary Impedance Measurement System (PIMMS) (continued) | | | | |
| Voltage Transmission Coefficient Magnitude (VTC) in waveguide systems | | | | |
| Waveguide size | VTC and corresponding insertion loss (dB) | Frequency | Minimum uncertainty (VTC) | Teddington |
| R100 | 1 (0 dB) | 8.2 GHz to 12.4 GHz | 0.0010 | |
| R100 | 0.316 (10 dB) | 8.2 GHz to 12.4 GHz | 0.00040 | |
| R100 | 0.1 (20 dB) | 8.2 GHz to 12.4 GHz | 0.00020 | |
| R100 | 0.0316 (30 dB) | 8.2 GHz to 12.4 GHz | 0.00010 | |
| R100 | 0.01 (40 dB) | 8.2 GHz to 12.4 GHz | 0.00006 | |
| R100 | 0.00316 (50 dB) | 8.2 GHz to 12.4 GHz | 0.00004 | |
| R140 | 1 (0 dB) | 12.4 GHz to 18 GHz | 0.0025 | |
| R140 | 0.316 (10 dB) | 12.4 GHz to 18 GHz | 0.00075 | |
| R140 | 0.1 (20 dB) | 12.4 GHz to 18 GHz | 0.00030 | |
| R140 | 0.0316 (30 dB) | 12.4 GHz to 18 GHz | 0.00010 | |
| R140 | 0.01 (40 dB) | 12.4 GHz to 18 GHz | 0.00006 | |
| R140 | 0.00316 (50 dB) | 12.4 GHz to 18 GHz | 0.00004 | |
| R220 | 1 (0 dB) | 18 GHz to 26.5 GHz | 0.0030 | |
| R220 | 0.316 (10 dB) | 18 GHz to 26.5 GHz | 0.00075 | |
| R220 | 0.1 (20 dB) | 18 GHz to 26.5 GHz | 0.00030 | |
| R220 | 0.0316 (30 dB) | 18 GHz to 26.5 GHz | 0.00010 | |
| R220 | 0.01 (40 dB) | 18 GHz to 26.5 GHz | 0.00006 | |
| R220 | 0.00316 (50 dB) | 18 GHz to 26.5 GHz | 0.00004 | |
| R320 | 1 (0 dB) | 26.5 GHz to 40 GHz | 0.0030 | |
| R320 | 0.316 (10 dB) | 26.5 GHz to 40 GHz | 0.00075 | |
| R320 | 0.1 (20 dB) | 26.5 GHz to 40 GHz | 0.00030 | |
| R320 | 0.0316 (30 dB) | 26.5 GHz to 40 GHz | 0.00010 | |
| R320 | 0.01 (40 dB) | 26.5 GHz to 40 GHz | 0.00006 | |
| R320 | 0.00316 (50 dB) | 26.5 GHz to 40 GHz | 0.00004 | |
| Mechanically-derived characteristic impedance of the following coaxial lines: | | | | |
| 7-16 | 49.8 Ω to 50.2 Ω | 0.009 Ω | Based on measurements of the diameters of airline conductors, these and associated uncertainties will also be reported. These measurements are made using air gauging techniques. | |
| 14 mm | 49.8 Ω to 50.2 Ω | 0.010 Ω | | |
| Type-N | 27.7 Ω to 28.3 Ω | 0.018 Ω | | |
| Type-N or GPC-7 | 49.6 Ω to 50.4 Ω | 0.016 Ω | | |
| Type-N | 74.4 Ω to 75.6 Ω | 0.031 Ω | | |
| Type-N | 99.2 Ω to 100.8 Ω | 0.078 Ω | | |
| 3.5 mm | 49.2 Ω to 50.8 Ω | 0.038 Ω | | |
| 2.92 mm | 48.9 Ω to 50.9 Ω | 0.048 Ω | | |
| 2.4 mm | 48.9 Ω to 51.4 Ω | 0.063 Ω | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---|-------------------|
| LENGTH | | | | Teddington |
| Gauge blocks: millimetre <i>Service Reference: LD01</i> | As BS EN ISO 3650:1999 0.5 mm to 25 mm 25 mm to 50 mm 50 mm to 75 mm 75 mm to 100 mm | 0.020 μm 0.020 μm to 0.022 μm 0.022 μm to 0.025 μm 0.025 μm to 0.029 μm | Measurement of central length by interferometry. Measured twice, wrung to a platen by each of the two measuring faces in turn, and the mean of these two measurements stated on the certificate. | |
| Gauge blocks: inch <i>Service Reference: LD01</i> | As BS 4311:2007 0.01 in to 0.4 in 0.4 in to 1 in 2 in 3 in 4 in | 0.76 $\mu\text{ inch}$ 0.76 $\mu\text{ inch}$ to 0.78 $\mu\text{ inch}$ 0.86 $\mu\text{ inch}$ 0.98 $\mu\text{ inch}$ 1.14 $\mu\text{ inch}$ | | |
| Gauge blocks: millimetre <i>Service Reference: LD01</i> | As BS EN ISO 3650:1999 0.5 mm to 10 mm 10 mm to 25 mm 25 mm to 50 mm 50 mm to 75 mm 75 mm to 100 mm | 0.032 μm to 0.033 μm 0.033 μm to 0.037 μm 0.037 μm to 0.050 μm 0.050 μm to 0.065 μm 0.065 μm to 0.082 μm | Measurement of central length by mechanical comparison. | |
| Gauge blocks: inch <i>Service Reference: LD01</i> | As BS 4311:2007 0.01 to 0.4 in 0.4 in to 1 in 2 in 3 in 4 in | 1.26 to 1.30 $\mu\text{ inch}$ 1.30 to 1.47 $\mu\text{ inch}$ 1.97 $\mu\text{ inch}$ 2.60 $\mu\text{ inch}$ 3.29 $\mu\text{ inch}$ | | |
| 1-D artefacts [Long gauge blocks and length bars]: length millimetre (and inch) <i>Service Reference: LD02</i> | As BS EN ISO 3650:1999 Grades K, 0 and 1; As BS 5317:1976 Reference and calibration grades 100 mm to 1200 mm | $Q[0.00012, 1.4 \times 10^{-7}L]$ mm | Measurement of length by CMM substitution method of gauges of length L (in mm) . | |
| 1-D artefacts [Long gauge blocks and length bars]: length millimetre (and inch) <i>Service Reference: LD02</i> | As BS EN ISO 3650:1999 Grades K, 0 and 1; As BS 5317:1976 Reference and calibration grades 100 mm to 1200 mm | $Q[0.00041, 2.0 \times 10^{-6}L]$ mm | Measurement of length by CMM method of gauges of length L (in mm) . | |
| Long gauge blocks: millimetre <i>Service Reference: LD05</i> | As BS EN ISO 3650:1999 Grades K, 0 and 1 100 mm to 1000 mm As BS EN ISO 3650:1999 Grades K, 0 and 1 100 mm to 1000 mm | $\sqrt{0.049^2 + (0.083 \times L)^2}$ μm i.e., 0.050 μm to 0.100 μm 0.116 + (0.263 $\times L$) μm i.e., 0.143 to 0.379 μm | Measurement of length by absolute interferometry of long gauge blocks of length L (in metres) to the stated standards. Measurement of length by interferometric comparison of long gauge blocks of length L (in metres) to the stated standards. | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---|-------------------|
| LENGTH (continued) | | | | Teddington |
| Length bars: millimetre <i>Service Reference: LD02</i> | As BS 5317:1976; Reference and calibration grades | | | |
| | 10 mm to 100 mm | $\sqrt{0.066^2 + (0.21 \times L)^2}$ μm i.e. 0.061 μm to 0.064 μm | | |
| | 100 mm to 1200 mm | $\sqrt{0.049^2 + (0.083 \times L)^2}$ μm i.e. 0.050 μm to 0.111 μm | | |
| | Above 100 mm up to 1200 mm | 0.116 + (0.263 x L) μm i.e. 0.143 to 0.432 μm | Measurement of length by interferometric comparison of length bars of length L (in metres) to the stated standards. | |
| Length bars: Inch <i>Service Reference: LD05</i> | As BS 1790:1961; Reference and calibration grades | | | |
| | 0.5 in to 4 in | $\sqrt{2.36^2 + (0.21 \times L)^2}$ μ inch i.e. 2.40 μ inch to 2.53 μ inch | Measurement of length by absolute interferometry of length bars of length L (in inches) to the stated standards. | |
| | 4 in to 48 in | $\sqrt{1.9^2 + (0.083 \times L)^2}$ μ inch i.e. 1.93 μ inch to 4.42 μ inch | | |
| | 4 in to 48 in | 4.57 + (0.263 x L) μ inch i.e. 5.62 μ inch to 17.2 μ inch | Measurement of length by interferometric comparison of long gauge blocks of length L (in inches) to the stated standards. | |
| Thermal expansion coefficient at 20 °C <i>Service Reference: LD03</i> | Expansion coefficient $9 \times 10^{-6} \text{ K}^{-1}$ to $13 \times 10^{-6} \text{ K}^{-1}$ | $(0.004 + 11/L + 0.000\ 007L)$ $\times 10^{-6} \text{ K}^{-1}$, where L is length of gauge block or length bar in millimetres. | The uncertainty applies to the measurement of the linear coefficient of thermal expansion, at 20 °C, of long series gauge blocks and length bars above 100 mm, up to 1200 mm (4 inch to 48 inch) which comply with the following standards: Reference and calibration grades of BS 1790:1961 (inch). Reference and calibration grades of BS 5317:1976 (millimetre). Grades K, 0 of ISO 3650:1998. | |
| Step gauges <i>Service Reference: LD04</i> | 210 mm to 1020 mm | $(100 + 232 L)$ nm, where L is length in metres | | |
| Thread measuring cylinders <i>Service Reference: LD07</i> | 0.05 mm to 5 mm diameter | 0.080 + $(0.0010 \times \text{diameter in mm}) \mu\text{m}$ | As BS 3777:1964 BS 5590:1978 and specials | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|---|---|-------------------|
| LENGTH (continued) | | | | Teddington |
| External cylinder Plain plug gauges (parallel) reference cylinders and rollers <i>Service Reference: LD07</i> | 0.1 mm to 100 mm diameter | 0.070 + (0.0011 x diameter in mm) μm | | |
| | 100 mm to 150 mm diameter | 0.05 + (0.0014 x diameter in mm) μm | | |
| Plain setting rings (parallel) <i>Service Reference: LD07</i> | 3 mm to 250 mm diameter | 0.070 + (0.0005 x diameter in mm) μm | As BS 4064:1966 and BS 4065:1966 Grade AA, and equivalent quality setting rings | |
| Stage micrometers and graticules <i>Service Reference: LR04</i> | 0 mm to 50 mm 50mm to 100mm 100mm to 150mm | 0.20 μm 0.30 μm 0.40 μm | | |
| Linewidth standards <i>Service Reference: LR03</i> | 0.5 μm to 10 μm 10 μm to 50 μm | 0.050 μm 0.10 μm | | |
| Reference stage graticules for image analysers <i>Service Reference: LR07</i> | Grid sizes 0 to 400 μm x 400 μm Spot sizes 3 μm to 48 μm | 0.10 μm 0.10 μm | | |
| Reference master screw plug and ring gauges (taper) to API specifications 5B and 7 <i>Service Reference: LD06</i> | 0 in to 20 in diameter | Pitch diameters 0.00037 in Major diameter 0.0002 in Minor diameter 0.0006 in Pitch 0.00012 in Taper 0.00011 in Flank angle 3.0 minutes of arc Stand off 0.00034 in | | |
| Receiver and position gauges, jigs and fixtures <i>Service Reference: LD10</i> | 1200 x 1000 x 700 | 0.36 + L/866 μm , where L is in mm 0.70 μm (using substitution method) 0.40 μm (using reversal method) | Measurements made using a coordinate measuring machine | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---|-------------------|
| ANGLE <i>Service Reference: LD08</i> Indexing tables Precision polygons Combination angle gauges Autocollimators <i>Visual and photoelectric</i> | From 0° to 360° 4 sides to 12 sides, excluding 7 and 11 sides 0° to 45° 0 minutes of arc to 10 minutes of arc | 0.040 seconds of arc 0.11 seconds of arc 0.30 seconds of arc 0.060 seconds of arc | As MOY/SCMI/18 and MOY/SCMI/45 | Teddington |
| FORM <i>Service Reference: LD07</i> Roundness reference standards Reference Sphere Diameter Back vertex focal length or power of a lens. <i>Service Reference: LR02</i> | 0 mm to 100 mm diameter 10 mm to 50 mm diameter ± 0.01 D to ± 25 D | 0.0050 µm 0.11 µm 0.0010 D to 0.010 D | Zygo interferometer and Hewlett Packard length measuring interferometer traceable to dimensional standards used to measure vertex of the back surface of a lens to the corresponding focus. | |
| Radius of curvature and sphericity of optical quality surfaces. <i>Service Reference: LR02</i> | 4 mm to 1000 mm radius of curvature. | 0.0020 mm, 26 nm for sphericity. | Test items are calibrated for radius of curvature and departure from spherical form using a Mark GPIxps Zygo interferometer in conjunction with a commercial laser length measuring interferometer. | |
| Power of small angle prisms. <i>Service Reference: LR02</i> | 0 to 20 prism dioptres (0° to 12° deviation). | 0.010 prism dioptres. | Measurements of small angle prisms are carried out using the a Mark GPIxps Zygo interferometer, auxiliary mirrors, a Hilger and Watts clinometer and calibration test pieces when necessary. | |
| Optical flatness <i>Service Reference: LR01</i> | 5 mm to 150 mm OD. 5 to 33mm 33 to 100mm 100 to 150mm | 14 nm 17 nm 20 nm | Flatness of optical quality surfaces using a Zygo Mark GPIxps interferometer housing a reference flat traceable to a liquid surface. | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|---|-------------------|
| OTHER MEASURING INSTRUMENTS, EQUIPMENT AND MACHINES | | | | |
| Laser frequency (Vacuum wavelength) <i>Service Reference: LL01</i> | Nominal wavelengths 500 nm to 2.0 μm | 1 part in 10^{13} | | Teddington |
| Laser interferometer systems <i>Service Reference: LL01</i> | 0 to 45 m <i>Compensated</i> <i>Uncompensated</i> | $\sqrt{0.08^2 + (0.2 \times L)^2} \mu\text{m}$ $\sqrt{0.08^2 + (0.12 \times L)^2} \mu\text{m}$ | $L = \text{length (m)}$ $L = \text{length (m)}$ | |
| Extensometer calibration rigs <i>Service Reference: MF06</i> | Displacements 0 mm to 1000 mm | $31 + (3.1 \times R)$ nm where R is the extension in mm | As BS EN ISO 9513:2012 | |
| INFRA-RED <i>Service Reference: OT05</i> | <i>Thermal Radiation</i> | 0.0020 (, <0.1) 0.0010 (, >0.9) | | |
| Wavenumber, ν for QA checks on mid-IR spectrophotometers | <i>Nominal Values:</i> 3060.0 cm^{-1} 2849.5 cm^{-1} 1942.9 cm^{-1} 1601.2 cm^{-1} 1583.0 cm^{-1} 1154.5 cm^{-1} 1028.3 cm^{-1} | 0.30 cm^{-1} 0.40 cm^{-1} 0.40 cm^{-1} 0.30 cm^{-1} 0.30 cm^{-1} 0.30 cm^{-1} 0.30 cm^{-1} | Calibrated Artefact: Matt polystyrene film nominally 0.04 mm thick. Each film is individually calibrated at all seven selected transmittance minima. Films are measured in a spectrophotometer with a nominal sample compartment temperature of $30 \text{ }^\circ\text{C} \pm 5 \text{ }^\circ\text{C}$ | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | |
|--|--|--|--|------------------|---|
| PHOTOMETRY <i>Service Reference: OT03</i> Luminous intensity (tungsten lamps) | 1 to 100 cd 100 to 1000 cd 1000 to 10000 cd | 0.70 % 0.60 % 0.70 % | The actual measurement uncertainty quoted on certificates depends critically on the lamp repeatability or the meter performance. The CMC relates to that which can be achieved using specially designed transfer standards and, in the case of sources, assumes that the correlated colour temperature or spectral power distribution is known. For illuminance/luminance meters, the calibration only applies for a tungsten source at a correlated colour temperature of 2856 K. | Teddington | |
| Illuminance (tungsten lamps and illuminance meters) | 0.1 to 500 lux 500 to 5000 lux 5000 to 20000 lux 20000 to 50000 lux | 0.90 % 0.80 % 0.90 % 1.0 % | | | |
| Luminance (tungsten sources and luminance meters) | 1 to 100 cd m ⁻² 100 to 1000 cd m ⁻² 1000 to 10000 cd m ⁻² | 1.3 % 1.2 % 1.3 % | | | |
| Correlated colour temperature (tungsten lamps and colour temperature meters) | 2000 to 3200 K | 10 K | | | |
| Absolute Scale Realisation of Spectral Irradiance | | | | | |
| Luminous flux (tungsten lamps) | 1 to 100 lumen 100 to 20000 lumen | 0.90 % 0.70 % | | | |
| SPECTRORADIOMETRY OF SOURCES <i>Service Reference: OT03</i> Absolute spectral irradiance of tungsten-based sources | 250 nm to 300 nm >300 nm to 400 nm >400 nm to 800 nm >800 nm to 1600 nm >1600 nm to 2400 nm >2400 nm to 2500 nm | 1.4 % 1.4 % 0.5 % 0.4 % 0.8 % 1.2 % | | | |
| Relative spectral power distribution and absolute spectral irradiance of tungsten and tungsten halogen sources | 250 nm to 300 nm >300 nm to 400 nm >400 nm to 800 nm >800 nm to 1600 nm >1600 nm to 2400 nm >2400 nm to 2500 nm | 2.1 % 1.8 % 1.1 % 1.0 % 1.3 % 2.0 % | | | Based on measurement against the ultra high temperature blackbody (UHTBB) and applicable to test lamps operating at approximately 2800 K to approximately 3250 K only. Nominal lamp power ratings in the range 500 W – 5,000 W. Includes the calculation of chromaticity coordinates and correlated colour temperature. |
| | | | | | For lamps rated 50 W to 5000 W (nominal), correlated colour temperature in the range 2600 K to 3250 K. Includes the calculation of chromaticity coordinates and correlated colour temperature. Note Calibration measurement capability is based on measurement against "primary" level standards and is applicable to test lamps operating at approximately 2856 K and approximately 3100 K only. |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | |
|--|--|--|--|-------------------|--|
| SPECTRORADIOMETRY OF SOURCES (continued) | | | | | |
| Relative and absolute spectral radiance of tungsten and tungsten halogen based sources | 300 nm to 400 nm 400 nm to 800 nm 800 nm to 1600 nm 1600 nm to 2500 nm | 3.5 % to 2.7 % 2.7 % to 1.8 % 1.8 % to 2.1 % 2.1 % to 3.0 % | For lamps rated 50 W to 5000 W (nominal), correlated colour temperature in the range 2600 K to 3250 K. | Teddington | |
| RADIOMETRY Service Reference: OT12 | | | | | |
| Spectral responsivity of laser power meters | 100 pW to 1 mW 350 nm to 1600 nm | 0.040 % | At laser wavelength or peak wavelength of bandpass filter. | | |
| Spectral responsivity | 200 nm to 240 nm 240 nm to 380 nm 380 nm to 405 nm 405 nm to 920 nm 920 nm to 1800 nm | 1.0 % 0.80 % 0.30 % 0.10 % 0.30 % | | | |
| COLORIMETRY AND SPECTROPHOTOMETRY - REGULAR TRANSMITTANCE Service Reference: OT05 | | | | | |
| Regular transmittance | 0.001 %T to 100 %T Wavelength range (nm): $200 \leq \lambda \leq 210$ $210 < \lambda \leq 310$ $310 < \lambda \leq 800$ $800 < \lambda \leq 3000$ | Absolute uncertainty for 90 %T, 60 %T and 0.001 %T: 0.50 %, 0.30 % and 0.000027 % 0.30 %, 0.15 % and 0.000018 % 0.30 %, 0.15 % and 0.000011 % 0.30 %, 0.30 % and 0.000022 % | | | |
| Optical density | 0 D to 5.0 D Wavelength range (nm): $200 \leq \lambda \leq 210$ $210 < \lambda \leq 310$ $310 < \lambda \leq 800$ $800 < \lambda \leq 3000$ | Absolute uncertainty for 0.05 D, 0.22 D and 5.0 D: 0.0024 D, 0.0022 D and 0.012 D 0.0014 D, 0.0011 D and 0.0077 D 0.0014 D, 0.0011 D and 0.0048 D 0.0014 D, 0.0022 D and 0.0097 D | Optical density is equivalent to absorbance (A) and is calculated from regular transmittance using the formula $D = \log_{10}(100/\%T)$. | | |
| Wavelength of absorption peaks | 200 to 3000 nm | 0.10 nm | | | |
| Colour data: CIELAB L* a* b* | 0 to 100 - 200 to + 200 - 200 to + 200 | 0.050 0.050 0.050 | Colour data is normally given for the CIE 2° and 10° observers and CIE Standard Illuminants A, C and D65. Other Standard Illuminants on request. | | |
| Colour data: CIE y, u', v' x, | 0 to 1 | 0.00020 | | | |
| Luminous transmittance Y | 0 %Y to 100 %Y | 0.15 % for 60 %Y | | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---|-------------------|
| COLORIMETRY AND SPECTROPHOTOMETRY - DIFFUSE REFLECTANCE <i>Service Reference: OT04</i> | | | | |
| Spectral diffuse reflectance; specular included and specular excluded geometries | 0 %R to 100 %R <i>Wavelength range (nm):</i> $200 \leq \lambda \leq 315$ $315 < \lambda \leq 460$ $460 < \lambda \leq 800$ $800 < \lambda \leq 2000$ $2000 < \lambda \leq 2500$ | Absolute uncertainty: 1.6 % for white, 0.60 % for black (0.050 + 0.0055R) % (0.050 + 0.0035R) % 1.6 % for white, 0.35 % for black 2.1 % for white, 0.65 % for black | The CMCs are for measurement against similar NPL masters, and examples are given covering the range from 'white' samples to 'black' samples. Higher uncertainties may apply where no similar master exists. | Teddington |
| Spectral radiance factor (see Notes 1 and 2) | 0% to 102%R <i>Wavelength range (nm):</i> $200 \leq \lambda \leq 315$ $315 < \lambda \leq 800$ $800 < \lambda \leq 2000$ $2000 < \lambda \leq 2500$ | Absolute uncertainty: 1.6 % for white, 0.60 % for black (0.050 + 0.0070R) % 2.3 % for white, 0.30 % for black 2.8 % for white, 0.70 % for black | Radiance factor results are expressed relative to the perfect reflecting (Lambertian) diffuser. A result >100 % implies that the sample reflects more radiation at 45° than a Lambertian diffuser. See also the remark above. | |
| Colour data: CIELAB L* a* b* | 0 to 100 - 200 to + 200 - 200 to + 200 | 0.15 0.10 0.10 | Colour data is normally given for the CIE 2° and 10° observers and CIE Standard Illuminants A, C and D65. Other Standard Illuminants on request. | |
| Colour data: CIE y, u', v' x, | 0 to 1 | 0.00020 | Colour data is normally given for the CIE 2° and 10° observers and CIE Standard Illuminants A, C and D65. Other Standard Illuminants on request. | |
| Luminous transmittance Y | 0 %Y to 100 %Y | 0.55 % for white, 0.10 % for black | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|--|---------------|
| TEMPERATURE | | | | |
| Standard resistance thermometers, fixed point calibrations <i>Service Reference: PM02</i> | -196 °C to 0.01 °C -189.3442 °C -38.8344 °C 0.01 °C 0 °C to 29.7646 °C 0 °C to 156.5985 °C 231.928 °C 419.527 °C 0 °C to 419.527 °C 419.527 °C to 660.323 °C | 0.0017 °C to 0.00016 °C 0.00050 °C 0.00035 °C 0.00011 °C 0.00016 °C to 0.00030 °C 0.00030 °C to 0.00070 °C 0.00070 °C 0.00090 °C 0.0010 °C 0.0010 °C to 0.0025 °C | Comparison at LN ₂ Calibrations at measurement current For HTSPRTs | Teddington |
| Resistance thermometers, calibration by comparison <i>Service Reference: PM04</i> | -196 °C -100 °C to -80 °C -80 °C to 0 °C 0 °C to 30 °C 30 °C to 100 °C | 0.0050 °C 0.010 °C 0.0060 °C 0.0030 °C 0.0050 °C | Comparison at LN ₂ and in acetone. Oil and water baths | |
| Resistance thermometers, by dry block calibration <i>Service Reference: PM04</i> | 50 °C to 150 °C 150 °C to 420 °C | 0.040 °C 0.040 °C to 0.10 °C | | |
| Temperature indicators with resistance sensor <i>Service Reference: PM04</i> | -196 °C to 420 °C | As for sensor | | |
| Fixed Point Cells <i>Service Reference: PK01</i> | | | | |
| Triple point of Argon | -189.3442 °C | 0.0005 °C | Cell compared with NPL reference cell during several realisations of the fixed point temperature using Standard Platinum Resistance Thermometers | |
| Triple point of Mercury | -38.8344 °C | 0.0002 °C | | |
| Melting point of Gallium | 29.7646 °C | 0.0002 °C | | |
| Freezing point of Indium | 156.5985 °C | 0.0007 °C | | |
| Freezing point of Tin | 231.928 °C | 0.0006 °C | | |
| Freezing point of Zinc | 419.527 °C | 0.0009 °C | | |
| Freezing point of Aluminium | 660.323 °C | 0.0025 °C | | |
| Freezing point of Silver | 961.78 °C | 0.004 °C | | |
| Water triple point cells | 0.01 °C | 0.000070 °C 0.000058 °C | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|--|-------------------|
| TEMPERATURE (continued) | | | | Teddington |
| Thermocouples <i>Service Reference: PM03</i> | | | | |
| Noble metal type Pt-Rh | 420 °C 962 °C, 1085 °C 1324 °C 1492 °C | 0.13 °C 0.21 °C 0.53 °C 0.72 °C | ITS-90 fixed points Secondary fixed point Co-C Secondary fixed point Pd-C derived from ITS-90 | |
| | 0 °C to 1100 °C 1100 °C to 1330 °C 1330 °C to 1500 °C | 0.30 °C 0.30 °C to 0.55 °C 0.55 °C to 0.72 °C | Polynomial interpolation with improved homogeneity | |
| | 1064.18 °C 1554.8 °C 0 °C to 1100 °C with 1100 °C to 1600 °C | 0.57 °C 0.85 °C 1.0 °C 1.0 °C to 1.5 °C | Wire Bridge Method Au Pd Interpolation based upon Au and Pd wire bridge measurements | |
| Pt-Rh (type B only) | 1768.2 °C 400 °C to 1100 °C 1100 °C to 1800 °C | 1.1 °C 0.3 °C 0.3 °C to 1.2 °C | Wire bridge method Pt Based upon Zn and Ag fixed points and Pt wire bridge | |
| Thermocouples noble metal type Pt-Pd | 420 °C 962 °C, 1085 °C 1324 °C 1492 °C | 0.10 °C 0.070 °C 0.53 °C 0.72 °C | ITS-90 fixed points Secondary fixed point Co-C Secondary fixed point Pd-C derived from ITS-90 | |
| | 0 °C to 1100 °C 1100 °C to 1330 °C 1100 °C to 1500 °C | 0.2 °C 0.2 °C to 0.55 °C 0.2 °C to 0.72 °C | Interpolation | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|--|-------------------|
| TEMPERATURE (continued) | | | | Teddington |
| Thermocouples, Noble metal type Au-Pt | 420 °C, 660 °C, 962 °C 0 °C to 1000 °C | 0.050 °C 0.050 °C | Where Zn and Ag fixed points used | |
| Thermocouples, base metal types | -196 °C -80 °C to 0 °C 0 °C to 50 °C 50 °C to 100 °C | 0.50 °C 0.10 °C 0.050 °C 0.10 °C | Comparison with LN ₂ and in oil and water baths | |
| Thermocouples, by dry block calibration | 50 °C to 700 °C | 0.75 °C | Calibration via comparison to the integrated reference PRT of a dry block calibrator | |
| Temperature indicators with thermocouple sensor <i>Service Reference: PM04</i> | -196 °C to +100 °C | As for sensor | | |
| Compensating and extension cables <i>Service Reference: PM03</i> | -25 °C to 100 °C | As for base metals thermocouples | By comparison. | |
| Thermocouple fixed point cells <i>Service Reference: PK01</i> Cu fixed point cell (freeze) | 1084 °C | 0.031 °C | Certification of fixed point cells by measurement (with Pt/Pd thermocouples) against NPL National Standard fixed point cells | |
| Co-C fixed point cell (melt) Pd-C fixed point cell (melt) | 1324 °C 1492 °C | 0.44 °C 0.65 °C | | |
| Disappearing filament pyrometers <i>Service Reference: PM06</i> | 700 °C to 800 °C 800 °C to 1700 °C 1700 °C to 2800 °C | 5.0 °C to 2.0 °C 2.0 °C 2.0 °C to 8.0 °C | | |
| Infrared Thermometers <i>Service Reference: PM06</i> | -40 °C to +50 °C 15 °C to 45 °C 50 °C to 260 °C 260 °C to 600 °C 600 °C to 1000 °C 1000 °C to 3000 °C | 0.10 °C 0.050 °C 0.10 °C 0.20 °C 0.30 °C 0.050 % of Celsius temperature | Including tympanic thermometers For temperatures above 1324 °C Eutectic Fixed Point can be used | |
| Blackbody Sources <i>Service Reference: PM06</i> | -40 °C to +260 °C 260 °C to 600 °C 600 °C to 1000 °C 962 °C, 1064 °C, 1085 °C 1000 °C to 3000 °C | 0.20 °C 0.24 °C 0.30 °C 0.060 °C 0.050 % of Celsius temperature | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---|-------------------|
| HUMIDITY | | | The accreditation covers other humidity units directly related to dew-point, e.g. vapour pressure, PPM weight, PPM volume, g/kg etc | Teddington |
| Dew-point <i>Service Reference: MH01</i> | +90 °C to +95 °C +90 °C to +75 °C +75 °C to +60 °C +60 °C to -60 °C -60 °C to -75 °C -75 °C to -90 °C | 0.10 °C to 0.12 °C 0.10 °C to 0.050 °C 0.050 °C to 0.033 °C 0.033 °C 0.033 °C to 0.10 °C 0.10 °C to 0.50 °C | | |
| Dew point in air and nitrogen at elevated pressure <i>Service Reference: MH07</i> | -60 °C to +10 °C | 0.070 °C | At pressures up to 1 MPa | |
| Dew point in various gases at elevated pressure <i>Service Reference: MH07</i> | -60 °C to +15 °C 1 µmol/mol to 100 µmol/mol | 0.12 °C 0.04 µmol/mol to 1.25 µmol/mol | At pressures up to 3 MPa in air, inert gases, methane and premade cylinder gas blends | |
| Relative Humidity <i>Service Reference: MH02/MH03</i> | 0.5 %rh to 98 %rh at temperatures from -40 °C to +100 °C | 0.60 % of reading + 0.10 %rh | | |
| Temperature Sensors incorporated in humidity instruments <i>Service Reference: MH02/MH03</i> | -40 °C to < -20 °C -20 °C to +50 °C > +50 °C to +100 °C | 0.080 °C 0.040 °C 0.080 °C | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|---|--|-------------------|
| NEUTRON DOSIMETRY NEUTRON SOURCE EMISSION RATE <i>Service Reference: RN05</i> Emission rate from radionuclide neutron sources Anisotropy of emission from radionuclide neutron sources NEUTRON FLUENCE Thermal neutron fluence <i>Service Reference: RN01</i> | Source emission rate 10^5 s^{-1} to $2 \times 10^9 \text{ s}^{-1}$ Source emission rate 10^2 s^{-1} to $2 \times 10^6 \text{ s}^{-1}$ Source emission rate 10^5 s^{-1} to 10^9 s^{-1} Anisotropy factor 0.5 to 1.2 Energy: thermal Thermal neutron beam Fluence rates: $10^3 \text{ cm}^2 \text{ s}^{-1}$ to $4 \times 10^4 \text{ cm}^2 \text{ s}^{-1}$ | 1.0 % to 1.2 % depending on source 1.2 % to 1.5 % depending on source 1.0 % to 0.50 % depending on source 1.2 % for Wescott fluence 4.0 % for 'true' fluence | Service reference number: RN05 Induced ^{56}Mn activity measured using sodium iodide detectors Service reference number: RN05 Relative measurement performed using a moderating detector assembly Service reference number: RN05. Measurements performed using a precision long counter in a low- scatter environment Service reference number: RN01 Fast neutrons moderated in a graphite pile. Beam of thermal neutrons extracted. Fluence standard - gold foil activation Service conforms to ISO Standard 8529 Parts 1 to 3. | Teddington |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|--|-------------------|
| NEUTRON FLUENCE (continued) | | | | |
| Thermal neutron fluence <i>Service Reference: RN01</i> | Energy: thermal Thermal neutrons in cavity -Fluence rates; $10^4 \text{ cm}^{-2} \text{ s}^{-1}$ to $3 \times 10^7 \text{ cm}^{-2} \text{ s}^{-1}$ | 1.0 % for Westcott fluence | Service reference number: RN01. Fast neutrons moderated in graphite pile. Thermal neutrons in a small 150 cm^3 cavity. Fluence standard - gold foil activation. | Teddington |
| Fast neutron fluence <i>Service Reference: RN02</i> | Energy: 70 keV to 17 MeV Monoenergetic fields Fluence rates: $1 \text{ cm}^{-2} \text{ s}^{-1}$ to $1500 \text{ cm}^{-2} \text{ s}^{-1}$ at 1 m from source | 4.0 % | Service reference number: RN02. Neutrons are produced using beams of protons or deuterons from a 3.5 MV Van de Graaff accelerator. Fluences measured using precision long counter. Service conforms to ISO Standard 8529 Parts 1 to 3. | |
| Fast neutron fluence <i>Service Reference: RN04</i> | Energy: broad energy range from sources: $^{241}\text{Am-Be}$, ^{252}Cf , $^{241}\text{Am-B}$, $^{241}\text{Am-Li}$, $^{241}\text{Am-F}$ Fluence rates: $1 \text{ cm}^{-2} \text{ s}^{-1}$ to $400 \text{ cm}^{-2} \text{ s}^{-1}$ at 1 m from source | 1.3 % | Service reference number: RN04. Fields are produced using radionuclide neutron sources of known emission rate. Actual fluence rate depends on particular source Service conforms to ISO Standard 8529 Parts 1 to 3. | |
| NEUTRON DOSE EQUIVALENT | | | | |
| Thermal and fast neutron dose equivalents <i>Service Reference: RN01, RN02 & RN04</i> | Energy: thermal Thermal neutron beam | 5.0 % | Service reference numbers: RN01, RN02, and RN04. Fluences are converted to ambient dose equivalent or personal dose equivalent using accepted conversion coefficients. For broad energy range neutron fields from sources the uncertainties in the neutron dose equivalent values reflect uncertainties in the source spectra rather than the conversion coefficients, which are assumed to be exact. Service conforms to: ISO 8529 2001 Part 1 and 2. ISO 8529 1998 Part 3. | |
| | Energy: 70 keV to 17 MeV Monoenergetic fields | 4.0 % | | |
| | Energy: broad energy range from radionuclide sources Dose equivalent rates: $^{241}\text{Am-Be}$: $1 \mu\text{Sv h}^{-1}$ to $400 \mu\text{Sv h}^{-1}$ at 1m from the source | 8.1 % | | |
| | ^{252}Cf : $1.6 \mu\text{Sv h}^{-1}$ to 3.2 mSv h^{-1} at 1m from the source | 2.4 % | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---|------------------|
| NEUTRON DOSE EQUIVALENT (continued) | | | | |
| Thermal and fast neutron dose equivalents <i>Service Reference: RN04</i> | 241Am-B: 5 $\mu\text{Sv h}^{-1}$ at 1m from the source | 8.5% | | Teddington |
| | 241Am-F: 1.6 $\mu\text{Sv h}^{-1}$ at 1m from the source | 9.0% | | |
| | 241Am-Li: 1.8 $\mu\text{Sv h}^{-1}$ at 1m from the source | 9.0% | | |
| Fast neutron personal dosimeter proficiency testing <i>Service Reference: RN04</i> | Energy: broad energy range from radionuclide sources Dose equivalent rates: 241Am-Be: 1 $\mu\text{Sv h}^{-1}$ to 400 $\mu\text{Sv h}^{-1}$ at 1m from the source | 8.1 % | Service reference number: RN04. In accordance with HSE Measurement Protocol for Performance Testing of Dosimetry Services for External, Whole Body Fast Neutron Radiation, June 2001. | |
| | 252Cf: 1.6 $\mu\text{Sv h}^{-1}$ to 3.2Sv h^{-1} at 1m from the source | 2.4 % | | |
| <u>Protection level dosimeters</u> Air kerma rate | | | | |
| X-rays <i>Service Reference: RD02</i> | ISO 4037 narrow spectrum (generating potential 8 keV to 250 keV) 350 $\mu\text{Gy h}^{-1}$ to 100 mGy h^{-1} | 1.5 % | Service reference number: RD02 Calibration of protection level ionisation chamber with volumes ranging from 35 cm^3 to 10 litres connected to a suitable secondary standard electrometer. | |
| γ -radiation <i>Service Reference: RD02</i> | ⁶⁰ Co 1 $\mu\text{Gy h}^{-1}$ to 0.3 Gy h^{-1} ¹³⁷ Cs 1 $\mu\text{Gy h}^{-1}$ to 0.3 Gy h^{-1} | 1.2 % 1.2 % | | |
| | ²⁴¹ Am 8 $\mu\text{Gy h}^{-1}$ to 0.3 mGy h^{-1} | 1.5 % | | |
| X-ray generating potential <i>Service Reference: RD04</i> | Generating potential 25 kVp to 150 kVp | 1.5 % | Service reference number: RD04. Calibration of penetrameters and kVp meters. | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---|-------------------|
| <u>Therapy level dosimeters</u> | | | | Teddington |
| Measurement of air kerma rate | | | | |
| X-rays <i>Service Reference: RD01</i> | Half value layers 0.024 mm Al to 20 mm Al (generating potential 8 kVp to 280 kVp) | 1.2 % | Service reference number: RD01 Calibration of NE2561, NE2611 and Farmer type and soft x-ray ionisation chambers. | |
| γ -radiation <i>Service Reference: RD01</i> | ^{60}Co | 1.2 % | Calibration of NE2561, NE2611, for Farmer type ionisation with a suitable secondary standard electrometer, if supplied. | |
| Measurement of absorbed dose to water | | | | |
| γ -radiation <i>Service Reference: RD01</i> | ^{60}Co | 1.4 % | Calibration of NE2561, NE2611, for Farmer type ionisation with a suitable secondary standard electrometer, if supplied. | |
| Photons <i>Service Reference: RD01</i> | TPR ₁₀ ²⁰ : 0.568 to 0.800 Nominal beam energy ^{60}Co , 4-25MV | 1.4 % | | |
| Electrometer (charge) <i>Service Reference: RD16</i> | 10 pC to 1 μC | 0.1% to 0.9% | Service reference number RD16: Calibration of suitable secondary standard electrometer | |
| Electrometer (charge) <i>Service Reference: RD16</i> | 10pA to 2 μA | 0.1 % to 0.9% | Service reference number RD16: Calibration of suitable secondary standard electrometer | |
| <u>High dose dosimetry</u> | | | | |
| Absorbed dose to water <i>Service Reference: RD07</i> | ^{60}Co Dose: >2 Gy | 2.2 % | Service reference number: RD07 High dose irradiation service. | |
| | ^{60}Co Dose: 2 kGy to 55 kGy | 2.2 % to 2.9 % depending on the dose | Service reference number: RD05 Dichromate dosimetry service. | |
| | ^{60}Co , ^{137}Cs , photons generated above 2 MeV and electrons generated above 4 MeV. Dose: 20 Gy to 100 kGy | 2.6 % | Service reference number: RD06 Alanine dosimetry service | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|--|-------------------|
| RADIOACTIVITY METROLOGY RR20 – Standards of Radioactivity (activity per unit mass, Bq g⁻¹): Solutions (or solid substrates spiked directly with such solutions) of α -particle, β -particle, X-ray and γ -ray emitting radionuclides measured by secondary techniques. <i>Service Reference: RR20</i> | 4 kBq g ⁻¹ to 15 GBq g ⁻¹ , depending on radionuclide | 0.30 % | Procedures directly supporting this work are: RMS005 – Standards of Radioactivity Solutions RMT006 - Measurement of a Gamma Spectrometry Sample RMT007 - Analysis of a Gamma Spectrometry Sample RSP014 - Ionisation Chamber Calibration Factors RMT031 - Activity Assay Using Ionisation Chambers | Teddington |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---|-------------------|
| RADIOACTIVITY METROLOGY (continued) | | | | |
| RR20 - Standards of Radioactivity (activity per unit mass, Bq g⁻¹): Solutions (or solid substrates spiked directly with such solutions) of α -particle, β -particle and X-ray emitting radionuclides measured by secondary liquid scintillation techniques <i>Service Reference: RR20</i> | 10 Bq g ⁻¹ to 500 kBq g ⁻¹ | 0.20 % | Procedures directly supporting this work are: RMS005 - Standards of Radioactivity Solutions RMT009 - Secondary Standardisation of Radionuclides using CIEMAT/NIST Technique RMT010 - Standardisation of Radionuclides by Alpha LSC RMT012 - Dilution check by liquid scintillation counting Results for gamma emitters may be certificated as $\text{gammas s}^{-1} \text{g}^{-1}$ by multiplication of the measured activity per unit mass by published emission probabilities. | Teddington |
| RR/0201 – Standard of Radioactivity (activity concentration, Bq m⁻³): Gaseous radionuclides other than radon <i>Service Reference: RR02</i> | 0.04 MBq m ⁻³ to 30 GBq m ⁻³ | 1.0 % | Procedures directly supporting this work are: RMT034 – Standardisation of Gaseous Radionuclides Other Than Radon and Carbon-11 | |
| RR/0203 - Instrument Calibration (response to activity concentration, Bq m⁻³): Customer supplied radioactivity -in-air monitors (other than radon) <i>Service Reference: RR02</i> | 40 kBq m ⁻³ to 30 GBq m ⁻³ | 4.0 % | Procedures directly supporting this work are: RMT03 - Calibration of Tritium-In-Air Monitors | |
| RR/0301 – Wide Area Reference Source Calibration (surface particle emission rate, particles s⁻¹): Customer supplied radioactive surface contamination sources <i>Service Reference: RR03</i> | 10 particles s ⁻¹ to 10000 particles s ⁻¹ | 0.50 % for alpha 1.0 % for beta | Procedures directly supporting this work are: RMS008 - RR0300 Calibration Service RMT004 - Measurement of a Wide Area Reference Source by the Primary Large Area Proportional Counter RQC004 - Quality Checks of the Large Area Proportional Counter RSP008 - Setting of the Alpha and Beta Counting Thresholds | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|--|------------------|
| RADIOACTIVITY METROLOGY (continued) | | | | |
| <p>RR20- Standards of Radioactivity (activity per unit mass, Bq g⁻¹ or γ emission rate per unit mass, s⁻¹ g⁻¹): Solutions (or solid substrates spiked directly with such solutions) of Environmental level standards of radioactivity <i>Service Reference: RR20</i></p> | 0.001 Bq g ⁻¹ to 100 kBq g ⁻¹ , 0.01 s ⁻¹ g ⁻¹ to 1000 s ⁻¹ g ⁻¹ depending on radionuclide | 0.30 % | <p>Procedures directly supporting this work are:</p> <p>RMS007 - Production of the NPL Mixed Radionuclide Solution RMS009 - Production of Environmental Radioactivity Standards</p> | Teddington |
| <p>RR/0601 - Standards of Radioactivity (activity concentration, Bq g⁻¹): Solutions of ³H standards for ³H measurement proficiency test <i>Service Reference: RR06</i></p> | 10 Bq g ⁻¹ to 1 kBq g ⁻¹ | 1.0 % | <p>In accordance with HSE Measurement Protocol for Performance Testing of the Determination of Tritium in Water (1993)</p> <p>Procedures directly supporting this work are:</p> <p>RMS012 - Performance Testing of the Determination of Tritium in Water RMS013 - Data Analysis of Assigned Value Comparison Exercise Results for Performance Testing of the Determination of Tritium in Water</p> | |
| <p>RR/0701 - Artefact calibration (activity content, Bq or Activity per unit mass, Bq g⁻¹): Gelatine capsules (¹³¹I only), brachytherapy wires (¹⁹²Ir only), brachytherapy seeds (¹²⁵I only) or solutions (or solid substrates spiked directly with such solutions) of α-particle, β-particle, X-ray and γ-ray emitting radionuclides measured by secondary techniques. <i>Service Reference: RR07</i></p> | 400 Bq to 15 GBq, 400 Bq g ⁻¹ to 15 GBq g ⁻¹ , depending on radionuclide | 0.30 % | <p>Procedures directly supporting this work are:</p> <p>RMT031 - Activity Assay using Ionisation Chambers RSP014 - Ionisation Chamber Calibration Factors RMS001 - Calibration of Customer Supplied Sources (Gamma Emitters) RMS002 - Calibration of Customer Supplied Sources (Beta Emitters) RMT006 - Measurement of a Gamma Spectrometry Sample RMT007 - Analysis of a Gamma Spectrometry Sample RMT009 - Secondary Standardisation of Radionuclides using CIEMAT/NIST Technique RMT012 - Dilution check by liquid scintillation counting</p> | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd

Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|--|-------------------|
| RADIOACTIVITY METROLOGY (continued) | | | | |
| RR/1201 - Calibration of impurity content of radionuclide solutions by high resolution gamma spectrometry (activity per unit mass, Bq g ⁻¹ or activity concentration, Bq ml ⁻¹) <i>Service Reference: RR12</i> | 0.5 Bq g ⁻¹ to 100 kBq g ⁻¹ or 0.5 Bq ml ⁻¹ to 100 kBq ml ⁻¹ Energy range 60–1100 keV Energy range 1100-1500 keV | 2.50 % 10 % | Procedures directly supporting this work are: RMS003 - Customer Impurity checks RMT006 - Measurement of a Gamma Spectrometry Sample RMT007 - Analysis of a Gamma Spectrometry Sample | Teddington |
| MASS <i>Service Reference: MM01</i> Specific values | Nominal value (g) | (mg) | The stated CMCs relate to measurements made on standards that are constructed in accordance with the principles contained in OIML Recommendation III for weights of Class E1. Intermediate values of weights can be calibrated to an uncertainty equal to the greater of the uncertainties associated with the next higher and lower nominal values in the table. | |
| | 50 000 | 2.5 | | |
| | 20 000 | 1.0 | | |
| | 10 000 | 0.40 | | |
| | 5 000 | 0.20 | | |
| | 3 000 | 0.20 | | |
| | 2 000 | 0.070 | | |
| | 1 000 | 0.030 | | |
| | 500 | 0.015 | | |
| | 300 | 0.010 | | |
| | 200 | 0.0070 | | |
| | 100 | 0.0040 | | |
| | 50 | 0.0030 | | |
| | 30 | 0.0030 | | |
| | 20 | 0.0015 | | |
| | 10 | 0.0010 | | |
| | 5 | 0.00060 | | |
| | 3 | 0.00060 | | |
| | 2 | 0.00030 | | |
| | 1 to 0.001 | 0.00020 | | |
| | 0.0005 to 0.00005 | 0.00030 | | |
| DENSITY <i>Service Reference: MM03</i> | Artefacts 1100-9000 kg/m ³ | | Using the following apparatus: | |
| | 1 g – 100 g | 4.0 – 0.5 kg/m ³ | 100g hydrostatic weighing apparatus | |
| | >100 g – 1 kg | 0.5 – 0.25 kg/m ³ | 1kg hydrostatic weighing apparatus | |
| Density of solid materials | >1 kg – 20 kg | 3.0 – 1.0 kg/m ³ | 20kg hydrostatic weighing apparatus | |
| | Artefacts >9000 kg/m ³ | | | |
| | 1 g – 100 g | 4.0 – 1.0 kg/m ³ | | |
| | >100 g – 1 kg | 1.0 kg/m ³ | | |
| | >1 kg – 20 kg | 5.0 – 2.0 kg/m ³ | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---------------|--|--------------------------|-------------------------------|
| <p>NPL Primary Reference Materials (NPL PRMs) AND NPL CALIBRATED GAS MIXTURES (NPL CGMs) <i>Service Reference: QE11, QE12 and QE13</i></p> <p>NPL Primary Reference Materials (NPL PRMs) Preparation of synthetic gas mixtures by gravimetry in accordance with ISO 6142:2015; verification by analysis.</p> <p>NPL CALIBRATED GAS MIXTURES (NPL CGMs) Certification of synthetic gas mixtures by analysis.</p> <p>The laboratory also has ISO Guide 34 accreditation for production of NPL Primary Reference Materials (NPL PRMs). Accredited certified reference material producer number 4002 http://www.ukas.com/wp-content/uploads/schedule_uploads/00014/4002%20RMP%20Single.pdf</p> <p>Gas mixtures can be produced and/or calibrated as listed below or in the BIPM CMC tables: http://kcdb.bipm.org/appendix/QM/GB/QM_GB_4.pdf</p> <p>Gas mixtures may be produced and/or calibrated for other amount fractions and/or other combinations of the listed gases.</p> | | | | Teddington |
| SYNTHETIC NATURAL GAS MIXTURES | | Amount fraction %mol/mol | Amount fraction %mol/mol | |
| Nitrogen | 0.02 to 25.2 | 0.18 % relative + 0.00038 | NPL PRMs and NPL CGMs | |
| Carbon dioxide | 0.04 to 25.0 | 0.20 % relative + 0.00045 | | |
| Methane | 55.0 to 99.9 | 0.018 % relative + 0.0020 | | |
| Ethane | 0.008 to 18 | 0.28 % relative + 0.000080 | | |
| Propane | 0.008 to 8.0 | 0.30 % relative + 0.000080 | | |
| <i>i</i> -Butane | 0.004 to 1.7 | 0.40 % relative + 0.000040 | | |
| <i>n</i> -Butane | 0.004 to 1.7 | 0.40 % relative + 0.000040 | | |
| <i>neo</i> -Pentane | 0.0005 to 0.5 | 0.80 % relative + 0.000015 | | |
| <i>i</i> -Pentane | 0.0025 to 0.6 | 0.40 % relative + 0.000030 | | |
| <i>n</i> -Pentane | 0.0025 to 0.6 | 0.40 % relative + 0.000030 | | |
| <i>n</i> -Hexane | 0.0008 to 0.5 | 0.40 % relative + 0.000018 | | |
| Helium | 0.001 to 0.5 | 0.95 % relative + 0.000050 | | |
| Oxygen | 0.05 to 1.0 | 1.0% relative | | |
| | | Amount fraction μ mol/mol | | Amount fraction μ mol/mol |
| Benzene | 5 to 500 | 1.1 % relative + 0.030 | | |
| Toluene | 5 to 250 | 1.1 % relative + 0.030 | | |
| Cyclohexane | 10 to 400 | 1.1 % relative + 0.030 | | |
| Methylcyclohexane | 10 to 400 | 1.1 % relative + 0.030 | | |
| <i>n</i> -Heptane | 10 to 500 | 1.1 % relative + 0.040 | | |
| <i>n</i> -Octane | 5 to 10 | 1.3 % relative + 0.025 | | |
| | 10 to 200 | 1.1 % relative + 0.040 | | |
| <i>n</i> -Nonane | 1 to 10 | 1.6 % relative + 0.0090 | | |
| | 10 to 120 | 1.2 % relative + 0.048 | | |
| <i>n</i> -Decane | 1 to 20 | 1.6 % relative + 0.013 | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|-------------------------------|--|---------------------------------|-------------------|
| NPL PRIMARY REFERENCE MATERIALS (NPL PRMs) AND NPL CALIBRATED GAS MIXTURES (NPL CGMs) (continued) | | | | Teddington |
| SYNTHETIC FUEL GAS MIXTURES | Amount fraction %mol/mol | Amount fraction %mol/mol | NPL PRMs and NPL CGMs | |
| Nitrogen | 0.1 to 95 | 0.30 % relative + 0.0020 | | |
| Carbon monoxide | 0.1 to 11 | 0.48 % relative + 0.0016 | | |
| Carbon dioxide | 0.3 to 8 | 0.48 % relative + 0.0016 | | |
| Oxygen | 0.2 to 2.5 | 0.78 % relative + 0.0005 | | |
| Hydrogen | 1 to 70 | 0.38 % relative + 0.0025 | | |
| Helium | 1 to 70 | 0.40 % relative + 0.0025 | | |
| Methane | 1 to 85 | 0.33 % relative + 0.0015 | | |
| Ethane | 0.3 to 35 | 0.35 % relative + 0.0010 | | |
| Ethene | 0.1 to 20 | 0.35 % relative + 0.00030 | | |
| Ethyne | 0.025 to 2 | 0.40 % relative + 0.00025 | | |
| Propane | 0.1 to 18 | 0.40 % relative + 0.00020 | | |
| Propene | 0.04 to 10 | 0.45 % relative + 0.00010 | | |
| i-Butane | 0.1 to 4 | 0.40 % relative + 0.00025 | | |
| n-Butane | 0.1 to 6 | 0.40 % relative + 0.00025 | | |
| 1-Butene | 0.015 to 1.55 | 0.45 % relative + 0.00020 | | |
| i-Butene | 0.018 to 1.2 | 0.50 % relative + 0.00020 | | |
| t-2-Butene | 0.015 to 0.85 | 0.45 % relative + 0.00013 | | |
| c-2-Butene | 0.015 to 0.35 | 0.45 % relative + 0.00013 | | |
| 1,3-Butadiene | 0.01 to 3 | 0.55 % relative + 0.00015 | | |
| i-Pentane | 0.05 to 0.8 | 0.45 % relative + 0.00020 | | |
| n-Pentane | 0.05 to 0.8 | 0.45 % relative + 0.00020 | | |
| SULPHUR ODORANT GAS MIXTURES | Amount fraction μ mol/mol | Amount fraction μ mol/mol | Matrix gas: Methane or nitrogen | |
| Hydrogen sulphide | 0.4 to 5,000 | <u>NPL PRM</u> | | |
| Carbonyl sulphide | 0.4 to 5,000 | 1.0 % relative + 0.0050 | | |
| Carbon disulphide | 0.4 to 200 | (All components) | | |
| Dimethyl sulphide | 0.4 to 200 | | | |
| Ethyl methyl sulphide | 0.4 to 200 | | | |
| Diethyl sulphide | 0.4 to 200 | <u>NPL CGM</u> | | |
| Methyl mercaptan [Methanethiol] | 0.4 to 200 | 1.2 % relative + 0.0050 | | |
| Ethyl mercaptan [Ethanethiol] | 0.4 to 200 | (All components) | | |
| i-propyl mercaptan [2-propanethiol] | 0.4 to 200 | | | |
| n-propyl mercaptan [1-propanethiol] | 0.4 to 200 | | | |
| Tert-butyl mercaptan [2-methyl-2-propanethiol] | 0.4 to 200 | | | |
| Tetrahydrothiophene [THT] | 0.4 to 200 | | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 **Issue date:** 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---|-------------------|
| NPL PRIMARY REFERENCE MATERIALS (NPL PRMs) AND NPL CALIBRATED GAS MIXTURES (NPL CGMs) (continued) | | | | Teddington |
| BINARY GAS MIXTURES <i>Service Reference: QE85-0000</i> | Amount fraction | | NPL CGM | |
| Nitric oxide in nitrogen | 100 nmol/mol to 10 μ mol/mol | 3.0 % relative | Calibrated in accordance with NPL procedure QPDQM/B/517 | |
| OZONE PHOTOMETERS <i>Service Reference: QE85-0000</i> | | | | |
| Ozone in synthetic Air | 0 nmol/mol to 100 nmol/mol 100 nmol/mol to 1 μ mol/mol | 3.0 nmol/mol 3.0 % relative | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | |
|--|---|--|---|-------------------|--|
| ULTRASONICS <i>Service Reference: AW07</i> | | | | | |
| End-of-cable loaded sensitivity of a hydrophone | 1 MHz to 8 MHz 9 MHz to 12 MHz 13 MHz to 16 MHz 17 MHz to 20 MHz 21 MHz to 30 MHz 31 MHz to 40 MHz | 6.0 % 7.0 % 8.0 % 11 % 12 % 15 % | Free field sensitivity determined through substitution using a secondary hydrophone in a non-linearly distorted sound field | Teddington | |
| End-of-cable loaded sensitivity of a hydrophone in fine frequency range | 1 MHz to 8 MHz 9 MHz to 12 MHz 13 MHz to 16 MHz 17 MHz to 20 MHz | 8.0 % 9.0 % 10 % 11 % | Free field sensitivity determined through substitution using a secondary hydrophone in a quasi-linear tone-burst acoustic field. Lowest frequency resolution is 10 kHz. | | |
| UNDERWATER ACOUSTICS | | | | | |
| Calibration of hydrophones and projectors <i>Service Reference: AW15</i> | | | According to IEC 60565:2006 | | |
| End of cable hydrophone receive sensitivity <i>Service Reference: AW10</i> | 25 Hz to 400 Hz | 0.50 dB | By comparison to a microphone using an air-pistonphone | | |
| Free field sensitivity of reference measuring hydrophones/projectors <i>Service Reference: AW15</i> | 1 kHz to 2 kHz 2 kHz to 500 kHz | 0.70 dB 0.50 dB | Using three-transducer spherical wave reciprocity method in a laboratory tank | | |
| Free field sensitivity of reference measuring hydrophones <i>Service Reference: AW10</i> | 1 kHz to 2 kHz 2 kHz to 1 MHz | 0.90 dB 0.70 dB | By comparison with NPL reference hydrophone in a laboratory tank | | |
| Directional response of transducers and hydrophones <i>Service Reference: AW10</i> | 1 kHz to 1 MHz | 0.21 dB | Normalised response versus angle. XY, XZ and YZ responses available. Performed in a laboratory tank | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|--|---------------|
| UNDERWATER ACOUSTICS (continued) | | | | |
| Calibration of hydrophones and projectors <i>Service Reference: AW15</i> | | | According to IEC 60565:2006 | Wraysbury |
| Projector sensitivity | 250 Hz to 500Hz 500 Hz to 1kHz 1 kHz to 350 kHz From noise limit to 210 dB re 1 μ Pa/V at 1 m | 1.2dB 1.0 dB 0.9 dB | Using calibrated hydrophone method in an open-water test facility | |
| Hydrophone sensitivity | 250 Hz to 500Hz 500 Hz to 1kHz 1 kHz to 350 kHz From -250 dB to -100 dB re. 1 V/ μ Pa | 1.2dB 1.0 dB 0.9 dB | Using calibrated projector method in an open-water test facility | |
| Complex admittance conductance susceptance capacitance | 250 Hz to 350 kHz | 2.0 % + 10 μ S 2.0 % + 10 μ S 2.0 % + 20 μ S | For underwater electro acoustic transducers only. Undertaken in open-water test facility. | Teddington |
| FORCE | | | | |
| Proving devices, load cells and other force-measuring devices in compression and tension modes <i>Service Reference: MF01</i> | <i>50 N Machine</i> 1.5 N to 50 N | 0.0020 % | Calibrations can be performed in accordance with ASTM E74-18, ISO 376:2011 and BS 8422:2003 standard and supplementary calibrations A, B, E, L and R. Forces can be applied incrementally and decrementally thus permitting the determination of hysteresis errors. | |
| | <i>500 N Machine</i> 25 N to 500 N | 0.0020 % | | |
| | <i>2.5 kN Machine</i> 25 N to 2.5 kN | 0.0010 % | | |
| | <i>20 kN Machine</i> 0.5 kN to 20 kN | 0.0010 % | | |
| | <i>120 kN Machine</i> 2.5 kN to 120 kN | 0.0010 % | | |
| | <i>1.2 MN Machine</i> 10 kN to 1.2 MN | 0.0010 % | | |
| | <i>5 MN Machine</i> 100 kN to 200 kN 200 kN to 5 MN | 0.1 % 0.050 % | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---|-------------------|
| FORCE (continued) | | | | Teddington |
| Strain Cylinders <i>Service Reference: MF03</i> | 200 kN to 2 MN | See remarks | Strain cylinders compared with a reference strain cylinder in a 3 MN compression machine in accordance with BS EN 12390-4:2000. | |
| Voltage Ratio | | | | |
| Calibration of DC voltage ratio meters used with strain gauge force transducers <i>Service Reference: MF04</i> | 0.01 mV/V to 0.05 mV/V 0.05 mV/V to 1.0 mV/V 1.0 mV/V to 2.5 mV/V 2.5 mV/V to 10 mV/V | 0.010 % 0.0050 % 0.0070 % 0.0050 % | Ratio meters are compared to a reference resistance network using a precision digital voltmeter to measure the voltage ratios generated | |
| PRESSURE <i>Service Reference: MP03</i> <u>Gas Pressure (absolute)</u> | | | | |
| Determination of effective area of deadweight testers | 3.5 kPa to 16 kPa 16 kPa to 700 kPa 700 kPa to 7 MPa | 0.0023 % 0.0021 % 0.0025 % + 0.25 ppm/MPa | | |
| Calibration of pressure indicating instruments | 75 kPa to 110 kPa | 5.0 Pa | | |
| <u>Gas Pressure (gauge)</u> | | | | |
| Determination of effective area of deadweight testers | 3.5 kPa to 16 kPa 16 kPa to 700 kPa 700 kPa to 7 MPa 7 MPa to 21 MPa | 0.0023 % 0.0021 % 0.0025 % + 0.25 ppm/MPa 0.0028 % + 0.50 ppm/MPa | | |
| <u>Oil Pressure (gauge)</u> | | | | |
| Determination of effective area of deadweight testers | 500 kPa to 200 MPa 200 MPa to 500 MPa | 0.0031 % + 0.24 ppm/MPa 0.0038 % + 0.24 ppm/MPa | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|--|-------------------|
| FIBRE OPTICS | | | Measurements carried out at 23 °C ± 2 °C, unless stated otherwise | Teddington |
| Mode field diameter <i>Service Reference: OT06-1020</i> | 3.5 µm to 13 µm 0 % to 1 % | 0.60 % 0.1 % | Single-mode fibre from 1250 nm to 1625 nm. Far field scan method Petermann II definition | |
| Mode field noncircularly <i>Service Reference: OT06-1020</i> | | | | |
| Effective area <i>Service Reference: OT06-1050</i> | 30 µm ² to 130 µm ² | 2.0 % | Far field scan method. Hankel transform | |
| Dispersion in single-mode optical fibre <i>Service Reference: OT06-1010</i> | | | Fibre length: 2 km to 50 km | |
| Dispersion | 0 ps.nm ⁻¹ to 1.3 x 10 ⁻⁵ ps.nm ⁻¹ | 1.5 % added in quadrature with 0.010 ps.nm ⁻¹ .km ⁻¹ | Laser based system | |
| Zero dispersion wavelength | 1250 nm to 1650 nm | 0.10 nm | Laser based system | |
| Dispersion slope at zero dispersion wavelength | -100 to +100 ps.nm ⁻¹ .km ⁻¹ | 1.5 % | Laser based systems | |
| Optical length <i>Service Reference: OT06-1110</i> | 0.1 km to 15 km measured in the wavelength range 1270 nm to 1650 nm. | (0.040 + 1.7 x 10 ⁻⁵ L) | Single-mode optical fibre, pulsed time of flight technique. | |
| | 15 km to 105 km measured at wavelengths of 1310 nm, 1550 nm & 1625 nm | (0.10 + 1.7 x 10 ⁻⁵ L) | L is optical length in metres | |
| Fibre attenuation coefficient uniformity <i>Service Reference: OT06-1100</i> | 0.17 db/km to 0.43 db/km | 0.0060 dB/km | Single-mode optical fibre (length 4 km to 14 km). Measured using an optical time domain reflectometer (OTDR), 1300 nm and 1550 nm wavelength windows | |
| Spectral attenuation of single mode fibre <i>Service Reference: OT06-1060</i> | 0.1 dB to 35 dB | 0.020 dB | Cut-back technique Wavelength range 1200 nm to 1650 nm Measurements carried out over the temperature range 18 °C to 23 °C. | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---|-------------------|
| FIBRE OPTICS (continued) | | | | Teddington |
| Spectral attenuation of multimode fibre <i>Service Reference: OT06-1060</i> | 0.1 dB to 35 dB | 0.020 dB | Cut-back technique Wavelength range 800 nm to 900 nm 1250 nm to 1350 nm Measurements carried out over the temperature range 18 °C to 23 °C. | |
| Cut-off wavelength of optical fibre and cable <i>Service Reference: OT06-1070</i> | 800 nm to 1600 nm | 2.0 nm | Transmitted power technique Measurements carried out over the temperature range 18 °C to 23 °C. | |
| <u>Fibre optic test equipment</u> <i>Service Reference: OT06-1010</i> | | | Measurements carried out at 20 °C ± 2 °C, unless otherwise stated | |
| Absolute responsivity of fibre optic power meters with FC/PC connectors | <i>Power level: -10 dBm to +23 dBm</i> 850 nm ± 30 nm 1300 nm ± 25 nm | 0.90 % 0.70 % | Minimum customer meter resolution 2 % of stated power levels. Multimode fibre | |
| | 980 nm ± 10 nm 1300 nm ± 25 nm 1500 nm ± 30 nm 1550 nm ± 20 nm 1620 nm ± 20 nm | 0.90 % 0.70 % 0.80 % 0.70 % 0.70 % | Single mode fibre | |
| Absolute responsivity of fibre optic power meters with SC/PC connectors | <i>Power level: -10 dBm to +23 dBm</i> 850 nm ± 30 nm 980 nm ± 10 nm 1300 nm ± 25 nm | 1.5 % 1.5 % 1.0 % | Minimum customer meter resolution 2 % of stated power levels Multimode fibre Single mode fibre Single mode and multi mode fibre | |
| | <i>Power level: -10 dBm to +23 dBm</i> 1500 nm ± 30 nm 1550 nm ± 20 nm 1620 nm ± 20 nm | 1.0 % 1.0 % 1.0 % | Minimum customer meter resolution 2 % of stated power levels Single mode fibre | |
| Absolute responsivity of fibre optic power meters with FC/APC or SC/APC connectors | <i>Power level: -10 dBm to +23 dBm</i> 1500 nm ± 30 nm 1550 nm ± 20 nm 1620 nm ± 20 nm | 1.6 % 1.6 % 1.6 % | Minimum customer meter resolution 2 % of stated power levels Single mode fibre | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | |
|---|--|--|--|-------------------|--|
| <u>Fibre optic test equipment</u> (continued) | | | | | |
| Linearity in fibre optic power meters with FC/PC, SC/PC, FC/APC and SC/APC connectors - Comparison technique | 20 dBm to 10 dBm +10 dBm to -90 dBm | 0.70 % 0.30 % | Wavelength range: 830 nm to 1620 nm Single mode and multi mode fibre | Teddington | |
| Linearity in fibre optic power meters with FC/PC, SC/PC, FC/APC and SC/APC connectors - Superposition technique | +15 dBm to -90 dBm | 0.050 % | Wavelength range: 1275 nm to 1640 nm Single mode fibre | | |
| Effective centre wavelength of fibre optic light source with spectral line width <5 nm | 800 nm to 1700 nm | 0.30 nm | FC/PC connectorised fibre output | | |
| Effective centre wavelength of fibre optic light source with spectral line width in the range 5 nm to 50 nm | 800 nm to 1700 nm | 1.2 nm | FC/PC connectorised fibre output | | |
| Spectral line width (FWHM) of sources | 0.07 nm to 50 nm | 0.10 nm | 800 nm to 1700 nm | | |
| Output power stability of fibre optic light sources | +10 dBm to -50 dBm | 0.0040 dB | Wavelength range 800 nm to 1700 nm | | |
| PARTICLE COUNTERS <i>Service Reference: QE15-1010</i> | | | | | |
| <u>Airborne particle number concentration</u> | | | | | |
| Calibration factor for condensation particle counters | Concentration range 1000 cm ⁻³ to 100,000 cm ⁻³ | 7.0 % | Comparison with an aerosol electrometer | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | |
|--|--|--|---|-------------------------|--|
| INSTRUMENTS FOR AIR QUALITY MONITORING <i>Service Reference: QE85-0000</i> | | | | | |
| Analysers Calibration | NO _x 200 ppb to 2 ppm SO ₂ 150 ppb to 1 ppm O ₃ 30 ppb to 1 ppm CO 0.5 ppm to 45 ppm | 4.0 % | Two point (zero and span) calibration. An assessment of uncertainty due to analyser repeatability and linearity is also undertaken. | customers' sites | |
| Determination of onsite standard concentration | NO _x 200 ppb to 2 ppm (NO and NO ₂) SO ₂ 150 ppb to 1 ppm CO 0.5 ppm to 45 ppm | 4.0 % | | | |
| NO ₂ molybdenum converter efficiency test | 100 ppb to 250 ppb NO ₂ | 1.5 % | Reaction of NO with O ₃ | | |
| Sample system collection efficiency | NO ₂ 50 ppb to 150 ppb SO ₂ 50 ppb to 150 ppb O ₃ 50 ppb to 150 ppb CO 6 ppm to 12 ppm | 1.5 % absolute 1.0 % absolute 1.5 % absolute 1.0 % absolute | | | |
| Analysers span noise test | Range as analyser calibration | 2.0 ppb | | | |
| Analysers zero noise test | NO _x , NO, SO ₂ , O ₃ , CO | 1.0 ppb | | | |
| Particulate analyser calibration | 0 mg.m ⁻³ to 1 mg.m ⁻³ | 1.5 % | Using 4 pre-weighed masses | | |
| Particulate analyser flow rate test | 1 slm to 10 slm 10 slm to 40 slm | 1.5 % 2.0 % | Volumetric and mass flow | | |
| BINARY GAS MIXTURES <i>Service Reference: QE85-0000</i> | | | | | |
| Nitric oxide in nitrogen | 200 nmol/mol to 2 µmol/mol | 3.0 % relative | | | |
| END | | | | | |



0478

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NPL Management Ltd
Issue No: 093 Issue date: 16 November 2018

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or*
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.*

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
 - As an explicit function of the measurand or of a parameter (see below).
 - As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
 - As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples, and an indication of how they are to be interpreted, are shown below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V:

Over the range 100 mV to 1 V, the CMC is 0.0025 % \cdot V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 % \cdot p + (0.12 \cdot 10⁻⁶ \cdot p \cdot 10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 \cdot 0.01 \cdot i, where i is the instrument indication.