

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0503</b> Accredited to <b>ISO/IEC 17025:2005</b>	<b>LMS Ltd</b>	
	<b>Issue No: 020    Issue date: 18 December 2018</b>	
	<b>The Modern Forge</b> Riverhead Sevenoaks Kent TN13 2EL	<b>Contact: Mr M D Jeal</b> Tel: +44 (0)1732 451866 Fax: +44 (0)1732 450127 E-Mail: sales@lms.ltd.uk Website: www.lms.ltd.uk
<b>Calibration performed by the Organisations at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer's Premises, e.g., Hospitals, Laboratories, and Manufacturing Plants (*including those of LMS)	Temperature	S



0503  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**LMS Ltd**  
**Issue No: 020 Issue date: 18 December 2018**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Temperature controlled, incubators, freezers, refrigerators, ovens, environmental chambers, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)				
TEMPERATURE	- 30 °C to 100 °C 100 °C to 250 °C	0.19 °C      using 0.53 °C      resistance sensors	Single and multipoint monitoring probes temperature profiling	S
	- 80 °C to 400 °C 400 °C to 800 °C 800 °C to 1000 °C	1.9 °C      using 4.0 °C      thermocouple 6.1 °C      sensors		
END				