


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|--|---|
|  0509 Accredited to ISO/IEC 17025:2017 | Ministry of Defence Issue No: 045 Issue date: 28 June 2021 | |
| | Calibration Centre Bolkiah Garrison BB3510 Negara Brunei Darussalam | Contact: Mr Sofri Rahman Tel: +673-2-386475 Fax: +673-2-380643 E-Mail: cal_lab@mindef.gov.bn |

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details | Activity | Location code |
|--|---|---------------|
| Address Calibration Centre Bolkiah Garrison BB3510 Negara Brunei Darussalam | Local contact Mr Sofri Rahman +673-2-386475 | Laboratory |
| | Electrical, DC and LF Electrical, RF and microwave Mass Temperature Pressure Humidity Dimensional Volume Torque | |

Site activities performed away from the locations listed above:

| Location details | Activity | Location code |
|--|--|------------------|
| The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. | Local contact Mr Sofri Rahman +673-2-386475 | Customers' Sites |
| | Mass Temperature | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|--|------------------|
| Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. the method used is by direct comparison unless otherwise stated in the remarks column | | | | |
| ELECTRICAL | | | | Laboratory |
| DC VOLTAGE | 10 V Reference | 0.50 ppm | This uncertainty can be realised with voltage standards within 20 ppm of the nominal voltage and only if they have their own temperature controlled enclosure of appropriate thermal stability | |
| Decade Values | 10 μ V, 100 μ V and 1 mV 10 mV 100 mV 1 V 10 V 100 V 1 kV | 0.50 μ V 70 ppm 10 ppm 2.5 ppm 1.5 ppm 3.0 ppm 2.5 ppm | The stated CMCs are for values that lie within 0.5 % of those listed. | |
| Other values | 0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1 kV | 1.0 μ V 8.0 ppm 7.0 ppm 11 ppm 18 ppm | | |
| DC RESISTANCE | | | | |
| Specific values | | | | |
| Generation | 0.1 Ω 1 Ω 1.9 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 19 k Ω 100 k Ω 1 M Ω 10 M Ω 19 M Ω 100 M Ω 1 G Ω | 4.0 ppm 4.0 ppm 16 ppm 4.0 ppm 4.0 ppm 4.0 ppm 4.0 ppm 10 ppm 3.5 ppm 4.0 ppm 3.5 ppm 35 ppm 9.0 ppm 170 ppm | | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|---|--|------------------|
| ELECTRICAL (continued) DC RESISTANCE (continued) Specific Values (continued) Measurement | 0.1 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1.9 Ω 19 k Ω 19 M Ω | 7.0 ppm 5.5 ppm 5.5 ppm 5.0 ppm 5.0 ppm 5.0 ppm 5.0 ppm 6.0 ppm 12 ppm 20 ppm 12 ppm 8.0 ppm 30 ppm | The stated CMCs are for values that lie within 10 % of those listed. | Laboratory |
| Other values Current carrying resistors | 0 m Ω to 1 m Ω 1 m Ω to 10 m Ω 10 m Ω to 100 m Ω | 0.050 % + 0.70 $\mu\Omega$ 0.090 % + 3.0 $\mu\Omega$ 240 ppm + 30 $\mu\Omega$ | At 5 A DC At 5 A DC At 1 A DC | |
| | 0 Ω to 0.1 Ω 0.1 Ω to 1 Ω 1 Ω to 5 Ω 5 Ω to 12 Ω 12 Ω to 50 Ω 50 Ω to 120 Ω 120 Ω to 120 k Ω 120 k Ω to 500 k Ω 500 k Ω to 1.2 M Ω 1.2 M Ω to 5 M Ω 5 M Ω to 12 M Ω 12 M Ω to 120 M Ω 120 M Ω to 1 G Ω | 60 $\mu\Omega$ 610 ppm 66 ppm 29 ppm 54 ppm 24 ppm 18 ppm 33 ppm 26 ppm 130 ppm 86 ppm 830 ppm 0.90 % | | |
| DC CURRENT | 0 μ A to 1 μ A 1 μ A to 100 mA 100 mA to 10 A 10 A to 20 A | 75 pA 65 ppm 60 ppm 110 ppm | | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (<i>k</i> = 2) | Remarks | Location Code |
|--|---|--|---------|------------------|
| ELECTRICAL (continued) AC VOLTAGE Specific values Specific frequencies | <i>40 Hz, 1 kHz, 10 kHz, 20 kHz and 50 kHz</i> 10 mV 20 mV 100 mV 200 mV 600 mV 1 V 2 V 6 V 10 V 20 V 60 V 100 V 200 V 1 kV 10 mV <i>100 Hz and 400 Hz</i> 20 mV <i>500 Hz</i> 100 mV <i>100 Hz and 400 Hz</i> 200 mV and 600 mV <i>500 Hz</i> 1 V 100 Hz and 400 Hz 2 V and 6 V 500 Hz 10 V 100 Hz and 400 Hz 20 V and 60 V 500 Hz 100 V 100 Hz and 400 Hz 200 V 500 Hz 1000 V 400 Hz and 500 Hz | 210 ppm 170 ppm 90 ppm 90 ppm 90 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 75 ppm 210 ppm 170 ppm 90 ppm 90 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 65 ppm 75 ppm | | Laboratory |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|--|---------|------------------|
| ELECTRICAL (continued) AC VOLTAGE Specific values (continued) Specific frequencies | 2 mV to 10 mV 5 mV to 12 mV 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 12 mV to 50 mV 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz | 0.35 % 0.16 % 0.14 % 0.14 % 0.18 % 0.60 % 670 ppm 550 ppm 570 ppm 670 ppm 0.11 % | | Laboratory |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|---|---------|------------------|
| ELECTRICAL (continued) AC VOLTAGE (continued) Other values (continued) | 50 mV to 120 mV 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 120 mV to 500 mV 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 500 mV to 1.2 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 1.2 V to 5 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 5.0 V to 12 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 12 V to 50 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 50 V to 120 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz | 300 ppm 170 ppm 220 ppm 440 ppm 970 ppm 130 ppm 230 ppm 270 ppm 410 ppm 950 ppm 150 ppm 130 ppm 190 ppm 360 ppm 950 ppm 410 ppm 230 ppm 270 ppm 410 ppm 950 ppm 160 ppm 130 ppm 190 ppm 360 ppm 930 ppm 460 ppm 320 ppm 320 ppm 460 ppm 0.15 % 270 ppm 250 ppm 250 ppm 420 ppm 0.14 % | | Laboratory |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---------|------------------|
| ELECTRICAL (continued) AC VOLTAGE (continued) Other values (continued) | 120 V to 500 V 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz | 620 ppm 810 ppm 0.15 % | | Laboratory |
| | 500 V to 1 kV 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz | 590 ppm 790 ppm 0.15 % | | |
| AC CURRENT | 10 mA to 12 mA 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz | 0.25 % 0.11 % 0.060 % | | |
| | 12 mA to 50 mA 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz | 0.32 % 0.22 % 0.21 % | | |
| | 50 mA to 120 mA 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz | 0.26 % 0.11 % 0.070 % | | |
| | 120 mA to 500 mA 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz | 0.33 % 0.24 % 0.26 % | | |
| | 500 mA to 1 A 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz | 0.27 % 0.14 % 0.18 % | | |
| | 1 A to 10 A 40 Hz to 1 kHz | 0.070 % | | |
| | 10 A to 20 A 40 Hz to 400 Hz | 0.10 % | | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code | |
|--|--|--|--|------------------|--|
| DISTORTION | | | | | |
| Distortion Factor | 0.1 % to 0.25 % 0.2 V to 0.5 V 0.5 V to 2 V 2 V to 300 V | 0.083 % distortion factor 0.043 % distortion factor 0.023 % distortion factor | The capabilities for distortion factor relate to fundamental components in the frequency range 20 Hz to 100 kHz. | Laboratory | |
| | 0.25 % to 0.4 % 0.2 V to 0.5 V 0.5 V to 2 V 2 V to 300 V | 0.068 % distortion factor 0.068 % distortion factor 0.032 % distortion factor | | | |
| | 0.4 % to 1.0 % 0.2 V to 0.5 V 0.5 V to 2 V 2 V to 300 V | 0.17 % distortion factor 0.090 % distortion factor 0.080 % distortion factor | | | |
| | 1.0 % to 100 % 0.2 V to 0.5 V 0.5 V to 2 V 2 V to 300 V | 0.90 % distortion factor 0.80 % distortion factor 0.80 % distortion factor | | | |
| FREQUENCY | | | | | |
| Specific values | 100 kHz 1 MHz 5 MHz 10 MHz | 2.7 parts in 10^{11} 2.7 parts in 10^{11} 5.4 parts in 10^{12} 3.0 parts in 10^{12} | Can be reported as elapsed time for repetitive events. 1/f | | |
| Other Values | dc to 10 kHz 10 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 15 GHz | 26 μ Hz 160 μ Hz 1.6 mHz 16 mHz 160 mHz 1.6 Hz 3 in 10^{12} | | | |
| RPM | | | | | |
| | 10 RPM to 12 000 RPM | 0.10 RPM | Mechanical Tachometer. | | |
| | 10 RPM to 100 000 RPM | 0.60 RPM | Optical Tachometers and Calibrators | | |
| TIME INTERVAL | | | | | |
| Stopwatch calibration | 10 s to 24 Hrs | 40 ms | Manually Triggered | | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---|------------------|
| RF POWER Signal sources | 300 kHz to 4.2 GHz + 20 dBm to - 20 dBm | 0.27 dB | The stated CMCs relate to the calibration of stable 50 Ω coaxial sources having an output VSWR of 1.01 or less and fitted with Type N connectors. | Laboratory |
| Specific frequencies | 2 MHz, 10 MHz, 15 MHz, 25 MHz, 30 MHz, 43 MHz, 50 MHz, 60 MHz, 88 MHz, 100 MHz, 125 MHz, 180 MHz, 200 MHz, 250 MHz, 350 MHz, 400 MHz and 500 MHz 1 W to 100 W | 3.0 % | | |
| | 550 MHz 600 MHz 650 MHz, 700 MHz, 750 MHz, 800 MHz, 850 MHz, 900 MHz, 950 MHz 1 GHz 1 W to 100 W | 4.0 % | | |
| RF ATTENUATION | 0 dB, to 40 dB 50 MHz to 6 GHz | 0.10 dB | The uncertainties for RF attenuation and VRC refer to a 50Ω coaxial system using type N precision connectors | |
| VOLTAGE REFLECTION COEFFICIENTSWR | 1.0 to 1.20 to 0.1 50 MHz to 61 GHz 1 GHz to 3 GHz 3 GHz to 6 GHz | 0.050 0.10 0.15 | | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|---------------------------------|------------------|
| RF CALIBRATION FACTOR | | | | |
| Substitution Method | 100 kHz 300 kHz 500 kHz 1 MHz 3 MHz 5 MHz 10 MHz 30 MHz 100 MHz 300 MHz 500 MHz 1 GHz 1.5 GHz 2 GHz 2.6 GHz | 2.2 % 1.9 % 1.9 % 1.9 % 1.8 % 1.8 % 1.8 % 1.8 % 1.8 % 1.8 % 1.8 % 1.8 % 1.8 % 1.9 % 1.9 % 2.1 % | Referenced to 1 mW at 50 MHz | Laboratory |
| Splitter method | 100 kHz 100 kHz to 4.2 GHz 300 kHz 500 kHz 1 MHz 3 MHz 10 MHz 50 MHz 100 MHz 300 MHz 1 GHz 2 GHz 3 GHz 4.2 GHz | 2.9 % 1.7 % 1.4 % 1.3 % 1.3 % 1.2 % 1.0 % 1.1 % 1.2 % 1.3 % 1.3 % 1.5 % 2.1 % | Referenced to 1 mW at 50 MHz | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | | Remarks | Location Code |
|--|--|--|-------------------|--|------------------|
| CAPACITANCE and DISSIPATION FACTOR | | | | | |
| Specific frequencies | 1 pF | C_p (pF) | D (tan δ) | For the calibration of standard four terminal pair capacitors. The CMCs quoted for dissipation factor apply to D values between zero and 0.002 | Laboratory |
| | 1 kHz to 100 kHz | 0.0037 | 0.0027 | | |
| | 1 MHz | 0.0043 | 0.0029 | | |
| | 10 pF | 0.0032 | 0.00049 | | |
| 1 kHz to 100 kHz | 0.0028 | 0.0012 | | | |
| 1 MHz | | | | | |
| 100 pF | 0.030 | 0.00033 | | | |
| 1 kHz to 100 kHz | 0.056 | 0.00036 | | | |
| 1 MHz | | | | | |
| 1000 pF | 0.62 | 0.00035 | | | |
| 1 kHz to 100 kHz | 0.71 | 0.00035 | | | |
| 1 MHz | | | | | |
| OSCILLOSCOPE CALIBRATION | | | | | |
| TIME INTERVAL (Horizontal deflection coefficients) | 1 μ s | 2.5 % | | | |
| | 5 μ s | 2.8 % | | | |
| | 20 μ s | 2.8 % | | | |
| | 500 μ s | 2.2 % | | | |
| | 1 ms | 2.3 % | | | |
| | 5 ms | 2.2 % | | | |
| | 10 ms | 2.3 % | | | |
| | 50 ms | 2.2 % | | | |
| | 100 ms | 2.3 % | | | |
| | DC AMPLITUDE (Vertical deflection coefficients) | 10 mV | 1.5 % | | |
| 20 mV | | 1.1 % | | | |
| 50 mV | | 0.94 % | | | |
| 100 mV | | 0.91 % | | | |
| 200 mV | | 1.1 % | | | |
| 500 mV | | 1.2 % | | | |
| 1 V | | 1.2 % | | | |
| 2 V | | 1.5 % | | | |
| 5 V | | 1.2 % | | | |
| 10 V | | 1.2 % | | | |
| RISETIME | 1 ns to 10 ns | 230 ps | | Nominal 25 mV | |
| | 1 ns to 10 ns | 228 ps | | Nominal 250 mV | |
| | 1 ns to 10 ns | 226 ps | | Nominal 1 V | |
| BANDWIDTH | 50 kHz to 300 MHz | 3.1 % | | Digital Oscilloscopes | |
| | | 5.0 % | | Analogue Oscilloscopes | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|--|-------------------------------|
| MASS | Nominal value (g) | (mg) | | |
| Artefacts | 26 000 25 000 20 000 10 000 5 000 2 000 1 000 500 200 100 50 20 10 5 2 1 0.5 0.2 0.1 0.05 0.02 0.01 0.005 0.002 0.001 | 20 20 10 5.4 2.7 1.0 0.53 0.27 0.10 0.053 0.033 0.027 0.020 0.017 0.013 0.010 0.0083 0.0067 0.0053 0.0040 0.0033 0.0027 0.0020 0.0020 0.0020 | Substitution Method E2 from 1 mg to 26 kg | Laboratory |
| Non-Automatic Weighing Machines | 5g 10g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 50 kg 100 kg 200 kg 250 kg 500 kg | 0.023 mg 0.032 mg 0.044 mg 0.073 mg 0.13 mg 0.27 mg 0.69 mg 1.3 mg 3.8 mg 9.6 mg 19 mg 306 mg 770 mg 1.7 g 5.2 g 5.6 g 8.9 g | Methods consistent with EURAMET CG18 Weights are available in OIML Class E2 from 1 mg to 20 kg F1 from 1 mg to 20 kg Max grouped load 62 kg M1 From 1 kg to 20 kg Max. grouped load 260 kg Other loads within the overall listed range may also be used Max. grouped load 500 kg | Laboratory & Customers' Sites |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|--|---|------------------|
| TEMPERATURE | | | Calibration by comparison with reference instruments | |
| Temperature indicators with sensors | -60 °C to -30 °C -30 °C to 0 °C 0 °C 0 °C to 200 °C 200 °C to 300 °C | 0.13 °C 0.055 °C 0.038 °C 0.055 °C 0.065 °C | In liquid bath | Laboratory |
| Liquid in glass thermometers | -60 °C to -30 °C -30 °C to 0 °C 0 °C 0 °C to 90 °C 90 °C to 250 °C | 0.17 °C 0.085 °C 0.060 °C 0.075 °C 0.080 °C | | |
| Calibration of temperature probes in air | 0 °C to 70 °C | 0.20 °C | | |
| Temperature controlled, chambers, environmental cabinets and ovens and similar apparatus | 25 °C to 50 °C 50 °C to 200 °C | 1.2 °C 1.4 °C | | Site |
| RELATIVE HUMIDITY | | | Calibration by comparison with reference salt solutions | Laboratory |
| Hygrometers | 35 %rh 50 %rh 80 %rh | 1.2 %rh 1.2 %rh 1.3 %rh | For the temperature range 15 °C to 30 °C | |
| TORQUE | | | | |
| Hand torque tools | To BS EN ISO 6789:2:2017 1.4 N·m to 1356 N·m To BS EN ISO 6789:2003 (withdrawn) 1356 N·m to 6780 N·m 680 N·m to 1356 N·m 340 N·m to 680 N·m 135 N·m to 340 N·m 56 N·m to 135 N·m 3 N·m to 56 N·m | 10% 50 N·m 18 N·m 5.0 N·m 2.5 N·m 1.0 N·m 0.40 N·m | The uncertainty quoted is for both the application of the calibration torque and the characteristics of the device being calibrated. Calibrations may also be given in lbf-in and lbf-ft. | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence
Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---|------------------|
| LENGTH Measuring Instruments and Machines Micrometers External | As BS 870:2008 0 mm to 25 mm | 2.0 μ m between any two points | | Laboratory |
| PRESSURE <u>Hydraulic Pressure (Gauge)</u> "Pressure equivalent" calibration of dead weight testers. | 500 kPa to 3.5 MPa 3.5 MPa to 7 MPa 7 MPa to 70 MPa 70 MPa to 140 MPa | 0.012 % + 42 Pa 0.0067 % + 42 Pa 0.0067 % + 42 Pa 0.0072 % + 42 Pa | Methods consistent with EURAMET CG3 and CG17 | |
| Calibration of pressure indicating instruments and gauges | 500 kPa to 3.5 MPa 3.5 MPa to 7 MPa 7 MPa to 70 MPa 70 MPa to 140 MPa | 0.012 % + 42 Pa 0.0067 % + 42 Pa 0.0067 % + 42 Pa 0.0072 % + 42 Pa | Calibration of pressure measuring devices with an electrical output may be undertaken. | |
| <u>Pneumatic Pressure (Gauge)</u> Calibration of pressure indicating instruments and gauges | -95 kPa to 16 kPa 16 kPa to 172 kPa 172 kPa to 345 kPa 345 kPa to 7 MPa | 0.013 % + 4 Pa 0.0082 % 0.0079 % 0.0068 % | Pressure measurements may be expressed in other units of pressure as required. | |
| <u>Pneumatic Pressure (Absolute)</u> Calibration of pressure indicating instruments and gauges | 16 kPa to 172 kPa 172 kPa to 345 kPa 345 kPa to 7 MPa | 0.0082 % + 1.9 Pa 0.0079 % + 1.9 Pa 0.0068% + 1.9 Pa | | |
| VOLUME Measuring cylinders | 100 ml to 2 l 2 l to 20 l | 0.18 ml 6.5 ml | Gravimetric method | |
| END | | | | |



0509
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Ministry of Defence

Issue No: 045 Issue date: 28 June 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
 - As an explicit function of the measurand or of a parameter (see below).
 - As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
 - As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %-p + (0.12 \cdot 10⁻⁶-p \cdot 10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 \cdot 0.01 \cdot i, where i is the instrument indication.