

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0518 Accredited to ISO/IEC 17025:2017	Springfields Fuels Limited Issue No: 027 Issue date: 18 August 2021	
	Chemical and Metallurgical Services Department Building A396 Springfields Salwick Preston PR4 0XJ	Contact: Mrs Andrea J Condron Tel: +44 (0)1772 763942 Fax: +44 (0)1772 762888 E-Mail: condroaj@westinghouse.com
Calibration performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Chemical and Metallurgical Services Department Building A396 Springfields Salwick Preston PR4 0XJ	Local contact Mrs A J Condron	Dimensional: Length, Angle, Form, Measuring Instruments and Machines A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Any Customer's Premises Mrs A J Condron	Dimensional: Form	B



0518
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Springfields Fuels Limited
Issue No: 027 Issue date: 18 August 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
LENGTH			NOTES	
Length Gauges, Flat and Spherical-ended (excluding length bars)	0 to 1000	1.0 + (8.0 x length in m)	1. All linear calibrations may be given in inch units.	A
Plain Plug Gauges (parallel) cylindrical setting standards and rollers	1 to 50 diameter 50 to 100 diameter 100 to 150 diameter 150 to 200 diameter 200 to 300 diameter	1.0 1.3 1.7 On diameter 2.1 3.1	2. The uncertainty quoted is for the distance separating the two parallel plane lines which, just enclose the profile under consideration.	A
Plain ring gauges (parallel) and setting standards	10 to 25 diameter 25 to 50 diameter 50 to 100 diameter 100 to 150 diameter 150 to 250 diameter	0.80 1.0 On diameter 1.6 2.1 3.2	3. Measuring instruments incorporating a digital scale may also be calibrated.	A
Plain gap gauges (parallel)	2 to 50 diameter 50 to 100 diameter 100 to 200 diameter 200 to 300 diameter 300 to 600 diameter	2.2 5.0 8.0 12 15	4. Features and associated parts of these gauges can be measured to the uncertainties given for equivalent items listed in this schedule.	A
ANGLE				
Squares Blade type	BS 939:2007 50 to 300 300 to 450	3.0 On squareness 5.0 See Note 2		A
Angle plates and box angle plates (See Note 2)	BS 5535:1978 50 to 450	Squareness: 3.0 + (1.0 per 100 mm) Parallelism: 1.0 + (1.0 per 100 mm)		A



0518
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Springfields Fuels Limited
Issue No: 027 Issue date: 18 August 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
FORM				
Surface plates Granite and Cast iron	BS 817:2008 160 x 100 to 1600 x 1000	1.5 + (0.80 x diagonal in m) See Note 2		A, B
MEASURING INSTRUMENTS AND MACHINES				
Micrometers (see note 3)				
External	BS 870:2008 0 to 600	Heads: 2.0 between any two points		A
Internal	BS 959:2008 0 to 900	Setting and extension rods		
Depth	BS 6468:2008 0 to 300	1.0 + (8.0 x length in m)		
Micrometer Heads	BS 1734:1951 0 to 100	1.0		A
Height Setting Micrometers (see note 3)	0 to 300	Heads: 1.5 between any two points Stepped column: 2.5 Overall performance: 3.0		A
Riser Blocks for Above Item	150 300 600	2.5 3.0 5.0		A
Parallels	BS 906:Part 1:1972 5 to 50 x 100 x 400	1.5 to 5.0		A
Vee blocks	BS 3731:1987 20 to 150	3.0 to 5.0		A
Vernier calliper gauges	BS 887:2008 0 to 1000	Overall performance 10 + (30 x length in m) (See Note 3)		A
END				



0518
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Springfields Fuels Limited
Issue No: 027 Issue date: 18 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$