


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0524 Accredited to ISO/IEC 17025:2017	Forensic Science Northern Ireland Issue No: 017 Issue Date: 10 August 2021	
	151 Belfast Road Carrickfergus County Antrim Belfast Northern Ireland BT38 8PL	Contact: Quality Manager Tel: +44 (0)28 9036 1888 Fax: +44 (0)28 9036 1900 E-Mail: quality.enquiries@fsni.gov.uk Website: www.fsni.gov.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Premises	Evidential Breath Testing Instruments Station Distances	A
Any Suitable Location	Highway Reference Distances Vehicle Speed Metering Equipment	B



0524
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Forensic Science Northern Ireland
Issue No: 017 Issue date: 10 August 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Cleartone vehicle speed metering equipment				
Distance and time	Fixed point 60 miles/hour	0.64 miles/hour	Cleartone	B
Highway reference distances	1 mile ¼ mile	0.12 % 0.12 %		B
Station distances	25 m to 100 m	14 + (0.040 x Length in m) mm		A
Lion Intoxilyzer 6000	0 µg ethanol /100 ml to 140 µg ethanol/100 ml simulated breath	2.0 µg ethanol/100 ml simulated breath		A
	Effect of interfering substances acetone 15 µg methanol 4 µg		Confirmation of compliance with Home Office Quality Framework Document	
END				



0524
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Forensic Science Northern Ireland
Issue No: 017 Issue date: 10 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$