


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> CALIBRATION</p> <p>0556</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>Horiba MIRA Limited</b></p> <p>Issue No: 027    Issue date: 18 August 2020</p>	
	<p><b>Calibration Centre</b> Watling Street Nuneaton Warwickshire CV10 0TU</p>	<p><b>Contact: Mr Philip Macleod</b> Tel: +44 (0)2476 355643 Fax: +44 (0)2476 358225 E-Mail: calibration.centre@horiba-mira.com Website: www.horiba-mira.com</p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<p>PRESSURE</p> <p><u>Gas Pressure Gauge</u></p> <p>Calibration of pressure indicating instruments and gauges</p> <p><u>Gas Pressure Absolute</u></p> <p>Calibration of pressure indicating instruments and gauges</p> <p><u>Hydraulic Pressure Gauge</u></p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>- 90 kPa to 0 Pa 0 to 2 MPa</p> <p>10 kPa to 200 kPa 200 kPa to 2.1 MPa</p> <p>0 to 50 MPa</p>	<p>0.024 % + 0.59 kPa 0.024 % + 0.59 kPa</p> <p>0.15 kPa 0.022 % + 0.59 kPa</p> <p>0.016 % + 0.11 MPa</p>	<p>Methods consistent with EURAMET CG3.</p> <p>Calibration of pressure devices with an electrical output can be undertaken</p>
<p>Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. the method used is by direct comparison unless otherwise stated in the remarks column.</p>			
<p>ELECTRICAL</p> <p>DC VOLTAGE</p> <p>Generation</p> <p>Measurement</p>	<p>0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1000 V</p> <p>0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V</p>	<p>20 ppm + 1.0 <math>\mu</math>V 11 ppm + 2.0 <math>\mu</math>V 12 ppm + 20 <math>\mu</math>V 18 ppm + 150 <math>\mu</math>V 18 ppm + 1.5 mV</p> <p>47 ppm + 1.2 <math>\mu</math>V 31 ppm + 4.0 <math>\mu</math>V 31 ppm + 80 <math>\mu</math>V 47 ppm + 600 <math>\mu</math>V 51 ppm + 6.0 mV</p>	



0556

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Horiba MIRA Ltd**

**Issue No: 027 Issue date: 18 August 2020**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC VOLTAGE Generation	1 mV to 33 mV 0.2 Hz to 10 Hz 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz  33 mV to 330 mV 0.2 Hz to 10 Hz 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz  0.33 V to 3.3 V 0.2 Hz to 10 Hz 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz  3.3 V to 10.23 V 0.2 Hz to 10 Hz  3.3 V to 33 V 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz  33 V to 330 V 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz  330 V to 1020 V 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	800 ppm + 100 $\mu$ V 800 ppm + 100 $\mu$ V 150 ppm + 100 $\mu$ V 200 ppm + 100 $\mu$ V 0.10 % + 100 $\mu$ V 0.35 % + 110 $\mu$ V 0.80 % + 150 $\mu$ V  300 ppm + 120 $\mu$ V 300 ppm + 120 $\mu$ V 150 ppm + 120 $\mu$ V 160 ppm + 120 $\mu$ V 350 ppm + 120 $\mu$ V 800 ppm + 140 $\mu$ V 0.20 % + 190 $\mu$ V  300 ppm + 400 $\mu$ V 300 ppm + 400 $\mu$ V 150 ppm + 400 $\mu$ V 190 ppm + 400 $\mu$ V 300 ppm + 400 $\mu$ V 700 ppm + 460 $\mu$ V 0.24 % + 1.2 mV  300 ppm + 4.5 mV  300 ppm + 4.0 mV 150 ppm + 4.0 mV 240 ppm + 4.0 mV 350 ppm + 4.0 mV 900 ppm + 5.0 mV  190 ppm + 30 mV 200 ppm + 30 mV 250 ppm + 30 mV 300 ppm + 30 mV 0.20 % + 80 mV  300 ppm + 120 mV 250 ppm + 120 mV 300 ppm + 120 mV	



0556

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Horiba MIRA Ltd**

**Issue No: 027 Issue date: 18 August 2020**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC VOLTAGE (cont'd) Measurement	20 mV to 200 mV 0.2 Hz to 10 Hz	0.40 %	
	1 mV to 200 mV 10 Hz to 50 Hz	0.30 % + 30 $\mu$ V	
	50 Hz to 100 Hz	0.080 % + 30 $\mu$ V	
	100 Hz to 30 kHz	0.060 % + 30 $\mu$ V	
	30 kHz to 50 kHz	0.070 % + 30 $\mu$ V	
	50 kHz to 100 kHz	0.36 % + 30 $\mu$ V	
	100 kHz to 200 kHz	0.91 % + 50 $\mu$ V	
	200 kHz to 1 MHz	2.0 % + 200 $\mu$ V	
	1 MHz to 2 MHz	5.0 % + 400 $\mu$ V	
	0.2 V to 2 V 0.2 Hz to 10 Hz	0.40 %	
	10 Hz to 50 Hz	0.31 % + 300 $\mu$ V	
	50 Hz to 100 Hz	0.080 % + 300 $\mu$ V	
	100 Hz to 30 kHz	0.060 % + 300 $\mu$ V	
	30 kHz to 50 kHz	0.060 % + 300 $\mu$ V	
	50 kHz to 100 kHz	0.36 % + 300 $\mu$ V	
	100 kHz to 200 kHz	0.91 % + 500 $\mu$ V	
200 kHz to 1 MHz	2.0 % + 2.0 mV		
1 MHz to 2 MHz	5.0 % + 4.0 mV		
2 V to 20 V 0.2 Hz to 10 Hz	0.40%		
10 Hz to 50 Hz	0.30 % + 3.0 mV		
50 Hz to 100 Hz	0.080 % + 3.0 mV		
100 Hz to 2 kHz	0.070 % + 3.0 mV		
2 kHz to 10 kHz	0.10 % + 3.0 mV		
30 kHz to 50 kHz	0.060 % + 30 $\mu$ V		
50 kHz to 100 kHz	0.30 % + 300 $\mu$ V		
10 kHz to 30 kHz	0.15 % + 3.0 mV		
30 kHz to 50 kHz	0.16 % + 3.0 mV		
50 kHz to 100 kHz	0.36 % + 3.0 mV		
100 kHz to 200 kHz	0.90 % + 5.0 mV		
200 kHz to 1 MHz	4.0 % + 40 mV		
1 MHz to 2 MHz	7.0 % + 40 mV		
20 V to 200 V 0.2 Hz to 10 Hz	0.40%		
10 Hz to 50 Hz	0.31 % + 30 mV		
50 Hz to 100 Hz	0.08 % + 30 mV		
100 Hz to 2 kHz	0.08 % + 30 mV		
2 kHz to 10 kHz	0.11 % + 30 mV		
10 kHz to 30 kHz	0.15 % + 30 mV		
30 kHz to 50 kHz	0.16 % + 30 mV		
50 kHz to 100 kHz	0.37 % + 30 mV		
100 kHz to 200 kHz	0.90 % + 5.0 mV		
200 kHz to 1 MHz	4.0 % + 40 mV		
1 MHz to 2 MHz	4.0 % + 40 mV		
200 V to 750 V 50 Hz to 100 Hz	0.39 % + 110 mV		
100 Hz to 2 kHz	0.19 % + 110 mV		
2 kHz to 10 kHz	0.23 % + 110 mV		
10 kHz to 25 kHz	0.28 % + 110 mV		



0556

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Horiba MIRA Ltd**

**Issue No: 027 Issue date: 18 August 2020**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
DC CURRENT			
Generation	0 $\mu$ A to 330 $\mu$ A 330 $\mu$ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20.5 A	150 ppm + 25 nA 110 ppm + 100 nA 100 ppm + 2.0 $\mu$ A 100 ppm + 20 $\mu$ A 210 ppm + 80 $\mu$ A 390 ppm + 1.0 mA 510 ppm + 2.5 mA 0.10 % + 4.0 mA	For the calibration of toroidal clamp-on ammeters
Measurement	10 A to 16.5 A 16.5 A to 150 A 150 A to 1025 A	0.25 % + 3.0 mA 0.25 % + 1.0 A 0.25 % + 1.5 A	
AC CURRENT			
Generation	0 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	610 ppm + 5.0 nA 480 ppm + 40 nA 480 ppm + 0.40 $\mu$ A 600 ppm + 4.0 $\mu$ A 0.11 % + 40 $\mu$ A	
Generation	29 $\mu$ A to 330 $\mu$ A 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0.20 % + 730 nA 0.15 % + 700 nA 0.13 % + 700 nA 0.30 % + 700 nA 0.80 % + 870 nA 1.6 % + 2.0 $\mu$ A	
	330 $\mu$ A to 3.3 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0.20 % + 2.0 $\mu$ A 0.13 % + 2.0 $\mu$ A 0.10 % + 2.0 $\mu$ A 0.20 % + 2.0 $\mu$ A 0.50 % + 2.2 $\mu$ A 1.0 % + 2.5 $\mu$ A	
	3.3 mA to 33 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0.18 % + 10 $\mu$ A 0.090 % + 10 $\mu$ A 0.040 % + 10 $\mu$ A 0.080 % + 10 $\mu$ A 0.20 % + 15 $\mu$ A 0.40 % + 15 $\mu$ A	
	33 mA to 330 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0.18 % + 100 $\mu$ A 0.090 % + 100 $\mu$ A 0.040 % + 100 $\mu$ A 0.10 % + 120 $\mu$ A 0.20 % + 200 $\mu$ A 0.40 % + 300 $\mu$ A	



0556

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Horiba MIRA Ltd**

**Issue No: 027 Issue date: 18 August 2020**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC CURRENT (cont'd) Generation (cont'd)	330 mA to 1.1 A 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.18 % + 780 $\mu$ A 0.050 % + 780 $\mu$ A 0.60 % + 1.8 mA 2.5 % + 6.0 mA	
	1.1 A to 3 A 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.18 % + 1.3 mA 0.060 % + 1.3 mA 0.60 % + 2.2 mA 2.5 % + 6.2 mA	
	3 A to 11 A 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.060 % + 9.0 mA 0.10 % + 9.0 mA 3.0 % + 9.0 mA	
	11 A to 20 A 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.12 % + 15 mA 0.15 % + 15 mA 3.0 % + 15 mA	
	10 A to 16.5 A 45 Hz to 65 Hz 65 Hz to 440 Hz	0.28 % + 3.0 mA 0.79 % + 3.0 mA	For the calibration of toroidal clamp-on ammeters
	16.5 A to 150 A 45 Hz to 65 Hz 65 Hz to 440 Hz	0.28 % + 13 mA 0.79 % + 13 mA	For the calibration of toroidal clamp-on ammeters
	150 A to 1025 A 45 Hz to 65 Hz 65 Hz to 440 Hz	0.28 % + 40 mA 0.79 % + 40 mA	For the calibration of toroidal clamp-on ammeters
Measurement	0 $\mu$ A to 200 $\mu$ A 40 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz to 10 kHz	0.42 % + 30 nA 0.24 % + 30 nA 0.48 % + 30 nA 0.57 % + 30 nA	
	200 $\mu$ A to 2 mA 40 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz 50 kHz to 100 kHz	0.36 % + 300 nA 0.18 % + 300 nA 0.15 % + 300 nA 0.15 % + 300 nA 0.30 % + 300 nA 0.50 % + 300 nA	
	2 mA to 20 mA 40 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz 50 kHz to 100 kHz	0.36 % + 3.0 $\mu$ A 0.18 % + 3.0 $\mu$ A 0.15 % + 3.0 $\mu$ A 0.15 % + 3.0 $\mu$ A 0.25 % + 3.0 $\mu$ A 0.30 % + 3.0 $\mu$ A 0.50 % + 3.0 $\mu$ A	



0556

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Horiba MIRA Ltd**

**Issue No: 027 Issue date: 18 August 2020**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC CURRENT (cont'd) Measurement (cont'd)	20 mA to 200 mA 40 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz 50 kHz to 100 kHz  0.2 A to 2 A 40 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz	0.36 % + 30 $\mu$ A 0.18 % + 30 $\mu$ A 0.15 % + 30 $\mu$ A 0.21 % + 30 $\mu$ A 0.50 % + 30 $\mu$ A 1.0 % + 30 $\mu$ A 3.0 % + 30 $\mu$ A  0.42 % + 300 $\mu$ A 0.24 % + 300 $\mu$ A 0.36 % + 300 $\mu$ A 0.52 % + 300 $\mu$ A 1.5 % + 300 $\mu$ A 4.0 % + 300 $\mu$ A	
DC RESISTANCE Generation	0 $\Omega$ to 11 $\Omega$ 11 $\Omega$ to 33 $\Omega$ 33 $\Omega$ to 110 $\Omega$ 110 $\Omega$ to 330 $\Omega$ 330 $\Omega$ to 1.1 k $\Omega$ 1.1 k $\Omega$ to 3.3 k $\Omega$ 3.3 k $\Omega$ to 11 k $\Omega$ 11 k $\Omega$ to 33 k $\Omega$ 33 k $\Omega$ to 110 k $\Omega$ 110 k $\Omega$ to 330 k $\Omega$ 330 k $\Omega$ to 1.1 M $\Omega$ 1.1 M $\Omega$ to 3.3 M $\Omega$ 3.3 M $\Omega$ to 11 M $\Omega$ 11 M $\Omega$ to 33 M $\Omega$ 33 M $\Omega$ to 110 M $\Omega$ 110 M $\Omega$ to 330 M $\Omega$ 330 M $\Omega$ to 1.1 G $\Omega$	40 ppm + 2.0 m $\Omega$ 30 ppm + 3.1 m $\Omega$ 28 ppm + 4.0 m $\Omega$ 28 ppm + 6.3 m $\Omega$ 28 ppm + 20 m $\Omega$ 28 ppm + 51 m $\Omega$ 28 ppm + 150 m $\Omega$ 28 ppm + 480 m $\Omega$ 28 ppm + 550 m $\Omega$ 32 ppm + 14 $\Omega$ 32 ppm + 24 $\Omega$ 60 ppm + 110 $\Omega$ 130 ppm + 360 $\Omega$ 250 ppm + 4.1 k $\Omega$ 500 ppm + 19 k $\Omega$ 0.30 % + 230 k $\Omega$ 1.5 % + 1.2 M $\Omega$	
Measurement	0 $\Omega$ to 20 $\Omega$ 20 $\Omega$ to 200 $\Omega$ 200 $\Omega$ to 2 k $\Omega$ 2 k $\Omega$ to 20 k $\Omega$ 20 k $\Omega$ to 200 k $\Omega$ 200 k $\Omega$ to 2 M $\Omega$ 2 M $\Omega$ to 20 M $\Omega$ 20 M $\Omega$ to 200 M $\Omega$ 200 M $\Omega$ to 1 G $\Omega$	94 ppm + 6.1 m $\Omega$ 67 ppm + 7.4 m $\Omega$ 59 ppm + 14 m $\Omega$ 69 ppm + 80 m $\Omega$ 120 ppm + 900 m $\Omega$ 190 ppm + 9.0 $\Omega$ 0.11 % + 90 $\Omega$ 2.4 % + 20 k $\Omega$ 4.0 % + 200 k $\Omega$	



0556

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Horiba MIRA Ltd**

**Issue No: 027 Issue date: 18 August 2020**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
CALIBRATION OF CHARGE AMPLIFIERS	1 pC to 33 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2% 930 ppm + 0.0060 pC 200 ppm + 0.0060 pC	
	33 pC to 330 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2% 360 ppm + 0.0080 pC 200 ppm + 0.0080 pC	
	330 pC to 3300 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2% 360 ppm + 0.050 pC 200 ppm + 0.060 pC	
	3300 pC to 33000 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2% 360 ppm + 0.65 pC 200 ppm + 0.60 pC	
	33000 pC to 100000 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2% 240 ppm + 2.0 pC 250 ppm + 6.0 pC	
Attenuation  With reference to a set value of 3 V	0 dB to 30 dB DC and 3 Hz to 10 kHz	0.020 dB	
TEMPERATURE SIMULATION			
Thermocouple type			
K	-200 °C to 1300 °C	0.30 °C	Including reference junction compensation
J	-210 °C to 1200 °C	0.26 °C	
N	-200 °C to 1300 °C	0.36 °C	
T	-250 °C to -150 °C	0.31 °C	Excluding reference junction compensation
T	-150 °C to 400 °C	0.25 °C	
K	-200 °C to 1300 °C	0.23 °C	
J	-210 °C to 1200 °C	0.16 °C	
N	-200 °C to 1300 °C	0.30 °C	
T	-250 °C to -150 °C	0.23 °C	
T	-150 °C to 400 °C	0.16 °C	



0556

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Horiba MIRA Ltd**

**Issue No: 027 Issue date: 18 August 2020**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
FREQUENCY			
Generation	10 mHz to 30 MHz	2.0 in $10^{11}$	
Measurement	1 mHz to 225 MHz	0.05 ppm	
ANGULAR VELOCITY			
Calibration of angular velocity sensors			Ambient temperature 18 °C to 22 °C.
Applied angular velocity	10 °/s to 2000 °/s	0.13 % of angular velocity	
Output voltage measurement	0 V to 200 mV 200 mV to 2 V 2 V to 20 V	560 ppm + 150 $\mu$ V 560 ppm + 150 $\mu$ V 560 ppm + 170 $\mu$ V	
END			





0556

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Horiba MIRA Ltd**

**Issue No: 027 Issue date: 18 August 2020**

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

**Expression of CMCs - symbols and units**

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %· $p$  + (0.12·10<sup>-6</sup>· $p$ ·10<sup>-6</sup>) + 4.0 Pa, where  $p$  is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 ·  $i$ , where  $i$  is the instrument indication.