


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0560 Accredited to ISO/IEC 17025:2017	Don Whitley Scientific Limited Issue No: 027 Issue date: 03 March 2025	
	Victoria Works Victoria Street Bingley West Yorkshire BD16 2NH	Contact: Mr David Boast Tel: +44 (0)1274 595728 Fax: +44 (0)1274 531197 E-Mail: info@dwscientific.co.uk Website: www.dwscientific.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Victoria Works Victoria Street Bingley West Yorkshire BD16 2NH	Temperature (Temperature Indicators)	Lab

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Premises, eg, Hospitals, Laboratories, and Manufacturing Plants (*including those of DWS)	Temperature (Sterilizers, etc, Thermal Products, and Temperature Indicators) Non-Automatic weighing machines	Site



0560
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Don Whitley Scientific Limited
Issue No: 027 Issue date: 03 March 2025

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				
Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, environmental chambers, Cold Rooms, fridges/refridgerators, freezers and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	-80 °C to -40 °C -40 °C to -25 °C -25 °C to +140 °C	0.85 °C 0.75 °C 0.34 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	Site
As above	-90 °C to +50 °C	0.29 °C	Multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping using temperature data loggers	Site
As above	130 °C to 260 °C 260 °C to 390 °C 390 °C to 520 °C 520 °C to 650 °C	0.62 °C 0.40 °C 0.71 °C 0.77 °C	Calibrations performed with thermocouple data loggers.	Site
Temperature indicators with sensors or recorders	-80 °C to -40 °C -40 °C to -25 °C -25 °C to +140 °C	0.85 °C 0.75 °C 0.14 °C	Calibration within a dry block	Lab & Site
Temperature indicators with sensors or recorders	130 °C to 260 °C 260 °C to 390 °C 390 °C to 520 °C 520 °C to 650 °C	0.33 °C 0.40 °C 0.48 °C 0.56 °C	Calibration within High Temperature dry block	Lab & Site
TIME INTERVAL				
Timers	10 s to 200 min	0.70 s	Calibration with a reference timer	Site
NON-AUTOMATIC WEIGHING MACHINES				
200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 9 kg		0.011 mg 0.015 mg 0.017 mg 0.023 mg 0.029 mg 0.037 mg 0.05 mg 0.08 mg 0.15 mg 0.36 mg 1.75 mg 3.50 mg 6.92 mg 18.9 mg 32.9 mg	Weights are available in OIML Class: E2 from 1 mg to 100 g, max. grouped load 211 g F1 from 1 mg to 5 kg, max. grouped load 9 kg Calibration method in line with the guidance within Euramet cg-18	Site
END				



0560
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Don Whitley Scientific Limited
Issue No: 027 Issue date: 03 March 2025

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$