


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0573</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Dounreay Site Restoration Limited</h3> <p>Issue No: 017 Issue date: 19 August 2021</p>	
	<p>Radiological Protection Instrument Calibration Laboratories</p> <p>Room 62 Building D1313 Dounreay Thurso Caithness KW14 7TZ</p>	<p>Contact: Mr W MacIvor</p> <p>Tel: +44 (0)1847 802 327</p> <p>Fax: +44 (0)1847 802330</p> <p>E-Mail: will.macivor@dounreay.com</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	
AIR KERMA RATE	¹³⁷ Cs 1 μGyh^{-1} to 10 μGyh^{-1} 10 μGyh^{-1} to 190 mGyh^{-1}	6.0% 4.0%	Instruments can be calibrated in accordance with the scheduled measured quantities and ranges	
	⁶⁰ Co 5 μGyh^{-1} to 25 μGyh^{-1} 400 μGyh^{-1} to 14 mGyh^{-1}	6.0% 4.0 %		
	²⁴¹ Am 15 μGyh^{-1} to 1 mGyh^{-1}	6.0%		
AMBIENT DOSE EQUIVALENT RATE	¹³⁷ Cs 1 μSvh^{-1} to 10 μSvh^{-1} 10 μSvh^{-1} to 200 mSvh^{-1}	6.0% 4.0 %		All Calibrations in compliance with GPG 14 including tests before use and periodic tests
	⁶⁰ Co 5 μSvh^{-1} to 30 mSvh^{-1} 400 μSvh^{-1} to 17 mSvh^{-1}	6.0% 4.0 %		
	²⁴¹ Am 35 μSvh^{-1} to 1.8 mSvh^{-1}	6.0%		
PERSONEL DOSE EQUIVALENT RATE	¹³⁷ Cs 1 μSvh^{-1} to 10 μSvh^{-1} 10 μSvh^{-1} to 200 mSvh^{-1}	6.0% 4.0 %		
	²⁴¹ Am 38 μSvh^{-1} to 2.0 mSvh^{-1}	6.0%		
SURFACE CONTAMINATION RESPONSE	Beta emitting nuclides ¹⁴ C, ⁹⁰ Sr, ³⁶ Cl	7%		
	Alpha emitting nuclides ²⁴¹ Am	8%		
END				



0573

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Dounreay Site Restoration Limited

Issue No: 017 Issue date: 19 August 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$