

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0589</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>G.T. Certification Limited</p> <p>Issue No: 024 Issue date: 19 August 2021</p>	
	<p>Unit A10 Link One Industrial Park George Henry Road Great Bridge Tipton West Midlands DY4 7BU</p>	<p>Contact: Mr G C Thompson Tel: +44 (0)121-522 3957 Fax: +44 (0)121-522 4958 E-Mail: sales@gtcertification.com Website: www.gtcertification.com</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
TORQUE			NOTES
Torque Wrenches	0.15 N·m to 2500 N·m to BS EN ISO 6789-2:2017	1.0 % See Notes 1 to 3	1 Calibrations may also be given in units of electrical signal output.
Torque Wrenches and Torque Drivers	0.15 N·m to 2500 N·m to BS EN ISO 6789:2003 (Withdrawn & superseded)	1.5 % See Notes 1 to 3	2 The uncertainty quoted is for both the application of the calibration torque and the characteristics of the device being calibrated.
Static Torque Transducers in clockwise and/or anti-clockwise direction in increasing and/or decreasing Torque using Reference Transducer	0.20 N·m to 1500 N·m to BS 7882:2017	0.90 % of reading See Notes 1 to 4	3 Calibration results may also be given in units of lbf-in and lbf-ft.
Static Torque Transducers in clockwise and/or anti-clockwise direction in increasing and/or decreasing Torque using Beam / Wheel and Masses	0.005 N·m to 1500 N·m to BS 7882:2017	0.045 % of reading See Notes 1 to 4	4 Calibrated statically using un-supported beam and masses or torque measuring transducer.
Britool 8000 Torque Testing Machine	10 N·m to 550 N·m to BS 7882:2017	0.75 % of reading See Notes 1 to 4	5 Calibrated using masses.
FORCE			The calibration may be performed in the following units: Newton(N), ton-force(tonf), pound-force(lbf) or kilogram-force(kgf).
Calibration of push pull force measuring devices in tension and compression	1 N to 2700 N	0.12 % see Note 5	
<p>END</p>			



0589

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

G.T. Certification Limited
Issue No: 024 Issue date: 19 August 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$