


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0590</p> <p>Accredited to ISO/IEC 17025:2005</p>	EffecTech Limited	
	Issue No: 045 Issue date: 01 September 2017	
	Dove House Dove Fields Uttoxeter Staffordshire ST14 8HU	Contact: Dr Gavin Squire Tel: +44 (0)1889 569229 Fax: +44 (0)1889 569220 E-Mail: gavin.squire@effectech.co.uk Website: www.effectech.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Dove House Dove Fields Uttoxeter Staffordshire ST14 8HU	Local contact Dr Gavin Squire Tel: +44 (0)1889 569229 Fax: +44 (0)1889 569220 email: gavin.squire@effectech.co.uk	Gas Calibration Process Gas Analysers Uttoxeter
Address N-163 MIDC Tarapur Boisar District Palghar - 401506 Maharashtra India	Local contact Padmakar Tillu Tel: +91 (0)2525 276137 Fax: +91 (0)2525 276827 email: padmakar.tillu@effectech.co.in	Gas Calibration Tarapur
Address QP West Support Services Area Ghuwairiya Street IR # 1 Ras Laffan Qatar	Local contact Biju Davis Tel: +974 5589 8625 Fax: +974 4451 5319	Gas Calibration Qatar



0590

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffecTech Limited

Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customers' sites or premises The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Process Gas Analysers	Customers' sites



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffecTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
PRIMARY REFERENCE GAS MIXTURES (PRGM) Preparation of synthetic gas mixtures by high-precision gravimetry in accordance with ISO 6142-1:2015 (Class I mixtures individually verified by analysis)				Uttoxeter
SYNTHETIC NATURAL GAS MIXTURES	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM016/UT	
nitrogen	0.02 to 25	0.12 % relative + 0.00034	Preparation of primary reference gas mixtures (PRGM) according to ISO 6142-1:2015 by high precision gravimetry	
carbon dioxide	0.05 to 25	0.10 % relative + 0.00006		
methane	34 to 100	0.055 - 0.05 % relative		
ethane	0.1 to 35	0.12 % relative + 0.00026		
propane	0.05 to 20	0.15 % relative + 0.00002		
iso-butane	0.01 to 2	0.15 % relative + 0.00011		
n-butane	0.01 to 2	0.15 % relative + 0.00011		
neo-pentane	0.001 to 0.5	0.35 % relative + 0.00005		
iso-pentane	0.002 to 0.6	0.25 % relative + 0.00005		
n-pentane	0.002 to 0.6	0.25 % relative + 0.00005		
n-hexane	0.001 to 0.5	0.50 % relative + 0.00005		
2-methylpentane	0.001 to 0.35	0.65 % relative + 0.00003		
3-methylpentane	0.001 to 0.35	0.65 % relative + 0.00003		
2,2-dimethylbutane	0.001 to 0.35	0.65 % relative + 0.00003		
benzene	0.001 to 0.2	0.65 % relative + 0.00003		
cyclohexane	0.001 to 0.2	0.65 % relative + 0.00003		
n-heptane	0.001 to 0.2	0.65 % relative + 0.00003		
toluene	0.001 to 0.1	0.65 % relative + 0.00003		
methylcyclohexane	0.001 to 0.1	0.65 % relative + 0.00003		
n-octane	0.0005 to 0.05	0.65 % relative + 0.00003		
n-nonane	0.0001 to 0.02	0.65 % relative + 0.00003		
n-decane	0.0001 to 0.005	0.65 % relative + 0.00003		
helium	0.005 to 0.2	0.85 % relative + 0.00022		
hydrogen	0.005 to 0.2	0.80 % relative + 0.0002		



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffecTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
PRIMARY REFERENCE GAS MIXTURES (PRGM) (continued)				
SYNTHETIC FUEL GAS MIXTURES	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM016/UT	Urtoxeter
nitrogen	0.1 to 60	0.12 % relative + 0.00033	Preparation of primary reference gas mixtures (PRGM) according to ISO 6142-1:2015 by high precision gravimetry	
carbon dioxide	0.1 to 30	0.35 % relative		
hydrogen	1 to 40 40 to 70	0.15 % relative + 0.015 0.075		
carbon monoxide	0.1 to 30	0.13 % relative + 0.0038		
methane	1 to 70	0.04		
ethane	0.5 to 28	0.13 % relative + 0.005		
ethene	0.5 to 12	0.6 % relative + 0.0025		
propane	0.1 to 1 1 to 15	0.01 0.2 % relative + 0.0065		
propene	0.1 to 5	0.25 % relative + 0.001		
SULPHUR GAS MIXTURES	amount fraction (ppm mol/mol)	amount fraction (ppm mol/mol)		
hydrogen sulphide	0.2 to 200	1 % relative + 0.01	Preparation of primary reference gas mixtures (PRGM) according to ISO 6142-1:2015 by high precision gravimetry	
carbonyl sulphide	0.2 to 200	1 % relative + 0.01		
carbon disulphide	0.2 to 200	1 % relative + 0.01		
methanethiol (methyl mercaptan)	0.2 to 200	1 % relative + 0.01		
ethanethiol (ethyl mercaptan)	0.2 to 200	1 % relative + 0.01		
dimethyl sulphide	0.2 to 200	1 % relative + 0.01		
1-propanethiol (n-propyl mercaptan)	0.2 to 200	1 % relative + 0.01		
2-propanethiol (iso-propyl mercaptan)	0.2 to 200	1 % relative + 0.01		
ethyl methyl sulphide (methyl ethyl sulphide)	0.2 to 200	1 % relative + 0.01		
1-butanethiol (n-butyl mercaptan)	0.2 to 200	1 % relative + 0.01		
2-methyl-2-propanethiol (tert-butyl mercaptan)	0.2 to 200	1 % relative + 0.01		



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffecTech Limited

Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
PRIMARY REFERENCE GAS MIXTURES (PRGM) (continued)				
SULPHUR GAS MIXTURES (continued)	amount fraction (ppm mol/mol)	amount fraction (ppm mol/mol)	In-house method TM016/UT matrix gas : methane or nitrogen (continued)	Uttoxeter
2-methyl-1-propanethiol (iso-butyl mercaptan)	0.2 to 200	1 % relative + 0.01		
1-methyl-1-propanethiol (sec-butyl mercaptan)	0.2 to 200	1 % relative + 0.01		
diethyl sulphide	0.2 to 200	1 % relative + 0.01		
n-hexyl mercaptan	0.2 to 200	1 % relative + 0.01		
tetrahydrothiophene (THT)	0.2 to 200	1 % relative + 0.01		
BINARY GAS MIXTURES	amount fraction (mol/mol)	amount fraction (mol/mol)	In-house method TM016/UT	
carbon monoxide/nitrogen	10 ppm to 200 ppm 200 ppm to 1000 ppm	1.3 % to 0.45 % relative 0.35 % to 0.30 % relative	Preparation of primary reference gas mixtures (PRGM) according to ISO 6142-1:2015 by high precision gravimetry	
carbon dioxide/nitrogen	0.1 % to 15 %	0.45 % to 0.15 % relative		
oxygen/nitrogen	0.5 % to 25 %	1.0 % to 0.2 % relative		
nitric oxide/nitrogen	5 ppm to 5000 ppm	0.6 % to 0.25 % relative		
nitrogen dioxide/synthetic air	5 ppm to 500 ppm	2.0 % relative		
sulphur dioxide/nitrogen	5 ppm to 2000 ppm	1.5 % to 0.5 % relative		
methane/nitrogen	0.1 % to 5 %	0.4 % to 0.15 % relative		
methane/synthetic air	0.1% to 2.5 %	0.4 % to 0.15 % relative		
propane/nitrogen	1 ppm to 1000 ppm 0.1 % to 2 %	2 % to 0.4 % relative 0.4 % to 0.15 % relative		
propane/synthetic air	1 ppm to 1000 ppm 0.1 % to 1.1 %	2 % to 0.4 % relative 0.4 % to 0.15 % relative		
PROPANE BALANCE GAS MIXTURES	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM016/UT	
nitrogen	0.1 to 3	0.15 % relative + 0.001	Preparation of primary reference gas mixtures (PRGM) according to ISO 6142-1:2015 by high precision gravimetry	
ethane	0.25 to 3	0.15 % relative + 0.001		
propane	92 to 99.5	0.055 - 0.05 % relative		
iso-butane	0.03 to 1	0.15 % relative + 0.00025		



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffectTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code	
PRIMARY REFERENCE GAS MIXTURES (PRGM) (continued)					
PROPANE BALANCE GAS MIXTURES (continued)	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM016/UT (continued)	Uttoxeter	
n-butane	0.03 to 1	0.15 % relative + 0.00025			
iso-pentane	0.02 to 0.08	0.6 % relative + 0.00005			
n-pentane	0.02 to 0.08	0.6 % relative + 0.00005			
SECONDARY REFERENCE GAS MIXTURES (SRGM) Calibration of synthetic gas mixtures by analysis					
SYNTHETIC NATURAL GAS MIXTURES	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM001/UT		
nitrogen	0.1 to 22	0.25 % relative + 0.0005	Calibration of gas mixtures in accordance with ISO 6143:2001 using gas chromatography with thermal conductivity detection (GC-TCD)		
carbon dioxide	0.05 to 15	0.18 % relative + 0.0001			
methane	34 to 100	0.11 - 0.10 % relative			
ethane	0.1 to 35	0.25 % relative			
propane	0.05 to 15	0.3 % relative			
iso-butane	0.01 to 0.15 0.15 to 2	0.00045 0.3 % relative			
n-butane	0.01 to 0.15 0.15 to 2	0.00045 0.3 % relative			
neo-pentane	0.002 to 0.35	0.7 % relative + 0.0001			
iso-pentane	0.005 to 0.35	0.5 % relative + 0.0001			
n-pentane	0.005 to 0.35	0.5 % relative + 0.0001			
n-hexane	0.001 to 0.35	1.0 % relative + 0.0001		Calibration of gas mixtures using gas chromatography with flame ionisation detection (GC-FID)	
2-methylpentane	0.001 to 0.35	1.3 % relative + 0.00005			
3-methylpentane	0.001 to 0.35	1.3 % relative + 0.00005			
2,2-dimethylbutane	0.001 to 0.35	1.3 % relative + 0.00005			
benzene	0.001 to 0.2	1.3 % relative + 0.00005			
cyclohexane	0.001 to 0.2	1.3 % relative + 0.00005			
n-heptane	0.001 to 0.2	1.3 % relative + 0.00005			
toluene	0.001 to 0.1	1.3 % relative + 0.00005			



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffecTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
SECONDARY REFERENCE GAS MIXTURES (SRGM) (continued)				
SYNTHETIC NATURAL GAS MIXTURES (continued)	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM001/UT (continued)	Uttoxeter
methylcyclohexane	0.001 to 0.1	1.3 % relative + 0.00005		
n-octane	0.0005 to 0.05	1.3 % relative + 0.00005		
n-nonane	0.0005 to 0.02	1.3 % relative + 0.00005		
n-decane	0.0005 to 0.005	1.3 % relative + 0.00005		
C ₆₊	0.001 to 0.35	1.0 % relative + 0.0001	C ₆₊ is the sum of all hydrocarbons containing six carbon atoms or greater	
oxygen	0.005 to 1	5 % relative	Calibration of gas mixtures using gas chromatography with thermal conductivity detection (GC-TCD)	
helium	0.005 to 0.2	1.7 % relative + 0.0004		
hydrogen	0.005 to 0.2	1.7 % relative + 0.0002		
			In-house method TM005/TA	Tarapur
nitrogen	0.1 to 12	0.25 % relative + 0.0005	Calibration of gas mixtures in accordance with	
carbon dioxide	0.05 to 8	0.18 % relative + 0.0001	ISO 6143:2001 using gas chromatography with thermal conductivity detection (GC-TCD)	
methane	64 to 100	0.11 - 0.10 % relative		
ethane	0.1 to 14	0.25 % relative		
propane	0.05 to 8	0.3 % relative		
iso-butane	0.01 to 0.15 0.15 to 1.2	0.00045 0.3 % relative		
n-butane	0.01 to 0.15 0.15 to 1.2	0.00045 0.3 % relative		
neo-pentane	0.002 to 0.35	0.7 % relative + 0.0001		
iso-pentane	0.005 to 0.35	0.5 % relative + 0.0001		
n-pentane	0.005 to 0.35	0.5 % relative + 0.0001		
n-hexane	0.001 to 0.35	1.0 % relative + 0.0001	Calibration of gas mixtures using gas chromatography with flame ionisation detection (GC-FID)	
2-methylpentane	0.001 to 0.35	1.3 % relative + 0.00005		
3-methylpentane	0.001 to 0.35	1.3 % relative + 0.00005		
2,2-dimethylbutane	0.001 to 0.35	1.3 % relative + 0.00005		



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffectTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
SECONDARY REFERENCE GAS MIXTURES (SRGM) (continued)				
SYNTHETIC NATURAL GAS MIXTURES (continued)	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM005/TA (continued)	Tarapur
benzene	0.001 to 0.2	1.3 % relative + 0.00005		
cyclohexane	0.001 to 0.2	1.3 % relative + 0.00005		
n-heptane	0.001 to 0.2	1.3 % relative + 0.00005		
toluene	0.001 to 0.1	1.3 % relative + 0.00005		
methylcyclohexane	0.001 to 0.1	1.3 % relative + 0.00005		
n-octane	0.0005 to 0.05	1.3 % relative + 0.00005		
n-nonane	0.0005 to 0.02	1.3 % relative + 0.00005		
n-decane	0.0005 to 0.005	1.3 % relative + 0.00005		
oxygen	0.005 to 1	5 % relative	Calibration of gas mixtures using gas chromatography with thermal conductivity detection (GC-TCD)	Qatar
nitrogen	0.1 to 12	0.35 % relative + 0.0005	In-house method TM022/QA Calibration of gas mixtures in accordance with	
carbon dioxide	0.05 to 8	0.20 % relative + 0.0005	ISO 6143:2001 using gas chromatography with thermal conductivity detection (GC-TCD)	
methane	64 to 100	0.11 – 0.1 % relative		
ethane	0.1 to 14	0.25 % relative		
propane	0.05 to 8	0.30 % relative		
iso-butane	0.01 to 0.15 0.15 to 1.2	0.0005 0.35 % relative		
n-butane	0.01 to 0.15 0.15 to 1.2	0.0005 0.35 % relative		
neo-pentane	0.004 to 0.35	0.70 % relative + 0.0001		
iso-pentane	0.005 to 0.1 0.1 to 0.35	0.0010 0.9 % relative + 0.0006		
n-pentane	0.005 to 0.18 0.18 to 0.35	0.001 0.9 % relative + 0.0006		
n-hexane	0.1 to 0.35	1.2 % relative		



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffecTech Limited

Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
SECONDARY REFERENCE GAS MIXTURES (SRGM) (continued)				
SYNTHETIC NATURAL GAS MIXTURES (continued)	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM022/QA (continued)	Qatar
n-hexane	0.001 to 0.1	1.7 % relative + 0.00005	Calibration of gas mixtures using gas chromatography with flame ionisation detection (GC-FID)	
2-methylpentane	0.001 to 0.35	1.7 % relative + 0.00005		
3-methylpentane	0.001 to 0.35	1.7 % relative + 0.00005		
2,2-dimethylbutane	0.001 to 0.35	1.7 % relative + 0.00005		
benzene	0.001 to 0.2	1.7 % relative + 0.00005		
cyclohexane	0.001 to 0.2	1.7 % relative + 0.00005		
n-heptane	0.001 to 0.2	1.7 % relative + 0.00005		
toluene	0.001 to 0.1	1.7 % relative + 0.00005		
methylcyclohexane	0.001 to 0.1	1.7 % relative + 0.00005		
n-octane	0.0005 to 0.05	1.7 % relative + 0.00005		
n-nonane	0.0005 to 0.02	1.7 % relative + 0.00005		
n-decane	0.0005 to 0.005	1.7 % relative + 0.00005		
Calculated values for:	Calculations valid for gas mixtures with amount fractions (% mol/mol)		Calculated values according to ISO 6976:1995 including amendment No 1, May 1998	All Sites
calorific value (superior)		0.1 % relative		
calorific value (inferior)		0.1 % relative		
relative density	nitrogen < 30 %	0.1 % relative		
density	carbon dioxide < 15 %	0.1 % relative		
Wobbe index	ethane < 15 %	0.1 % relative		
mean molecular mass	other components < 5 %	0.1 % relative		
compression factor	methane no restriction	0.1 % relative		
gross heating value		0.1 % relative		
net heating value	no compositional restrictions specified	0.1 % relative		
relative density		0.1 % relative		
compressibility factor		0.1 % relative		
gross heating value		0.1 % relative	Calculated values according to methods given in GPA 2172-09 (2009) using data tables from GPA 2145-09	
net heating value	no compositional restrictions specified	0.1 % relative	Calculated values according to methods given in ASTM D3588-98 (2011) using data tables from GPA 2145-09	
relative density		0.1 % relative		
density		0.1 % relative		
compressibility factor		0.1 % relative		
		0.1 % relative		



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffectTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
SECONDARY REFERENCE GAS MIXTURES (SRGM) (continued)				
SULPHUR GAS MIXTURES	amount fraction (ppm mol/mol)	amount fraction (ppm mol/mol)	In-house method TM002/UT matrix gas : methane or nitrogen	Uttoxeter
hydrogen sulphide	0.2 to 10	2 % relative + 0.03	Calibration of gas mixtures using gas chromatography with sulphur chemiluminescence detection (GC-SCD)	
carbonyl sulphide	0.2 to 10	2 % relative + 0.03		
methanethiol (methyl mercaptan)	0.2 to 10	2 % relative + 0.03		
ethanethiol (ethyl mercaptan)	0.2 to 10	2 % relative + 0.03		
dimethyl sulphide	0.2 to 10	2 % relative + 0.03		
1-propanethiol (n-propyl mercaptan)	0.2 to 10	4 % relative + 0.03		
2-propanethiol (iso-propyl mercaptan)	0.2 to 10	2 % relative + 0.03		
ethyl methyl sulphide (methyl ethyl sulphide)	0.2 to 10	2 % relative + 0.03		
1-butanethiol (n-butyl mercaptan)	0.2 to 10	4 % relative + 0.03		
2-methyl-2-propanethiol (tert-butyl mercaptan)	0.2 to 10	2 % relative + 0.03		
2-methyl-1-propanethiol (iso-butyl mercaptan)	0.2 to 10	4 % relative + 0.03		
1-methyl-1-propanethiol (sec-butyl mercaptan)	0.2 to 10	4 % relative + 0.03		
diethyl sulphide	0.2 to 10	2 % relative + 0.03		
n-hexyl mercaptan	0.2 to 10	4 % relative + 0.03		
tetrahydrothiophene (THT)	0.2 to 10	2 % relative + 0.03		
BLAST FURNACE GAS MIXTURES	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM004/UT	
nitrogen	40 to 50	1 % relative	Calibration of gas mixtures in accordance with ISO 6143:2001 using gas chromatography with thermal conductivity detection (GC-TCD)	
carbon dioxide	20 to 30	1 % relative		
hydrogen	1 to 8	1 % relative		
carbon monoxide	20 to 30	1 % relative		



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffecTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
SECONDARY REFERENCE GAS MIXTURES (SRGM) (continued)				
PROPANE BALANCE GAS MIXTURES	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM020/UT	Uttoxeter
nitrogen	0.1 to 3	0.3 % relative + 0.002	Calibration of gas mixtures using gas chromatography with thermal conductivity detection (GC-TCD)	
ethane	0.25 to 3	0.3 % relative + 0.002		
propane	92 to 99.5	0.11 - 0.10 % relative		
iso-butane	0.03 to 1	0.3 % relative + 0.0005		
n-butane	0.03 to 1	0.3 % relative + 0.0005		
iso-pentane	0.02 to 0.08	1.2 % relative + 0.0001		
n-pentane	0.02 to 0.08	1.2 % relative + 0.0001		
BINARY EMISSION GAS MIXTURES	amount fraction (mol/mol)	amount fraction (mol/mol)	In-house method TM025/UT	Uttoxeter
propane in nitrogen	1 ppm to 50 ppm	0.6 % relative	Calibration of gas mixtures in accordance with ISO/DIS 12963 using gas chromatography with flame ionisation detection (GC-FID)	
propane in synthetic air	1 ppm to 50 ppm	0.6 % relative		
BINARY EMISSION GAS MIXTURES	amount fraction (mol/mol)	amount fraction (mol/mol)	In-house method TM014	All Sites
carbon monoxide in nitrogen or synthetic air	10 ppm to 200 ppm 200 ppm to 1000 ppm	1.5 % to 0.9 % relative 0.7 % to 0.6 % relative	Calibration of gas mixtures in accordance with ISO/DIS 12963 using dynamically generated reference gases in accordance with ISO 6145 Part 7: Thermal Mass Flow Controllers	
carbon dioxide in nitrogen or synthetic air	0.1 % to 1 % 1 % to 15 %	0.9 % to 0.6 % relative 0.6 % to 0.3 % relative		
oxygen in nitrogen	0.5 % to 3 % 3 % to 25 %	2.1 % to 0.7 % relative 0.7 % to 0.4 % relative		
nitric oxide in nitrogen	10 ppm to 60 ppm 60 ppm to 600 ppm	1.1 % to 0.9 % relative 0.6 % to 0.5 % relative		
nitrogen dioxide in synthetic air	5 ppm to 500 ppm	4 % relative		
sulphur dioxide in nitrogen or synthetic air	5 ppm to 100 ppm 100 ppm to 1000 ppm	3.0 % to 1.2 % relative 1.2 % to 1.0 % relative		
methane in nitrogen	0.1 % to 1 % 1 % to 5 %	0.7 % to 0.4 % relative 0.4 % to 0.3 % relative		
methane in synthetic air	0.1 % to 1 % 1 % to 2.5 %	0.7 % to 0.4 % relative 0.4 % to 0.3 % relative		



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffectTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
GAS ANALYSERS Calibration of gas analysers using reference gas mixtures				
NATURAL GAS ANALYSERS	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM003	Customers' sites
nitrogen	0.1 to 22	0.3 % relative + 0.002	Calibration of gas analysers used for natural gas analysis in accordance with ISO 10723:2012	
carbon dioxide	0.05 to 15	0.25 % relative + 0.0005		
methane	34 to 100	0.07 % mol/mol		
ethane	0.1 to 23	0.25 % relative + 0.0015		
propane	0.05 to 10	0.3 % relative + 0.0005		
iso-butane	0.01 to 2.0	0.3 % relative + 0.0005		
n-butane	0.01 to 2.0	0.3 % relative + 0.0005		
neo-pentane	0.005 to 0.35	0.7 % relative + 0.0005		
iso-pentane	0.005 to 0.35	0.5 % relative + 0.0005		
n-pentane	0.005 to 0.35	0.5 % relative + 0.0005		
n-hexane	0.001 to 0.35	1.0 % relative + 0.0005		
n-heptane	0.001 to 0.20	1.3 % relative + 0.0001		
n-octane	0.0005 to 0.05	1.3 % relative + 0.0001		
n-nonane	0.0005 to 0.02	1.3 % relative + 0.0001		
n-decane	0.0005 to 0.005	1.3 % relative + 0.0001		
OTHER FUEL GAS ANALYSERS	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM006	
C ₁ - C ₃	0.0008 to 100	amount fractions from 1 % to 100 % ± 0.5 % relative	Calibration of gas analysers based on ISO 10723:2012	
C ₄	0.001 to 50			
C ₅	0.001 to 9	amount fractions from 0.1 % to 1 % ± 1 % relative		
C ₆	0.001 to 1.5			
C ₇	0.001 to 0.5	amount fractions from 0.0008 % to 0.1 % ± 2 % relative		
C ₈	0.001 to 0.2			
C ₉	0.001 to 0.2			
C ₁₀	0.001 to 0.05			
benzene	0.001 to 1			



0590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

EffectTech Limited
Issue No: 045 Issue date: 01 September 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
GAS ANALYSERS (continued)				Customers' sites
OTHER FUEL GAS ANALYSERS (continued)	amount fraction (% mol/mol)	amount fraction (% mol/mol)	In-house method TM006 (continued)	
toluene	0.001 to 0.4	amount fractions from 1 % to 100 % ± 0.5 % relative		
xylene (m, p and o)	0.001 to 0.1			
argon	0.1 to 100	amount fractions from 0.1 % to 1 % ± 1 % relative		
carbon dioxide	0.03 to 100			
carbon monoxide	0.001 to 100	amount fractions from 0.0008 % to 0.1 % ± 2 % relative		
helium	0.1 to 100			
hydrogen	0.08 to 100			
nitrogen	0.1 to 100			
oxygen	0.05 to 100			
OTHER GAS ANALYSERS	amount fraction (ppm mol/mol)	amount fraction (ppm mol/mol)		
hydrogen sulphide	1 to 10	2 % relative + 0.03		
END				