


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0604 Accredited to ISO/IEC 17025:2017	Young Calibration Limited Issue No: 038 Issue date: 30 July 2021	
	5 Cecil Pashley Way Shoreham Airport Shoreham-by-Sea West Sussex BN43 5FF	Contact: Mr A Young Tel: +44 (0)1273 455572 E-Mail: ay@youngcalibration.co.uk Website: www.youngcalibration.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code	
Address 5 Cecil Pashley Way Shoreham Airport Shoreham-by-Sea West Sussex BN43 5FF	Local contact Mr A Young	Pressure Electrical Temperature indicators - Electrical simulation Air velocity Air flow Water flow Hydrocarbon flow Temperature Humidity	Lab

Site activities performed away from the locations listed above:

Location details	Activity	Location code	
The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Local contact Mr A Young	Pressure Electrical Temperature indicators - Electrical simulation Temperature Humidity	Site



0604
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Young Calibration Limited
Issue No: 038 Issue date: 30 July 2021

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
FLOW				
Hydrocarbon oils				
Calibration of flow meters using gravimetric and reference meter methods				
Volume flow rate	0.5 l/min to 440 l/min	0.40 %	Calibrations are carried out with fluids within the viscosity range 5 to 20cSt at fluid temperatures of up to 60 °C	Lab
Quantity of fluid passed	0.33 l to 361.7 l 361.7 l to 4 400 l (at flow rates of 0.5 l/min to 440 l/min)	0.40 % 1.0 %		
Mass flow rate	0.5 kg/min to 367.4 kg/min	0.40 %	Calibrations are carried out at pressures of up to 10 bar	Lab
Mass of fluid passed	0.4 kg to 302 kg 302 kg to 3 674 kg (at flow rates of 0.5 kg/min to 367.4 kg/min)	0.40 % 1.0 %		
Water				
Calibration of flow meters using gravimetric and reference meter methods				
Volume flow rate	1.0 ml/hr to 2400 ml/hr 1.0 ml/hr to 2400 ml/hr	0.10 % + 0.0040 ml/hr 0.55 % + 0.36 ml/hr	Calibrations are carried out at pressures of up to 4 bar	Lab
Quantity of fluid passed	0.04 l/min to 1 000 l/min	0.15 %		
Mass flow rate	0.25 l to 801.6 l 801.6 l to 7 000 l (at flow rates of 0.04 l/min to 1 000 l/min)	0.15 % 1.0 %		
Mass of fluid passed	1 g/hr to 2400 g/hr 1 g/hr to 2400 g/hr	0.10 % + 0.0040 g/hr 0.55 % + 0.36 g/hr		
	0.04 kg/min to 998 kg/min	0.15 %		
	0.25 g to 800 kg 800 kg to 9 980 kg (at flow rates of 0.04 kg/min to 998 kg/min)	0.15% 1.0 %		



0604
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Young Calibration Limited
Issue No: 038 Issue date: 30 July 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
Gas				
Calibration of flow meters using Sonic nozzle method			Calibration of flow meters with an electrical or pressure output can be undertaken	
Volume flow rate	1 ml/min to 1 300 l/min	0.52 % + 0.3 ml/min	Calibrations are carried out at pressures of up to 8 bar	Lab
Quantity of gas passed	300 ml to 26 000 l (at flow rates of 1 ml/min to 1 300 l/min)	0.52 % + 0.3 ml		
Mass flow rate	0.001 2 g/min to 1.56 kg/min	0.52 % + 0.0003 g/min		
Mass of gas passed	0.36 g to 31.2 kg (at flow rates of 0.001 2 g/min to 1.56 kg/min)	0.52 % + 0.0003 g		
Calibration of flow meters using Turbine meter method			Calibrations are carried out at ambient conditions	
Volume flow rate	10 l/s to 450 l/s	0.90 %		
Quantity of gas passed	300 l to 540 kl (at flow rates of 10 l/s to 450 l/s)	0.90 %		
Mass flow rate	12 g/s to 0.54 kg/s	0.90 %		
Mass of gas passed	0.36 kg to 643 kg (at flow rates of 12 g/s to 0.54 kg/s)	0.90 %		
Calibration of flow meters using LDA method			Calibrations are carried out at ambient conditions using a laser doppler anemometer	Lab
Volume flow rate	40 l/s to 1 250 l/s	0.70 % + 0.030 l/s		
Quantity of gas passed	12 kl to 375 kl (at flow rates of 40 l/s to 1 250 l/s)	0.70 % + 9.0 l		
Mass flow rate	48 g/s to 1.488 kg/s	0.70 % + 0.036 g/s		
Mass of gas passed	14.28 kg to 1 785 kg (at flow rates of 48 g/s to 1.488 kg/s)	0.70 % + 11 g		



0604
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Young Calibration Limited
Issue No: 038 Issue date: 30 July 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
AIR VELOCITY				
Pitot tubes	1.0 m/s to 80 m/s	0.15 % + 0.000 20 m/s	1) Calibrations are performed against a laser doppler anemometer, or a secondary standard where requested. 2) Pitot tube uncertainty dependant on pitot differential pressure range 3) Air velocity instruments up to 480 x 120 mm diameter (working area): uncertainty is dependent on design of instrument under test.	Lab
Thermal and ultrasonic anemometers	0.05 m/s to 80 m/s	0.16 % + 0.000 30 m/s		
Vane anemometers	0.1 m/s to 40 m/s	0.20 % + 0.0010 m/s		
Rotating cup anemometers	1.0 m/s to 21 m/s	0.23 %		
PRESSURE			Methods consistent with EURAMET CG17	
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges	5 kPa to 130 kPa	22 Pa	Devices with an electrical output can also be calibrated.	Lab
Gas pressure (gauge)				
Calibration of pressure indicating instruments and gauges	-90 kPa to 0 Pa 0 Pa to 2.5 kPa 2.5 kPa to 5 kPa 5 kPa to 6.9 kPa 6.9 kPa to 20 kPa 20 kPa to 270 kPa 270 kPa to 2.1 MPa 2.1 MPa to 3.5 MPa	38 Pa 0.045 % + 0.060 Pa 0.046 % + 0.31 Pa 9 Pa 16 Pa 0.029 % 0.023 % 1.1 kPa	Absolute pressure calibrations can be undertaken using gauge pressure generation and the associated barometric pressure with the additional absolute pressure uncertainty as listed	Lab
Hydraulic pressure (gauge)				
Calibration of pressure indicating instruments and gauge	0.55 MPa to 55 MPa 55 MPa to 110 MPa	0.019 % + 90 Pa 0.018 %		Lab
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges	5.0 kPa to 130 kPa	22 Pa		Site
Gas pressure (gauge)				
Calibration of pressure indicating instruments and gauges	-90 kPa to 0 Pa 0 Pa to 2.5 kPa 2.5 kPa to 5.0 kPa 5.0 kPa to 6.9 kPa 6.9 kPa to 34.5 kPa 34.5 kPa to 250 kPa 2.1 MPa to 3.5 MPa	38 Pa 0.045 % + 0.060 Pa 0.046 % + 0.31 Pa 9 Pa 16 Pa 82 Pa 1.1 kPa		Site



0604
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Young Calibration Limited
Issue No: 038 Issue date: 30 July 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL				Lab and site
DC Voltage Generation	0 V to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 1000 V	0.012 % + 6.0 μ V 0.0070 % + 33 μ V 0.0070 % + 370 μ V 0.0073 % + 29 mV	Using multifunction calibrator.	
Measurement	0 V to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	0.0047 % + 5.5 μ V 0.0034 % + 34 μ V 0.0033 % + 330 μ V 0.0048 % + 3.3 mV 0.0051 % + 31 mV	Using digital multimeter.	
Mains Voltage Measurement Mains Voltage Frequency	200 V to 300 V Nominal 50 Hz	1.0 % 2.5 %		
DC Current Generation	0 mA to 33 mA 33 mA to 330 mA 330 mA to 2.2 A 2.2 A to 11 A	0.016 % + 0.35 μ A 0.017 % + 4.5 μ A 0.041 % + 190 μ A 0.079 % + 460 μ A	Using multifunction calibrator.	
Measurement	0 mA to 30 mA 0 mA to 100 mA 0 mA to 10A	0.017 % + 6.0 μ A 0.060 % + 7.5 μ A 0.21 % + 1.2 mA	Using digital multimeter.	
DC Resistance	0 Ω to 33 Ω 33 Ω to 330 Ω 330 Ω to 3.3 k Ω 3.3 k Ω to 33 k Ω 33 k Ω to 110 k Ω 110 k Ω to 330 k Ω 330 k Ω to 3.3 M Ω 3.3M Ω to 11 M Ω 11 M Ω to 33 M Ω 33 M Ω to 110 M Ω 110 M Ω to 330 M Ω	20 m Ω 60 m Ω 500 m Ω 5.0 Ω 25 Ω 60 Ω 800 Ω 10 k Ω 43 k Ω 800 k Ω 2.2 M Ω	Using multifunction calibrator.	
Measurement	0 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	20 m Ω 150 m Ω 1.5 Ω 15 Ω 140 Ω 5.0 k Ω 1.1 % + 23 k Ω 3.2 % + 7.4 M Ω	Using digital multimeter.	



0604
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Young Calibration Limited
Issue No: 038 Issue date: 30 July 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued)				Lab and site
Frequency			Using GPS disciplined oscillator and frequency counter.	
Specific values	1 MHz, 5 MHz and 10 MHz	0.12 ppm		
Ranges source and measure	0.5 Hz to 100 kHz	3.7 ppm	Frequency can be expressed in other units, for example RPM, at equivalent uncertainties.	
Source only	250 MHz to 1 GHz	0.13 ppm		
Elapsed time	10 s to 24 hours	0.13 ppm + 54 ms	For calibration of timers and stopwatches.	
Electrical calibration of temperature indicators				
Ambient	15 °C to 30 °C	0.41 °C	For reporting reference junction temperature.	
Base Metal Thermocouples	-270 °C to 0 °C 0 °C to 1370 °C	0.67 °C 0.55 °C	Including reference junction compensation	
Noble Metal Thermocouples	-50 °C to +399 °C 400 °C to 1760 °C	0.96 °C 0.70 °C		
PRTs				
Generate resistance (Pt 100)	-100 °C to +850 °C	0.43 °C		
Measure resistance (Pt 100)	-100 °C to +850 °C	0.27 °C		
TEMPERATURE			By comparison with Reference thermometers	
Resistance thermometers and electronic probes with indicators	-40 °C to +20 °C 20 °C to 200 °C 0 °C	0.35 °C 0.059 °C 0.043 °C	In Liquid bath	Lab
Base metal thermocouples	-40 °C to +20 °C 20 °C to 200 °C	0.55 °C 0.70 °C		
Block Calibrators	-40 °C to +200 °C	0.045 °C		Lab and site
Liquid baths	-40 °C to +200 °C	0.045 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	
Air temperature	0 °C to 60 °C 0 °C to 60 °C	0.19 °C 0.21 °C	In air chamber In air chamber	Lab Site



0604
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Young Calibration Limited
Issue No: 038 Issue date: 30 July 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
HUMIDITY				
Dew point	-32 °C to +60 °C	0.31 °C	By comparison with chilled mirror hygrometer and reference thermometers	Lab
Relative humidity	Temperature range 0 °C to 60 °C 5 %rh to 20 %rh 20 %rh to 40 %rh 40 %rh to 60 %rh 60 %rh to 80 %rh 80 %rh to 95 %rh	0.19 °C 0.20 %rh 0.31 %rh 0.70 %rh 1.4 %rh 1.9 %rh		Lab
	Temperature range 0 °C to 60 °C 5 %rh to 20 %rh 20 %rh to 40 %rh 40 %rh to 60 %rh 60 %rh to 80 %rh 80 %rh to 95 %rh	0.21 °C 1.3 %rh 1.4 %rh 1.7 %rh 2.2 %rh 2.5 %rh	By comparison with reference hygrometer	Site
END				



0604
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Young Calibration Limited
Issue No: 038 Issue date: 30 July 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10⁻⁶·p·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.